

	<b>DOCUMENT TITLE:</b> Equipment Requirements Form	<b>Revision:</b> A
	<b>DOCUMENT #:</b> FM2-QM.SLQ050	<b>Effective Date:</b> 30Jul2022

## Equipment Requirements

<b>Equipment ID #:</b>	ST-012	<b>Equipment Description:</b>	Weighing Scale
<b>Model #:</b>	AB54-S	<b>S/N:</b>	1125311515
<b>Manufacturer:</b>	Mettler Toledo	<b>Operating Range:</b>	0 – 51 g (0.1 mg readability)
<b>Type (check one):</b>	<input type="checkbox"/> Production / <input type="checkbox"/> Engineering / <input checked="" type="checkbox"/> Quality / <input type="checkbox"/> Other _____		

Calibration Requirements (check type of calibration):		
<input checked="" type="checkbox"/> Calibration Required	<input type="checkbox"/> Limited Calibration	<input type="checkbox"/> No Calibration
Internal / External (circle one)  Interval: <u>Annual</u>  Internal calibration procedure #: <u>N/A</u>	For limited calibration, document nature of limited calibration (how it can / cannot be used)	For no calibration required, document rationale for reason(s) equipment does not require calibration
<b>Requirements or Rationale for No Calibration:</b> Calibrate for full operating range per manufacturer's specification.		
Preventive Maintenance Requirements (check type of calibration):		
<input type="checkbox"/> Preventive Maintenance Required	<input checked="" type="checkbox"/> No Preventive Maintenance	
Interval: _____  Internal PM procedure #: _____	For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
<b>Requirements or Rationale for No PM:</b> The scale housing and weighing pan are made of high grade durable materials. All commonly used cleaning agents can be used to clean surfaces, as needed. Periodic preventive maintenance is not required.		

Approval		
	Signature	Date
<b>Quality:</b>		27 Sep 2022
<b>Engineering:</b>		27 Sept 2022