

	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-001	Equipment Description:	Portable Exchange Deionizers (PEDI) System
Model #:	D-350-MB	S/N:	N/A
Manufacturer:	Culligan	Operating Range:	Maximum operating parameters: • Pressure: 80 psig • Temperature: 95°F (35°C) • Turbidity: <5 NTU • Color: <5 Units • Organics: <3 mg/L • Free Chlorine: <0.2 mg/L • Total Dissolved Solids (TDS): <600 mg/L
Type (check one):	☒ Production / ☐ Engineering / ☐ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☐ Calibration Required Internal / External (circle one) Interval: _____ Internal calibration procedure #: _____	☐ Limited Calibration For limited calibration, document nature of limited calibration (how it can / cannot be used)	☒ No Calibration For no calibration required, document rationale for reason(s) equipment does not require calibration
Requirements or Rationale for No Calibration: There are no electro-mechanical components that require calibration for this water deionizing system. Water quality is periodically tested to ensure water outputted from this system meets requirements.		
Preventive Maintenance Requirements (check type of calibration):		
☐ Preventive Maintenance Required Interval: _____ Internal PM procedure #: _____	☒ No Preventive Maintenance For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
Requirements or Rationale for No PM: Upon installation of a new tank, fittings are inspected for any leaks. There are no electro-mechanical components that require preventive maintenance.		

Approval		
	Signature	Date
Quality:		14 Sep 2022
Engineering:		14SEPT2022