

	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements


Equipment ID #:	ST-005	Equipment Description:	Diafiltration System, Repligen KrosFlo KTF-2000
Model #:	SYTF-2000	S/N:	190603-001
Manufacturer:	Repligen	Operating Range:	N/A
Type (check one):	☒ Production / ☐ Engineering / ☐ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☒ Calibration Required	☐ Limited Calibration	☐ No Calibration
Internal / External (circle one) Interval: <u>Every four (4) years</u> Internal calibration procedure #: <u>Repligen Doc # IF.KTF-POP.1146</u>	For limited calibration, document nature of limited calibration (how it can / cannot be used)	For no calibration required, document rationale for reason(s) equipment does not require calibration

Requirements or Rationale for No Calibration:

The manufacturer recommends the system be serviced annually. This includes calibration of three pressure sensors, pump RPM check, flowmeter test, scale accuracy verification, and general preventive maintenance on the system. Using risk assessment, there is no risk to the safety or efficacy of the product (suspension) for a calibration and PM frequency to be performed annually. Accuracy of the pump RPM's does not affect the suspension or its ability to meet specification. If the pump RPM's are not accurate, this would only affect the speed of the manufacturing process. Flowmeters are not used in the manufacturing process of the suspension. The pertinent scale is being calibrated independently by an accredited third-party calibration service provider on an annual basis. Additionally, the diafiltration system is expected to be used once every three months. Based on this rationale, extending the calibration and preventive maintenance on the system to every four (4) years is warranted.

Preventive Maintenance Requirements (check type of calibration):	
☒ Preventive Maintenance Required	☐ No Preventive Maintenance
Interval: <u>Every four (4) years</u> Internal PM procedure #: <u>Repligen Doc # IF.KTF-POP.1146</u>	For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance
Requirements or Rationale for No PM: See rationale for PM frequency above.	

Approval		
	Signature	Date
Quality:	Bryan Wong 	21 Nov 2024
Engineering:	Brian McVerry 	21 NOV 2024

Note to file: The original equipment requirements form for ST-005 was approved in September 2022. Upon further evaluation of the equipment's usage and its impact to product quality, the team determined to update the calibration and PM frequency as noted above.

BW
26 Nov 2024