


	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-007	Equipment Description:	Floor Scale, Suspension
Model #:	318506	S/N:	318506-2108
Manufacturer:	Global Industrial	Operating Range:	0 – 1000 lb
Type (check one):	☒ Production / ☐ Engineering / ☐ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☒ Calibration Required	☐ Limited Calibration	☐ No Calibration
Internal / External (circle one) Interval: <u>Annual</u> Internal calibration procedure #: <u>N/A</u>	For limited calibration, document nature of limited calibration (how it can / cannot be used)	For no calibration required, document rationale for reason(s) equipment does not require calibration
Requirements or Rationale for No Calibration: Calibrate for full operating range per manufacturer's specification.		
Preventive Maintenance Requirements (check type of calibration):		
☐ Preventive Maintenance Required	☒ No Preventive Maintenance	
Interval: _____ Internal PM procedure #: _____	For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
Requirements or Rationale for No PM: There are no mechanical components that require preventive maintenance. Annual calibration would include any necessary adjustments to the scale for proper operation and functionality.		

Approval		
	Signature	Date
Quality:		14 Sep 2022
Engineering:		14SEPT2022