

	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-008	Equipment Description:	Peristaltic Pump, Large Flow
Model #:	M6-12L	S/N:	11006080100
Manufacturer:	U.S. Solid	Operating Range:	Flow rate range: 0.69 – 12,000 mL/min
Type (check one):	☒ Production / ☐ Engineering / ☐ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☐ Calibration Required Internal / External (circle one) Interval: _____ Internal calibration procedure #: _____	☐ Limited Calibration For limited calibration, document nature of limited calibration (how it can / cannot be used)	☒ No Calibration For no calibration required, document rationale for reason(s) equipment does not require calibration

Requirements or Rationale for No Calibration:


The flow rate of the peristaltic pump does not need to be precise for its function in the manufacturing process. The pump is used to pump fluids to/from various containers/tanks.

Preventive Maintenance Requirements (check type of calibration):	
☒ Preventive Maintenance Required Interval: <u>Annual</u> Internal PM procedure #: <u>N/A</u>	☐ No Preventive Maintenance For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance

Requirements or Rationale for No PM:

PM checklist:

- 1) Check the pump pipeline for excessive wear. Turn the pump wheel by hand to check the overall flexibility.
- 2) If the bearing has dry wear, add bearing lubricating oil to the bearing body, as needed.
- 3) Check whether the tightening bolts on the pump head of the peristaltic pump are tightened. Tighten bolts, as needed.

Approval		
	Signature	Date
Quality:		14 Sep 2022
Engineering:		14SEPT2022