


	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-011	Equipment Description:	Fume Hood
Model #:	XPM04060	S/N:	N/A
Manufacturer:	Johnson Controls	Operating Range:	Flow rate: >350 CFM
Type (check one):	☒ Production / ☐ Engineering / ☐ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☒ Calibration Required	☐ Limited Calibration	☐ No Calibration
Internal / External (circle one) Interval: <u>Annual</u> Internal calibration procedure #: <u>N/A</u>	For limited calibration, document nature of limited calibration (how it can / cannot be used)	For no calibration required, document rationale for reason(s) equipment does not require calibration
Requirements or Rationale for No Calibration: Annually, confirm the flow rate setting meets specification. Flow rate shall be set/calibrated with drape curtains fully closed. Adjust baffling/flow, as needed.		
Preventive Maintenance Requirements (check type of calibration):		
☐ Preventive Maintenance Required	☒ No Preventive Maintenance	
Interval: _____ Internal PM procedure #: _____	For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
Requirements or Rationale for No PM: The fume hood is 10' x 10' galvanized steel hood with a 14" diameter exhaust outlet. There is no preventive maintenance on the fume hood or the flow controller required to ensure the fume hood operates as intended.		

Approval		
	Signature	Date
Quality:	Bryan Wong 	07 Nov 2022
Engineering:	Brian McVerry 	08 Nov 2022