

	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-014	Equipment Description:	UV/Visible Spectrophotometer
Model #:	7205	S/N:	85652
Manufacturer:	Jenway	Operating Range:	198 - 800 nm
Type (check one):	☐ Production / ☐ Engineering / ☒ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☐ Calibration Required Internal / External (circle one) Interval: <u> N/A </u> Internal calibration procedure #: <u> N/A </u>	☐ Limited Calibration For limited calibration, document nature of limited calibration (how it can / cannot be used)	☒ No Calibration For no calibration required, document rationale for reason(s) equipment does not require calibration
Requirements or Rationale for No Calibration: Equipment quality control qualification is performed before use with the VIS Photometric Accuracy Calibration Standard (Part Number: TF.SLQ001) in accordance with document QC-E.SLQ015.		
Preventive Maintenance Requirements (check type of calibration):		
☐ Preventive Maintenance Required Interval: <u> N/A </u> Internal PM procedure #: <u> N/A </u>	☒ No Preventive Maintenance For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
Requirements or Rationale for No PM: There are no preventive maintenance requirements recommended by the manufacturer. The equipment external surfaces should be kept clean and free from dust. Refer to the equipment manual for routine maintenance, cleaning, and lamp replacement instructions.		

Approval		
	Signature	Date
Quality:		11 Oct 2023
Engineering:		11 Oct 2023