



Date of Assessment: January 16th & 17th 2025

Client Name: SILQ Technologies Corporation **Client Address:** 323 Sunny Isles Blvd. (Seventh Floor); Sunny Isles Beach, FL 33160

Company Audited: SILQ Technologies Corporation **Address:** 323 Sunny Isles Blvd. (Seventh Floor); Sunny Isles Beach, FL 33160

Telephone Number: 786-590-5300

Name of Senior Company Official: D. V. Sharma

Title: CEO / President

Name of Head of Quality: Bryan Wong

Title: Director of QA

Name of Head of Regulatory: Sigi Caron

Title: VP of RA & Clinical Affairs

Is Organization Chart Available? Yes
(Attach Copy of Org Chart)

Number of Employees at this Location: Twelve (12)

Number in QA: Three (3)

Number of Buildings: Two (2) **Approximate Square Footage:** Sunny Isles (office space only) / Chico, CA (custom components)

Number of Years at this Location: Six (6)

Name of Registrar/Notified Body: N/A

Certificate Number(s): N/A
(Attach Copy)

Expiration Date(s): N/A

FDA Facility Registration Number: 3014650073
(Attach Copy of Registration)

Date of Last Inspection? Jan 4th 2023

Number of 483's: Zero (0)

FDB License (CA only): N/A

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	4 Quality Management System		
	4.1 General Requirements		
4.1 – 1	Has Organization established, documented, implemented and maintained a QMS and maintained (continually improved) its effectiveness in accordance with ISO 13485? Version?	Yes	As scripted, the Quality Management System (QMS) implemented by SILQ complies with ISO 13485:2016 and 21 CFR, Part 820 requirements
4.1 – 2	Where has the organization identified the processes needed for the QMS and their application throughout the organization?	Yes	Processes employed by SILQ, in support of quality and product realization activities have been implemented, where applicable.
4.1 – 3	Have all roles within the organization been adequately documented (e.g. manufacturer, EU Rep, etc.)?	Yes	SILQ has implemented a current organizational chart; QM.SLQ034.E – Organization Chart.
4.1 – 4	Has the establishment applied a risk-based approach for all processes employed in support of the QMS?	Yes	Two documents have been implemented in support of risk management activities: (a) QM.SLQ012.B – Risk Management; and (b) QM.SLQ013.B – Risk Analysis.
4.1 – 5	Has each QMS process been adequately defined: (a) criteria; (b) method for monitoring and measuring; (c) availability of resources; and (d) collection and maintenance of records?	Yes	Quality processes needed to ensure the QMS is fully-functional have been implemented.
4.1 – 6	When the establishment outsources processes the impact product conformity, are the controls instituted in-line with the risk to product, regulatory requirements, or the QMS.	Yes	The following processes are outsourced: • Contract Manufacturing; • Calibration; • Sterilization; • Inventory Management; • Distribution.
4.1 – 7	As the establishment scripted procedures for the validation and application of computer software used in support of the QMS?	Yes	File Hold Software, implemented in support of the QMS, has been validated.
	4.2 Documentation Requirements		
	4.2.1 General		

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
4.2 – 1	Does the organization have documented statements of a quality policy and quality objectives?	Yes	The SILQ Quality Policy has been documented within QM.SLQ035.C – Quality Policy. <i>“Excellent patient care is the focus of all SILQ activities. We endeavor to meet or exceed our customers’ satisfaction and expectations in providing safe and effective coated medical devices of the highest quality and that meets all applicable regulatory requirements.”</i> <i>“Each SILQ employee is personally committed to achieving this goal and continually improving our ability to maintain customer focus by operating a highly effective quality system that meets the requirements of applicable global standards and regulatory requirements for medical devices”.</i>
4.2 – 2	Does the organization have the documented procedures required by ISO 13485?	Yes	As previously noted, SILQ has implemented policies and procedures required by ISO 13485:2016 and 21 CFR, Part 820.
4.2 – 3	Are adequate documents in place to ensure the effective planning, operation and control of the organization’s processes?	Yes	Quality planning activities are performed in accordance with QM.SLQ025.A – Quality Planning.
4.2.2 Quality Manual			
4.2 – 4	Does the organization have a quality manual?	Yes	A quality manual has been implemented by SILQ; QM.SLQ027.D – Quality Manual.
4.2 – 5	Where in the quality manual is the scope of the QMS identified, including details of and justification for exclusions and/or requirements that don’t apply?	Yes	In accordance with 2. 2 of QM.SLQ027.D; the following ISO 13485:2016 clauses have been identified as Not Applicable: • Clause 7.5.3 – Installation Activities; • Clause 7.5.4 – Servicing activities; • Clause 7.5.9.2 – Particular Requirements for Implantable Medical Devices; and • Clause 7.5.10 – Customer Property.
4.2.3 Medical Device File			
4.2 – 6	Has the establishment created a medical file for each product or each product family?	Yes	In accordance with 4.2.5 of QM.SLQ027.D; the Device Master Record (DMR) is considered synonymous with the Medical Device File (MDF).

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
4.2 – 7	Does the Medical Device File contain: • General description; • Intended use; • Labeling; • IFU; • Product specification; • Specifications/procedures for manufacturing, packaging, storage, handling, & distribution; • Procedures for measuring and monitoring; and • As appropriate procedures for installation and servicing.	Yes	In support of this assessment, the following DMR(s) were reviewed (Pathway Medtech, LLC): • SLQ-211410SPT – SILQ 2-Way Cleartract SPT (14FR); • SLQ-211605SPT – SILQ 2-Way Cleartract SPT (16FR); and • SLQ-211810SPT – SILQ 2-Way Cleartract SPT (18FR).
4.2.4 Control of Documents			
4.2 – 8	How are the documents required by the QMS controlled?	Yes	Two procedures have been implemented in support of document control activities: • QM.SLQ001.A – Document Control; and • QM SLQ014.B – Electronic Document System Work Instruction.
4.2 – 9	Does the documented procedure that defines the controls for <u>each</u> of the following? • Review and approve documents for adequacy prior to issue? • Review and update as necessary and re-approve documents? • Ensure that changes and the current revision status of documents are identified? • Ensure that relevant versions of applicable documents are available at points of use? • Ensure that documents remain legible and readily identifiable? • Ensure that documents of external origin are identified and their distribution controlled? • Prevent the unintended use of obsolete documents, and to apply suitable identification to them if they are retained for any purpose.	Yes	In support of this assessment, the following DCOs were reviewed and found to be acceptable: • DCO072 (SOP); • DCO073 (SOP); • DCO074 (SOP); • DCO075 (SOP); • DCO076 (Specification); • DCO077 (WIs); • DCO078 (Test Protocol); • DCO079 (SOP); • DCO080 (Test Protocol); • DCO081 (Test Protocol); and • DCO082 (Specification). <i>Note: For 2024, there were eleven (11) DCOs processed.</i>

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	4.2.5 Control of Records		
4.2 – 10	What records exist that provide evidence of conformity to requirements and of the effective operation of the QMS?	Yes	The following types of records were reviewed in support of this assessment: • FDA Establishment Registration; • Quality Manual; • Quality Policy; • Organizational Chart; • Quality Objectives; • SOPs; • Work Instructions; • DCOs; • DMRs; • Training; • Job Descriptions; • Management Review; • DHF; • DHRs; • Risk Management; • Supplier Certificates; • Supplier Questionnaires; • Supplier Audit; • Calibration; • Complaints; and • NCRs.
4.2 – 11	Are records legible, readily identifiable and retrievable?	Yes	Good Documentation Practice (GDP) is being managed in accordance with QM.SLQ002.B – Good Documentation Practices.
4.2 – 12	Does the organization have a documented procedure for defining the controls needed for the identification, storage, protection, retrieval, retention time and disposition of records?	Yes	In accordance with 19.1 of QM.SLQ001.A; records are retained electronically through the use of File Hold Software. In accordance with 19.6 of QM.SLQ001.A; the record retention period is 5-years.
	5 Management Responsibility		
	5.1 Management Commitment		
5.1 – 1	How does top management communicate the importance of meeting customer and legal requirements to the organization?	Yes	The importance of meeting customer requirements is communicated by the CCO, in accordance with 5.2 of QM.SLQ036.E – Sales Order SOP.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
5.1 – 2	What are the quality objectives established by top management?	Yes	In accordance with QM.SLQ037.A – Quality Objectives, objectives for 2024/25 are stated as: - Operations / Incoming quality / Quality Assurance management / The lot acceptance rate of incoming material used in finished product shall be greater than 80%. - Operations / Finished product quality / Chief Technology Officer / The complaint rate of failed catheters shall be less than 5% of the overall distributed catheters.
5.1 – 3	Does top management conduct management reviews?	Yes	Management reviews are managed in accordance with QM.SLQ018.A – Management Review. In accordance with 6.1.1 of QM.SLQ018.A; Management Review is to be held annually, at a minimum.
5.2 Customer Focus			
5.2 – 1	How does top management ensure that customer requirements are determined and met?	Yes	The importance of meeting customer requirements is communicated by the CCO, in accordance with 5.2 of QM.SLQ036.E – Sales Order SOP
5.3 Quality Policy			
5.3 – 1	Has a company quality policy been established?	Yes	The SILQ Quality Policy has been documented within QM.SLQ035.C – Quality Policy. <i>“Excellent patient care is the focus of all SILQ activities. We endeavor to meet or exceed our customers’ satisfaction and expectations in providing safe and effective coated medical devices of the highest quality and that meets all applicable regulatory requirements.”</i> <i>“Each SILQ employee is personally committed to achieving this goal and continually improving our ability to maintain customer focus by operating a highly effective quality system that meets the requirements of applicable global standards and regulatory requirements for medical devices”.</i>
5.3 – 2	Is the quality policy: - Applicable to organization purpose; - A commitment to comply requirements; - Provide a framework for establishing quality objectives; - Communicated and understood throughout the organization; and - Reviewed for continued suitability.	Yes	As scripted, SILQ quality policy is appropriate for the organization and meets the intent of 21 CFR, Part 820 and ISO 13485:2016 requirements.
5.4 Planning			
5.4.1 Quality Objectives			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
5.4 – 1	Has top management established quality objectives (including those needed to meet requirements for product) at relevant functions and levels within the organization?	Yes	In accordance with QM.SLQ037.A – Quality Objectives, objectives for 2024/25 are stated as: - Operations / Incoming quality / Quality Assurance management / The lot acceptance rate of incoming material used in finished product shall be greater than 80%. - Operations / Finished product quality / Chief Technology Officer / The complaint rate of failed catheters shall be less than 5% of the overall distributed catheters.
	5.4.2 Quality Management System Planning		
5.4 – 2	How does the organization ensure that the integrity of the QMS is maintained when changes to the QMS are planned and implemented?	Yes	QMS planning is managed in accordance with QM.SLQ025.A – Quality Planning.
	5.5 Responsibility, Authority and Communication		
	5.5.1 Responsibility and Authority		
5.5 – 1	How are responsibilities and authorities defined, documented and communicated within the organization?	Yes	In accordance with 7.5.1 of QM.SLQ027.D; responsibility and authority are delineated through the use of the SILQ Organizational Chart (QM.SLQ037).
	5.5.2 Management Representative		
5.5 – 2	Who is the management representative for quality?	Yes	In accordance with QM.SLQ034.E – Organization Chart; Tavi Cruz-Urbe (VP of QA) is the designated management representative.
	5.5.3 Internal Communication		
5.5 – 3	How is information regarding the effectiveness of the QMS communicated within the organization?	Yes	In accordance with 7.5.3 of QM.SLQ027.D; <i>“SILQ ensures effective communication between its various levels and functions regarding the processes of the quality management system and their effectiveness, suitability, and adequacy through periodic company meetings and Management Reviews. It is also the responsibility of each Manager or department head to ensure all relevant information is passed on to their functions.”</i>
	5.6 Management Review		
	5.6.1 General		
5.6 – 1	What is the frequency of management reviews?	Yes	Management reviews are managed in accordance with QM.SLQ018.A – Management Review. In accordance with 6.1.1 of QM.SLQ018.A; Management Review is to be held annually, at a minimum.
5.6 – 2	Does the organization have documented evidence that management reviews are being held in accordance with the published schedule?	Yes	The last management review was held on 02 February 2024 for FY 2023.
	5.6.2 Review Input		
5.6 – 3	What kind of information is input into management reviews?	Yes	Management Review Input is managed in accordance with 6.3 of Q.SLQ018.A.
	5.6.3 Review Output		

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
5.6 – 4	What decisions, actions, and/or outcomes have resulted from management reviews?	Yes	Management Review Output is managed in accordance with QM.SLQ018.A.
	6 Resource Management		
	6.1 Provision of Resources		
6.1 – 1	What resources has the organization provided to implement and maintain the QMS and continually improve its effectiveness?	Yes	SILQ has implemented an effective quality organization (three team members) in support of the establishment's quality.
6.1 – 2	What resources has the organization provided to ensure that customer and regulatory requirements are met?	Yes	SILQ has retained a qualified VP of Regulatory and Clinical Affairs to support Regulatory Affairs.
	6.2 Human Resources		
6.2 – 1	What are the education, training, skills and experience required by this job/task?	Yes	In accordance with 8.2 of QM.SLQ027.D; “Senior Management strives to ensure personnel performing work affecting product quality are competent on the basis of appropriate education, training, skills, and experience.” “Personnel performing work affecting product quality are selected based on required competencies, appropriate education, training (internal and/or external), skills, and experience.” “Management identifies training needs and provides training for all personnel performing tasks affecting quality as required. Training and effectiveness of training is conducted in accordance with established procedures by qualified individuals. Appropriate records of education, experience, training, and skills are maintained.”
6.2 – 2	How does the organization determine the necessary education, training, skills and experience for people performing work affecting product quality?	Yes	In accordance with 8.2 of QM.SLQ027.D; <i>“personnel performing work affecting product quality are selected based on required competencies, appropriate education, training (internal and/or external), skills, and experience.”</i>
6.2 – 3	What training or other actions does the organization provide to satisfy the needs of personnel?	Yes	Employee training is managed in accordance with QM.SLQ003.B – Employee Training.
6.2 – 4	When the organization provides training or other actions to satisfy competence needs, how does it evaluate the effectiveness of those actions?	Yes	In accordance with 6.1 of QM.SLQ003.B; <i>“all SILQ personnel and contractors / consultants are to be determined competent to perform their job based on appropriate education, experience, skills and training. Management and at least one other departmental representative conduct candidate interviews and complete reference checks as applicable. Documentation of the employee's and contractor's / consultant's qualifications (e.g., resume) is filed in the respective employee file.”</i>

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
6.2 – 5	Where does the organization maintain records of education, training, skills and experience?	Yes	Training records are managed in accordance with Section 10 of QM.SLQ003.B. In support of this assessment, the following training records were reviewed: • C. Spiridon – QM.SLQ027.D; • C. Spiridon – QM-SLQ015.C; • C. Spiridon – QMSLQ037.A; • E. Rao – QM.SLQ027.D; • E. Rao – QM-SLQ015.C; • E. Rao – QMSLQ037.A; • C. Turner – QM.SLQ027.D; • C. Turner – QM-SLQ015.C; • C. Turner – QMSLQ037.A; • T. Downey – QM.SLQ027.D; • T. Downey – QM-SLQ015.C; and • T. Downey – QMSLQ037.A. The following job descriptions were reviewed in support of this assessment: • Administrative Director – C. Spiridon; • Director/Manager R & D – E. Rao; • Director/Manager Manufacturing – C. Turner; and • Chief Financial officer – T. Downey.
	6.3 Infrastructure		
6.3 – 1	Are the buildings, workspace, and utilities appropriate to meet product requirements?	NR	Not Reviewed: This assessment was remote; therefore, the facility and equipment were not assessed during this audit. <i>Note: the primary facility is essentially office space.</i>
6.3 – 2	Is the process equipment (hardware & service) appropriate for the manufacture of medical devices that are safe and effective in their intended use?	NR	Not Reviewed: This assessment was remote; therefore, the facility and equipment were not assessed during this audit. <i>Note: the primary facility is essentially office space.</i>
6.3 – 3	Are the requirements for maintenance activities being appropriately documented and the records maintained?	NR	Not Reviewed: This assessment was remote; therefore, the facility and equipment were not assessed during this audit. <i>Note: the primary facility is essentially office space.</i>
	6.4 Work Environment & Contamination Control		
	6.4.1 Work Environment		
6.4 – 1	What kind of work environment is required to achieve conformity to product requirements?	Yes	Manufacturing occurs at a critical subcontractor facility. The manufacturing environment is required to be in a Clean Room (ISO 14644, Class 7). Note: The gowning room is also managed as a Class 7.
6.4 – 2	How is this environment managed and maintained?	Yes	Manufacturing occurs at a critical subcontractor facility. The manufacturing environment is required to be in a Clean Room (ISO 14644, Class 7). Note: The gowning room is also managed as a Class 7.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
6.4 – 3	Are requirements for employee health, cleanliness, and clothing been documented in a procedure?	Yes	In accordance with 8.5 of QM.SLQ027.D; “CMOs that provide product or services are selected based on their ability to fulfill applicable requirements for work environment and contamination control.”
6.4.2 Contamination Control			
6.4 – 4	Has the organization established a procedure for the control of contaminated product?	N/A	Not Applicable. SILQ does not handle contaminated product.
6.4 – 5	Does the organization have a CER? If so, does the organization have procedures that define the monitoring and measuring of environmental conditions that may impact finished medical devices (i.e., particle counts, temperature, RH differential pressure, surface contamination, & product bio-burden)?	N/A	Not Applicable. SILQ does not maintain a Controlled Environment Room (CER).
7 Product Realization			
7.1 Planning of Product Realization			
7.1 – 1	Where are the processes needed for product realization identified?	Yes	Process requirements for successful product realization are delineated within the DMR.
7.1 – 2	Is the planning of product realization consistent with the requirements of the other processes of the QMS?	Yes	Product realization activities are performed by Pathway Medical. Pathway’s QMS is aligned with the requirements delineated within ISO 13485:2016 and 21 CFR, Part 820.
7.1 – 3	Where in the product realization process does the organization determine the quality objectives and requirements for products?	Yes	Objectives and requirements are defined within the DMR.
7.1 – 4	When planning for product realization, how does the organization establish processes, documents, and provide resources specific to the product?	Yes	Processes needed for successful product realization are defined within the DMR and executed by Pathway.
7.1 – 5	How does the organization determine verification, validation, monitoring, inspection and test activities specific to the product, and the criteria for product acceptance?	Yes	Requirements for verification, validation, monitoring, inspection, and test activities, necessary to facilitate product realization, are developed in conjunction with Pathway
7.1 – 6	What records exist depicting that both the realization processes and the product meet requirements?	Yes	A Device History Record (DHR) is compiled for each lot manufactured by Pathway.
7.1 – 7	What are the outputs of product realization planning?	Yes	The output of product realization is a finished SILQ device that is safe and effective in its intended use.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	7.2 Customer-Related Processes		
	7.2.1 Determination of Requirements Related to the Product		
7.2 – 1	How does Organization determine <u>each</u> of the following requirements? - Customer Specified - Legal & Regulatory - Organizational	Yes	Requirements for SILQ devices are determined during the design and development.
	7.2.2 Review of Requirements Related to the Product		
7.2 – 2	What kind of review is performed to ensure that the organization has the ability to meet requirements before committing to supply product?	Yes	Customer orders and contracts are reviewed in accordance with 6.3 of QM.SLQ036.E – Sales Order.
7.2 – 3	How does the organization ensure that product requirements are defined, documented, and reviewed before committing to supply product?	Yes	In accordance with 6.1.4 of QM.SLQ036.E; <i>“if there are any requirements not stated by the customer but necessary for specified or intended use, a SILQ associate will confirm these requirements with the customer before accepting the order.”</i>
7.2 – 4	How does the organization ensure that contract or order requirements differing from those previously expressed are resolved before committing to supply product?	Yes	In accordance with 6.3.1.3 of QM.SLQ036.E; when changes are noted in customer orders differing from previously accepted orders, these changes are required to be resolved prior to order acceptance.
7.2 – 5	Does the organization have records of the product-requirement review results and actions resulting from them?	Yes	Product requirements/order acceptance is managed through the use of sales-order forms.
7.2 – 6	When customers do not have documented requirements, how does the organization confirm their requirements before accepting orders?	Yes	Requirements reviewed as part of the sales order review process are delineated within 6.1.2 of QM.SLQ036.E.
7.2 – 7	When product requirements are changed, how does the organization ensure that relevant documents are changed and that relevant personnel are made aware of the changes?	Yes	Changes made to product(s) are managed as part of the design and development change process.
	7.2.3 Customer Communication		
7.2 – 8	What method(s) are used to communicate with customers regarding - Product information? - Enquiries, contracts, or order handling, including amendments? - Feedback, including customer complaints? - Advisory notices?	Yes	In accordance with 6.1.1 of QMSLQ036.E; <i>“if the customer requests a product that is a SILQ current part number, a Sales/Customer Service associate (or designee) receives the customer's order from phone, fax, or other means and initiates the sales order.”</i>

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	7.3 Design and Development		
	7.3.1 General		
7.3 – 1	Has the organization established procedures in support of design and development activities?	Yes	The following SOPs have been implemented in support of design and development activities: • QM.SLQ004.A – Design Control Program; • QM.SLQ005.B – Design Project Planning; • QM.SLQ006.A – Design Input; • QM.SLQ007.A – Design Output; • QM.SLQ008.A – Design Review; • QM.SLQ009.A – Design Verification & Validation; • QM.SLQ010.A – Design Transfer; • QM.SLQ011.A – Statistical techniques – Work Instruction; and • QM.SLQ012.B – Risk Management.
	7.3.2 Design and Development Planning		
7.3 – 2	Does the organization have a documented procedure for Design & Development Planning?	Yes	SILQ has implemented QM.SLQ005.B – Design Project Planning, in support of design and development activities. <i>Note: SILQ has one Class 2 device cleared by FDA (reference K192034).</i>
7.3 – 3	How does Organization ensure effective communication and clear assignment of responsibility between different groups involved in design and development?	Yes	Project managers are tasked with communicating roles, responsibilities, and project requirements to the project team, as defined within 5.3 of QM.SLQ005.B – Design Project Planning.
7.3 – 4	As product design and development progresses, how are the planning outputs documented and updated?	Yes	Design outputs are documented within the DHF.
	7.3.3 Design and Development Inputs		
7.3 – 5	What are the design inputs the organization typically considers?	Yes	Requirements for design inputs are delineated in QM.SLQ006.A – Design Input.
7.3 – 6	When identified, where are the design and development inputs recorded?	Yes	Design inputs are recorded in the DHF and incorporated into product specification(s). In support of this audit, PS-0006. Rev G – Product Specification / Traceability Matrix – SILQ Foley Products (Pathway) was reviewed.
7.3 – 7	How and when are the design and development inputs reviewed and approved for adequacy?	Yes	Design and development inputs are reviewed during the design review process.
7.3 – 8	How does the organization ensure that requirements are complete, unambiguous and don't conflict with each other?	Yes	In accordance with 1.2 of QM.SLQ006.A; requirements for the resolution of incomplete, ambiguous, or conflicting product requirements are defined.
	7.3.4 Design and Development Outputs		
7.3 – 9	How are design and development outputs verified against the design inputs?	Yes	Design outputs are managed in accordance with QM.SLQ007.A – Design Output.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.3 – 10	Are design and development outputs approved prior to their release?	Yes	Design outputs are identified during design phase reviews.
7.3 – 11	What outputs include information for purchasing, production and service?	Yes	Purchasing will utilize design outputs in support of identifying potential requirements and suppliers needed for successful product realization.
7.3 – 12	Where is the product acceptance criteria specified?	Yes	Acceptance criteria are documented in the product specification and in inspection documentation.
7.3 – 13	Where are product characteristics needed for safe and proper use specified?	Yes	Product characteristics needed for the safe and effective use of SILQ devices are placed into the Instructions for Use (IFU).
7.3 – 14	Where are records of design and development maintained?	Yes	Records of design and development activities are retained in the Design History File (DHF).
7.3.5 Design and Development Review			
7.3 – 15	At what stages of design and development does the organization perform reviews to evaluate if the results meet requirements?	Yes	Design Reviews are being pursued in accordance with QM.SLQ008.A – Design Review.
7.3 – 16	What functions (including specialists) are represented at these reviews?	Yes	In accordance with 6.6 and 6.7 of QM.SLQ008.A; “At a minimum, a representative from Quality Assurance, R&D/Engineering, and Manufacturing must attend the design review. When appropriate, additional representatives from other departments or other subject matter experts are to attend if requested or deemed necessary.” “An independent reviewer, not directly associated with the design stage being reviewed, is to attend all design reviews. The independent reviewer is to be identified by Quality Assurance management.”

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.3 – 17	Where are the records associated with the Design and Development Reviews maintained?	Yes	In support of this assessment, the following phase review documentation was reviewed (Project 00053): • TR-SLQ12, Rev A – Test report, Biocompatibility, SLQ Clear Tract Catheter (genotoxicity & Bacterial Endotoxin Testing / ISO 10993-3); • Design Transfer – Commercial Release Review 05-26-23; • Project 00053, Change Assessment (05-13-23); • SILQ-4007, Rev A – IFU 2-Way 100% Silicone Clear Tract SPT Catheter; • SLQ-LB-4008, Rev A Label Pouch SILQ Clear Tract SPT Catheter; • RM-0021, Rev F – PFMEA SILQ Clear Tract Foley Catheter; • PS-0006, Rev G – Product Specification / Traceability Matrix – SILQ Foley Products (Pathway) was reviewed; • RM-0019, Rev D - Hazard Analysis (HA) – Risk Management Report, SILQ Clear Tract Foley Catheter; • RM-0018, Rev D - Risk Management Plan – SILQ Foley Catheter
7.3.6 Design and Development Verification			
7.3 – 18	What verification activities are performed to ensure that the design and development outputs have met the input requirements?	Yes	Design Verification (DV) activities are pursued in accordance with QM.SLQ009.A – Design Verification & Validation.
7.3 – 19	Are records containing the results of verification activities and their resulting actions collected and maintained?	Yes	DV records (protocols & reports) are reviewed and approved by SILQ and executed by Pathway.
7.3.7 Design and Development Validation			
7.3 – 20	What design and development validation activities are performed to ensure that the product is capable of meeting the requirements for the intended use?	Yes	Design Validation (DVAL) activities are performed in accordance with QM.SLQ009.A – Design Verification & Validation. <i>In accordance with 7.2 of QM.SLQ009.A; “Design validation activities are intended to provide objective evidence that the final product, when manufactured according to the Device Master Record, conforms to established user needs and intended use. Design validation testing must be performed on pre-production or production units, lots, batches or their equivalent. If design validation testing is not performed on initial production, the equivalence to initial production must be documented.”</i>
7.3 – 21	Do records show that validation is performed prior to product shipment?	Yes	DVAL records (protocols & reports) are reviewed and approved by SILQ and executed by Pathway.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.3 – 22	If not, is the justification recorded?	Yes	In accordance with 7.9 of QM.SLQ009. A; “When testing is completed, a final report is written to include the data analysis, results, conclusions and any deviations. Attach all original data sheets and any attachments required to support the test data and conclusions reached in the final report.” “The completed final report is reviewed and a decision is made as to whether unit under test met the acceptance criteria.”
7.3.8 Design and Development Transfer			
7.3 – 23	Does the organization have a documented procedure for Design & Development Transfer?	Yes	Design transfer activities are performed in accordance with QM.SLQ010.A – Design Transfer
7.3.9 Control of Design and Development Changes			
7.3 – 24	How are design and development changes identified?	Yes	Design changes are performed in accordance with section 16.1 through 16.4 of QM.SLQ004.A.
7.3 – 25	Where are the records maintained?	Yes	Pathway is managing design control activities. SILQ retains access to the DHFs.
7.3 – 26	Are changes reviewed, verified, validated, and approved before implementation?	Yes	In accordance with 16.2 of QM .SLQ004.A; “in cases where major changes to a product line are made, an updated or supplemental Design Project Scope Form (FM1-QM.SLQ004) is to be completed to determine the extent of design control activities required.”
7.3 – 27	How does the organization document the review of design and development changes including evaluation of the effect on component parts and products in the field?	Yes	In accordance with 16.4 of QM.SLQ004.A; “design changes are to be handled through the document change order process as described in the QM.SLQ001, Document Control SOP.”
7.3.10 Design and Development Files			
7.3 – 28	Does the organization have a documented procedure for managing design and development files (e.g., DHF)?	Yes	The DHFs are managed in accordance with 15.1 through 15.3 of QM.SLQ004.A.
7.4 Purchasing			
7.4.1 Purchasing Process			
7.4 – 1	Does the organization have a written procedure for the Purchasing Process?	Yes	Purchasing activities are managed in accordance with QM.SLQ020.D – Purchasing Controls.
7.4 – 2	How does the organization ensure that purchased product conforms to specified requirements?	Yes	Receiving inspection activities are pursued in accordance with 9.2 of QM.SLQ020.D.
7.4 – 3	How does the organization determine the type and extent of control applied to the supplier and the purchased product?	Yes	In accordance with 7.7 of QM.SLQ020.D; “products and services are only to be purchased from suppliers, contractors and consultants listed on the Approved Supplier List (ASL) and only for those product and service types for which they have been qualified (as specified on the ASL).”

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.4 – 4	How does the organization evaluate and select suppliers?	Yes	SILQ suppliers are evaluated and selected in accordance with QM.SLQ015.B – Supplier Quality Assurance.
7.4 – 5	Is the criterion for supplier selection, evaluation and re-evaluation of suppliers documented by procedure?	Yes	In accordance with 7.1 through 7.5 of QM.SLQ015.B; suppliers are placed into one of five categories promised on risk. In support of this assessment the following suppliers were reviewed: • Pathway, LLC (Cat 1): ISO 13485:2016 (08-15-25); • Richman Chemical (Cat 1): ISO 9001:2015 (04-01-27); • Steripax (Cat 1): ISO 9001:2015 (11-12-25); • Micro Precision (Cat III): ISO/IEC 17025:2017 (10-31-26); and • Repligen (Cat III): ISO 9001:2015 (01-23-26).
7.4 – 6	What type of records are collected and maintained in support of supplier evaluation and management activities?	Yes	In support of this assessment, the following purchasing/supplier quality records were reviewed: • ASL (FM-6 QMS.SLQ015.B – Approved Suppliers List); • Supplier Audit Report (Pathway 11/28 & 11/28 23); • ISO Certificates; • Supplier Questionnaires; and • Quality Agreements.
7.4.2 Purchasing Information			
7.4 – 7	Do orders/contracts include requirements for approval of product, procedures, processes and equipment?	Yes	In accordance with 7.1 of QM.SLQ020.B; purchase orders are used, reference: <i>“purchase Orders (P.O.) constitute a contract between SILQ and a supplier, contractor or consultant and may be issued only by authorized purchasing personnel, with department specific approvals as described in this procedure.”</i>
7.4 – 8	Does the organization require special qualifications for supplier personnel?	Yes	Special supplier requirements, including supplier qualifications, will be delineated through the use of purchase orders, contracts, drawings, and relevant attachments.
7.4 – 9	If special qualifications are required, where is the requirement documented?	Yes	Special supplier requirements, including supplier qualifications, will be delineated through the use of purchase orders, contracts, drawings, and relevant attachments.
7.4 – 10	Does the organization have QMS requirements for their suppliers?	Yes	Special supplier requirements, including supplier qualifications, will be delineated through the use of purchase orders, contracts, drawings, and relevant attachments.
7.4.3 Verification of Purchased Product			
7.4 – 11	What inspection or other activities are used to ensure that purchased product meets the organization's purchasing requirements?	Yes	When required, receiving inspection is performed in accordance with QM.SLQ039.B – Receiving Inspection.
7.4 – 12	Does the organization ever perform product verification (source inspection) at the supplier's site?	N/A	Not Applicable. SILQ retains the right of access for their supplier base; however, SILQ does not typically perform source inspection.
7.5 Production and Service Provision			
7.5.1 Control of Production and Service Provision			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.5 – 1	When carrying out production or service what controlled conditions in place? - Documented Procedures - Work Instructions - Suitable Equipment - Gages - Measuring Equipment - Packaging & Labeling	Yes	Pathway is tasked with the application of appropriate tooling to ensure the success of product realization.
7.5 – 2	Is batch management performed in support of product traceability requirements?	Yes	In accordance with 7.6 of QM.SLQ019.C – Identification and Traceability; <i>“batch or lot numbers that have been assigned by the manufacturer (supplier) are acceptable to maintain and reference in the DHR, without having to assign a new identifying number.”</i> In support of this assessment, the following DHR was reviewed; SLQ-810205125241 (expires 0-5-15-27).
7.5.2 Cleanliness of Product			
7.5 – 3	Does the organization have established and documented requirements for cleanliness of product?	Yes	Pathway is tasked with performing product realization activities in accordance with the SILQ DMRs.
7.5 – 4	Are products cleaned prior to sterilization?	Yes	SILQ products are assembled in a Class 7 Cleanroom by Pathway. SILQ finished devices are cleaned prior to shipment to sterilization through the use of EO.
7.5.3 Installation Activities			
7.5 – 5	If appropriate, does the organization have established documented requirements which contain acceptance criteria for installing and verifying the installation of the medical device?	N/A	Not Applicable. Clause 7.5.3 has been declared as “N/A” in the quality manual (QM.SLQ027.D).
7.5 – 6	If the agreed upon customer requirements allow installation to be performed other than by the organization or its authorized agent, does the organization provide documented requirements for installation and verification?	N/A	Not Applicable. Clause 7.5.3 has been declared as “N/A” in the quality manual (QM.SLQ027.D).
7.5 – 7	Where are the records of installation and verification performed by the organization or its authorized agent maintained?	N/A	Not Applicable. Clause 7.5.3 has been declared as “N/A” in the quality manual (QM.SLQ027.D).
7.5.4 Servicing Activities			
7.5 – 8	If servicing is a specified requirement, does the organization have established and documented procedures, work instructions, reference materials, and reference measurement procedures, as necessary, for performing servicing activities?	N/A	Not Applicable. Clause 7.5.4 has been declared as “N/A” in the quality manual (QM.SLQ027.D).

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.5 – 9	Where are the records of servicing activities, carried out by the organization, maintained?	N/A	Not Applicable. Clause 7.5.4 has been declared as “N/A” in the quality manual (QM.SLQ027.D).
	7.5.5 Particular Requirements for Sterile Medical Devices		
7.5 – 10	Does the organization maintain records of the process parameters for the sterilization process, which is employed for each sterilization batch?	Yes	Sterilization of SILQ product is performed using EO. Sterilization records are managed through the use of a batch number.
7.5 – 11	Are sterilization records traceable to each production batch of medical devices?	Yes	Sterilization records are managed through the use of a batch number. The sterile load records are retained in the DHR.
	7.5.6 Validation of Processes for Production and Service Provision		
7.5 – 12	Does the organization have any production or service processes where the resulting output cannot be verified later?	Yes	Process Validation (PV) activities are performed (performed by Pathway) in accordance with QM.SLQ047.A – Process Validation.
7.5 – 13	Are records available that demonstrate that the validation performed has met the requirements?	Yes	There have been zero PVs performed since the previous internal Audit.
7.5 – 14	How are special processes reviewed and approved?	Yes	Processes employed in support of product realization are required to be validated or verified through inspection and/or test.
7.5 – 15	Are records of personnel and equipment qualification collected and maintained?	Yes	Personnel and equipment qualification information will be retained in the PV Reports. Note: PV reports require the approval of SILQ.
7.5 – 16	When changes are made to processes, how does the organization revalidate the changes?	Yes	Process changes are assessed for the need to repeat a partial or complete PV test.
7.5 – 17	Does the organization have established and documented procedures for the validation of the application of computer software?	Yes	Software validation is pursued in accordance with QM.SLQ032.A – Software Validation.
	7.5.7 Particular Requirements for Validation of Processes for Sterilization and Sterile Barrier Systems		
7.5 – 18	Has the organization established documented procedures for the validation of sterilization and sterile barrier systems?	Yes	TR-0075, Rev B – Sterilization Product Adoption, HDX 2-Way Foley Catheter <i>Note: there have been zero need(s) to update the adoption report in 2024. No significant material changes have been made to SILQ finished devices.</i>
7.5 – 19	Where are the records for validation of sterilization processes maintained?	Yes	TR-0075, Rev B – Sterilization Product Adoption, HDX 2-Way Foley Catheter <i>Note: there have been zero need(s) to update the adoption report in 2024. No significant material changes have been made to SILQ finished devices.</i>

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	7.5.8 Identification		
7.5 – 20	Where appropriate, does the organization identify the product by suitable means throughout product realization?	Yes	In accordance with 7.6 of QM.SLQ019.C – Identification and Traceability; <i>“batch or lot numbers that have been assigned by the manufacturer (supplier) are acceptable to maintain and reference in the DHR, without having to assign a new identifying number.”</i>
7.5 – 21	Does the organization have an established and documented procedure for product identification?	Yes	Identification and traceability are managed in accordance with QM.SLQ019.C – Identification and Traceability.
7.5 – 22	Does the organization have established and documented procedures to ensure that medical devices returned to the organization are identified and distinguished from conforming product?	Yes	SILQ medical devices can be returned to SILQ or Pathway for failure investigations. SILQ devices are returned to Pathway. Pathway will perform device decontamination and failure investigation. Upon completion of the investigation, Pathway will provide a copy of the failure reports.
	7.5.9 Traceability		
7.5 – 23	Does the organization have established and documented procedures for traceability?	Yes	Identification and traceability are managed in accordance with QM.SLQ019.C – Identification and Traceability.
7.5 – 24	Do the procedures define the extent of product traceability and the records required?	Yes	Finished device traceability is managed in accordance with 10.1 and 10.2 of QM.SLC019.C.
7.5 – 25	Where traceability is a requirement, does the organization control and record the unique identification of the product?	Yes	Pathway is tasked with the assignment of batch numbers.
7.5 – 26	In defining the records required for traceability, does the organization include records of all components, materials, and work environment conditions, if these could cause the medical device not to satisfy its specified requirements?	Yes	Records of materials procured for medical device manufacturing are managed in accordance with 7.1 of QM.SLQ019.C.
7.5 – 27	Does the organization require that its agents or distributors maintain records of the distribution of medical devices to allow traceability; and ensure these records are available for inspection?	Yes	There are two flows for the distribution of finished devices into commercialization: (a) Pathway; and (b) Health Products for You - Consignment.
7.5 – 28	Are records of the name and address of the shipping package consignee maintained?	Yes	A distribution log is maintained in support of tracking product distribution.
	7.5.10 Customer Property		
7.5 – 29	Does the organization exercise care with customer property while it is under the organization's control or being used by the organization?	N/A	Not Applicable. Clause 7.5.10 has been declared as “N/A” in the quality manual (QM.SLQ027.D).
7.5 – 30	Does the organization identify, verify, protect, and safeguard customer property provided for use or incorporation into the product?	N/A	Not Applicable. Clause 7.5.10 has been declared as “N/A” in the quality manual (QM.SLQ027.D).
7.5 – 31	If customer property is lost, damaged, or otherwise found to be unsuitable for use, does the organization report the issues to the customer?	N/A	Not Applicable. Clause 7.5.10 has been declared as “N/A” in the quality manual (QM.SLQ027.D).
	7.5.11 Preservation of Product		

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.5 – 32	Does the organization have established and documented procedures or documented work instructions for preserving the conformity of product during internal processing and delivery to the intended destination?	Yes	Handling, Storage, Packaging, Preservation and Delivery are managed in accordance with 12.6 of QM.SLQ027.D.
7.5 – 33	Does this preservation include identification, handling, packaging, storage, and protection?	Yes	Specific processes employed in support of product realization and instructions for the preservation of product are the responsibility of Pathway.
7.5 – 34	Does the organization have established and documented procedures or documented work instructions for the control of product with a limited shelf-life or requiring special storage conditions?	Yes	Material requiring shelf-life or other special handling and storage requirement are managed by Pathway.
7.5 – 35	Are special storage conditions controlled and recorded?	Yes	Material requiring shelf-life or other special handling and storage requirement are managed by Pathway.
7.6 Control of Monitoring and Measuring Devices			
7.6 – 1	How does the organization determine the measurements to be taken and the measuring equipment needed to demonstrate conformity with requirements?	Yes	Monitoring and measuring devices, needed to support product realization, are selected and managed by Pathway.
7.6 – 2	Does the organization have established and documented procedures to ensure that monitoring and measurement can be carried out and are carried out in a manner that is consistent with the monitoring and measurement requirements?	Yes	Calibration of monitoring and measuring devices is managed in accordance with QM.SLQ050.A – Calibration and Preventive Maintenance.
7.6 – 3	How does the organization ensure that measuring and test equipment is calibrated or verified proper frequencies with NIST traceable standards?	Yes	In accordance with 11.3 of QM.SLQ050.A; NIST traceability is a requirement.
7.6 – 4	What process is used to adjust or re-adjust measuring and test equipment when needed?	Yes	The adjusting and re-adjusting of monitoring and measuring devices will be managed by Pathway as part of their calibration system.
7.6 – 5	How are measuring tools identified to enable the calibration status to be determined?	Yes	In accordance with 11.1 of QM.SLQ050.A; measuring and monitoring equipment shall be identified with a calibration label. <i>“Calibration labels will include: calibration date, initials of person who performed the calibration, and calibration due date.”</i>
7.6 – 6	How does the organization safeguard measuring equipment from adjustments that would invalidate the measurement results?	Yes	Only SILQ personnel that have been trained to use monitoring and measuring devices shall be permitted to operate said devices.
7.6 – 7	How does the organization ensure that measuring its test equipment is protected from damage and deterioration during handling, maintenance and storage?	Yes	In support of preventing damage, the handling of monitoring and measuring devices is managed in accordance with 12.1 of QM.SLQ050.A.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.6 – 8	Does the organization assess and record the validity of the previous measuring results when the equipment is found not to conform to requirements?	Yes	The validity of calibration results will be verified in accordance with 11.9.1 of QM.SLQ050.A.
7.6 – 9	What action does the organization pursue when equipment and product are affected when equipment non-conformances are identified?	Yes	Monitoring and measuring devices found to be out-of-tolerance, non-conforming and/or other operational discrepancies will be managed in accordance with 11.9 of QM. SLQ050.A.
7.6 – 10	Where are the records documenting the results of calibration and verification maintained?	Yes	In support of this assessment, the following calibration records were reviewed: • ST-005 Repligen Diafiltration System: Due 11-01-26; • ST-006 Optima Floor Scale: Due 12-11-25; • ST-011 Johnson Controls Fume Hood: Due 12-12-25; • ST-007 Global Scale: Due 12-11-25; • ST-015 VIS Photometric Accuracy Calibration Standard: Due 10-30-25; and • ST-017 Temp/Humidity Monitor: Due: 09-25-26.
7.6 – 11	Does the organization employ computer software for monitoring and measurement	N/A	Not Applicable. SILQ does not employ software in support of the metrology function.
7.6 – 12	If the organization employs computer software for monitoring and measuring, how is the validity of the software confirmed?	N/A	Not Applicable. SILQ does not employ software in support of the metrology function.
7.6 – 13	Does the organization employ and external metrology laboratory for calibration services?	Yes	Micro Precision (California) is the primary provider of metrology support.
7.6 – 14	If an external metrology laboratory is employed for metrology services, is the facility EN ISO 17025 certified?	Yes	Micro Precision is an ISO/IEC 17025:2017 accredited facility.
8 Measurement, Analysis and Improvement			
8.1 General			
8.1 – 1	How does the organization plan and implement measurement, analysis, and improvement processes needed to demonstrate conformity of the product?	Yes	In accordance with 13.7 of QM.SLQ027.D.; the following metrics are being compiled and reviewed as part of management review, for example: • Customer feedback; • Conformity to product requirements; • Characteristics and trends of processes and products; • Internal audit results; • Supplier quality performance; and • CAPA.
8.1 – 2	How does the organization plan and implement measurement, analysis, and improvement processes needed to ensure conformity of the quality management system?	Yes	SILQ compiles and analyzes metrics from multiple areas. The data collected is used to drive continuous improvement.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.1 – 3	How does the organization plan and implement measurement, analysis and improvement processes needed to continuously improve and maintain the effectiveness of the quality management system?	Yes	Tools used in support of measurement, analysis, and improvement are: (a) complaints; (b) audits (internal & critical subcontractor); (c) management review; and (d) CAPA.
8.1 – 4	How does the organization determine what monitoring measurement, and analysis methods to use, including statistical techniques?	Yes	The application of statistical methods is managed in accordance with QM.SLQ011.A – Statistical Techniques Work Instruction.
8.2 Monitoring and Measurement			
8.2.1 Feedback			
8.2 – 1	How does the organization obtain information about whether customer requirements have been achieved?	Yes	Table 1.0 of QM.SLQ033.A – Post-Market Surveillance; delineates post-market data collected, in support of post-market surveillance.
8.2 – 2	Has the organization established a documented procedure for a feedback system to provide early warning of quality problems and for input into the corrective and preventive action processes?	Yes	Feedback activities are managed in accordance with 6.3 of QM.SLQ033.A – Post-Market Surveillance.
8.2 – 3	If national or regional regulations require the organization to gain experience from the post-production phase, does the review of this experience form part of the feedback system?	Yes	Post-market surveillance activities are pursued in accordance with QM.SLQ033.A.
8.2.2 Complaint Handling			
8.2 – 4	Has the organization established a procedure(s) for the timely handling of complaints?	Yes	Complaints received by SILQ are managed in accordance with QM.SLQ021.D – Product Complaint System
8.2 – 5	As a minimum, the complaint procedure shall include: • Receiving and recording of complaint information; • Evaluation of information to ensure if it constitutes a complaint; • Investigation of complaints; • Determination if a complaint is reportable; • Handling of complaint-related product; and • The need to initiate CAPA as appropriate.	Yes	As scripted, QM.SLQ021, Rev D meets the requirements delineated within 21 CFR, Part 820.198 and ISO 13485:2016, Clause 8.2.2.
8.2 – 6	Has the organization established documented procedures for the issuance and implementation of advisory notices?	Yes	Advisory notices are managed in accordance with QM.SLQ030.A – Advisory Notices and Recalls.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.2 – 7	Does the organization maintain records of all customer complaint investigations?	Yes	In support of this assessment, the following compliant files were reviewed: • PCF 2024001: Opened 03-13-24 / Closed 05-02-24; • PCF 2024002: Opened 04-07-24 / Closed 02-22-24; • PCF 2024003: Opened 06-10-24 / Pending Closure; and • PCF 2024004: Opened 12-17-24 / Closed 01-16-25. <i>Note: There were four complaints in 2024.</i>
8.2 – 8	If an investigation determines that the activities outside the organization contributed to the customer complaint, is relevant information exchanged between the organizations involved?	Yes	Information gleaned during a complaint investigations will be shared with reporting parties, as appropriate
8.2 – 9	If any customer complaint is not followed by corrective and/or preventive action, is the reason and authorization documented?	Yes	Currently SILQ is investigating all complaints received impacting SILQ finished devices.
8.2 – 10	Has the organization issued an advisory notice or processed a market withdraw in FY 2023	N/A	Not Applicable. SILQ has never issued an advisory notice
8.2.3 Reporting to Regulatory Authorities			
8.2 – 11	If deemed to be appropriate, is the complaint reported to applicable regulatory bodies; i.e., MDRs, vigilance reports, etc.?	Yes	Four (4) procedures scripted for reporting to regulatory authorities include: • QM.SLQ022.A – Medical Device Reporting; • QM.SLQ023.A – eMDR Submission Work Instruction; • QM.SLQ030.A – Advisory Notices and Recalls; and • QM.SLQ033.A – Post-Market Surveillance.
8.2 – 12	If national or regional regulations require notification of adverse events that meet specified reporting criteria, does the organization have documented procedures to support notification to appropriate regulatory authorities?	Yes	US FDA centric procedures for reporting adverse events include: • QM.SLQ022.A – Medical Device Reporting; and • QM.SLQ023.A – eMDR Submission Work Instruction.
8.2.4 Internal Audit			
8.2 – 13	Does the organization have a documented procedure for internal audits?	Yes	The Internal Audit Program is managed in accordance with QM.SLQ017.A – Internal Audits.
8.2 – 14	Do audits determine whether the QMS conforms to the requirements of EN ISO 13485 and to the other requirements established by the organization?	Yes	The Internal Audit Program assesses all aspects of the SILQ QMS.
8.2 – 15	Do audits determine whether the QMS is effectively implemented and maintained?	Yes	The audit program is used to gage the level of QMS compliance versus applicable quality, regulatory, and statutory requirements.
8.2 – 16	Are audits currently being performed (on-time) in accordance with the published audit schedule?	Yes	This assessment was the third audit of the SILQ QMS by Dr. Devine. Please note; this procedure was delayed by 30-days due to DGII's December schedule being impacted by another client's FDA inspection. DGII promises all of clients' immediate support for agency inspections.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.2 – 17	Where are the audit criteria, scope, frequency, and methods defined?	Yes	In accordance with 2.1 of QM.SLQ017.A; “internal audits per this procedure are limited to those elements of the organization subject to the Quality Management System. Qualified external auditors may be utilized to perform part of, or all, functions of this procedure.”
8.2 – 18	Can the organization demonstrate that selection of auditors and the conduct of audits are objective and impartial, and that auditors don’t audit their own work?	Yes	Dr. Devine is a trained and qualified auditor with 44-years of industry experience. Note: DGII auditors cannot assess the adequacy of the audits that they perform. However, DGII can attest to the fact that audits are being performed in accordance with ISO 19011 requirements and in compliance with ISO 13485:2016, 21 CFR, Part 820, and applicable quality, regulatory, and statutory requirements. It is the responsibility of the client to ascertain if this audit meets their requirements.
8.2 – 19	Who in the organization ensures that actions are taken to eliminate detected nonconformities and their causes?	Yes	In accordance with 10.5 of QM.SLQ017.A; CAPAs will be opened and assigned to department managers for mitigation of non-conformances identified during the audit.
8.2 – 20	What activities are performed to verify the actions taken and the subsequent reporting of actions taken?	Yes	In accordance with 10.2 of QM.SLQ017.A; “all observations from the audit shall be communicated/discussed to ensure correct interpretation and to allow any open issues to be resolved prior to issuing the final audit report, FM3-QM.SLQ017.”
8.2.5 Monitoring and Measurement of Processes			
8.2 – 21	What methods are used to monitor and measure the QMS processes?	Yes	Examples of tools used in support of the monitoring and measurement of the QMS include: • Management review; • Audits (internal & external); • Complaints; • Quality objectives; and • CAPAs.
8.2 – 22	When the desired results are not achieved, what actions are taken to ensure that the product meets requirements?	Yes	CAPA will be pursued to mitigate QMS performance issues.
8.2.6 Monitoring and Measurement of Product			
8.2 – 23	What characteristics are checked to verify that product requirements have been met?	Yes	Pathway performs a final inspection premised on requirements delineated within the DMR.
8.2 – 24	At what stages of the product realization process do monitoring and measuring activities take place?	Yes	The performance of monitoring and measuring occurs throughout product realization.
8.2 – 25	How is evidence of conformity with acceptance criteria maintained?	Yes	Product conformance with the DMR is maintained through the compilation of the DHR.
8.2 – 26	Do records indicate the person(s) authorizing release of product?	Yes	SILQ is tasked with the formal release of finished device. Distribution and shipment of SILQ devices is performed by Pathway.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.2 – 27	How does the organization ensure that product is not released until the all requirements have been met?	Yes	SILQ that pass final acceptance inspection, are permitted to be released into commercialization.
	8.3 Control of Nonconforming Product		
	8.3.1 General		
8.3 – 1	Is non-conforming product being appropriately identified and controlled?	Yes	Nonconforming product is being appropriately quarantined, documented, and released once disposition has been assigned.
8.3 – 2	Has the organization established a procedure for the control of non-conforming product?	Yes	Non-conforming products are managed in accordance with QM.SLQ040.B – Nonconforming Materials
8.3 – 3	Are records associated with non-conforming product being maintained?	Yes	In accordance with 11.2 of QM.SLQ040.B; <i>“a copy of the closed NCMR is to be included, or referenced, in the appropriate files and batch records / DHR’s as required by each specific disposition.”</i> In support of this assessment, the following NCR’s originating from Pathway were reviewed: • NCR 24-027 – Opened 05-14-24 / Closed 05-15-24 (UAI); and • NCR-24-004 – Opened 01-09-24 / Closed 07-02-24 (Scrap).
	8.3.2 Actions in Response to Nonconforming Product Detected Before Delivery		
8.3 – 4	Does the organization deal with non-conformances in one or more of the following ways: • Taking action to eliminate the detected nonconformity; • Taking action to preclude its original intended use or application; and/or • Authorizing its use/release under concession?	Yes	Nonconforming product identified, prior to shipment, will be reworked, if possible.
	8.3.3 Actions in Response to Nonconforming Product Detected After Delivery		
8.3 – 5	What actions are pursued when a nonconformance is detected after shipment?	Yes	Non-conforming devices (escapes) detected after shipment will be recalled in accordance with QM.SLQ030.A – Advisory Notices and Recalls.
8.3 – 6	Does the organization have an established procedure for the issuance of advisory notices?	Yes	Advisory notice issuance is managed in accordance with QM.SLQ030.A – Advisory Notices and Recalls.
	8.3.4 Rework		
8.3 – 7	Are rework activities being pursued in accordance with documented procedures and instructions?	Yes	In accordance with 10.6.4.1 of QM.SLQ040.B; <i>“rework and re-test and/or re-inspection of product, parts, or materials are to be performed in accordance with instructions provided in NCMR or established manufacturing procedures or to an approved and released Temporary Deviation.”</i>
8.3 – 8	Are all rework products inspected upon the completion of rework?	Yes	Rework is required to be passed through the inspection process.
	8.4 Analysis of Data		

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.4 – 1	Has the organization established documented procedures to determine, collect, and analyze appropriate data to demonstrate the suitability and effectiveness of the quality management system and to evaluate where continual improvement of the effectiveness of the quality management system can be made?	Yes	The QM.SLQ027.D contains sufficient granularity documenting the data collected from key functional areas and trended to drive continuous improvement.
8.4 – 2	What data is collected and analyzed to demonstrate the suitability and effectiveness of the QMS and to evaluate where continual improvement of its effectiveness can be made?	Yes	Premised on the data reviewed during the review of the 2 Feb 2024 Management Review, the data being compiled and analyzed is appropriate for the SILQ QMS.
8.4 – 3	What information does this analysis provide relating to feedback?	Yes	Feedback received from customers from the complaint system will be used to drive product improvement(s).
8.4 – 4	What information does this analysis provide relating to conformity to product requirements?	Yes	Production yield information is used to determine the level of compliance achieved by Pathways.
8.4 – 5	What information does this analysis provide relating to suppliers?	Yes	Supplier performance is gaged versus a supplier's ability to meet SILQ requirements.
8.4 – 6	What information does this analysis provide relating to characteristics and trends of processes and products?	Yes	Data collected and analyzed from the customer base is used gage SILQ's ability meet customer requirements.
8.4 – 7	Where are the records documenting the results of the analysis of data maintained?	Yes	The results of analysis performed on data collected is presented and reviewed as part of management review and retained by QA.
8.5 Improvement			
8.5.1 General			
8.5 – 1	Can the organization demonstrate that the effectiveness of the QMS continuously improves?	Yes	SILQ has revised multiple SOPs since the November 2023 audit, which reflects continuous improvement in support of the QMS.
8.5 – 2	Can the organization demonstrate that it identifies and implements changes to ensure continued QMS effectiveness?	Yes	SILQ has revised multiple SOPs since the November 2023 audit, which reflects continuous improvement in support of the QMS.
8.5 – 3	What tools does the organization employ in pursuit of improvement activities?	Yes	Examples of tools that can be applied to drive continuous improvement of the QMS will be: (a) training; (b) management review; (c) audits (internal & supplier); (d) complaints and customer feedback; and (e) CAPA.
8.5.2 Corrective Action			
8.5 – 4	Do corrective actions records identify and address root cause(s)?	N/A	Not Applicable. No corrective actions were pursued in 2024.
8.5 – 5	Are actions taken appropriate to the severity of the problem?	N/A	Not Applicable. No corrective actions were pursued in 2024.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.5 – 6	Does the organization have a documented procedure defining requirements for each of the following activities? - Reviewing nonconformities (including customer complaints) - Determining the causes of nonconformities - Evaluating the need for action to ensure that nonconformities do not recur - Determining and implementing action needed - Records of the results of any investigation and of action taken - Reviewing the corrective action taken and its effectiveness	Yes	QM.SLQ016.C – Corrective and Preventive Action
8.5 – 7	Does the organization maintain records for investigation and corrective actions taken?	N/A	Not Applicable. No corrective actions were pursued in 2023.
	8.5.3 Preventive Action		
8.5 – 8	How does the organization determine potential nonconformities to take action on?	Yes	Preventive action activities will be identified premised on their ability to prevent nonconformities from occurring.
8.5 – 9	Do preventive action records identify and address root cause(s)?	N/A	Not Applicable. No preventive actions were pursued in 2024.
8.5 – 10	Are preventive actions taken appropriate to the severity of the problem?	N/A	Not Applicable. No preventive actions were pursued in 2024.
8.5 – 11	Does the organization have a documented procedure defining requirements for each of the following activities? - determining potential nonconformities and their causes - evaluating the need for action to prevent occurrence of nonconformities - determining and implementing action needed - (records) recording of the results of any investigation and of action taken - reviewing preventive action taken and its effectiveness	Yes	QM.SLQ016.C – Corrective and Preventive Action
8.5 – 12	Where does the organization maintain records of preventive action taken?	N/A	Not Applicable. No preventive actions were pursued in 2024.

Respectfully Submitted,

A handwritten signature in dark ink, appearing to read "Chris Devine", with a long horizontal flourish extending to the right.

Date: 17 January 2025

Christopher J. Devine, Ph.D.

President

Devine Guidance International, Inc.