

PHASE 0 - CAPA INITIATION

Initiated By:	Jacqueline Kim	Date Initiated:	03/11/2025
CAPA Description:	Equipment lacks: 1. Traceability 2. Verification/Validation Planning 3. Labeling 4. IQ, PM, & CAL Records/Requirements 5. Conditions identified for revalidation/verification 6. Tolerance specifications for manufacturing AUD-23-012 The calibration frequency for pin gauge set (ES-0047-01) and temperature gauge (ES-0119-03) do not comply with requirements outlined in QP-020 Table 7.1.5. There is no documentation or impact analysis to justify change in calibration frequency AUD-23-006 Minor #2 Minor finding 2: QP-012 rev C, Table 8.1, does not fully conform with 21 CFR 820.75(c) in that the table does not include equipment/line being moved as a situation requiring revalidation, or at least requiring that the equipment/process be reviewed for potential need to revalidate, depending on the complexity and sensitivity of the equipment in question. AUD-23-007 Minor finding 4: No documented evidence that Magnetic Needle and Flowpath are clean and meet the requirements of ASTM F3127. To comply with ASTM F3127, either cleaning for magnetic needle/flowpath needs to be instituted or a justification (such as the components are received in a clean state, processed in a cleanroom, visually inspected for cleanliness) needs to be documented. Our preference would be to incorporate a rinse after functional check with DI/purified water or alcohol/flush with compressed air. If not, some testing on each lot may be required. AUD-23-011 Pathway confirmed they did not possess equipment notebooks to record use and any records relating to use of a balance. During the facility tour, the clean room was inspected. Balance, PRT-10003T, was noted. A request was made for the most recent calibration record for the balance. Pathway confirmed they did not have the most recent record. They also confirmed that the balance was not set up properly in the eQMS as there was no schedule set up for the next calibration. In addition, they clarified that they did not have any equipment logs to record use or modifications to equipment in general. This lack of traceability presents a significant risk for non-conformances to occur. Pathway has confirmed the absence of both a Master Validation Plan and an Installation Qualification document for the O-Link fixture. AUD-23-011 During a review of Process Validation document QP-012 Rev D, a request was made for the IOQ protocol and report for both the glue robot and the o-link fixture. While the documentation for the glue robot was readily available, the corresponding materials for the o-link fixture were not provided. Subsequently, an inquiry was made regarding the availability of the MVP for the Biora product line; however, Pathway indicated that this document was also unavailable. Perimeter AUD-24-002 Major #6		

	Control of monitoring and measuring equipment is not implemented as required. 1) Equipment is stored without labeling to indicate its calibration status. Equipment used for production and for Engineering (clarified during the audit to mean "R&D") are stored together without labels clearly identifying calibration status and/or exemption from calibration requirements due to R&D use only status. 2) Tool tolerance levels are not assessed to ensure the required tolerance level of the build is met. Any caliper can be selected for a build. 3) Calibrated equipment is stored in a tool cage upstairs and must be carried manually downstairs each time, some of which is quite heavy which increases the likelihood for the tool to fall as there is no safeguard in place to ensure it is somewhat secured during transport. 4) In case equipment is compromised prior to the next scheduled calibration (e.g., if the equipment were to be dropped during transport downstairs), there is no documented process to evaluate the impact of the event (e.g., verification of calibration prior to continuing to use the equipment)		
Source Types: (check all that apply)	☒ Audits ☐ Complaints ☐ Continuous Improvements ☐ Documentation ☐ Incoming Inspection ☐ Management Reviews ☐ Other Nonconformances ☐ Potential Nonconformance ☐ Purchasing ☐ Receiving Inspection ☐ SCARs ☐ Suppliers ☐ Other: <<Insert here>>		
Source #: (e.g., NCR-24-001, Audit-24-001)	AUD-23-012 AUD-23-006 Minor #2 AUD-23-007 AUD-23-011 AUD-24-002 Major #6	Part #:	N/A
Lot #:	N/A	Qty:	N/A
Correction / Containment: (indicate the immediate action(s) taken to correct the nonconformity)	1. Impacted equipment was identified and quarantined/labeled 2. CAL log was updated to include impacted equipment identified in audits 3. AUD-23-012 a. Created memo to address equipment that are purchased already calibrated but does not adhere to Pathway's calibration procedure; Attachment A b. Updated QP-020 Calibration procedure to accept calibration frequencies based on the supplied calibration certificate; ECO-001646 on 04/02/2024 4. AUD-23-006 a. Revision to QP-012 was released per ECO-001404 and effective on 10/23/2023 5. AUD-23-007 a. PFMEA RM-0098 updated to rev C to add #s 30, 44, 48 - fluid line contamination; ECO-001477 and effective 11/03/2023 b. IP-NGN-0001 released to rev A; ECO-001478 and effective 10/26/2023 c. Letter to File / MEMO written to address cleaning justification dated 09/06/2023 and in the DHF 6. AUD-23-011 a. Process Validation i. Master Validation Plan ii. Master Validation Execution and report iii. IQ O-Link Fixture that includes CAL and PM activities b. Balance i. Entered balance, PRT-10003T into equipment system to be calibrated and create document on how to use balance. 7. AUD-24-002 a. No immediate action taken		
Interim Disposition:	☐ Inspection to New Revision ☐ Rework-to-Accept/Scrap ☐ Scrap ☐ Sort-to-Accept-As-Is ☐ Sort-to-Scrap or Return to Supplier ☐ Use-As-is ☐ R&D Use Only ☒ Other: QMS Documentation; Equipment returned to Engineering/Manufacturing under Use-As-Is after initial corrections were made		

Product Risk Assessment (to be completed by Customer) (☒ N/A if not applicable)

Product Severity Rating:	☐ (1) No Patient Effect ☐ (2) Minor Injury ☐ (3) Moderate Injury ☐ (4) Serious Injury ☐ (5) Life Threatening Injury/Death/Disability
Probability Rating:	☐ (1) Improbable ☐ (2) Remote ☐ (3) Occasional ☐ (4) Likely ☐ (5) Frequent
Detection Rating:	☐ (1) Very High ☐ (2) High ☐ (3) Moderate ☐ (4) Low ☐ (5) Remote

Quality Management System Process Risk Assessment (to be completed by Pathway) (☐ N/A if not applicable)

QMS Rating & Risk Level:	☐ (1) Minor ☐ (2) Minor ☐ (3) Minor ☒ (4) Major ☐ (5) Major
QMS Severity Type:	☐ (1) Business Process Risk ☐ (2) Internal Compliance ☐ (3) Document Compliance ☒ (4) Quality System Compliance ☐ (5) Regulatory Approval or Warning Risk
CAPA Determination:	☒ CAPA is Required – Proceed to Phase 1 ☐ CAPA is not Required – Provide Justification prior to CAPA closure: --
Assigned CAPA Owner:	Amber O'Brein
Phase 1 Target Due Date:	03/28/2025

PHASE 0 – APPROVAL SIGNATURES

CAPA Owner (Name/Title):	Amber O'Brein	Signature/Date:	 Amber O'Brien (24-Mar-2025 09:07 PDT)
Executive Management (Name/Title):	Aaron Rogers	Signature/Date:	 Aaron Rogers (13-Mar-2025 11:26 PDT)
Quality Assurance (Name/Title):	Larry Dacoron	Signature/Date:	 Larry Dacoron (14-Mar-2025 09:12 PDT)

CAPA-23-0008 Phase 0

Final Audit Report

2025-03-24

Created:	2025-03-12 (Pacific Daylight Time)
By:	Jacqueline Kim (jkim@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAAozCcjmZQvWrtOpRgvCXfEW07GX960bmZ
Number of Documents:	1
Document page count:	3
Number of supporting files:	0
Supporting files page count:	0

"CAPA-23-0008 Phase 0" History

-  Document created by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-03-12 - 3:01:48 PM PDT
-  Document emailed to Aaron Rogers (arogers@pathwaymedtech.com) for approval
2025-03-12 - 3:02:47 PM PDT
-  Document emailed to ldacoron@pathwaymedtech.com for approval
2025-03-12 - 3:02:47 PM PDT
-  Document emailed to aobrien@pathwaymedtech.com for approval
2025-03-12 - 3:02:47 PM PDT
-  Agreement viewed by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-03-12 - 3:03:33 PM PDT
-  Email viewed by aobrien@pathwaymedtech.com
2025-03-12 - 3:13:38 PM PDT
-  aobrien@pathwaymedtech.com authenticated with Email OTP by verifying one-time code sent via email
2025-03-12 - 3:14:27 PM PDT
-  Agreement viewed by aobrien@pathwaymedtech.com
2025-03-12 - 3:14:28 PM PDT
-  Email viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-03-13 - 11:23:27 AM PDT

-  Aaron Rogers (arogers@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email
2025-03-13 - 11:23:47 AM PDT
-  Agreement viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-03-13 - 11:23:48 AM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC
2025-03-13 - 11:26:19 AM PDT
-  Document approved by Aaron Rogers (arogers@pathwaymedtech.com)
Approval Date: 2025-03-13 - 11:26:19 AM PDT - Time Source: server
-  Email viewed by Idacoron@pathwaymedtech.com
2025-03-14 - 9:06:31 AM PDT
-  Idacoron@pathwaymedtech.com authenticated with Email OTP by verifying one-time code sent via email
2025-03-14 - 9:10:13 AM PDT
-  Agreement viewed by Idacoron@pathwaymedtech.com
2025-03-14 - 9:10:14 AM PDT
-  Signer Idacoron@pathwaymedtech.com entered name at signing as Larry Dacoron
2025-03-14 - 9:11:59 AM PDT
-  Larry Dacoron (Idacoron@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC
2025-03-14 - 9:12:01 AM PDT
-  Document approved by Larry Dacoron (Idacoron@pathwaymedtech.com)
Approval Date: 2025-03-14 - 9:12:01 AM PDT - Time Source: server
-  New document URL requested by aobrien@pathwaymedtech.com
2025-03-24 - 9:04:47 AM PDT
-  Email viewed by aobrien@pathwaymedtech.com
2025-03-24 - 9:05:18 AM PDT
-  aobrien@pathwaymedtech.com authenticated with Email OTP by verifying one-time code sent via email
2025-03-24 - 9:06:02 AM PDT
-  Signer aobrien@pathwaymedtech.com entered name at signing as Amber O'Brien
2025-03-24 - 9:07:20 AM PDT

 Amber O'Brien (aobrien@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC

2025-03-24 - 9:07:22 AM PDT

 Document approved by Amber O'Brien (aobrien@pathwaymedtech.com)

Approval Date: 2025-03-24 - 9:07:22 AM PDT - Time Source: server

 Agreement completed.

2025-03-24 - 9:07:22 AM PDT