

	Design Review		Doc No.: FM-0052
	<i>This document contains confidential, proprietary information of Pathway LLC and shall not be copied or reproduced without prior written permission there from.</i>		Rev.: B
			Page 1 of 2

Project No.	Product Name	Part No./ Rev	Review Phase	Review Date
00049	SILQ ClearTract 16 Fr 10ml	SLQ-1610 / A	Line Extension Commercial Release	03/11/2022
Subject of Review				
#	Topic	Details	Reference Documents	
1	Change Assessment / Plan	The document was reviewed and complete.	FILE-008918 03/01/2022	
2	Device Master Record (DMR)	The documents were reviewed and complete. • Assembly Specifications / BOM: SLQ-1610 rev A, SLQ-8004 rev A • Components: SLQ-3000 rev A, SLQ-5000 rev A, SLQ-5001 rev A, SLQ-5002 rev A, SLQ-5004 rev A, SLQ-5005 rev A, SLQ-5006 rev A, SLQ-50011 rev A, SLQ-5012 rev A, SLQ-6004 rev A • Inspection Forms: SLQ-8004, SLQ-6004, SLQ-5001, SLQ-5000, SLQ-3000 • Labeling: SLQ-4000 rev A, SLQ-4001 rev A, SLQ-4006 rev A • Manufacturing Procedures: MP-SLQ-0001 rev A, MP-SLQ-0002 rev A	Documents are noted in "Details"	
3	Product Performance	No product performance testing was required. The 1610 is equivalent to 160510.	Refer to Change Assessment #1	
4	Sterilization Adoption	There is no impact to the sterilization validation performed because there were no sterile barrier or pallet configuration changes that affect sterility. No further action is required. The 1610 is equivalent to 160510.	Refer to Change Assessment #1	
5	Labeling Content Review	The previous labeling content review applies. The only changes to the labeling were renumbering their document numbers from HDX-4XXX to SLQ-4XXX. There were no content changes.	FILE-008308 12/01/2021, ECO-000899	
6	Risk Management Plan	The document was reviewed and up to date.	RM-0018 rev C	
7	Product Specification/Trace Matrix	No new requirements. Document is up to date. Trace matrix is up to date and complete.	PS-0006 rev D	
8	Risk – Hazard Analysis	The document was reviewed and up to date. No new risks discovered. Complete.	RM-0019 rev C	
9	Risk – DFMEA	The document was reviewed and up to date. No new risks discovered.	PS-0020 rev C	
10	Process Validation	The documents were reviewed. • MVP: TP-HDX-0007 rev B, TR-HDX-0007 rev A. • Mfg Line Evaluation – TP-HDX-0008 rev B, TR-HDX-0008 rev B Lot 2 and 3 were completed and report updated.	Documents are noted in "Details"	
11	Risk – PFMEA	The document was reviewed and up to date. No new risks discovered.	RM-0021 rev C	
12	Market Clearance	Letter to File Memo was reviewed and complete.	FILE-008919 03/02/2022	

	Design Review		Doc No.: FM-0052
	<i>This document contains confidential, proprietary information of Pathway LLC and shall not be copied or reproduced without prior written permission there from.</i>		Rev.: B
			Page 2 of 2

Project No.	Product Name		Part No./ Rev	Review Phase	Review Date
00049	SILQ ClearTract 16 Fr 10ml		SLQ-1610 / A	Line Extension Commercial Release	03/11/2022
13	Review of Line Extension Deliverables	Each document was reviewed and discussed. Project team in attendance has concluded that the deliverables are complete and satisfy the Design Control requirements in accordance with QP-007 rev F for this design change (line extension). It is agreed upon to release the 1610 for commercial distribution and close the project design history file..		Refer to #1-#13, QP-007 rev F.	
Action Items					
#	Description			Assigned To	Due Date
N/A	N/A			N/A	N/A
Reviewers					
Functional Group	Required?	Name	Title	Signature	Date
Customer - SILQ Technologies Corp	☒	Brian McVerry	CTO	 Brian McVerry (Mar 11, 2022 12:30 PST)	Mar 11, 2022
Project Manager	☒	Michael Klumph	Project Manager	 Michael Klumph (Mar 11, 2022 13:48 PST)	Mar 11, 2022
Research & Development / Manufacturing	☒	Arthur Deptala	CEO, Managing Partner	 Arthur Deptala (Mar 11, 2022 13:58 PST)	Mar 11, 2022
Quality Assurance / Regulatory Affairs	☒	Aaron Rogers	Director of Regulatory and Quality	 Aaron Rogers (Mar 11, 2022 14:46 PST)	Mar 11, 2022
Quality Assurance	☒	Larry Dacoron	Sr. Quality Engineer	 Larry Dacoron (Mar 11, 2022 14:08 PST)	Mar 11, 2022
Independent Reviewer	☒	Karim Natafji	Project Manager (not directly responsible for this project)	 Karim Natafji (Mar 11, 2022 14:08 PST)	Mar 11, 2022