

DESIGN REVIEW FOR SILQ TECHNOLOGIES CORP

MAY 26, 2023

	Design Review		Doc No.: FM-0052
	<i>This document contains confidential, proprietary information of Pathway LLC and shall not be copied or reproduced without prior written permission there from.</i>		Rev.: B
			Page 1 of 2

Project No.	Product Name	Part No./ Rev	Review Phase	Review Date
00053	SILQ ClearTract SPT	SLQ-211410SPT / A SLQ-211605SPT / A SLQ-211610SPT / A SLQ-211810SPT / A	Line Extension SPT Commercial Release	05/26/2023
Subject of Review				
#	Topic	Details	Reference Documents	
1	Change Assessment / Plan	The document was reviewed and complete.	FILE-013630 05/11/2023	
2	Biocompatibility Testing	The documents were reviewed and complete.	TR-SLQ-0012 rev A	
3	Updated Labeling	New labeling documents created for SPT products. - IFU – SLQ-4007 rev A - Pouch Label – SLQ-LB-4008 rev A - Box Label – SLQ-LB-4009 rev A - Shipper Label – SLQ-LB-4010 rev A	Documents are noted in “Details”	
4	Label Content Review	The document was reviewed and complete.	FILE-013631 05/11/2023	
5	Update Product Specification / Trace Matrix	The document was updated, reviewed and complete.	PS-0006 rev E	
6	Device Master Record (DMR)	The documents were reviewed and complete. - Assembly Specifications / BOM: SLQ-211410 rev A, SLQ-211605 rev A, SLQ-211610 rev A, SLQ-211810 rev A; SLQ-8100 rev A, SLQ-8101 rev A, SLQ-8103 rev A, SLQ-8104 rev A - Inspection Forms: SLQ-8100 rev A, SLQ-8101 rev A, SLQ-8103 rev A, SLQ-8104 rev A - Labeling: See #3	Documents are noted in “Details”	
7	Risk – PFMEA	The document was updated for new models, reviewed and up to date. No new risks discovered.	RM-0021 rev D	
8	Market Clearance	K222118 received clearance on 12/01/22 for this product change.	K222118 12/01/2022	
9	Review of Line Extension Deliverables	Each document was reviewed and discussed. Project team in attendance has concluded that the deliverables are complete and satisfy the Design Control requirements in accordance with QP-007 rev E for this design change (line extension). It is agreed upon to release the SPT products for commercial distribution.	Refer to #1-#8, QP-007 rev E.	

	Design Review		Doc No.: FM-0052
	<i>This document contains confidential, proprietary information of Pathway LLC and shall not be copied or reproduced without prior written permission there from.</i>		Rev.: B
			Page 2 of 2

Project No.	Product Name	Part No./ Rev	Review Phase	Review Date
00053	SILQ ClearTract SPT	SLQ-211410SPT / A SLQ-211605SPT / A SLQ-211610SPT / A SLQ-211810SPT / A	Line Extension SPT Commercial Release	05/26/2023
Action Items				
#	Description		Assigned To	Due Date
N/A	N/A		N/A	N/A

	Design Review		Doc No.: FM-0052
	<i>This document contains confidential, proprietary information of Pathway LLC and shall not be copied or reproduced without prior written permission there from.</i>		Rev.: B
			Page

Project No.	Product Name		Part No./ Rev	Review Phase	Review Date
00053	SILQ ClearTract SPT		SLQ-21XXXXSPT / A	Line Extension SPT Commercial Release	05/26/2023
Reviewers					
Functional Group	Required?	Name	Title	Signature	Date
Customer, Eng, Mkt	☒	Brian McVerry	Chief Technology Officer	<u>Brian McVerry</u> Brian McVerry (May 26, 2023 13:55 PDT)	May 26, 2023
Customer, RAQA	☒	Bryan Wong	Director, QA	<u>Bryan Wong</u> Bryan Wong (May 26, 2023 13:51 PDT)	May 26, 2023
Project Manager	☒	Mikey Klumph	Project Manager	<u>Mikey Klumph</u> Mikey Klumph (May 26, 2023 10:35 PDT)	May 26, 2023
Engineering, Mfg	☒	Will Hail	Research and Development Engineer	<u>Will Hail</u> Will Hail (May 26, 2023 11:19 PDT)	May 26, 2023
Independent Reviewer	☒	Karim Natafji	Project Manager	<u>Karim Natafji</u> Karim Natafji (May 26, 2023 10:46 PDT)	May 26, 2023
Regulatory, Quality	☒	Aaron Rogers	Director of Regulatory and Quality	<u>Aaron Rogers</u> Aaron Rogers (May 26, 2023 10:40 PDT)	May 26, 2023
	☐				
	☐				
	☐				
	☐				