

The complaint process is defined in QM.SLQ021 Product Complaint System. The MDR process is defined in QM.SLQ022 Medical Device Reporting. This log is used to assign numbers, to facilitate tracking, and provide a complete list of all complaints and their status.

Color Key:	Information required per QM.SLQ021 Product Complaint System	Information required per QM.SLQ022 MDR reportable event information														
Obtain PCF Number																Completion
YXXX (if MDR event, add "-MDR" to end of PCF number)	Customer/Company Site Name	Brief Description of Event/Problem	Date Complaint Initiated	Initiated by	Date of Event	Date Reported	FDA Due Date	Date submitted to FDA	FDA CORE ID#	port Number (16 digit serial)	Product/Device Name	Device Model No.	Serial Number	Batch / Lot No.	Date Closed	Closed By
2021001	Providence, Santa Monica	Foley catheter falling out of patient	2-Nov-21	Dr. Jennifer Linehan	12-Oct-21	26-Oct-21	N/A	N/A	N/A	N/A	Hydrophilix 16Fr x 5ml Catheter	1605	N/A	HDX8001-080720	17-Nov-21	Bryan Wong
2022001	Anita Deeds/ Tri Valley Urology Medical Group	Rip in catheter near the hub	11-Oct-22	Charles Greiner	11-Oct-22	11-Oct-22	N/A	N/A	N/A	N/A	Hydrophilix 16 Fr x 5 ml Catheter	160510	N/A	HDX-8006110821	15-Nov-22	Bryan Wong
2022002	Sandy Luijens/Rancho Los Amigos	Balloon port broke off while attempting to push saline in	16-Nov-22	Sandy Luijens	12-Jul-22	19-Jul-22	N/A	N/A	N/A	N/A	Hydrophilix 16 Fr x 5 ml Catheter	160510	N/A	HDX-8006110821	21-Dec-22	Bryan Wong
2022003	Silvia Rodriguez/Rancho Los Amigos	Catheter slipped out of patient. Confirmed leak in the balloon.	30-Nov-22	Bryan Wong	28-Sep-22	30-Nov-22	N/A	N/A	N/A	N/A	Hydrophilix 16 Fr x 5 ml Catheter	160510	N/A	HDX-8006110821	5-Apr-23	Bryan Wong
2023001	Anita Deeds/ Tri Valley Urology Medical Group	Blockage in catheter; adverse event; blood coming out of catheter, patient experienced pain and constipation	29-Mar-23	Eeke Romo	29-Mar-23	29-Mar-23	28-Apr-23	15-Dec-25		c12512152254 46.05fc80f03 099458ea604a 1b0e631fc6ef	Hydrophilix 16 Fr x 10 ml Catheter	161010	N/A	SLQ1610-012022	11-Apr-23	Bryan Wong
2023002	Dulce Garcia / Rancho Los Amigos National Rehab Center	Catheter balloon deflating (two instances). 1) Balloon deflate at point of insertion into patient. Catheter unstable. 2) Balloon deflated after one day of use. Patient returned to ER to get replacement.	15-May-23	Brian McVerry	15-May-23	15-May-23	N/A	N/A	N/A	N/A	Hydrophilix 16 Fr x 10 ml Catheter	161010	N/A	SLQ1610-012022	18-Sep-23	Bryan Wong
2023003	Sandy Andres / Rancho Los Amigos National Rehab Center	Report of catheter leak. Upon investigation, not a device leak/malfunction.	15-Nov-23	Eeke Romo	8-Nov-23	15-Nov-23	N/A	N/A	N/A	N/A	Hydrophilix 16 Fr x 10 ml Catheter	161010	N/A	SLQ1610-012022	11-Dec-23	Bryan Wong
2023004	West Los Angeles VA Medical Center	Patient experienced pain and discomfort. Ruptured balloon on Silq catheter rendering catheter inoperative.	5-Dec-23	Eeke Romo	5-Dec-23	5-Dec-23	4-Jan-24	15-Dec-25		0251215225630 6340b8a23d34b 6d935a55bcd98 d9e9e	Hydrophilix 16 Fr x 10 ml Catheter	161010	N/A	SLQ1610-012022	13-Dec-23	Bryan Wong
2024001	Oliver Estrada / Rancho Los Amigos National Rehab Center	Dark spot on distal end of catheter	13-Mar-24	Charles Greiner	11-Mar-24	11-Mar-24	N/A	N/A	N/A	N/A	Cleartract 16 Fr x 10 ml Catheter	211610SPT	N/A	SLQ-81041201231	2-May-24	Bryan Wong
2024002	Tousif Hussain Sheikh / HPFY	Port broke off	3-Apr-24	Charles Greiner	3-Apr-24	3-Apr-24	N/A	N/A	N/A	N/A	Cleartract 18 Fr x 10 ml Catheter	211810SPT	N/A	SLQ-81020403231	22-May-24	Bryan Wong
2024003	Gina Koehler / Cleveland Clinic	Luer lock connection malfunction	10-Jun-24	Ethan Rao	21-May-24	22-May-24	N/A	N/A	N/A	N/A	Cleartract 16 Fr x 10 ml Catheter	211610SPT	N/A	SLQ-81040413231	7-Jan-25	Bryan Wong
2024004	Gurav Pandey, HPFY	Customer observed a crack in device	17-Dec-24	Charles Greiner	17-Dec-24	17-Dec-24	N/A	N/A	N/A	N/A	Cleartract 18 Fr x 10 ml Catheter	211810SPT	N/A	SLQ-81020515241	16-Jan-25	Bryan Wong
2025001	Olusegun Adetaja / University of Virginia Department of Urology	Luer lock connection malfunction	25-Jul-25	Ethan Rao	8-Jul-25	17-Jul-25	N/A	N/A	N/A	N/A	Cleartract 16 Fr x 10 ml Catheter	211610SPT	N/A	SLQ-81040413231	15-Aug-25	Ethan Rao