


	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-005	Equipment Description:	Diafiltration System, Repligen KrosFlo KTF-2000
Model #:	SYTF-2000	S/N:	190603-001
Manufacturer:	Repligen	Operating Range:	N/A
Type (check one):	☒ Production / ☐ Engineering / ☐ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☒ Calibration Required	☐ Limited Calibration	☐ No Calibration
Internal / External (circle one) Interval: <u>Biennial</u> Internal calibration procedure #: <u>Repligen Doc # IF.KTF-POP.1146</u>	For limited calibration, document nature of limited calibration (how it can / cannot be used)	For no calibration required, document rationale for reason(s) equipment does not require calibration
Requirements or Rationale for No Calibration: The manufacturer recommends the system be serviced annually. This includes calibration of three pressure sensors, pump RPM check, flowmeter test, scale accuracy verification, and general preventive maintenance on the system. Using risk assessment, there is no risk to the safety or efficacy of the product (suspension) for a calibration and PM frequency to be performed biennially. Accuracy of the pump RPM's does not affect the suspension or its ability to meet specification. If the pump RPM's are not accurate, this would only affect the speed of the manufacturing process. Flowmeters are not used in the manufacturing process of the suspension. The pertinent scale is being calibrated independently by an accredited third-party calibration service provider on an annual basis. Given the projected usage of the system, extending the preventive maintenance on the system to every 2 years is warranted.		
Preventive Maintenance Requirements (check type of calibration):		
☒ Preventive Maintenance Required	☐ No Preventive Maintenance	
Interval: <u>Biennial</u> Internal PM procedure #: <u>Repligen Doc # IF.KTF-POP.1146</u>	For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
Requirements or Rationale for No PM: See rationale for biennial frequency for preventive maintenance above.		

Approval		
	Signature	Date
Quality:		14 Sep 2022
Engineering:		14SEPT2022