



	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-009	Equipment Description:	Pneumatic Drum Stirrer
Model #:	35005K64	S/N:	V10708301
Manufacturer:	McMaster-Carr	Operating Range:	Mixer speed: 300 – 1750 rpm
Type (check one):	☒ Production / ☐ Engineering / ☐ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☐ Calibration Required	☐ Limited Calibration	☒ No Calibration
Internal / External (circle one) Interval: _____ Internal calibration procedure #: _____	For limited calibration, document nature of limited calibration (how it can / cannot be used)	For no calibration required, document rationale for reason(s) equipment does not require calibration
Requirements or Rationale for No Calibration: The pneumatic drum stirrer is used to stir the concentrate/fluids in the concentrate reactor. The mixing speed of the pneumatic drum stirrer does not need to be precise for its function in the manufacturing process.		
Preventive Maintenance Requirements (check type of calibration):		
☒ Preventive Maintenance Required	☐ No Preventive Maintenance	
Interval: <u>Annual</u> Internal PM procedure #: <u>N/A</u>	For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
Requirements or Rationale for No PM: PM checklist: 1) Check the components for excessive wear. 2) If the bearing has dry wear, add bearing lubricating oil to the bearing body, as needed. 3) Check whether the tightening bolts on the drum stirrer are tightened. Tighten bolts, as needed.		

Approval		
	Signature	Date
Quality:		14 Sep 2022
Engineering:		14SEPT2022