



MICRO PRECISION CALIBRATION, INC.
22835 INDUSTRIAL PLACE
GRASS VALLEY CA 95949
530-268-1860

Certificate of Calibration

Date: Sep 9, 2022

Cert No. 551220085329508

Customer:

SILQ TECHNOLOGIESC

311 OTTERSON DRIVE SUITE 60
CHICO CA 95928

Work Order #: SAC-70124170

MPC Control #: ST-006

Serial Number: 10C190611004

Asset ID: N/A

Department: N/A

Gage Type: PLATFORM SCALE

Performed By: GLENN HIGGINS

Manufacturer: OPTIMA

Received Condition: IN TOLERANCE

Model Number: XK3150-EX

Returned Condition: IN TOLERANCE

Size: 1000 LBS

Cal. Date: September 07, 2022

Temp/RH: 70.0°F / 40.0%

Cal. Interval: 12 MONTHS

Location: Calibration performed at Customer's facility

Cal. Due Date: September 07, 2023

Calibration Notes:

Test Points

Seq.	Description	Standard	Tolerance -	Tolerance +	As Found	As Left	UOM	Result
1	Weight Applied:	200.0	195.0	205.0	199.8	199.8	LBS	Passed
2	Weight Applied:	400.0	395.0	405.0	399.8	399.8	LBS	Passed
3	Weight Applied:	600.0	595.0	605.0	599.4	599.4	LBS	Passed
4	Weight Applied:	800.0	795.0	805.0	799.2	799.2	LBS	Passed
5	Weight Applied:	1,000.0	995.0	1005.0	999.0	999.0	LBS	Passed

Standards Used to Calibrate Equipment

I.D.	Description.	Model	Serial	Manufacturer	Cal. Due Date	Traceability #
M1190	PRECISION WEIGHT SET	CLASS F	N/A	CENCO / RICE LAKE	Mar 31, 2024	551220084119519

Procedures Used in this Event

Procedure Name

Description

MPC-WEI-001 Rev. 07

Weighing Instruments, General, rev.07, Jul-07-2021

Calibrating Technician:

GLENN HIGGINS

QC Approval:

ROBERT MEANS

STATEMENTS OF PASS OR FAIL CONFORMANCE: The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

THE CALIBRATION REPORT STATUS:

PASS: Term used when compliance statement is given, and the measurement result is PASS.

PASS²: Term used when compliance statement is given, and the measurement result is conditional passed or PASS².

FAIL: Term used when compliance statement is given, and the measurement result is FAIL.

FAIL²: Term used when compliance statement is given, and the measurement result is conditional failed or FAIL².

REPORT OF VALUE: Term used when reported measurement is not requiring compliance statement in report.

ADJUSTED: When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.

LIMITED: When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated.

This calibration report complies with ISO/IEC 17025:2017, ANSI/NCSL Z540.3-2006 and ANSI/NCSL Z540.1-1994. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.