

	DOCUMENT TITLE: Installation Qualification Protocol, Floor Scale, Suspension	Revision: A
	DOCUMENT #: VV.SLQ016	Effective Date: 29Nov2022

PROTOCOL ATTACHMENT 1 – IQ: Verification of the installation

Verification	Acceptance criteria	Results observed	Conformity C / NC
Brand	Global Industrial	Global Industrial	C
Model number	318506	318 506	C

Comments:	SI-007 Silq Technologies Equipment Tag No explicit Serial # present 01 JAN 2023 Equipment Serial Number: 318506-2108 01 JAN 2023
------------------	--

Completed By (Sign and Date):  09 DEC 2022

Reviewed By (Sign and Date):  02 Jan 2023

	DOCUMENT TITLE: Installation Qualification Protocol, Floor Scale, Suspension	Revision: A
	DOCUMENT #: VV.SLQ016	Effective Date: 29Nov2022

PROTOCOL ATTACHMENT 2 – IQ: Verification of the documentation

Documentation	Acceptance criteria	Results observed	Conformity C / NC
Equipment user manual	Presence of the user manual in English	English user manual present	C
Equipment calibration certificate	Presence of equipment calibration certificate	Calibration certificate present	C
Calibration requirements	Schedule / frequency of calibration determined	Calibration frequency determined	C

Comments:	
------------------	--

Completed By (Sign and Date):  09 DEC 2022

Reviewed By (Sign and Date):  02 Jan 2023

	DOCUMENT TITLE: Installation Qualification Protocol, Floor Scale, Suspension	Revision: A
	DOCUMENT #: VV.SLQ016	Effective Date: 29Nov2022

PROTOCOL ATTACHMENT 3 – IQ: Verification of utilities

Verification	Acceptance criteria	Results observed	Conformity C / NC
Power supply	Compliant if the floor scale is powered by AC adapter cable	AC Adapter cable Powered	C
Wiring	No damage to AC adapter cable	No visible damage	C
Operating temperature	-10 – 40°C	18.5°C	C
Operating relative humidity	< 90% RH	40% RH	C
Operation	Equipment can be powered on and is operational	Equipment operational	C

Comments:	
------------------	--

Completed By (Sign and Date):  19 DEC 2022

Reviewed By (Sign and Date):  02 Jan 2023

	DOCUMENT TITLE: Installation Qualification Protocol, Floor Scale, Suspension	Revision: A
	DOCUMENT #: VV.SLQ016	Effective Date: 29Nov2022

PROTOCOL ATTACHMENT 4 – Calibration Log

In the table below, record the calibration information for all equipment used during the test. If the equipment is “calibrate before use” then include or refer to the procedure used for calibration, record any equipment required and its calibration info, and record the results of the calibration (data).

Instrument ID	Description	Use	Calibration Date	Due Date	Verified By/Date
ST-OB	Temp/Humidity Monitor	Measure Temp/Humidity	10 NOV 2022	10 NOV 2024	19 DEC 2022

	DOCUMENT TITLE: Installation Qualification Protocol, Floor Scale, Suspension	Revision: A
	DOCUMENT #: VV.SLQ016	Effective Date: 29Nov2022

PROTOCOL ATTACHMENT 5 – Signature Log

In the table below, all personnel who have completed any section of this protocol must include the listed information.

Printed Name	Title	Initials	Signature	Company	Date
Chris Turner	Director of Manufacturing	CT		Silq Technologies	19 DEC 2022