

	DOCUMENT TITLE: Quality Control Procedure, Crystal Violet Absorbance, Silq Surface Treatment	Revision: A
	DOCUMENT #: OM.SLQ003	Effective Date: 06Jun2024

Attachment 1: Crystal Violet Absorbance Datasheet

Tester (Print name & Sign): Ethan Rao

Ethan Rao

Date of Test: 18 July 2024

General Product Information

Part Number:	SLQ-8104	Part Description:	SILQ 2-Way Cleartract SPT 16 Fr x 10mL Catheter
Part Revision:	B	Lot #:	SLQ-80040411241E

Supplies Traceability

Part Number	Description	Lot Number	Expiration Date
S.SLQ008	Crystal violet dye	SHBR0031	01 July 2027
S.SLQ009	Acetic acid	231511	01 May 2028

Equipment Traceability

Part Number	Description	Equipment ID	Last qualified on*
E.SLQ015	Jenway 7205 Spectrophotometer	E.SLQ015	18 July 2024

* Attach QC Qualification Datasheet


Summary of Results

Catheter Sample #	Result	Acceptance criteria	Meets Criteria?
1	0.458	≥ .400	☒ Yes ☐ No
2	0.467	≥ .400	☒ Yes ☐ No
3	0.415	≥ .400	☒ Yes ☐ No
4	0.493	≥ .400	☒ Yes ☐ No
5	0.416	≥ .400	☒ Yes ☐ No
6	0.429	≥ .400	☒ Yes ☐ No
7	0.444	≥ .400	☒ Yes ☐ No
8	0.433	≥ .400	☒ Yes ☐ No
9	0.499	≥ .400	☒ Yes ☐ No
10	0.481	≥ .400	☒ Yes ☐ No
11	0.496	≥ .400	☒ Yes ☐ No
12	0.452	≥ .400	☒ Yes ☐ No
13	0.469	≥ .400	☒ Yes ☐ No
14	0.412	≥ .400	☒ Yes ☐ No
15	0.449	≥ .400	☒ Yes ☐ No
16	0.487	≥ .400	☒ Yes ☐ No
17	0.467	≥ .400	☒ Yes ☐ No
18	0.493	≥ .400	☒ Yes ☐ No
19	0.514	≥ .400	☒ Yes ☐ No

	DOCUMENT TITLE: Quality Control Procedure, Crystal Violet Absorbance, Silq Surface Treatment	Revision: A
	DOCUMENT #: OM.SLQ003	Effective Date: 06Jun2024

Catheter Sample #	Result	Acceptance criteria	Meets Criteria?
20	0.465	$\geq .400$	☒ Yes ☐ No
21	0.438	$\geq .400$	☒ Yes ☐ No
22	0.482	$\geq .400$	☒ Yes ☐ No

OVERALL TEST RESULTS (circle one): PASS / FAIL

Reviewed By (Sign and Date):  13 Nov 2024