

	DOCUMENT TITLE: Quality Control Procedure, Coating Adherence Verification, Silq Surface Treatment	Revision: A
	DOCUMENT #: OM.SLQ004	Effective Date: 06Jun2024

Attachment 1: Coating Adherence Verification Datasheet

Tester (Print name & Sign): Ethan Rao



Date of Test: 18 July 2024

General Product Information

Part Number:	SLQ-8104	Part Description:	SILQ 2-Way Cleartract SPT 16 Fr x 10mL Catheter
Part Revision:	B	Lot #:	SLQ-80040411241E

Equipment Traceability

Part Number	Description	Equipment ID	Last qualified on*
E.SLQ015	Jenway 7205 Spectrophotometer	E.SLQ015	18 July 2024

* Attach QC Qualification Datasheet

Summary of Results

Catheter Sample #	Result	Acceptance criteria	Meets Criteria?
1	0.002	≤ .02	☒ Yes ☐ No
2	0.006	≤ .02	☒ Yes ☐ No
3	0.000	≤ .02	☒ Yes ☐ No
4	0.003	≤ .02	☒ Yes ☐ No
5	0.009	≤ .02	☒ Yes ☐ No
6	0.002	≤ .02	☒ Yes ☐ No
7	0.006	≤ .02	☒ Yes ☐ No
8	0.001	≤ .02	☒ Yes ☐ No
9	0.011	≤ .02	☒ Yes ☐ No
10	0.010	≤ .02	☒ Yes ☐ No
11	0.006	≤ .02	☒ Yes ☐ No
12	0.005	≤ .02	☒ Yes ☐ No
13	0.003	≤ .02	☒ Yes ☐ No
14	0.000	≤ .02	☒ Yes ☐ No
15	0.004	≤ .02	☒ Yes ☐ No
16	0.009	≤ .02	☒ Yes ☐ No
17	0.001	≤ .02	☒ Yes ☐ No
18	0.008	≤ .02	☒ Yes ☐ No
19	0.009	≤ .02	☒ Yes ☐ No
20	0.010	≤ .02	☒ Yes ☐ No
21	0.012	≤ .02	☒ Yes ☐ No
22	0.008	≤ .02	☒ Yes ☐ No

 OVERALL TEST RESULTS (circle one): **PASS** / FAIL

Reviewed By (Sign and Date):



19 Nov 2024