




MICRO PRECISION CALIBRATION, INC
2165 N. Glassell St.,
Orange, CA 92865
714-901-5659



Certificate of Calibration

Date: Aug 2, 2024

Cert No. 5523631031124783

Customer:

SILQ TECHNOLOGIES CORPORATION
607 CHARLES E YOUNG DR E
LOS ANGELES CA 90095

MPC Control #: LAF5769
Asset ID: LAF5769
Gage Type: ANALYTICAL BALANCE
Manufacturer: METTLER TOLEDO
Model Number: AB54-S
Size: 51 g
Temp/RH: 22.0°C / 48.0%
Location: Calibration performed at MPC facility

Work Order #: LA-90065228
Purchase Order #: 0000131
Serial Number: 1125311515
Department: N/A
Performed By: JUAN ACEVES
Received Condition: IN TOLERANCE
Returned Condition: IN TOLERANCE
Cal. Date: August 02, 2024
Cal. Interval: 12 MONTHS
Cal. Due Date: August 02, 2025

Calibration Notes:

See attached data sheet (2 pages).

Standards Used to Calibrate Equipment

I.D.	Description.	Model	Serial	Manufacturer	Cal. Due Date	Traceability #
DS3215	WEIGHT SET 29 PCS	E1 E2	589/J232/J234	FUYUE WEIGHT	Sep 30, 2024	2964214A

Procedures Used in this Event

Procedure Name	Description
MPC-WEI-001 Rev. 08	Weighing Instruments, General, Rev.08, Jul-15-2024

Calibrating Technician:

Juan A
JUAN ACEVES

QC Approval:

Ilya Vaks
ILYA VAKS

STATEMENTS OF PASS OR FAIL CONFORMANCE: The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

THE CALIBRATION REPORT STATUS:

PASS: Term used when compliance statement is given, and the measurement result is PASS.

PASS²: Term used when compliance statement is given, and the measurement result is conditional passed or PASS².

FAIL: Term used when compliance statement is given, and the measurement result is FAIL.

FAIL²: Term used when compliance statement is given, and the measurement result is conditional failed or FAIL².

REPORT OF VALUE: Term used when reported measurement is not requiring compliance statement in report.

ADJUSTED: When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.

LIMITED: When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated. This calibration report complies with ISO/IEC 17025:2017, ANSI/NCSL Z540.3-2006 and ANSI/NCSL Z540.1-1994. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.



Calibration Report of Mettler Toledo AB54-S Analytical Balance

MPC Control #:	LAF5769	Serial Number:	1125311515
Asset ID:	LAF5769	Calibration Date:	August 02, 2024

Repeatability

Range	Applied	Lower Limit	As Found	As Left	Upper Limit	Result	Uncertainty ($\pm$)	TUR
51.0000 g	20.0000 g	19.9998 g	20.0001 g	20.0001 g	20.0002 g	PASS	0.000089 g	2.2 : 1

The standard deviation was found to be 0.000 g


Linearity

Range	Applied	Lower Limit	As Found	As Left	Upper Limit	Result	Uncertainty ($\pm$)	TUR
51.0000 g	5.0000 g	4.9996 g	5.0001 g	5.0001 g	5.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
51.0000 g	12.0000 g	11.9996 g	12.0001 g	12.0001 g	12.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
51.0000 g	25.0000 g	24.9996 g	25.0000 g	25.0000 g	25.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
51.0000 g	38.0000 g	37.9996 g	38.0000 g	38.0000 g	38.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
51.0000 g	51.0000 g	50.9996 g	51.0001 g	51.0001 g	51.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$

Eccentricity

Position	Applied	Lower Limit	As Found	Deviation from Center	Upper Limit	Result	Uncertainty ($\pm$)	TUR
Center (1)	20.0000 g	19.9996 g	20.0000 g	0.0000 g	20.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
Front Left (2)	20.0000 g	19.9996 g	20.0001 g	0.0001 g	20.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
Back Left (3)	20.0000 g	19.9996 g	19.9999 g	0.0001 g	20.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
Back Right (4)	20.0000 g	19.9996 g	20.0000 g	0.0000 g	20.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$
Front Right (5)	20.0000 g	19.9996 g	20.0000 g	0.0000 g	20.0004 g	PASS	0.000089 g	$\geq 4.0 : 1$

Platform Eccentricity Loading Positions



Max Deviation from Center	Position(s)
0.0001 g	(2) (3)



Calibration Report of Mettler Toledo AB54-S Analytical Balance

MPC Control #: LAF5769
 Asset ID: LAF5769

Serial Number: 1125311515
 Calibration Date: August 02, 2024


Statements of Pass or Fail Conformance

The uncertainty of measurement has been taken into account when determining compliance with specification.

All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCCL Z540.3-2006

The status of compliance with the acceptance criteria is reported as:

- PASS** — Compliant with specification.
- PASS^z** — The measured value is within acceptance limits.
 However, a portion of the expanded uncertainty of measurement at 95% exceeds the specified tolerance.
- FAIL^z** — The measured value is not within the acceptance limits.
 However, a portion of the expanded uncertainty of measurement at 95% is within the specified tolerance.
- FAIL** — Not compliant with specification.



The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated.

This calibration report complies with ISO/IEC 17025:2017 and ANSI/NCCL Z540.3-2006, Method 6 --- Guard Bands Based on Test Uncertainty Ratio.

- End of Calibration Report -