

	DOCUMENT TITLE: Quality Control Procedure, Coating Adherence Verification, Silq Surface Treatment	Revision: A
	DOCUMENT #: OM.SLQ004	Effective Date: 06Jun2024

Tester (Print name & Sign): Ethan Rao *Ethan Rao*

Date of Test: 13 June 2025

General Product Information

Part Number:	SLQ-8104	Part Description:	SILQ 2-Way Cleartract SPT 16 Fr x 10mL Catheter
Part Revision:	B	Lot #:	SLQ-05012025

Equipment Traceability

Part Number	Description	Equipment ID	Last qualified on*
E.SLQ015	Jenway 7205 Spectrophotometer	E.SLQ015	13 June 2025

* Attach QC Qualification Datasheet

Summary of Results

Catheter Sample #	Result	Acceptance criteria	Meets Criteria?
1	0.006	$\leq .02$	☒ Yes ☐ No
2	0.003	$\leq .02$	☒ Yes ☐ No
3	0.007	$\leq .02$	☒ Yes ☐ No
4	0.001	$\leq .02$	☒ Yes ☐ No
5	0.005	$\leq .02$	☒ Yes ☐ No
6	0.000	$\leq .02$	☒ Yes ☐ No
7	0.008	$\leq .02$	☒ Yes ☐ No
8	0.006	$\leq .02$	☒ Yes ☐ No
9	0.001	$\leq .02$	☒ Yes ☐ No
10	0.009	$\leq .02$	☒ Yes ☐ No
11	0.010	$\leq .02$	☒ Yes ☐ No
12	0.011	$\leq .02$	☒ Yes ☐ No
13	0.006	$\leq .02$	☒ Yes ☐ No

OVERALL TEST RESULTS (circle one): **PASS** / FAIL

Reviewed By (Sign and Date): *[Signature]* **23 Jun 2025**

	DOCUMENT TITLE: Quality Control Procedure, Coating Adherence Verification, Silq Surface Treatment	Revision: A
	DOCUMENT #: OM.SLQ004	Effective Date: 06Jun2024

ATTACHMENT 1: QC Qualification Datasheet

Tester: Ethan Rao

Date of Test: 13 June 2025

Equipment Under Test (EUT)

Description	Equipment ID	Serial number
Jenway 7205 UV/Visible Spectrophotometer	E.SLQ015	85652

Equipment Traceability

Description	Equipment ID	Calibration Due Date
VIS Photometric Accuracy Calibration Standard, 400-700nm, HF-0.2 (TF.SLQ001)	ST-015	30 Oct 2025

EUT Absorbance Values

Wavelength (nm)	EUT Absorbance Value Reading	Acceptance criteria	Result (Pass / Fail)
400	0.410	0.395 – 0.415	Pass
440	0.240	0.230 – 0.250	Pass
585	0.251	0.240 – 0.260	Pass
700	0.219	0.210 – 0.230	Pass

OVERALL TEST RESULTS (circle one): **PASS** / FAIL

Completed By (Sign and Date): Ethan Rao 13 June 2025

Reviewed By (Sign and Date): Brian M. King 23 Jun 2025