

	DOCUMENT TITLE: Quality Control Procedure, Crystal Violet Absorbance, Silq Surface Treatment	Revision: A
	DOCUMENT #: OM.SLQ003	Effective Date: 06Jun2024

Tester (Print name & Sign): Ethan Rao *Ethan Rao*

Date of Test: 29 Nov 2024

General Product Information

Part Number:	SLQ-8102	Part Description:	SILQ 2-Way Cleartract SPT 18 Fr x 10mL Catheter
Part Revision:	B	Lot #:	SLQ-11182024

Supplies Traceability

Part Number	Description	Lot Number	Expiration Date
S.SLQ008	Crystal violet dye	SHBR0031	01 July 2027
S.SLQ009	Acetic acid	231511	01 May 2028

Equipment Traceability

Part Number	Description	Equipment ID	Last qualified on*
E.SLQ015	Jenway 7205 Spectrophotometer	E.SLQ015	29 Nov 2024

* Attach QC Qualification Datasheet

Summary of Results

Catheter Sample #	Result	Acceptance criteria	Meets Criteria?
1	0.412	≥ .400	☒ Yes ☐ No
2	0.436	≥ .400	☒ Yes ☐ No
3	0.461	≥ .400	☒ Yes ☐ No
4	0.471	≥ .400	☒ Yes ☐ No
5	0.464	≥ .400	☒ Yes ☐ No
6	0.481	≥ .400	☒ Yes ☐ No
7	0.432	≥ .400	☒ Yes ☐ No
8	0.455	≥ .400	☒ Yes ☐ No
9	0.440	≥ .400	☒ Yes ☐ No
10	0.465	≥ .400	☒ Yes ☐ No
11	0.429	≥ .400	☒ Yes ☐ No

OVERALL TEST RESULTS (circle one): **PASS** / FAIL

Reviewed By (Sign and Date): *[Signature]* **03 Dec 2024**

	DOCUMENT TITLE: Quality Control Qualification Procedure, Jenway 7205 Spectrophotometer	Revision: A
	DOCUMENT #: QC-E.SLQ015	Effective Date: 08Nov2023

ATTACHMENT 1: QC Qualification Datasheet

Tester: Ethan Rao

Date of Test: 29 Nov 2024

Equipment Under Test (EUT)

Description	Equipment ID	Serial number
Jenway 7205 UV/Visible Spectrophotometer	E.SLQ015	85652

Equipment Traceability

Description	Equipment ID	Calibration Due Date
VIS Photometric Accuracy Calibration Standard, 400-700nm, HF-0.2 (TF.SLQ001)	ST-015	30 Oct 2025

EUT Absorbance Values

Wavelength (nm)	EUT Absorbance Value Reading	Acceptance criteria	Result (Pass / Fail)
400	0.398	0.395 – 0.415	Pass
440	0.240	0.230 – 0.250	Pass
585	0.253	0.240 – 0.260	Pass
700	0.219	0.210 – 0.230	Pass

OVERALL TEST RESULTS (circle one) **PASS** / FAIL

Completed By (Sign and Date): Ethan Rao 29 Nov 2024

Reviewed By (Sign and Date): Brian M. King 03 Dec 2024