

	DOCUMENT TITLE: Quality Control Procedure, Coating Adherence Verification, Silq Surface Treatment	Revision: A
	DOCUMENT #: OM.SLQ004	Effective Date: 06Jun2024

Attachment 1: Coating Adherence Verification Datasheet

Tester (Print name & Sign): Ethan Rao *Ethan Rao*

Date of Test: 02 June 2023

General Product Information

Part Number:	SLQ-8102	Part Description:	SILQ 2-Way Cleartract SPT 18 Fr x 10mL Catheter
Part Revision:	B	Lot #:	SLQ-81020515241

Equipment Traceability

Part Number	Description	Equipment ID	Last qualified on*
E.SLQ015	Jenway 7205 Spectrophotometer	E.SLQ015	02 June 2024

* Attach QC Qualification Datasheet

Summary of Results

Catheter Sample #	Result	Acceptance criteria	Meets Criteria?
1	0.001	$\leq .02$	☒ Yes ☐ No
2	0.002	$\leq .02$	☒ Yes ☐ No
3	0.001	$\leq .02$	☒ Yes ☐ No
4	0.003	$\leq .02$	☒ Yes ☐ No
5	0.005	$\leq .02$	☒ Yes ☐ No
6	0.002	$\leq .02$	☒ Yes ☐ No
7	0.009	$\leq .02$	☒ Yes ☐ No
8	0.004	$\leq .02$	☒ Yes ☐ No
9	0.001	$\leq .02$	☒ Yes ☐ No

OVERALL TEST RESULTS (circle one): **PASS** / FAIL

Reviewed By (Sign and Date): *Brian M. King*

	DOCUMENT TITLE: Quality Control Qualification Procedure, Jenway 7205 Spectrophotometer	Revision: A
	DOCUMENT #: QC-E.SLQ015	Effective Date: 08Nov2023

ATTACHMENT 1: QC Qualification Datasheet

Tester: Ethan Rao

Date of Test: 02 Jun 2024

Equipment Under Test (EUT)

Description	Equipment ID	Serial number
Jenway 7205 UV/Visible Spectrophotometer	ST-014	85652

Equipment Traceability

Description	Equipment ID	Calibration Due Date
VIS Photometric Accuracy Calibration Standard, 400-700nm, HF-0.2 (TF.SLQ001)	ST-015	30 Oct 2025

EUT Absorbance Values

Wavelength (nm)	EUT Absorbance Value Reading	Acceptance criteria	Result (Pass / Fail)
400	0.401	0.395 – 0.415	Pass
440	0.238	0.230 – 0.250	Pass
585	0.250	0.240 – 0.260	Pass
700	0.219	0.210 – 0.230	Pass

OVERALL TEST RESULTS (circle one): **PASS** / FAIL

Completed By (Sign and Date): Ethan Rao 02 Jun 2024

Reviewed By (Sign and Date): Bruno M. V. 03 Jun 2024