


	DOCUMENT TITLE: Equipment Requirements Form	Revision: A
	DOCUMENT #: FM2-QM.SLQ050	Effective Date: 30Jul2022

Equipment Requirements

Equipment ID #:	ST-012	Equipment Description:	Weighing Scale
Model #:	AB54-S	S/N:	1125311515
Manufacturer:	Mettler Toledo	Operating Range:	0 – 51 g (0.1 mg readability)
Type (check one):	☐ Production / ☐ Engineering / ☒ Quality / ☐ Other _____		

Calibration Requirements (check type of calibration):		
☒ Calibration Required	☐ Limited Calibration	☐ No Calibration
Internal / External (circle one) Interval: <u>Biennial</u> Internal calibration procedure #: <u>N/A</u>	For limited calibration, document nature of limited calibration (how it can / cannot be used)	For no calibration required, document rationale for reason(s) equipment does not require calibration
Requirements or Rationale for No Calibration: Calibrate for full operating range per manufacturer's specification.		
Preventive Maintenance Requirements (check type of calibration):		
☐ Preventive Maintenance Required	☒ No Preventive Maintenance	
Interval: _____ Internal PM procedure #: _____	For no preventive maintenance required, document rationale for reason(s) equipment does not require preventive maintenance	
Requirements or Rationale for No PM: The scale housing and weighing pan are made of high grade durable materials. All commonly used cleaning agents can be used to clean surfaces, as needed. Periodic preventive maintenance is not required.		

Approval		
	Signature	Date
Quality:		27 Sep 2022
Engineering:		27 Sept 2022