



Date of Assessment: November 15th through 17th 2023

Client Name: SILQ Technologies Corporation **Client Address:** 323 Sunny Isles Blvd. (Seventh Floor); Sunny Isles Beach, FL 33160

Company Audited: SILQ Technologies Corporation **Address:** 323 Sunny Isles Blvd. (Seventh Floor); Sunny Isles Beach, FL 33160

Telephone Number: 786-590-5300

Name of Senior Company Official: D. V. Sharma

Title: CEO / President

Name of Head of Quality: Bryan Wong

Title: Director of QA

Name of Head of Regulatory: Sigi Caron

Title: VP of RA & Clinical Affairs

Is Organization Chart Available? Yes
(Attach Copy of Org Chart)

Number of Employees at this Location: Twelve (12)

Number in QA: Three (3)

Number of Buildings: One (1) **Approximate Square Footage:** Office Space Only

Number of Years at this Location: Four (4)

Name of Registrar/Notified Body: N/A

Certificate Number(s): N/A
(Attach Copy)

Expiration Date(s): N/A

FDA Facility Registration Number: 3014650073 **Date of Last Inspection?** N/A
(Attach Copy of Registration)

Number of 483's: N/A

FDB License (CA only): N/A

Q#	21 CFR, Part 803 & 806	Compliant Yes/No (R/Y/G)	Supporting Evidence
	Additional Audit Questions		
	21 CFR, Part 803 Medical Device Reporting		
1	Does the manufacturer have a documented procedure to support Medical Device Reporting and the filing of MDRs?	Yes	Two procedures, employed in support of reporting adverse events, are: • QM.SLQ022.A – Medical Device Reporting; and • QM.SLQ023.A – eMDR Submission Work Instruction.
2	Is the procedure compliant with 21 CFR, Part 803 requirements?	Yes	Procedures employed in support of the eMDR process meet the regulatory requirements delineated within 21 CFR, Part 803.
3	Is the manufacturer using electronic reporting (eMDR)?	Yes	MDRs will be filled electronically (eMDR) through the FDA's Gateway in accordance with QM.SLQ023.A – eMDR Submission Work Instruction.
4	Does the written procedure cover all aspects of adverse event reporting?	Yes	Written procedures adequately address the FDA's regulatory requirements delineated within 21 CFR, Part 803.
5	Is the manufacturer filing supplemental reports when appropriate?	N/A	Not Applicable. SILQ has not filed a MDR in 2023.
6	How many MDRs has the establishment submitted since the beginning of this year (01 January 2023)?	N/A	Not Applicable. SILQ has not filed a MDR in 2023.
7	Where any of the MDRs filed late?	N/A	Not Applicable. SILQ has not filed a MDR in 2023.
8	Where any opportunities for filing an MDR missed?	N/A	Not Applicable. SILQ has not filed a MDR in 2023.
9	Is there a link back to the manufacturer's complaint management system?	Yes	The SILQ complaint procedure (<i>QM.SLQ021.C – Product Complaint System</i>) has been referenced in QM.SLQ022.A (<i>reference 3.1.1</i>).
	21 CFR, Part 806 Medical Devices; Reports of Corrections and Removals		
10	Does the manufacturer have a documented procedure to support device corrections and removals?	Yes	SILQ has scripted QM.SLQ030.A – Advisory Notices and Recalls in support of corrections and removals.
11	Is the procedure compliant with 21 CFR, Part 806 requirements?	Yes	QM.SLQ030.A as scripted by SILQ complies with 21 CFR, Part 806 requirements.
12	Have any device corrections or removals been performed in the previous 12-months?	N/A	Not Applicable. SILQ has not had to perform a market withdraw in 2023.
13	Were the corrections or removals reported to the FDA in accordance with 806.10?	N/A	Not Applicable. SILQ has not had to perform a market withdraw in 2023.
14	If recalls have occurred, what were their classification (I, II, or III)?	N/A	Not Applicable. SILQ has not had to perform a market withdraw in 2023.
15	Are records of corrections and removals being maintained and made available to the FDA in accordance with 806.30?	N/A	Not Applicable. SILQ has not had to perform a market withdraw in 2023.

Respectfully Submitted,

A handwritten signature in dark ink, appearing to read 'C. Devine', is written over a horizontal line.

Date: 17 November 2023

Christopher J. Devine, Ph.D.
President
Devine Guidance International, Inc.