



Date of Assessment: November 15th through 17th 2023

Client Name: SILQ Technologies Corporation **Client Address:** 323 Sunny Isles Blvd. (Seventh Floor); Sunny Isles Beach, FL 33160

Company Audited: SILQ Technologies Corporation **Address:** 323 Sunny Isles Blvd. (Seventh Floor); Sunny Isles Beach, FL 33160

Telephone Number: 786-590-5300

Name of Senior Company Official: D. V. Sharma

Title: CEO / President

Name of Head of Quality: Bryan Wong

Title: Director of QA

Name of Head of Regulatory: Sigi Caron

Title: VP of RA & Clinical Affairs

Is Organization Chart Available? Yes
(Attach Copy of Org Chart)

Number of Employees at this Location: Twelve (12)

Number in QA: Three (3)

Number of Buildings: Two (2) **Approximate Square Footage:** Sunny Isles (office space only) / Chico, CA (custom components)

Number of Years at this Location: Four (4)

Name of Registrar/Notified Body: N/A

Certificate Number(s): N/A
(Attach Copy)

Expiration Date(s): N/A

FDA Facility Registration Number: 3014650073 **Date of Last Inspection?** Jan 4th 2023
(Attach Copy of Registration)

Number of 483's: Zero (0)

FDB License (CA only): N/A

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	4 Quality Management System		
	4.1 General Requirements		
4.1 – 1	Has Organization established, documented, implemented and maintained a QMS and maintained (continually improved) its effectiveness in accordance with ISO 13485? Version?	Yes	The SILQ Quality Management System (QMS) complies with 21 CFR, Part 820 and ISO 13485:2016 requirements.
4.1 – 2	Where has the organization identified the processes needed for the QMS and their application throughout the organization?	Yes	SILQ procedures scripted in support of the QMS processes have been clearly identified.
4.1 – 3	Have all roles within the organization been adequately documented (e.g. manufacturer, EU Rep, etc.)?	Yes	QM.SLQ034.E – Organization Chart
4.1 – 4	Has the establishment applied a risk-based approach for all processes employed in support of the QMS?	Yes	The SILQ QMS has been scripted to ensure risk management has been appropriately addressed within: • QM.SLQ012.B – Risk Management; and • QM.SLQ013.B – Risk Analysis.
4.1 – 5	Has each QMS process been adequately defined: (a) criteria; (b) method for monitoring and measuring; (c) availability of resources; and (d) collection and maintenance of records?	Yes	SILQ quality processes reviewed in support of this audit have been adequately defined.
4.1 – 6	When the establishment outsources processes the impact product conformity, are the controls instituted in-line with the risk to product, regulatory requirements, or the QMS.	Yes	The following processes are outsourced: • Contract Manufacturing; • Metrology; • QA & RA Support, as needed; • Sterilization; • Medical Device Distribution; and • Audits.
4.1 – 7	As the establishment scripted procedures for the validation and application of computer software used in support of the QMS?	Yes	Software employed by SILQ (File Hold), in support of the QMS, has been validated.
	4.2 Documentation Requirements		
	4.2.1 General		
4.2 – 1	Does the organization have documented statements of a quality policy and quality objectives?	Yes	The SILQ Quality Policy has been documented within QM.SLQ035.A – Quality Policy.
4.2 – 2	Does the organization have the documented procedures required by ISO 13485?	Yes	Policies and procedures required by 21 CFR, Part 820 and ISO 13485:2016 have been appropriately implemented, as applicable.
4.2 – 3	Are adequate documents in place to ensure the effective planning, operation and control of the organization's processes?	Yes	Quality planning activities are pursued in accordance with QM.SLQ025.A – Quality Planning.
	4.2.2 Quality Manual		

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
4.2 – 4	Does the organization have a quality manual?	Yes	SILQ has implemented a Quality Manual (QM.SLQ027.B – <i>Quality Manual</i>).
4.2 – 5	Where in the quality manual is the scope of the QMS identified, including details of and justification for exclusions and/or requirements that don't apply?	Yes	In accordance with 2. 2.1 through 2.2.4 of QM.SLQ027.B; the following clauses have been identified as Not Applicable: • Clause 7.5.3 of ISO 13485:2016 – Installation Activities; • Clause 7.5.4 of ISO 13485:2016 – Servicing activities; • Clause 7.5.9.2 of ISO 13485:2016 – Particular Requirements for Implantable Medical Devices; and • Clause 7.5.10 of ISO 13485:2016 – Customer Property.
4.2.3 Medical Device File			
4.2 – 6	Has the establishment created a medical file for each product or each product family?	Yes	In accordance with 4.2.5 of QM.SLQ027.A; The Medical Device File (MDF) requirement is achieved through the use of a DMR (reference: <i>a DMR is a compilation of records containing the procedures and specifications for a finished device. Also referred to as the Medical Device File</i>).
4.2 – 7	Does the Medical Device File contain: • General description; • Intended use; • Labeling; • IFU; • Product specification; • Specifications/procedures for manufacturing, packaging, storage, handling, & distribution; • Procedures for measuring and monitoring; and • As appropriate procedures for installation and servicing.	Yes	In support of this assessment, the following DMR (Note: Pathway controls the DMR) was reviewed: SLQ-211810.A (SILQ-2-Way Cleared 18Fr x 10Mi Catheter, Sterile, 10-packs). Note: the DMR is comprised of nine (9) documents.
4.2.4 Control of Documents			
4.2 – 8	How are the documents required by the QMS controlled?	Yes	Documents are managed in accordance with QM.SLQ001.A – Document Control. Note: <i>the electronic document management process is delineated within QM SLQ014.B – Electronic Document System Work Instruction.</i>

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
4.2 – 9	Does the documented procedure that defines the controls for <u>each</u> of the following? • Review and approve documents for adequacy prior to issue? • Review and update as necessary and re-approve documents? • Ensure that changes and the current revision status of documents are identified? • Ensure that relevant versions of applicable documents are available at points of use? • Ensure that documents remain legible and readily identifiable? • Ensure that documents of external origin are identified and their distribution controlled? • Prevent the unintended use of obsolete documents, and to apply suitable identification to them if they are retained for any purpose.	Yes	For 2023, 12 DCOs were processed. In support of this assessment, the following DCOs were reviewed: • DCO059; • DCO061; • DCO062; • DCO064; • DCO065; • DCO068;and • DCO069.
4.2.5 Control of Records			
4.2 – 10	What records exist that provide evidence of conformity to requirements and of the effective operation of the QMS?	Yes	The following types of records were reviewed in support of this assessment: • FDA Establishment Registration; • Organizational Chart; • Quality Manual; • SOPs; • WIs; • DCOs; • DMRs; • Risk Management Docs (RMP, RMR, & FMEA's); • Management Review (including data & analysis) • Sales Orders; • Supplier Certifications; • Supplier Surveys; • DHF Documents; • Calibration Certificates; • Process Validations; • NCRs (Pathway); • Complaints; • Fume Hood Testing; and • Workstation Practices Procedures.
4.2 – 11	Are records legible, readily identifiable and retrievable?	Yes	Good Documentation Practices (QM.SLQ002.B – <i>Good Documentation Practices</i>) has been implemented in support of sustaining GDP at SILQ.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
4.2 – 12	Does the organization have a documented procedure for defining the controls needed for the identification, storage, protection, retrieval, retention time and disposition of records?	Yes	In accordance with 19.1 of QM.SLQ001.A; records are retained in File Hold. In accordance with 19.6 of QM.SLQ001.A; <i>“all quality system documents (and records) are to be retained for a minimum of 5 years based on the maximum potential lifetime of the medical device products at SILQ as required by ISO 13485 (Section 4.2.5 Control of Records), 21 CFR Part 820 (Subpart M Records) and MDD (Annex II Section 6.1, Annex V/VI Section 5/1). See Appendix 3 for retention requirements for all quality record types.”</i>
5 Management Responsibility			
5.1 Management Commitment			
5.1 – 1	How does top management communicate the importance of meeting customer and legal requirements to the organization?	Yes	Management communicates the importance of meeting customer requirements In accordance with 5.2 of QM.SLQ036.C – Sales Order SOP. The Chief Commercial Officer is tasked with: <i>“Reviews initiated sales orders and determines if customer requirements and applicable regulatory requirements can be met. If requirements can be met, Operations/Manufacturing approves the sales order.</i> <i>Works with applicable CMO to ensure product is shipped in accordance with the approved sales order.</i> <i>Ensures that traceability of shipped product to consignees and other relevant shipping information is maintained.”</i>
5.1 – 2	What are the quality objectives established by top management?	C	Concern. As of this audit, the quality objectives are under development. This was also noted in the Dec 2022 Management Review Meeting Minutes.
5.1 – 3	Does top management conduct management reviews?	Yes	SILQ management reviews are managed in accordance with QM.SLQ018.A – Management Review. In accordance with 6.1.1 of QM.SLQ018.A; <i>“at a minimum, Senior Management will hold meetings annually to review product quality and the effectiveness, suitability and adequacy of the Quality System to meet the company quality goals. If serious quality issues arise, it may be necessary to meet more frequently.”</i>
5.2 Customer Focus			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
5.2 – 1	How does top management ensure that customer requirements are determined and met?	Yes	Management communicates the importance of meeting customer requirements in accordance with 5.2 of QM.SLQ036.C – Sales Order SOP. The Chief Commercial Officer is tasked with: “Reviews initiated sales orders and determines if customer requirements and applicable regulatory requirements can be met. If requirements can be met, Operations/Manufacturing approves the sales order. - Works with applicable CMO to ensure product is shipped in accordance with the approved sales order. - Ensures that traceability of shipped product to consignees and other relevant shipping information is maintained.”
5.3 Quality Policy			
5.3 – 1	Has a company quality policy been established?	Yes	The SILQ Quality Policy has been documented within QM.SLQ035.A – Quality Policy. The policy is stated as: “Patient care is the focus of all SILQ activities. We endeavor to meet or exceed our customers’ satisfaction and expectations with safe and effective coated medical devices of the highest quality and that meets all applicable regulatory requirements.” “Each SILQ employee is personally committed to achieving this goal and continually improving our ability to maintain customer focus by operating a highly effective quality system that meets the requirements of applicable global standards and regulatory requirements.”
5.3 – 2	Is the quality policy: - Applicable to organization purpose; - A commitment to comply requirements; - Provide a framework for establishing quality objectives; - Communicated and understood throughout the organization; and - Reviewed for continued suitability.	Yes	The SILQ quality policy is appropriate for the organization and meets the intent of 21 CFR, Part 820 and ISO 13485:2016 requirements.
5.4 Planning			
5.4.1 Quality Objectives			
5.4 – 1	Has top management established quality objectives (including those needed to meet requirements for product) at relevant functions and levels within the organization?	C	Concern. As of this audit, the quality objectives are under development. This was also noted in the Dec 2022 Management review Meeting Minutes.
5.4.2 Quality Management System Planning			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
5.4 – 2	How does the organization ensure that the integrity of the QMS is maintained when changes to the QMS are planned and implemented?	Yes	SILQ pursues QMS planning in accordance with QM.SLQ025.A – Quality Planning.
	5.5 Responsibility, Authority and Communication		
	5.5.1 Responsibility and Authority		
5.5 – 1	How are responsibilities and authorities defined, documented and communicated within the organization?	Yes	In accordance with 7.5.1 of QM.SLQ027.B; responsibility and authority are delineated through the use of the SILQ Organizational Chart.
	5.5.2 Management Representative		
5.5 – 2	Who is the management representative for quality?	Yes	In accordance with QM.SLQ034.E – Organization Chart; Tavi Cruz-Urbe (VP of QA) is the designated management representative.
	5.5.3 Internal Communication		
5.5 – 3	How is information regarding the effectiveness of the QMS communicated within the organization?	Yes	In accordance with 7.5.3 of QM.SLQ027.B; <i>“SILQ ensures effective communication between its various levels and functions regarding the processes of the quality management system and their effectiveness, suitability, and adequacy through periodic company meetings and Management Reviews. It is also the responsibility of each Manager or department head to ensure all relevant information is passed on to their functions.”</i>
	5.6 Management Review		
	5.6.1 General		
5.6 – 1	What is the frequency of management reviews?	Yes	In accordance with 6.1.1 of QM.SLQ018.A; management reviews are to be held annually, as a minimum.
5.6 – 2	Does the organization have documented evidence that management reviews are being held in accordance with the published schedule?	Yes	The initial SILQ Management Review was held on 21 December 2022. The review was well-attended by the SILQ management team.
	5.6.2 Review Input		

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
5.6 – 3	What kind of information is input into management reviews?	Yes	As a minimum, the following agenda items are delineated within 6.3 of QM.SLQ018.A: • “Review of minutes and action items from the previous Management Review (required at each meeting); • Review of the Quality Policy; • Review of quality objectives; • Results, actions and status of internal audits, regulatory agency inspections and notified body audits; • Status of the Corrective and Preventive Action System; • Review of Product Complaint trends; • Review of Production Data including nonconformities; • Review of Training System performance; • Post Market Surveillance / Customer Feedback; • Review of reporting to regulatory authorities; • Review of statistical techniques (sampling plans) used to evaluate the quality of components, materials and finished goods to assure continued applicability; • Review of supplier performance; • Changes that could affect the quality management system; • Recommendations for improvement; • New or revised standards and regulatory requirements; • Review of Risk Management Activities; • Review status of Design Control Activities; and • Review of Resource Requirements and current staffing levels.”
	5.6.3 Review Output		
5.6 – 4	What decisions, actions, and/or outcomes have resulted from management reviews?	Yes	Review outputs, delineated within 6.9 of QM.SLQ018.A, consist of: • “Improvements needed to maintain the effectiveness, suitability and adequacy of the quality management system and its processes; • Improvement of product related to customer requirements; • Changes needed to respond to applicable new or revised regulatory requirements; • Resource needs; and • Quality objectives established.”
	6 Resource Management		
	6.1 Provision of Resources		
6.1 – 1	What resources has the organization provided to implement and maintain the QMS and continually improve its effectiveness?	Yes	SILQ has an established quality organization (staff of 3) led by Tavi Cruz-Uribe.
6.1 – 2	What resources has the organization provided to ensure that customer and regulatory requirements are met?	Yes	SILQ has retained a qualified VP of Regulatory and Clinical Affairs to support the RA function.
	6.2 Human Resources		

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
6.2 – 1	What are the education, training, skills and experience required by this job/task?	Yes	In accordance with 8.2 (Human Resources) of QM.SLQ027.B; <i>“management identifies training needs and provides training for all personnel performing tasks affecting quality as required. Training and effectiveness of training is conducted in accordance with established procedures by qualified individuals. Appropriate records of education, experience, training, and skills are maintained.”</i>
6.2 – 2	How does the organization determine the necessary education, training, skills and experience for people performing work affecting product quality?	Yes	In accordance with 8.2 of QM.SLQ027.B; <i>“personnel performing work affecting product quality are selected based on required competencies, appropriate education, training (internal and/or external), skills, and experience.”</i>
6.2 – 3	What training or other actions does the organization provide to satisfy the needs of personnel?	Yes	SILQ manages employee training in accordance with QM.SLQ003.B – Employee Training.
6.2 – 4	When the organization provides training or other actions to satisfy competence needs, how does it evaluate the effectiveness of those actions?	Yes	In accordance with 6.1 of QM.SLQ003.B; <i>“all SILQ personnel and contractors / consultants are to be determined competent to perform their job based on appropriate education, experience, skills and training. Management and at least one other departmental representative conduct candidate interviews and complete reference checks as applicable. Documentation of the employee’s and contractor’s / consultant’s qualifications (e.g., resume) is filed in the respective employee file.”</i>
6.2 – 5	Where does the organization maintain records of education, training, skills and experience?	Yes	SILQ records for training are managed in accordance with Section 10 of QM.SLQ003.B. Training records are retained in File Hold. In support of this assessment, the following training records were reviewed: • Training Matrix: President/CEO; • QM.SLQ016.C – CAPA (01-31-23) – V. Sharma; • Training Matrix: QA Director; • QM.SLQ019.C – ID & Traceability (10-11-22) – B. Wong; • Training Matrix: Quality Specialist; and • QM.SLQ0001. A – Document Control (10-12-20) – D. Porter.
6.3 Infrastructure			
6.3 – 1	Are the buildings, workspace, and utilities appropriate to meet product requirements?	NR	Not Reviewed: This assessment was remote; therefore, the facility and equipment were not assessed during this audit. <i>Note: the primary facility is essentially office space.</i>
6.3 – 2	Is the process equipment (hardware & service) appropriate for the manufacture of medical devices that are safe and effective in their intended use?	NR	Not Reviewed: This assessment was remote; therefore, the facility and equipment were not assessed during this audit. <i>Note: the primary facility is essentially office space.</i>
6.3 – 3	Are the requirements for maintenance activities being appropriately documented and the records maintained?	NR	Not Reviewed: This assessment was remote; therefore, the facility and equipment were not assessed during this audit. <i>Note: the primary facility is essentially office space.</i>
6.4 Work Environment & Contamination Control			
6.4.1 Work Environment			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
6.4 – 1	What kind of work environment is required to achieve conformity to product requirements?	Yes	Manufacturing occurs at a critical subcontractor facility. The manufacturing environment is required to be in a Clean Room. <i>Note: manufacturing at the Chico facility is managed in accordance with Good Laboratory Practices (fume hoods are used). The Fume Hood is tested by Independent Air Group (HVAC Test & Balance Report). The following test report for EF-1 was reviewed (Dec 2022). Additionally, in accordance with section 13 of QM.SLQ049.A – Workstation Practices, requirements for working in the Chico facility (under the Fume Hood) is defined by procedure.</i>
6.4 – 2	How is this environment managed and maintained?	Yes	Manufacturing occurs at a critical subcontractor facility. The manufacturing environment is required to be in a Clean Room. <i>Note: manufacturing at the Chico facility is managed in accordance with Good Laboratory Practices (fume hood is used).</i>
6.4 – 3	Are requirements for employee health, cleanliness, and clothing been documented in a procedure?	Yes	In accordance with 8.5 of QM.SLQ027.A; “CMOs that provide product or services are selected based on their ability to fulfill applicable requirements for work environment and contamination control.”
6.4.2 Contamination Control			
6.4 – 4	Has the organization established a procedure for the control of contaminated product?	N/A	Not Applicable. SILQ does not handle contaminated product. Contaminated product shall be returned to the contract manufacturer for failure investigations.
6.4 – 5	Does the organization have a CER? If so, does the organization have procedures that define the monitoring and measuring of environmental conditions that may impact finished medical devices (i.e., particle counts, temperature, RH differential pressure, surface contamination, & product bio-burden)?	N/A	Not Applicable. SILQ does not maintain a Controlled Environment Room (CER) in the Florida facility (office space only).
7 Product Realization			
7.1 Planning of Product Realization			
7.1 – 1	Where are the processes needed for product realization identified?	Yes	SILQ product realization requirements are delineated within the Device Master Record (DMR). As previously noted in this checklist, the DMR doubles as the MDF. <i>Note: the DMR is retained by the critical subcontractor. As previously noted in this checklist, DMR SLQ-211810.A was reviewed.</i>
7.1 – 2	Is the planning of product realization consistent with the requirements of the other processes of the QMS?	Yes	Product realization activities are performed in accordance with requirements delineated within the SILQ QMS.
7.1 – 3	Where in the product realization process does the organization determine the quality objectives and requirements for products?	Yes	Requirements associated with SILQ products are delineated within the product specification and translated into requirements needed for product realization through the DMR.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.1 – 4	When planning for product realization, how does the organization establish processes, documents, and provide resources specific to the product?	Yes	Processes needed for successful product realization are developed in conjunction with the critical subcontractor.
7.1 – 5	How does the organization determine verification, validation, monitoring, inspection and test activities specific to the product, and the criteria for product acceptance?	Yes	Requirements for verification, validation, monitoring, inspection, and test activities, necessary to facilitate product realization, are developed in conjunction with the critical subcontractor.
7.1 – 6	What records exist depicting that both the realization processes and the product meet requirements?	Yes	Records associated with the performance of product realization activities are retained in the Device History Records (DHR).
7.1 – 7	What are the outputs of product realization planning?	Yes	The output of product realization is a finished SILQ device that is safe and effective in its intended use.
7.2 Customer-Related Processes			
7.2.1 Determination of Requirements Related to the Product			
7.2 – 1	How does Organization determine <u>each</u> of the following requirements? - Customer Specified - Legal & Regulatory - Organizational	Yes	Requirements for SILQ devices are determined during the design and development phase of product realization. The requirements, once determined, are documented within the product specification.
7.2.2 Review of Requirements Related to the Product			
7.2 – 2	What kind of review is performed to ensure that the organization has the ability to meet requirements before committing to supply product?	Yes	Customer orders are reviewed in accordance with 6.3 of QM.SLQ036.C – Sales Order.
7.2 – 3	How does the organization ensure that product requirements are defined, documented, and reviewed before committing to supply product?	Yes	In accordance with 6.1.4 of QM.SLQ036.C; “if there are any requirements not stated by the customer but necessary for specified or intended use, a SILQ associate will confirm these requirements with the customer before accepting the order.
7.2 – 4	How does the organization ensure that contract or order requirements differing from those previously expressed are resolved before committing to supply product?	Yes	In accordance with 6.3.1.3 of QM.SLQ036.C; requirements differing from previous orders are required to be resolved prior to order acceptance.
7.2 – 5	Does the organization have records of the product-requirement review results and actions resulting from them?	Yes	In support of this assessment, the following sales-order forms were reviewed (Sage): • SO # 0000050; • SO # 0000053; • SO # 0000059; • SO # 0000064; • SO # 0000067; • SO # 0000069; and • SO # 0000072.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.2 – 6	When customers do not have documented requirements, how does the organization confirm their requirements before accepting orders?	Yes	Requirements reviewed as part of the sales order review process are delineated within 6.1.2.1 through 6.1.2.10 of QM.SLQ036.C and consist of: • Customer contact information; • Recipient (e.g., Ship To) contact information; • Product and quantity requested, including cost; • Sales order number; • Customer purchase order number, if available; • Sales order date (e.g., date when sales order was initiated); • Name of SILQ Sales/Customer Service representative receiving the sales order; • Method of shipment; • Required delivery date; and • Any additional requirements specified by the customer, including requirements for delivery and post-delivery activities.
7.2 – 7	When product requirements are changed, how does the organization ensure that relevant documents are changed and that relevant personnel are made aware of the changes?	Yes	Changes made to product(s) are managed as part of the design and development process.
7.2.3 Customer Communication			
7.2 – 8	What method(s) are used to communicate with customers regarding - Product information? - Enquiries, contracts, or order handling, including amendments? - Feedback, including customer complaints? - Advisory notices?	Yes	In accordance with 6.1.1 of QMSLQ036.C; <i>“if the customer requests a product that is a SILQ current part number, a Sales/Customer Service associate (or designee) receives the customer’s order from phone, fax, or other means and initiates the sales order.”</i>
7.3 Design and Development			
7.3.1 General			
7.3 – 1	Has the organization established procedures in support of design and development activities?	Yes	Multiple procedures have been scripted in support of design and development activities: • QM.SLQ004.A – Design Control Program; • QM.SLQ005.B – Design Project Planning; • QM.SLQ006.A – Design Input; • QM.SLQ007.A – Design Output; • QM.SLQ008.A – Design Review; • QM.SLQ009.A – Design Verification & Validation; • QM.SLQ010.A – Design Transfer; and • QM.SLQ011.A – Statistical techniques – Work Instruction.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	7.3.2 Design and Development Planning		
7.3 – 2	Does the organization have a documented procedure for Design & Development Planning?	Yes	QM.SLQ005.B – Design Project Planning has been scripted in support of design and development planning. <i>Note: SILQ has one Class 2 device cleared by FDA (reference K192034).</i>
7.3 – 3	How does Organization ensure effective communication and clear assignment of responsibility between different groups involved in design and development?	Yes	In accordance with 5.3 of QM.SLQ005.B – Design Project Planning, project managers are tasked with communicating roles, responsibilities, and project requirements to the project team.
7.3 – 4	As product design and development progresses, how are the planning outputs documented and updated?	Yes	Project # 00053 – SILQ Clear Tract Foley Catheter Supra Pubic Indications for Use (USA) was audited in support of design control compliance.
	7.3.3 Design and Development Inputs		
7.3 – 5	What are the design inputs the organization typically considers?	Yes	SILQ requirements for design inputs are defined within QM.SLQ006.A – Design Input.
7.3 – 6	When identified, where are the design and development inputs recorded?	Yes	Design inputs are placed into the product specification. In support of this audit, PS-0006. Rev E – Product Specification / Traceability Matrix – SILQ Foley Products (Pathway) was reviewed. <i>Note: Reference Project 00053</i>
7.3 – 7	How and when are the design and development inputs reviewed and approved for adequacy?	Yes	The design and development inputs are reviewed and approved during design review.
7.3 – 8	How does the organization ensure that requirements are complete, unambiguous and don't conflict with each other?	Yes	In accordance with 1.2 of QM.SLQ006.A; <i>"this procedure describes the general process to identify, document and approve user needs and product requirements for a given product. In addition, this procedure details a system to resolve incomplete, ambiguous, or conflicting product requirements"</i>
	7.3.4 Design and Development Outputs		
7.3 – 9	How are design and development outputs verified against the design inputs?	Yes	Design outputs are managed in accordance with QM.SLQ007.A – Design Output.
7.3 – 10	Are design and development outputs approved prior to their release?	Yes	Design outputs are identified during phase reviews. Once the design team agrees on the required outputs, they are documented within the product specification.
7.3 – 11	What outputs include information for purchasing, production and service?	Yes	Purchasing will employ the design outputs in support of identifying suppliers needed to support product realization.
7.3 – 12	Where is the product acceptance criteria specified?	Yes	The criterion for product acceptance is documented in the product specification and in inspection documentation.
7.3 – 13	Where are product characteristics needed for safe and proper use specified?	Yes	The characteristic and user information needed for the safe and effective use of SILQ devices are placed into the IFU.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.3 – 14	Where are records of design and development maintained?	Yes	Records of design and development activities are retained in the Design History File (DHF).
	7.3.5 Design and Development Review		
7.3 – 15	At what stages of design and development does the organization perform reviews to evaluate if the results meet requirements?	Yes	Design Reviews are being pursued in accordance with QM.SLQ008.A – Design Review.
7.3 – 16	What functions (including specialists) are represented at these reviews?	Yes	“At a minimum, a representative from Quality Assurance, R&D/Engineering, and Manufacturing must attend the design review. When appropriate, additional representatives from other departments or other subject matter experts are to attend if requested or deemed necessary.” “An independent reviewer, not directly associated with the design stage being reviewed, is to attend all design reviews. The independent reviewer is to be identified by Quality Assurance management.”
7.3 – 17	Where are the records associated with the Design and Development Reviews maintained?	R	In support of this assessment, the following phase review was reviewed: - Project # 00053 – SILQ Clear Tract Foley Catheter Supra Pubic Indications for Use (USA) (reference PS-0006, Rev E). Note: Testing was performed by LGGS. The following documents were reviewed in support of this assessment: - RM-0018, Rev D – RMP, HDX Foley Catheter; - RM-0019, Rev D – Hazard Analysis (RMP) HDX Foley Catheter; - RM-0020, Rev D – dFMEA – RMR, HDX Foley Catheter; - SLQ.211605, Rev A- SILQ 2-Way Cleartract 16 Fr x 5ml – BOM; - RM-0021, Rev D – pFMEA – RMR, HDX Foley Catheter; and - TR.SLQ-0012, Rev A – Bio-compatibility Testing in Accordance with ISO 10993-1:2018. Recommendation: RM-0019, Rev D is referencing ISO 14971:2012 requirements in the tables.
	7.3.6 Design and Development Verification		
7.3 – 18	What verification activities are performed to ensure that the design and development outputs have met the input requirements?	Yes	DV activities are pursued in accordance with QM.SLQ009.A – Design Verification & Validation.
7.3 – 19	Are records containing the results of verification activities and their resulting actions collected and maintained?	Yes	DV protocols and reports are reviewed and approved by SILQ and retained by the critical subcontractor.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	7.3.7 Design and Development Validation		
7.3 – 20	What design and development validation activities are performed to ensure that the product is capable of meeting the requirements for the intended use?	Yes	DVAL activities are pursued in accordance with QM.SLQ009.A – Design Verification & Validation
7.3 – 21	Do records show that validation is performed prior to product shipment?	Yes	Protocols and reports reviewed by this auditor reflect the individuals performing and approving DVAL activities.
7.3 – 22	If not, is the justification recorded?	Yes	When DVAL activities are not performed, adoption reports are scripted providing adequate rationale as to why DVAL was not pursued; e.g., TR-0075, Rev B – Sterilization Product Adoption, HDX 2-Way Foley Catheter. There has been no changes required to be made to TR-0075, Rev B.
	7.3.8 Design and Development Transfer		
7.3 – 23	Does the organization have a documented procedure for Design & Development Transfer?	Yes	SILQ design transfer activities are managed in accordance with QM.SLQ010.A – Design Transfer
	7.3.9 Control of Design and Development Changes		
7.3 – 24	How are design and development changes identified?	Yes	Design changes are managed in accordance with section 16 (Design Changes) of QM.SLQ004.A.
7.3 – 25	Where are the records maintained?	Yes	Design control is managed by Pathway. The design control records are being managed in accordance with the Pathway QMS. <i>Note: Pathway is assessed annually by SILQ.</i>
7.3 – 26	Are changes reviewed, verified, validated, and approved before implementation?	Yes	In accordance with 16.2 of QM.SLQ004.A; <i>“in cases where major changes to a product line are made, an updated or supplemental Design Project Scope Form (FM1-QM.SLQ004) is to be completed to determine the extent of design control activities required.”</i>
7.3 – 27	How does the organization document the review of design and development changes including evaluation of the effect on component parts and products in the field?	Yes	In accordance with 16.4 of QM.SLQ004.A; <i>“design changes are to be handled through the document change order process as described in the QM.SLQ001, Document Control SOP.”</i>
	7.3.10 Design and Development Files		
7.3 – 28	Does the organization have a documented procedure for managing design and development files (e.g., DHF)?	Yes	In accordance with section 15 (Design History File) of QM.SLQ004.A; design and development files are retained in the DHF (<i>managed by Pathway</i>).
	7.4 Purchasing		
	7.4.1 Purchasing Process		
7.4 – 1	Does the organization have a written procedure for the Purchasing Process?	Yes	Procurement activities are managed in accordance with QM.SLQ020.B – Purchasing Controls.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.4 – 2	How does the organization ensure that purchased product conforms to specified requirements?	Yes	- In accordance with 7.7 of QM.SLQ020.B; “products and services are only to be purchased from suppliers, contractors and consultants listed on the Approved Supplier List (ASL) and only for those product and service types for which they have been qualified (as specified on the ASL).” Additionally, the use of purchase orders and receiving inspection support the procurement of conforming products and services. - In accordance with 7.8 of QM.SLQ020.B; “purchasing of products must be in accordance with released, document controlled product specifications (i.e., engineering drawing, source control drawing, etc.).”
7.4 – 3	How does the organization determine the type and extent of control applied to the supplier and the purchased product?	Yes	Supplier controls implemented by SILQ are determined premised on product and supplier risk.
7.4 – 4	How does the organization evaluate and select suppliers?	Yes	SILQ suppliers are selected and evaluated in accordance with QM.SLQ015.B – Supplier Quality Assurance.
7.4 – 5	Is the criterion for supplier selection, evaluation and re-evaluation of suppliers documented by procedure?	Yes	Suppliers are placed into one of five categories defined within 7.1 through 7.5 of QM.SLQ015.B. The categories are premised on risk. In support of this assessment the following suppliers were reviewed: - Cole-Parmer (Class IV) – Supplier Quality Self-Assessment Survey Form; - Firefly Scientific (Class III) – ISO/IEC 17025:2017; - Micro Precision (Class III) – ISO/IEC 17025:2017; - Pathway (Class I) – ISO 13485:2016; and - Repligen (Class III) – ISO 9001:2015.
7.4 – 6	What type of records are collected and maintained in support of supplier evaluation and management activities?	Yes	In support of this assessment, the following purchasing/supplier quality records were reviewed: - ASL (FM-6 QMS.SLQ015.B – Approved Suppliers List); - Supplier Agreements; - ISO Certifications; - Self-Assessment Surveys; and - On-site Surveys.
7.4.2 Purchasing Information			
7.4 – 7	Do orders/contracts include requirements for approval of product, procedures, processes and equipment?	Yes	The creation and use of purchase orders is defined within section 7 (Purchasing) of QM.SLQ020.B. In accordance with 7.1 of QM.SLQ020.B; “purchase Orders (P.O.) constitute a contract between SILQ and a supplier, contractor or consultant and may be issued only by authorized purchasing personnel, with department specific approvals as described in this procedure.”

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.4 – 8	Does the organization require special qualifications for supplier personnel?	Yes	Special requirements for supplier, including supplier qualifications, will be delineated through the use of purchase orders, contracts, drawings, and relevant attachments.
7.4 – 9	If special qualifications are required, where is the requirement documented?	Yes	Special requirements for supplier, including supplier qualifications, will be delineated through the use of purchase orders, contracts, drawings, and relevant attachments.
7.4 – 10	Does the organization have QMS requirements for their suppliers?	Yes	Special requirements for supplier, including supplier qualifications, will be delineated through the use of purchase orders, contracts, drawings, and relevant attachments.
7.4.3 Verification of Purchased Product			
7.4 – 11	What inspection or other activities are used to ensure that purchased product meets the organization's purchasing requirements?	Yes	SILQ has established a procedure for receiving inspection; QM.SLQ039.B – Receiving Inspection.
7.4 – 12	Does the organization ever perform product verification (source inspection) at the supplier's site?	N/A	Not Applicable. SILQ retains the right of access for their supplier base; however, SILQ does not typically perform source inspection.
7.5 Production and Service Provision			
7.5.1 Control of Production and Service Provision			
7.5 – 1	When carrying out production or service what controlled conditions in place? - Documented Procedures - Work Instructions - Suitable Equipment - Gages - Measuring Equipment - Packaging & Labeling	Yes	SILQ and SILQ's critical subcontractor employ all available tools to support product realization activities. However, as previously noted, a critical subcontractor is tasked with performing manufacturing activities.
7.5 – 2	Is batch management performed in support of product traceability requirements?	Yes	In accordance with 7.6 of QM.SLQ019.C – Identification and Traceability; <i>"batch or lot numbers that have been assigned by the manufacturer (supplier) are acceptable to maintain and reference in the DHR, without having to assign a new identifying number."</i>
7.5.2 Cleanliness of Product			
7.5 – 3	Does the organization have established and documented requirements for cleanliness of product?	Yes	Product cleanliness requirements are determined in conjunction with the critical subcontractor.
7.5 – 4	Are products cleaned prior to sterilization?	Yes	SILQ products are assembled in a controlled environment (Clean Room). Product is appropriately cleaned prior to shipment for terminal sterilization through the use of EO.
7.5.3 Installation Activities			
7.5 – 5	If appropriate, does the organization have established documented requirements which contain acceptance criteria for installing and verifying the installation of the medical device?	N/A	Not Applicable. Clause 7.5.3 has been declared as "N/A" in the quality manual (QM.SLQ027.B).

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.5 – 6	If the agreed upon customer requirements allow installation to be performed other than by the organization or its authorized agent, does the organization provide documented requirements for installation and verification?	N/A	Not Applicable. Clause 7.5.3 has been declared as “N/A” in the quality manual (QM.SLQ027.B).
7.5 – 7	Where are the records of installation and verification performed by the organization or its authorized agent maintained?	N/A	Not Applicable. Clause 7.5.3 has been declared as “N/A” in the quality manual (QM.SLQ027.B).
7.5.4 Servicing Activities			
7.5 – 8	If servicing is a specified requirement, does the organization have established and documented procedures, work instructions, reference materials, and reference measurement procedures, as necessary, for performing servicing activities?	N/A	Not Applicable. Clause 7.5.4 has been declared as “N/A” in the quality manual (QM.SLQ027.B).
7.5 – 9	Where are the records of servicing activities, carried out by the organization, maintained?	N/A	Not Applicable. Clause 7.5.4 has been declared as “N/A” in the quality manual (QM.SLQ027.B).
7.5.5 Particular Requirements for Sterile Medical Devices			
7.5 – 10	Does the organization maintain records of the process parameters for the sterilization process, which is employed for each sterilization batch?	Yes	Terminal sterilization is performed using EO. Sterilization records are managed through the use of a batch number. The sterile load records are retained in the DHR.
7.5 – 11	Are sterilization records traceable to each production batch of medical devices?	Yes	Terminal sterilization is performed using EO. Sterilization records are managed through the use of a batch number. The sterile load records are retained in the DHR.
7.5.6 Validation of Processes for Production and Service Provision			
7.5 – 12	Does the organization have any production or service processes where the resulting output cannot be verified later?	Yes	Process validation activities are managed in accordance with QM.SLQ047.A – Process Validation. <i>Note: Process Validation (PV) activities are the responsibility of the critical subcontractor; however, SILQ is tasked with the review and approval of PV activities, as appropriate.</i>
7.5 – 13	Are records available that demonstrate that the validation performed has met the requirements?	Yes	In support of this assessment, the following PV's were reviewed: • TR-017, Rev A – IQ Coater; • TR-HDX-0005, Rev A – OQ – Hydrophilix Foley Catheter Surface Treatment Using the Automation Machine; • TR-HDX-0006, Rev A – PQ – Hydrophilix Foley Catheter Surface Treatment process; • TR-0137, Rev A – IQ – Checklist Spectrophotometer; and • TR-0122, rev A – TMV – Measurement System Analysis, UV Light Absorbance Test, Hydrophilix Products.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.5 – 14	How are special processes reviewed and approved?	Yes	All processes employed in support of product realization are either validated or verified through inspection and/or test.
7.5 – 15	Are records of personnel and equipment qualification collected and maintained?	Yes	In support of this assessment, the following PV's were reviewed: • TR-017, Rev A – IQ Coater; • TR-HDX-0005, Rev A – OQ – Hydrophilix Foley Catheter Surface Treatment Using the Automation Machine; • TR-HDX-0006, Rev A – PQ – Hydrophilix Foley Catheter Surface Treatment process; • TR-0137, Rev A – IQ – Checklist Spectrophotometer; and • TR-0122, rev A – TMV – Measurement System Analysis, UV Light Absorbance Test, Hydrophilix Products.
7.5 – 16	When changes are made to processes, how does the organization revalidate the changes?	Yes	Changes made to processes are assessed for the need to repeat partial PV testing or the entire PV protocol.
7.5 – 17	Does the organization have established and documented procedures for the validation of the application of computer software?	Yes	SILQ validates software in accordance with the requirements delineated within QM.QLS032.A – Software Validation.
7.5.7 Particular Requirements for Validation of Processes for Sterilization and Sterile Barrier Systems			
7.5 – 18	Has the organization established documented procedures for the validation of sterilization and sterile barrier systems?	Yes	TR-0075, Rev B – Sterilization Product Adoption, HDX 2-Way Foley Catheter <i>Note: there have been zero need(s) to update the adoption report in 2023. No significant material changes have been made to SILQ finished devices.</i>
7.5 – 19	Where are the records for validation of sterilization processes maintained?	Yes	TR-0075, Rev B – Sterilization Product Adoption, HDX 2-Way Foley Catheter <i>Note: there have been zero need(s) to update the adoption report in 2023. No significant material changes have been made to SILQ finished devices.</i>
7.5.8 Identification			
7.5 – 20	Where appropriate, does the organization identify the product by suitable means throughout product realization?	Yes	In accordance with 7.6 of QM.SLQ019.C – Identification and Traceability; <i>"batch or lot numbers that have been assigned by the manufacturer (supplier) are acceptable to maintain and reference in the DHR, without having to assign a new identifying number."</i>
7.5 – 21	Does the organization have an established and documented procedure for product identification?	Yes	Product identification and traceability activities are managed in accordance with QM.SLQ019.C – Identification and Traceability.
7.5 – 22	Does the organization have established and documented procedures to ensure that medical devices returned to the organization are identified and distinguished from conforming product?	Yes	SILQ medical devices are not returned to SILQ, devices are returned to Pathway, LLC. Pathway performs device decontaminations, failure investigation activities, and provides SILQ with copies of the failure reports.
7.5.9 Traceability			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.5 – 23	Does the organization have established and documented procedures for traceability?	Yes	Product identification and traceability activities are managed in accordance with QM.SLQ019.C – Identification and Traceability.
7.5 – 24	Do the procedures define the extent of product traceability and the records required?	Yes	Finished device traceability is managed in accordance with section 10 (Finished Device Traceability) of QM.SLC019.C.
7.5 – 25	Where traceability is a requirement, does the organization control and record the unique identification of the product?	Yes	The critical subcontractor is tasked with the assignment of batch numbers.
7.5 – 26	In defining the records required for traceability, does the organization include records of all components, materials, and work environment conditions, if these could cause the medical device not to satisfy its specified requirements?	Yes	Records of materials procured for medical device manufacturing are managed in accordance with 7.1 of QM.SLQ019.C; <i>“all purchased products including raw materials, components, fabricated parts, finished goods, manufacturing materials, and supplies are to have unique part numbers assigned and a corresponding specification developed and released.”</i>
7.5 – 27	Does the organization require that its agents or distributors maintain records of the distribution of medical devices to allow traceability; and ensure these records are available for inspection?	Yes	In accordance with 13.5 of QM.SLQ027.B; <i>“Purchased product is subjected to incoming acceptance activities. In-process inspections are in the form of operator-based inspections. Acceptance activities are established and recorded to ensure product conformity. Finished products are subjected to a final inspection.”</i> <i>“Products are released by Quality Assurance of applicable CMO for distribution after all specified activities have been satisfactorily completed and required documentation of the product has been verified.”</i>
7.5 – 28	Are records of the name and address of the shipping package consignee maintained?	Yes	Records are available through Ship Station Software (managed by Pathway). The sales order is opened by SILQ and the product is shipped by Pathway.
7.5.10 Customer Property			
7.5 – 29	Does the organization exercise care with customer property while it is under the organization's control or being used by the organization?	N/A	Not Applicable. Clause 7.5.10 has been declared as “N/A” in the quality manual (QM.SLQ027.B).
7.5 – 30	Does the organization identify, verify, protect, and safeguard customer property provided for use or incorporation into the product?	N/A	Not Applicable. Clause 7.5.10 has been declared as “N/A” in the quality manual (QM.SLQ027.B).
7.5 – 31	If customer property is lost, damaged, or otherwise found to be unsuitable for use, does the organization report the issues to the customer?	N/A	Not Applicable. Clause 7.5.10 has been declared as “N/A” in the quality manual (QM.SLQ027.B).
7.5.11 Preservation of Product			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.5 – 32	Does the organization have established and documented procedures or documented work instructions for preserving the conformity of product during internal processing and delivery to the intended destination?	Yes	In accordance with 12.6 of QM.SLQ027.B: “Designated storage areas are established to prevent damage, deterioration or contamination of the product pending delivery. This preservation includes identification, handling, packaging, storage and protection. Preservation shall also apply to the component level.” “Finished product and components are used subject to FIFO (First-In First-Out) practices. Prior to shipping, product shall be assessed to ensure expiration dates, if applicable, have not been exceeded.” “Product known to be subject to deterioration over time shall be identified and assessed at defined intervals. If applicable, product with a limited shelf life is identified as such with an expiration date on receipt into stock.” “Components shall be maintained according to any special storage conditions indicated.”
7.5 – 33	Does this preservation include identification, handling, packaging, storage, and protection?	Yes	Specific processes employed in support of product realization and instructions for the product preservation are the responsibility of the critical subcontractor.
7.5 – 34	Does the organization have established and documented procedures or documented work instructions for the control of product with a limited shelf-life or requiring special storage conditions?	Yes	Materials that have a shelf-life or special handling and storage requirement are managed by the critical subcontractor.
7.5 – 35	Are special storage conditions controlled and recorded?	Yes	Materials that have a shelf-life or special handling and storage requirement are managed by the critical subcontractor.
7.6 Control of Monitoring and Measuring Devices			
7.6 – 1	How does the organization determine the measurements to be taken and the measuring equipment needed to demonstrate conformity with requirements?	Yes	Measuring and monitoring equipment necessary to support effective product realization is selected and managed by the CMO.
7.6 – 2	Does the organization have established and documented procedures to ensure that monitoring and measurement can be carried out and are carried out in a manner that is consistent with the monitoring and measurement requirements?	Yes	Metrology activities are managed in accordance with QM.SLQ050.A – Calibration and Preventive Maintenance
7.6 – 3	How does the organization ensure that measuring and test equipment is calibrated or verified proper frequencies with NIST traceable standards?	Yes	In accordance with 11.3 of QM.SLQ050.A; “calibration shall be done to a NIST traceable standard. Calibrations to other than NIST traceable standards shall have a documented rationale for use of that standard and be filed with the equipment file.”

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.6 – 4	What process is used to adjust or re-adjust measuring and test equipment when needed?	Yes	The adjusting and re-adjusting of measuring and monitoring equipment will be managed by the CMO as part of the calibration process.
7.6 – 5	How are measuring tools identified to enable the calibration status to be determined?	Yes	In accordance with 11.1 of QM.SLQ050.A; measuring and monitoring equipment shall be identified with a calibration label. <i>“Calibration labels will include: calibration date, initials of person who performed the calibration, and calibration due date.”</i>
7.6 – 6	How does the organization safeguard measuring equipment from adjustments that would invalidate the measurement results?	Yes	Only SILQ personnel that have been trained to use measuring and monitoring equipment shall be permitted to use measuring and monitoring equipment.
7.6 – 7	How does the organization ensure that measuring its test equipment is protected from damage and deterioration during handling, maintenance and storage?	Yes	In support of preventing damage, the handling of measuring and monitoring equipment is managed in accordance with 12.1 of QM.SLQ050.A. <i>“All measurement equipment shall be handled, stored, packaged, and transported (as applicable) in a manner that will not adversely affect the calibration or condition of the equipment.”</i>
7.6 – 8	Does the organization assess and record the validity of the previous measuring results when the equipment is found not to conform to requirements?	Yes	Verifying the validity of calibration results is managed in accordance with 11.9.1 of QM.SLQ050.A;
7.6 – 9	What action does the organization pursue when equipment and product are affected when equipment non-conformances are identified?	Yes	Out-of-tolerance conditions, non-conformances and/or other discrepancies associated with measuring and monitoring equipment shall be managed in accordance with 11.9 of QM. SLQ050.A. <i>Note: in accordance with 11.9.1 of QM.SLQ050.A; “Quality Assurance shall investigate and assess the validity of the previous measuring results. An evaluation shall be performed of any adverse impact on product quality and affected product shall be dispositioned appropriately</i>
7.6 – 10	Where are the records documenting the results of calibration and verification maintained?	C	In support of this assessment, the following calibration records were reviewed: • MS6300-M Empex Sealer (Pathway/Transcat) – Due 04/11/2024; • LAF5769 Scale (Past Due Cal) – Due 10/03/2023; • ST-006 Platform Scale (Past Due Cal) – Due 09/07/2023; • ST-007 Platform Scale (Past Due Cal) – Due 09/09/2023; and • FF-0.2 Firefly Science Filter – Due 10/30/2025. Concern: Tag/Out log out policy to prevent use of uncalibrated equipment needs to be enforced. <i>Note: there is no requirement to calibrate equipment within its stated cycle; however, if a decision is made not to calibrate equipment due to lack of need, the equipment shall be adequately identified as to its status as to prevent use.</i>
7.6 – 11	Does the organization employ computer software for monitoring and measurement	N/A	Not Applicable. SILQ does not employ software in support of the metrology function.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
7.6 – 12	If the organization employs computer software for monitoring and measuring, how is the validity of the software confirmed?	N/A	Not Applicable. SILQ does not employ software in support of the metrology function.
7.6 – 13	Does the organization employ and external metrology laboratory for calibration services?	Yes	Micro Precision (California) is the primary provider of metrology support.
7.6 – 14	If an external metrology laboratory is employed for metrology services, is the facility EN ISO 17025 certified?	Yes	Micro Precision is an ISO/IEC 17025:2017 accredited facility.
8 Measurement, Analysis and Improvement			
8.1 General			
8.1 – 1	How does the organization plan and implement measurement, analysis, and improvement processes needed to demonstrate conformity of the product?	Yes	In accordance with 13.7 (Analysis of Data), the following metrics are being compiled and reviewed as part of management review, for example: • Customer feedback; • Conformity to product requirements; • Characteristics and trends of processes and products; • Internal audit results; • Supplier quality performance; and • CAPA. Note: QM.SLQ011.A – Statistical Techniques Work Instruction
8.1 – 2	How does the organization plan and implement measurement, analysis, and improvement processes needed to ensure conformity of the quality management system?	Yes	SILQ assesses metrics compiled from multiple areas. The data collected is used to drive continuous improvement.
8.1 – 3	How does the organization plan and implement measurement, analysis and improvement processes needed to continuously improve and maintain the effectiveness of the quality management system?	Yes	Tools used in support of measurement, analysis, and improvement are be internal audits; complaints; management review and CAPA.
8.1 – 4	How does the organization determine what monitoring measurement, and analysis methods to use, including statistical techniques?	Yes	The employment of statistical modalities by SILQ is delineated within QM.SLQ011.A – Statistical Techniques Work Instruction.
8.2 Monitoring and Measurement			
8.2.1 Feedback			
8.2 – 1	How does the organization obtain information about whether customer requirements have been achieved?	Yes	Post-market data collected, in support of post-market surveillance is delineated within Table 1.0 (Examples of PMS Data) of QM.SLQ033.A – Post-Market Surveillance.
8.2 – 2	Has the organization established a documented procedure for a feedback system to provide early warning of quality problems and for input into the corrective and preventive action processes?	Yes	Feedback activities are managed in accordance with 6.3 of QM.SLQ033.A – Post-Market Surveillance; <i>“feedback information is analyzed and if necessary corrective and / or preventive actions are initiated per the CAPA procedure QM.SLQ016.”</i>

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.2 – 3	If national or regional regulations require the organization to gain experience from the post-production phase, does the review of this experience form part of the feedback system?	Yes	Post-market surveillance activities are pursued in accordance with QM.SLQ033.A. In accordance with 6.1 of QM.SLQ033.A; “a <i>Post-Market Surveillance (PMS) Plan</i> for each device shall be established. The requirements for the PMS Plan should be directly proportional to the risk associated with the device based on its intended use.”
8.2.2 Complaint Handling			
8.2 – 4	Has the organization established a procedure(s) for the timely handling of complaints?	Yes	Complaints received by SILQ are managed in accordance with QM.SLQ021.C – Product Complaint System
8.2 – 5	As a minimum, the complaint procedure shall include: • Receiving and recording of complaint information; • Evaluation of information to ensure if it constitutes a complaint; • Investigation of complaints; • Determination if a complaint is reportable; • Handling of complaint-related product; and • The need to initiate CAPA as appropriate.	Yes	As scripted, QM.SLQ021, Rev C meets the requirements delineated within 21 CFR, Part 820.198 and ISO 13485:2016, Clause 8.2.2.
8.2 – 6	Has the organization established documented procedures for the issuance and implementation of advisory notices?	Yes	Advisory notices are managed in accordance with QM.SLQ030.A – Advisory Notices and Recalls.
8.2 – 7	Does the organization maintain records of all customer complaint investigations?	R	In support of this assessment, the following compliant files were reviewed: • PCF 2023001 Opened 29 Mar 2023 / Closed 03 Apr 2023; • PCF 2023002 Opened 15 May 2023 / Closed 16 May 2023; and • PCF 2023003 Opened 15 Nov 2023 / Closed 16 Nov 2023. Recommendation: Regulators expect to see some level of clinical oversight for decisions made regarding adverse events and potential patient injury. The current practice currently enforced by FDA is to ensure an individual with a clinical background also review and approve decisions pertaining to the need to file an adverse event (eMDR) with FDA or pursue similar actions with other regulatory bodies.
8.2 – 8	If an investigation determines that the activities outside the organization contributed to the customer complaint, is relevant information exchanged between the organizations involved?	Yes	Information associated with a complaint investigation will be shared with reporting parties, if it has been requested.
8.2 – 9	If any customer complaint is not followed by corrective and/or preventive action, is the reason and authorization documented?	Yes	SILQ investigates all product complaints received impacting SILQ products.
8.2 – 10	Has the organization issued an advisory notice or processed a market withdraw in FY 2023	N/A	Not Applicable. SILQ has never issued an advisory.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
	8.2.3 Reporting to Regulatory Authorities		
8.2 – 11	If deemed to be appropriate, is the complaint reported to applicable regulatory bodies; i.e., MDRs, vigilance reports, etc.?	Yes	There are four (4) procedures employed in support of reporting to regulatory authorities: • QM.SLQ022.A – Medical Device Reporting; • QM.SLQ023.A – eMDR Submission Work Instruction; • QM.SLQ030.A – Advisory Notices and Recalls; and • QM.SLQ033.A – Post-Market Surveillance.
8.2 – 12	If national or regional regulations require notification of adverse events that meet specified reporting criteria, does the organization have documented procedures to support notification to appropriate regulatory authorities?	Yes	As of this assessment, the two procedures scripted are US FDA centric: • QM.SLQ022.A – Medical Device Reporting; and • QM.SLQ023.A – eMDR Submission Work Instruction.
	8.2.4 Internal Audit		
8.2 – 13	Does the organization have a documented procedure for internal audits?	Yes	The SILQ Internal Audit Program is documented within QM.SLQ017.A – Internal Audits
8.2 – 14	Do audits determine whether the QMS conforms to the requirements of EN ISO 13485 and to the other requirements established by the organization?	Yes	The SILQ Internal Audit Program is used to assess all aspects of the QMS.
8.2 – 15	Do audits determine whether the QMS is effectively implemented and maintained?	Yes	The audit program will gage the level of QMS with applicable quality, regulatory, and statutory requirements.
8.2 – 16	Are audits currently being performed (on-time) in accordance with the published audit schedule?	Yes	Note: This assessment was the second audit of the SILQ QMS by Dr. Devine.
8.2 – 17	Where are the audit criteria, scope, frequency, and methods defined?	Yes	In accordance with 2.1 of QM.SLQ017.A; <i>“internal audits per this procedure are limited to those elements of the organization subject to the Quality Management System. Qualified external auditors may be utilized to perform part of, or all, functions of this procedure.”</i>
8.2 – 18	Can the organization demonstrate that selection of auditors and the conduct of audits are objective and impartial, and that auditors don’t audit their own work?	Yes	Dr. Devine is a trained and qualified auditor with 42-years of industry experience. Note: DGII auditors cannot assess the adequacy of the audits that they perform. However, DGII can attest to the fact that audits are being performed in accordance with ISO 19011 requirements and in compliance with ISO 13485:2016, 21 CFR, Part 820, and applicable quality, regulatory, and statutory requirements. It is the responsibility of the client to ascertain if this audit meets their requirements.
8.2 – 19	Who in the organization ensures that actions are taken to eliminate detected nonconformities and their causes?	Yes	In accordance with 10.5 of QM.SLQ017.A; CAPAs will be opened and assigned to department managers for mitigation.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.2 – 20	What activities are performed to verify the actions taken and the subsequent reporting of actions taken?	Yes	In accordance with 10.2 of QM.SLQ017.A; "all observations from the audit shall be communicated/discussed to ensure correct interpretation and to allow any open issues to be resolved prior to issuing the final audit report, FM3-QM.SLQ017."
8.2.5 Monitoring and Measurement of Processes			
8.2 – 21	What methods are used to monitor and measure the QMS processes?	Yes	Tools employed in support of the monitoring and measurement of the QMS include: • Management reviews; • Complaints; • Internal audits; • CAPAs; and • Quality Objectives.
8.2 – 22	When the desired results are not achieved, what actions are taken to ensure that the product meets requirements?	Yes	When the QMS fails to meet performance expectations, CAPA will be pursued to mitigate QMS performance issues.
8.2.6 Monitoring and Measurement of Product			
8.2 – 23	What characteristics are checked to verify that product requirements have been met?	Yes	The CMO (Pathway) performs a final inspection premised on requirements delineated within the DMR.
8.2 – 24	At what stages of the product realization process do monitoring and measuring activities take place?	Yes	The performance of monitoring and measuring occurs throughout product realization.
8.2 – 25	How is evidence of conformity with acceptance criteria maintained?	Yes	Evidence of product conformity with the DMR is maintained through the DHR.
8.2 – 26	Do records indicate the person(s) authorizing release of product?	Yes	SILQ is tasked with the formal release of finished device. The distribution and shipment of finished devices is performed by the CMO.
8.2 – 27	How does the organization ensure that product is not released until the all requirements have been met?	Yes	SILQ finished devices, that successfully complete final acceptance, are permitted to be released into commercialization.
8.3 Control of Nonconforming Product			
8.3.1 General			
8.3 – 1	Is non-conforming product being appropriately identified and controlled?	Yes	In support of this assessment, the following Pathway NCRs were reviewed: • NCR 22-0009 Opened 11/01/2022 – Sort; • NCR 22-0010 Opened 11/01/2022 – Sort; • NCR 22-0011 Opened 11/01/2022 – Scrap; • NCR 22-0013 Opened 11/01/2022 – UAI; • NCR 22-0016 Opened 11/01/2022 – UAI; and • NCR 22-0017 Opened 11/01/2022 – UAI.
8.3 – 2	Has the organization established a procedure for the control of non-conforming product?	Yes	Non-conforming products are managed in accordance with QM.SLQ040.A – Nonconforming Materials

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.3 – 3	Are records associated with non-conforming product being maintained?	Yes	In accordance with 11.2 of QM.SLQ040.A; <i>“a copy of the closed NCMR is to be included, or referenced, in the appropriate files and batch records / DHR’s as required by each specific disposition.”</i> In support of this assessment, the following NCR’s originating from Pathway were reviewed: • NCR 22-0009 Opened 11/01/2022 – Sort; • NCR 22-0010 Opened 11/01/2022 – Sort; • NCR 22-0011 Opened 11/01/2022 – Scrap; • NCR 22-0013 Opened 11/01/2022 – UAI; • NCR 22-0016 Opened 11/01/2022 – UAI; and • NCR 22-0017 Opened 11/01/2022 – UAI.
8.3.2 Actions in Response to Nonconforming Product Detected Before Delivery			
8.3 – 4	Does the organization deal with non-conformances in one or more of the following ways: • Taking action to eliminate the detected nonconformity; • Taking action to preclude its original intended use or application; and/or • Authorizing its use/release under concession?	Yes	Nonconforming product identified, prior to shipment, will be reworked. If a concession request is made by the CMO, the concession will require review and approval of SILQ
8.3.3 Actions in Response to Nonconforming Product Detected After Delivery			
8.3 – 5	What actions are pursued when a nonconformance is detected after shipment?	Yes	Non-conformances detected after shipment will be recalled in accordance with QM.SLQ030.A – Advisory Notices and Recalls.
8.3 – 6	Does the organization have an established procedure for the issuance of advisory notices?	Yes	The issuance of advisory notices is managed in accordance with QM.SLQ030.A – Advisory Notices and Recalls.
8.3.4 Rework			
8.3 – 7	Are rework activities being pursued in accordance with documented procedures and instructions?	Yes	In accordance with 10.6.4.1 of QM.SLQ040.A; <i>“rework and re-test and/or re-inspection of product, parts, or materials are to be performed in accordance with instructions provided in NCMR or established manufacturing procedures or to an approved and released Temporary Deviation.”</i>
8.3 – 8	Are all rework products inspected upon the completion of rework?	Yes	All rework is required to be cycled back through the inspection process. However rework is not typical.
8.4 Analysis of Data			

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.4 – 1	Has the organization established documented procedures to determine, collect, and analyze appropriate data to demonstrate the suitability and effectiveness of the quality management system and to evaluate where continual improvement of the effectiveness of the quality management system can be made?	Yes	The quality manual contains sufficient granularity documenting the data collected from key functional areas and trended to drive continuous improvement. <i>Note: Typically this type of information is delineated in a stand-alone SOP due to its relevance to an organization and the need to trend data for determining continuous improvement activities.</i>
8.4 – 2	What data is collected and analyzed to demonstrate the suitability and effectiveness of the QMS and to evaluate where continual improvement of its effectiveness can be made?	Yes	Premised on the data delineated through the management review process; the following data was reviewed during management review: • Audits; • Regulator Inspections (FDA); • CAPAs (zero in 2022) • Complaints; • NCRs • Training • Post-Market Surveillance & Customer Feedback; • Sampling Plans (statistics); • Supplier Performance; • QMS Changes; • Recommendations for Improvement; • New and/or Revised Regulatory Requirements; • Risk Management; • Design Control; and • Resources.
8.4 – 3	What information does this analysis provide relating to feedback?	Yes	Feedback received from customers is garnered from the complaint system. Information compiled from complaints will be used to drive product improvement(s).
8.4 – 4	What information does this analysis provide relating to conformity to product requirements?	Yes	Production yield information collected from the CMO is used to ascertain compliance levels versus the DMR.
8.4 – 5	What information does this analysis provide relating to suppliers?	Yes	Supplier performance is measured in accordance with their ability to meet SILQ procurement requirements.
8.4 – 6	What information does this analysis provide relating to characteristics and trends of processes and products?	Yes	Collectively, all of the data collected and analyzed is used to determine SILQ's ability meet customer requirements.
8.4 – 7	Where are the records documenting the results of the analysis of data maintained?	Yes	The results of data trended and the subsequent analysis is retained as part of the management review.
8.5 Improvement			
8.5.1 General			
8.5 – 1	Can the organization demonstrate that the effectiveness of the QMS continuously improves?	Yes	SILQ has revised multiple SOPs since the 2022 audit, which reflects that the QMS is improving.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.5 – 2	Can the organization demonstrate that it identifies and implements changes to ensure continued QMS effectiveness?	Yes	SILQ has revised multiple SOPs since the 2022 audit, which reflects that the QMS is improving.
8.5 – 3	What tools does the organization employ in pursuit of improvement activities?	Yes	Tools that will be used to drive continuous improvement of the QMS will be: (a) training; (b) management review; (c) audits (internal & supplier); (d) complaints and customer feedback; and (e) CAPA.
	8.5.2 Corrective Action		
8.5 – 4	Do corrective actions records identify and address root cause(s)?	N/A	Not Applicable. No corrective actions were pursued in 2023.
8.5 – 5	Are actions taken appropriate to the severity of the problem?	N/A	Not Applicable. No corrective actions were pursued in 2023.
8.5 – 6	Does the organization have a documented procedure defining requirements for each of the following activities? - Reviewing nonconformities (including customer complaints) - Determining the causes of nonconformities - Evaluating the need for action to ensure that nonconformities do not recur - Determining and implementing action needed - Records of the results of any investigation and of action taken - Reviewing the corrective action taken and its effectiveness	Yes	QM.SLQ016.B – Corrective and Preventive Action
8.5 – 7	Does the organization maintain records for investigation and corrective actions taken?	N/A	Not Applicable. No corrective actions were pursued in 2023.
	8.5.3 Preventive Action		
8.5 – 8	How does the organization determine potential nonconformities to take action on?	Yes	Preventive action activities will be identified premised on their ability to prevent nonconformities from occurring.
8.5 – 9	Do preventive action records identify and address root cause(s)?	N/A	Not Applicable. No preventive actions were pursued in 2023.
8.5 – 10	Are preventive actions taken appropriate to the severity of the problem?	N/A	Not Applicable. No preventive actions were pursued in 2023.

Q#	ISO 13485:2016 Clause	Compliant Yes/No (R/G)	Supporting Evidence
8.5 – 11	Does the organization have a documented procedure defining requirements for each of the following activities? - determining potential nonconformities and their causes - evaluating the need for action to prevent occurrence of nonconformities - determining and implementing action needed - (records) recording of the results of any investigation and of action taken - reviewing preventive action taken and its effectiveness	Yes	QM.SLQ016.B – Corrective and Preventive Action
8.5 – 12	Where does the organization maintain records of preventive action taken?	N/A	Not Applicable. No preventive actions were pursued in 2023.

Respectfully Submitted,



Date: 17 November 2023

Christopher J. Devine, Ph.D.
President
Devine Guidance International, Inc.