

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Internal Quality Audit Checklist

Audit #: <u>22-01</u>		
Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
	General Requirements - ISO 13485:2016 Section 4.1; 21 CFR 820.5; MDD 93/42/EEC Annex II (1, 3.2, 6.1)	
1.	Review previous audit results and audit action effectiveness for this section.	N/A
2.	Using a risk-based approach, has the Organization established, documented, implemented and maintained a QMS and maintained its effectiveness in accordance with ISO 13485 and applicable regulatory requirements? (Questions in section 4.1 are verified throughout the audit)	Did not evaluate. Evaluate in the next supplier audit.
3.	Where has the Organization identified the processes needed for the QMS or regulatory requirements and their application throughout the organization?	Did not evaluate. Evaluate in the next supplier audit.
4.	Has the organization identified the criteria and methods needed to ensure the operation and control of these processes are effective?	Did not evaluate. Evaluate in the next supplier audit.
5.	How has the Organization implemented actions necessary to achieve planned results and maintain the effectiveness of processes needed for the QMS or regulatory requirements?	Did not evaluate. Evaluate in the next supplier audit.
6.	Are processes needed for the QMS managed by the Organization in accordance with the requirements of ISO 13485 and/or regulatory requirements?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: 22-01

Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
7.	How does the Organization ensure the availability of resources and information necessary to support the operation and monitoring of these processes?	Did not evaluate. Evaluate in the next supplier audit.
8.	How does the organization monitor, measure and analyze the identified processes?	Did not evaluate. Evaluate in the next supplier audit.
	Documentation Requirements – ISO 13485:2016 Section 4.2; 21 CFR 820.40; MDD 93/42/EEC Annex II	
	General – ISO 13485:2016 Section 4.2.1; 21 CFR 820.40, 820.181; MDD 93/42/EEC Annex II (3.1, 3.2, 4, 6.1)	
9.	Does the Organization have documented statements of a quality policy and quality objectives? (See 5.3, 5.4.1)	Did not evaluate. Evaluate in the next supplier audit.
10.	Does the Organization have a quality manual?	Did not evaluate. Evaluate in the next supplier audit.
11.	Are adequate documents, including records, in place to ensure the effective planning, operation and control of the Organization's processes?	Did not evaluate. Evaluate in the next supplier audit.
12.	Does documentation include the records required by 13485:2016? (See 4.2.5)	Did not evaluate. Evaluate in the next supplier audit.
13.	Are there provisions to include records for other documentation specified by applicable regulatory requirements?	Did not evaluate. Evaluate in the next supplier audit.
	General – ISO 13485:2016 Section 4.2.3; 21 CFR 820.40, 820.181; MDD 93/42/EEC Annex II (3.2, 6.1)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
14.	Does the Organization maintain a file / Device Master Record (DMR) for each type or model of device containing or identifying documents with product specifications, QMS requirements, and regulatory requirements (See 4.2.3)?	Did not evaluate. Evaluate in the next supplier audit.
15.	Review a file / DMR for one or more of the Organization's devices to verify it contains or references the following: • Device specifications (i.e., drawings, component specifications, source control drawings, etc.)? • Production process specifications (i.e., process work inspection, identification of equip., production methods, etc.)? • Quality control test procedures (i.e., identification of test methods, test equipment, acceptance criteria, etc.)? • Manufacturing, packaging, labeling and distribution specifications (i.e., labeling specification / drawings, identification of methods, materials, and processes)? • Equipment maintenance procedures? • Procedures for measuring and monitoring? • As appropriate, requirements for installation and servicing?	Did not evaluate. Evaluate in the next supplier audit.



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
16.	Is each File / DMR prepared and approved in accordance with FDA and ISO 13485 requirements?	Did not evaluate. Evaluate in the next supplier audit.
	Quality Manual – ISO 13485:2016 Section 4.2.2; 21 CFR 820.186; MDD 93/42/EEC Annex II (3.2)	
17.	Does the quality manual define the scope of the QMS, including details of and justification for exclusions and/or requirements that don't apply?	Did not evaluate. Evaluate in the next supplier audit.
18.	Does the quality manual contain or reference the documented procedures established and include a description of the interaction between the processes for the QMS?	Did not evaluate. Evaluate in the next supplier audit.
19.	Does the quality manual outline the documentation structure of the QMS?	Did not evaluate. Evaluate in the next supplier audit.
	Control of Documents – ISO 13485:2016 Section 4.2.4; 21 CFR 820.40; MDD 93/42/EEC Annex II (3.2, 6.1)	
20.	Is there a documented procedure that defines the controls for all documents?	Did not evaluate. Evaluate in the next supplier audit.
21.	Does the above procedure include controls for documents of external origin, such as standards, customer drawings and data related to quality system requirements? Does it ensure that these documents are identified and their distribution controlled?	Did not evaluate. Evaluate in the next supplier audit.
22.	Are controlled documents reviewed for adequacy and approved prior to use?	Did not evaluate. Evaluate in the next supplier audit.
23.	Are controls in place to review and update as necessary and re-approve documents?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: 22-01

Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
24.	Have authorized functional department and associated personnel been identified to review and approve changes?	Did not evaluate. Evaluate in the next supplier audit.
25.	Are records of changes maintained? Do they include description of change, affected documents, authorized approval signature(s) and date and effective date?	Did not evaluate. Evaluate in the next supplier audit.
26.	Do controls ensure that changes and the current revision status of documents are identified?	Did not evaluate. Evaluate in the next supplier audit.
27.	Are controls in place to ensure only the most current revision of applicable documents are available at points of use?	Did not evaluate. Evaluate in the next supplier audit.
28.	Are controls in place to remove obsolete documents from all points of use or otherwise controlled to prevent unintended use? Are obsolete documents readily available in the event they needed for any purpose?	Did not evaluate. Evaluate in the next supplier audit.
29.	Is a system to retain obsolete documents for specified retention times (in accordance with applicable regulatory requirements) in place?	Did not evaluate. Evaluate in the next supplier audit.
30.	Is there a master list or equivalent document control procedure identifying the current revision of documents?	Did not evaluate. Evaluate in the next supplier audit.
31.	Are document changes communicated to affected personnel in a timely manner?	Did not evaluate. Evaluate in the next supplier audit.
32.	Are controls in place to ensure that documents remain legible and readily identifiable?	Did not evaluate. Evaluate in the next supplier audit.



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
33.	Are controls in place to prevent documents from deterioration and loss? Including electronic files, where applicable?	Did not evaluate. Evaluate in the next supplier audit.
	Control of Records – ISO 13485:2016 Section 4.2.5; 21 CFR 820.180, 820.184; MDD 93/42/EEC Annex II (3.2, 6.1)	
34.	Does the Organization have a documented procedure defining the controls needed for the identification, storage, security, integrity and protection, retrieval, retention time and disposition of records?	Yes. Quality Records QP-003 Rev F
35.	Has the organization established documented procedures to define methods for protecting confidential health information contained in records in accordance with applicable regulatory requirements?	Yes. Established in Quality Records QP-003 Rev F
36.	Does the above procedure define the following with respect to quality records: • Identification • Preparation • Location • Retention periods • Maintenance • Disposition • Storage requirements	Yes



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A


DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
37.	What records exist that provide evidence of conformity to requirements and of the effective operation of the QMS? (Should be reviewed throughout the audit)	In compliance based on review of records throughout the audit.
38.	Are relevant subcontractor records considered to be quality data?	Yes. Example: calibration reports by subcontractor/suppliers.
39.	Does a Device History Record (DHR) exist for each work order which demonstrates the device was manufactured in accordance with the DMR and applicable regulatory requirements?	Yes. DHR for Lot #'s SLQ1610-012022 and SLQ1610-012122
40.	Does the DHR contain (at a minimum) or reference the location of the following information: • Work order number used? • Date(s) of manufacture? • Initials or signature of person(s) responsible? • Quantity manufactured? • Quantity released for distribution? • Acceptance records (which demonstrate the device was manufactured in accordance with the DMR)? • Primary product identification label and labeling used?	Yes. DHR for Lot #'s SLQ1610-012022 and SLQ1610-012122
41.	Is a Quality System Record (QSR) maintained and does it include or refer to the location of all the procedures and the documentation of activities required but are not specific to a particular device?	Yes. Arena PLM

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Quality Management System (ISO 13485:2016 Section 4; 21 CFR 820 Subpart A, B, D, M; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
42.	Are records legible, readily identifiable, retrievable and stored in a manner to prevent deterioration or damage? (Should be reviewed throughout the audit)	Yes. Based on records reviewed during this audit.
43.	Where applicable, are quality records made available for evaluation by the customer or the customer's representative for an agreed period?	Not applicable
	Records – General Requirements – No ISO Equivalent; 21 CFR 820.180; No MDD Equivalent	
44.	Are confidential records marked as such?	Yes. QP-003 Rev F Section 6.6
45.	Does the organization have a policy that defines which QMS records are not to be shared with an FDA investigator?	No policy specified.
Additional Observations / Comments: 		
Auditor Signature / Date:  19 Oct 2022		



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Management Responsibility (ISO 13485:2016 Section 5; 21 CFR 820 Subpart B; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
	Management Commitment – ISO 13485:2016 Section 5.1; 21 CFR 820.20; MDD 93/42/EEC Annex II (2, 3.1)	
1.	Review previous audit results and audit action effectiveness for this section.	N/A
2.	How does top management communicate the importance of meeting customer, applicable regulatory requirements and legal requirements to the organization?	Did not evaluate. Evaluate in the next supplier audit.
3.	Has a company quality policy been established? (See 5.3)	Did not evaluate. Evaluate in the next supplier audit.
4.	What are the quality objectives established by top management? (See 5.4.1)	Did not evaluate. Evaluate in the next supplier audit.
5.	Does top management conduct management reviews? (See 5.6)	Yes, per QP-01 Rev G
6.	How does top management ensure the availability of resources to support and continually improve the QMS?	Did not evaluate. Evaluate in the next supplier audit.
	Customer Focus – ISO 13485:2016 Section 5.2; No QSR Equivalent; No MDD Equivalent	
7.	How does top management ensure that customer requirements are determined and met?	Did not evaluate. Evaluate in the next supplier audit.
	Quality Policy – ISO 13485:2016 Section 5.3; 21 CFR 820.20; MDD 93/42/EEC Annex II (3.2)	
8.	Is the quality policy is appropriate to the goals of the Organization and needs of customers?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Management Responsibility (ISO 13485:2016 Section 5; 21 CFR 820 Subpart B; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
9.	Does the quality policy include a commitment to comply with requirements and to maintain QMS effectiveness?	Did not evaluate. Evaluate in the next supplier audit.
10.	Does quality policy provide a framework for the quality objectives?	Did not evaluate. Evaluate in the next supplier audit.
11.	Is the quality policy communicated and understood within the organization?	Did not evaluate. Evaluate in the next supplier audit.
12.	Is there an established process to review the quality policy for continuing suitability?	Did not evaluate. Evaluate in the next supplier audit.
	Planning – ISO 13485:2016 Section 5.4; No QSR Equivalent; MDD 93/42/EEC Annex II	
	Quality Objectives – ISO 13485:2016 Section 5.4.1; No QSR Equivalent; MDD 93/42/EEC Annex II (3.2)	
13.	Has top management established quality objectives (including those needed to meet applicable regulatory and product requirements) at relevant functions and levels within the organization?	Did not evaluate. Evaluate in the next supplier audit.
14.	Are the quality objectives measurable and consistent with the quality policy?	Did not evaluate. Evaluate in the next supplier audit.
15.	What are the objectives and how is progress tracked?	Did not evaluate. Evaluate in the next supplier audit.
	Quality Management System Planning – ISO 13485:2016 Section 5.4.2; No QSR Equivalent; MDD 93/42/EEC Annex II (3.2)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Management Responsibility (ISO 13485:2016 Section 5; 21 CFR 820 Subpart B; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
16.	How does the Organization ensure that the planning of the QMS is carried out in order to meet the requirements given in ISO 13485:2016, as well as the quality objectives?	Did not evaluate. Evaluate in the next supplier audit.
17.	How does the Organization ensure that the integrity of the QMS is maintained when changes to the QMS are planned and implemented?	Did not evaluate. Evaluate in the next supplier audit.
	Responsibility, Authority and Communication – ISO 13485:2016 Section 5.5; 21 CFR 820.20; MDD 93/42/EEC Annex II	
	Responsibility and Authority – ISO 13485:2016 Section 5.5.1; 21 CFR 820.20; MDD 93/42/EEC Annex II (3.2)	
18.	Does the Organization have an organizational chart which describes the interrelation of personnel?	Yes. FM-0053 Rev K
19.	Are personnel who manage, perform and verify work affecting quality have the independence and authority to perform these tasks?	Yes
20.	How is the organizational chart controlled?	Yes. FM-0053 Rev K
21.	Does the Organization have job descriptions which describe the functions of the position, responsibilities and authorities?	Yes. Job descriptions for R&D Engineer (JD-003 Rev C), Production Assembler (JD-006 Rev C), Production Manager (JD-007 Rev C), and Sr. Quality Engineer (JD-016 Rev C).
	Management Representative – ISO 13485:2016 Section 5.5.2; 21 CFR 820.20; No MDD Equivalent	
22.	Has the Organization appointed a management representative? Is this appointment documented?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Management Responsibility (ISO 13485:2016 Section 5; 21 CFR 820 Subpart B; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
23.	Does the management representative have responsibility and authority to: a) ensure that processes needed for the QMS are effectively established, implemented and maintained? b) report to top management on the performance of the QMS and any need for improvement? c) ensure the promotion of awareness of applicable regulatory and customer requirements and quality management system requirements throughout the organization? d) ensure the quality policy and quality objectives are understood, implemented and maintained at all levels?	Did not evaluate. Evaluate in the next supplier audit.
	Internal Communication – ISO 13485:2016 Section 5.5.3; No QSR Equivalent; No MDD Equivalent	
24.	How is information regarding the effectiveness of the QMS communicated within the organization?	Did not evaluate. Evaluate in the next supplier audit.
25.	Is the communication documented?	Did not evaluate. Evaluate in the next supplier audit.
	Management Review – ISO 13485:2016 Section 5.6; 21 CFR 820.20; MDD 93/42/EEC Annex II	
	General – ISO 13485:2016 Section 5.6.1; 21 CFR 820.20; MDD 93/42/EEC Annex II (3.2)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Management Responsibility (ISO 13485:2016 Section 5; 21 CFR 820 Subpart B; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
26.	Does the Organization have a documented procedure requiring top management to review the quality system at defined intervals and sufficient frequency?	Yes, Management Review SOP QP-001 Rev G
27.	What is the frequency that top management performs management reviews?	At least every 12 months
28.	Are management reviews conducted to assess the suitability, adequacy and effectiveness of the quality system with regards to requirements of applicable external regulations and internal quality procedures?	Yes. Agenda of Management review meeting May 6, 2022
29.	Do the reviews include assessing opportunities for improvement and the need for changes to the quality management system including quality policy and quality objectives?	Yes. Agenda of Management review meeting May 6, 2022
30.	Are records of management reviews maintained?	Yes
	Review Input – ISO 13485:2016 Section 5.6.2; No QSR Equivalent; MDD 93/42/EEC Annex II (3.2)	



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Management Responsibility (ISO 13485:2016 Section 5; 21 CFR 820 Subpart B; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
31.	Verify that <u>each</u> of the following were included in review(s), at a minimum: a) feedback b) complaint handling c) reporting to regulatory authorities d) results of audits, e) monitoring and measurement processes f) monitoring and measurement of product g) status of preventive and corrective actions, h) follow-up actions from previous management reviews, i) changes that could affect the quality management system, j) recommendations for improvement, and k) new or revised regulatory requirements.	Yes. Agenda of Management review meeting May 6, 2022
	Review Output - ISO 13485:2016 Section 5.6.3; No QSR Equivalent; MDD 93/42/EEC Annex II (3.2)	



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Management Responsibility
(ISO 13485:2016 Section 5; 21 CFR 820 Subpart B; MDD 93/42/EEC Annex II)

#	Audit Question / Item	Evidence / Comments
32.	Are records from management reviews maintained regarding any decisions and actions related to; and do they include, at a minimum: a) the dates of the review, b) attendees, c) improvement needed to maintain the suitability, adequacy and effectiveness of the quality management system and its processes, d) improvement of product related to customer requirements, and e) Changes needed to respond to revised regulatory requirements f) resource needs.	Not evaluated. Only the agenda of the management review was shared with the auditor.

Additional Observations / Comments:

Auditor Signature / Date:

19 Oct 2022

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Resource Management (ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
	Provision of Resources – ISO 13485:2016 Section 6.1; 21 CFR 820.20; MDD 93/42/EEC Annex II (3.2)	
1.	Review previous audit results and audit action effectiveness for this section.	N/A
2.	What resources has the Organization provided to implement and maintain the QMS effectiveness?	Pathway has retained sufficient resources to maintain an effective QMS. Reviewed organization chart FM-0053 Rev K
3.	What resources has the Organization provided to ensure that customer and regulatory requirements are met? (See 6.2, 6.3, 6.4)	Organization chart FM-0053 Rev K
	Human Resources – ISO 13485:2016 Section 6.2; 21 CFR 820.25; MDD 93/42/EEC Annex II (5.2)	
4.	How are the education, training, skills and experience required for a job/task defined and documented?	Education, training, skills, and experience required for a job is documented in the job description. Reviewed job descriptions for R&D Engineer (JD-003 Rev C), Production Assembler (JD-006 Rev C), Production Manager (JD-007 Rev C), and Sr. Quality Engineer (JD-016 Rev C).
5.	How does the organization document the processes for establishing competence, provide needed training and ensure awareness of personnel?	Documented in Training SOP QP-017 Rev. E



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Resource Management
(ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)

#	Audit Question / Item	Evidence / Comments
6.	Review sampling of records. How does person meet identified qualifications?	Reviewed CV's Will Hail and Larry Dacoron. Personnel met qualifications outlined in their respective job descriptions. Finding #1: No resumes/CV's on file for production assemblers Amalia Fuller and Eva Tamayo. Unable to determine their qualifications for their respective job descriptions.
7.	Does the Organization have a documented procedure for ensuring all personnel are trained to adequately perform their assigned responsibilities?	Documented in Training SOP QP-017 Rev. E
8.	Have training requirements been determined for the general employee population, as well as, for various job functions? Including those personnel performing work that affect product quality?	Yes, training requirements outlined in Training Matrix FM-0003 Rev R. a) Training record for Sean Tamayo (Production Manager) was reviewed (MP-SLQ-0001 Rev A). Finding per Section 6.1 of QP-017 Rev E: The training record for MP-SLQ-0001 Rev A was signed on 2/18/2022 after work was performed on 2/14/22 on Lots SLQ1610-012022 and SLQ1610-012122. b) Training records for Amalia Fuller (Production Assembler) were reviewed (QM-01 Rev E trained on 6/25/21, QP-003 Rev F trained on 6/25/21, QP-008 Rev C trained on 6/25/21, and MP-SLQ-0001 Rev A trained on 2/18/22). Finding per Section 6.1 of QP-017 Rev E: The training record for MP-SLQ-0001 Rev A was signed on 2/18/22 after work was performed on 1/20/22 and 1/21/22 on Lots SLQ1610-012022 and SLQ1610-012122,

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Resource Management (ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
		respectively. Section 6.1 of QP-017 Rev E c) Training records for Eva Tamayo (Production Assembler) were reviewed (QM-01 Rev E trained on 5/20/22, QP-003 Rev F trained on 5/20/22, QP-008 Rev C trained on 5/20/22, and MP-SLQ-0001 Rev A trained on 2/18/22). Finding per Section 6.1 of QP-017 Rev E: The training record for MP-SLQ-0001 Rev A was signed on 2/18/22 after work was performed on 1/20/22 and 1/21/22 on Lots SLQ1610-012022 and SLQ1610-012122, respectively. Section 6.1 of QP-017 Rev E d) Training records for Will Hail (R&D Engineer) were reviewed (QM-01 Rev E trained on 6/10/20, QP-003 Rev F trained on 6/10/20, QP-030 Rev D trained on 6/10/20, QP-012 Rev C trained on 6/10/20, and MP-SLQ-0002 Rev A trained on 5/27/22). Finding per Section 6.1 of QP-017 Rev E: The training record for MP-SLQ-0002 Rev A was signed on 5/27/22 after work was performed on 1/20/22 and 1/21/22 on Lots SLQ1610-012022 and SLQ1610-012122, respectively. Section 6.1 of QP-017 Rev E e) Training records for Larry Dacoron (Quality Engineer) were reviewed (QP-005 Rev G trained on 3/8/21, QP-008 Rev C trained on 3/8/21, QP-009 Rev D trained on 3/8/21, and QP-010 Rev F trained on 3/8/21). Observation #2: As reviewer of DHR's, recommend



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Resource Management
(ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)

#	Audit Question / Item	Evidence / Comments
		to train to procedures being utilized in the process (MP-SLQ-0001, MP-SLQ-0002, TM-HDX-0001, and TM-HDX-0002). Per 21 CFR 820.25 (b). Finding #2: Personnel performing work on product Lot #'s SLQ1610-012022 and SLQ1610-012122 per applicable procedures MP-SLQ-0001 Rev A and MP-SLQ-0002 Rev A were trained after the dates the work was performed. See MP-SLQ-0001 Rev A training records for Sean Tamayo, Amalia Fuller, and Eva Tamayo. See MP-SLQ-0002 Rev A training record for Will Hail
9.	Is there a provision for training temporary and new employees?	Yes, temporary employees are included in the definition of trainee. Provision for new employee training. QP-017 Rev E
10.	Does re-training occur after changes to job functions and/or procedure revisions?	Retraining occurs annually (Section 8.3.4), when significant changes to applicable standards/regulations (Section 8.4.2.1), or as issues occur against the QMS (e.g., customer complaint, corrective action, nonconformance) (Section 8.4.2.2).
11.	Does the training include awareness of device defects resulting from improper job performance?	Yes, Section 8.3.2 of QP-017 Rev E.
12.	Are there provisions for all personnel who are required to work under special environmental conditions or who perform special processes or functions are appropriately trained or supervised by a trained person? (See 6.4.1)	Yes, per Section 8.6.1 of QP-017 Rev E.
13.	Are training and qualification records maintained for all personnel?	Yes. Exception: qualifications for production assemblers were not available (see Finding details)



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Resource Management
(ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)

#	Audit Question / Item	Evidence / Comments
14.	Is the effectiveness of training evaluated or other action taken in order to ensure competency is evaluated? Is the methodology used to check effectiveness proportionate with the work for which the training or other action is being provided?	Yes, per Section 8.5 of QP-017 Rev E.
15.	How do individual activities contribute to the achievement of the quality policy and quality objectives?	Product built meets product specifications and therefore meets a component of customer satisfaction.
	Infrastructure - ISO 13485:2016 Section 6.3; 21 CFR 820.70; MDD 93/42/EEC Annex I (1)	
16.	Are the buildings, workspace, and utilities appropriate to meet product requirements, prevent mix-ups, maintain traceability, and assure orderly handling? How are they maintained?	Did not evaluate. Evaluate in the next supplier audit.
17.	Has a schedule been established and maintained for the periodic adjustment, cleaning, and other preventive maintenance of equipment to ensure manufacturing and quality specifications are met? Including hardware and software?	Did not evaluate. Evaluate in the next supplier audit.
18.	Are maintenance activities, including date and person(s) performing the activity, documented?	Did not evaluate. Evaluate in the next supplier audit.
19.	Is unscheduled maintenance documented, including date and person(s) performing the maintenance?	Did not evaluate. Evaluate in the next supplier audit.
20.	Are the maintenance activities being conducted as per the procedures? And are records being maintained?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Resource Management (ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
21.	What supporting services (such as transport or communication or information systems) are needed to ensure that product meets requirements? How are they maintained?	Did not evaluate. Evaluate in the next supplier audit.
	Work Environment- ISO 13485:2016 Section 6.4.1; 21 CFR 820.70; MDD 93/42/EEC Annex I (8)	
22.	Does the Organization have a documented procedure defining the requirements for the work environment?	Yes, QP-018 Rev N
23.	Where environmental conditions may have a potential adverse effect on product quality, have appropriate measures been put in place to ensure the control of those conditions? Check controls for: • Temperature/humidity • Cleanliness/contamination • Personnel health, cleanliness and clothing	Yes, per QP-018 Rev N
24.	Where environmental monitoring has been implemented, is the output regularly reviewed? Are appropriate corrective measures taken when necessary? Are reviews and corrective measures documented?	Yes. Daily Cleanroom Monitoring Log – Positive Pressure for Suite 104 cleanroom. Signed on 5/3/22. Reviewed by Bruce Tamayo.
25.	Is there a schedule for inspection of environmental controls? Is the schedule adhered to and is the equipment functioning properly? Are all activities documented?	Schedule established in QP-018 Rev N. Daily Cleanroom Monitoring Log – Positive Pressure for Suite 104 cleanroom. Signed on 5/3/22. Covers all work days in April 2022.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Resource Management (ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
	Contamination Control – ISO 13485:2016 Section 6.4.2; 21 CFR 820.70; MDD 93/42/EEC Annex I (8)	
26.	Does the organization have established procedures for the control of contaminated or potentially contaminated product in order to prevent contamination of the work environment, personnel, or product? (i.e. return of product from the clinical environment)	Yes, QP-018 Rev N
27.	For sterile medical devices, has the organization established procedures documenting the requirements for control of contamination with microorganisms or particulate matter and maintain the required cleanliness during assembly or packaging processes?	Yes, QP-018 Rev N
28.	Review environmental monitoring records	Reviewed Cleanroom Certification Report executed by CERTS for Cleanroom, Gown room Suite 104. Report dated 8/5/22. Certified to Class 7 per ISO 14644-1:2015.. Reviewed Microbial Test Results for lower level cleanroom Suite 104. Report dated 8/1/22. Microbial contact and air sampling. Reviewed Daily Cleanroom Monitoring Log – Positive Pressure for Suite 104 cleanroom. Signed on 5/3/22

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Resource Management (ISO 13485:2016 Section 6; 21 CFR 820 Subpart B, G; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
Additional Observations / Comments:		
Auditor Signature / Date: <i>Bryan Wang</i> 19 Oct 2022		

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
	Planning of Product Realization – ISO 13485:2016 Section 7.1; 21 CFR 820.20; MDD 93/42/EEC Annex I (2), Annex II (3.2)	
1.	Review previous audit results and audit action effectiveness for this section.	N/A
2.	Does the Organization have a documented procedure defining the processes needed for product realization?	Did not evaluate. Evaluate in the next supplier audit.
3.	Is the planning of product realization consistent with the requirements of the other processes of the QMS? (Verify there are no inconsistencies or conflicts between quality system procedures)	Did not evaluate. Evaluate in the next supplier audit.
4.	Where in the product realization process are the quality objectives and requirements for products determined?	Quality objectives and requirements for products are determined as needed throughout the product realization process.
5.	When planning for product realization, how processes, documents, and provide resources specific to the product established?	Planning for product realization is established in collaboration with client.
6.	How are processes established to provide resources specific to the product including infrastructure and work environment?	Planning for product realization including infrastructure and work environment needs is established in collaboration with client.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
7.	How are verification, validation, monitoring, inspection, test activities, handling, storage, distribution and traceability activities specific to the product and the criteria for product acceptance determined?	Established in collaboration with client
8.	What records exist showing that both the realization processes and the product meets these requirements?	Process and equipment qualifications and / or validations, DHR's
9.	What are the outputs of product realization planning?	Finished manufacturing procedures, routers/travelers, bill of materials
10.	Does the Organization have a documented procedure defining the requirements for risk management in the product realization process?	Did not evaluate. Evaluate in the next supplier audit.
11.	Review risk management file for sampling of products.	Did not evaluate. Evaluate in the next supplier audit.
	Customer-related Processes: Determination of Requirements Related to the product - ISO 13485:2016 Section 7.2.1; 21 CFR 820.30, 820.50; MDD 93/42/EEC Annex I, Annex II (3.2)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
12.	How does the Organization determine <u>each</u> of the following requirements: a) requirements specified by the customer, including the requirements for delivery and post-delivery activities, b) requirements not stated by the customer but necessary for specified or intended use, where known, c) statutory and regulatory requirements related to the product, and d) any user training needed to ensure specified performance and safe use of the medical device; e) any additional requirements determined by the organization.	Acceptance of work orders per WI-0006 Rev D. Observation #5: Recommend providing traceability to customer purchase order number in a work order. Reviewed DHRs for Lot #'s SLQ1610-012022 and SLQ1610-012122. Observation #6 In review of TM-HDX-0001 Rev A and TM-HDX-0002 Rev A, no acceptance criteria specified. However, Inspection Record Templates for SLQ-8004 A include acceptance criteria. Recommendation to update test method procedures to include appropriate acceptance criteria.
	Customer-related Processes: Review of Requirements Related to the Product – ISO 13485:2016 Section 7.2.2; 21 CFR 820.50, 820.160; No MDD Equivalent	
13.	Does the Organization have a documented procedure defining the processes / reviews to ensure that the organization has the ability to meet requirements, including regulatory requirements, before committing to supply product?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
14.	How does the Organization ensure that contract or order requirements differing from those previously expressed are resolved before committing to supply product?	Did not evaluate. Evaluate in the next supplier audit.
15.	Are records of the product requirement review results and actions resulting from them maintained?	Did not evaluate. Evaluate in the next supplier audit.
16.	When customers don't have documented requirements, how does the Organization confirm their requirements before accepting orders?	Did not evaluate. Evaluate in the next supplier audit.
17.	When product requirements are changed, how does the Organization ensure that relevant documents are changed and that relevant personnel are made aware of the changes?	Did not evaluate. Evaluate in the next supplier audit.
18.	Is user training identified in accordance with 7.2.1 above and is available or planned to be available?	Did not evaluate. Evaluate in the next supplier audit.
	Customer-related Processes: Customer Communication – ISO 13485:2016 Section 7.2.3; No QSR Equivalent; MDD 93/42/EEC Annex II (3.1)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
19.	What method(s) are used to communicate with customers regarding: • product information? • enquiries, contracts, or order handling, including amendments? • feedback, including customer complaints? (See section 8.2.1) • advisory notices? (See section 8.5.1)	N/A. Not responsibility of Pathway
20.	When customer communication involves recalls, Field Safety Notices and actions, does the organization have established procedures to communicate with regulatory authorities in accordance with applicable regulatory requirements?	N/A. Not responsibility of Pathway
	Design and Development - ISO 13485:2016 Section 7.3.1; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	
21.	Does the organization have established procedures for design and development?	Did not evaluate. Evaluate in the next supplier audit.
	Design and Development Planning - ISO 13485:2016 Section 7.3.2; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	
22.	Does the Organization have a documented procedure and plans for design and development?	Did not evaluate. Evaluate in the next supplier audit.
23.	What are the stages in the design and development process and where are they documented?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
24.	How are the review, verification, validation, and design transfer activities appropriate to each design and development stage determined?	Did not evaluate. Evaluate in the next supplier audit.
25.	How/where are design and development responsibilities, authorities and interfaces defined?	Did not evaluate. Evaluate in the next supplier audit.
26.	How does the Organization ensure effective communication and clear assignment of responsibility between different groups involved in design and development?	Did not evaluate. Evaluate in the next supplier audit.
27.	How does the organization ensure traceability of design and development outputs to design and development inputs? Is this documented?	Did not evaluate. Evaluate in the next supplier audit.
28.	Are resource needs, including necessary competence of personnel defined by the organization during design and development planning?	Did not evaluate. Evaluate in the next supplier audit.
29.	As product design and development progresses, how are the planning outputs documented and updated?	Did not evaluate. Evaluate in the next supplier audit.
	Design and Development Inputs – ISO 13485:2016 Section 7.3.3; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	
30.	Are design input procedures established that ensure that design requirements relating to a device are appropriate and include functional, performance, usability and safety requirements that address the intended use of the device, including the needs of the user and patient?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
31.	Do design input procedures include provisions for applicable regulatory requirements and standards?	Did not evaluate. Evaluate in the next supplier audit.
32.	Do design inputs include applicable output(s) of risk management?	Did not evaluate. Evaluate in the next supplier audit.
33.	Do design inputs include provisions for information derived from previous similar designs?	Did not evaluate. Evaluate in the next supplier audit.
34.	Are other requirements essential for design and development of product and processes considered as part of the design input process?	Did not evaluate. Evaluate in the next supplier audit.
35.	Are the design inputs reviewed for adequacy and approved?	Did not evaluate. Evaluate in the next supplier audit.
36.	How does the Organization ensure that requirements are complete, unambiguous able to be verified and validated and don't conflict with each other?	Did not evaluate. Evaluate in the next supplier audit.
	Design and Development Outputs - ISO 13485:2016 Section 7.3.4; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	
37.	Are design output procedures established that make reference to acceptance criteria and ensure that the outputs that are essential for the proper functioning of the device are identified?	Did not evaluate. Evaluate in the next supplier audit.
38.	Do the design and development outputs contain appropriate information for purchasing, production and service provisions where appropriate?	Did not evaluate. Evaluate in the next supplier audit.
39.	Is the design outputs documented, reviewed, and approved before release?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
40.	Is there a process to ensure the design outputs meet the input requirements?	Did not evaluate. Evaluate in the next supplier audit.
41.	Are the design outputs documented in the DHF?	Did not evaluate. Evaluate in the next supplier audit.
	Design and Development Review – ISO 13485:2016 Section 7.3.5; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	
42.	Are there procedures that ensure that formal documented design reviews of the design and development / design change are planned and conducted at the appropriate stages of the device's design development?	Did not evaluate. Evaluate in the next supplier audit.
43.	Do the design reviews evaluate the ability of the results of the design and development to meet requirements?	Did not evaluate. Evaluate in the next supplier audit.
44.	Do the design reviews include participants representative of functions concerned with the design and development stage being reviewed? including other specialist personnel?	Did not evaluate. Evaluate in the next supplier audit.
45.	Do the design reviews document the identification of the design, the date, actions, and the individuals performing the review?	Did not evaluate. Evaluate in the next supplier audit.
46.	Are design review documented in the design history file (DHF)?	Did not evaluate. Evaluate in the next supplier audit.
47.	Review records of the results of the reviews and verify any necessary actions have been taken.	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
	Design and Development Verification – ISO 13485:2016 Section 7.3.6; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	
48.	Are design verification requirements established in procedures that confirms how design output meets the design input requirements?	Did not evaluate. Evaluate in the next supplier audit.
49.	Are plans for verification, including methods, acceptance criteria and as appropriate, statistical techniques with rationale for sample size documented?	Did not evaluate. Evaluate in the next supplier audit.
50.	In cases where the intended use requires the medical device to be connected to, or to have an interface with, other medical device(s), does verification include confirmation that the design outputs meet design inputs when connected or interfaced?	Did not evaluate. Evaluate in the next supplier audit.
51.	Are the results of the design verification, including the identification of the design, method(s), the date, and the individuals performing the verification, documented in the DHF?	Did not evaluate. Evaluate in the next supplier audit.
52.	Review records of the results of the verification activities <u>and</u> resulting actions.	Did not evaluate. Evaluate in the next supplier audit.
	Design and Development Validation – ISO 13485:2016 Section 7.3.7; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
53.	Are design validation requirements established in procedures that outline how design validation is performed under defined operating conditions on production equivalent product? Is the rationale for the choice of product used for validation documented?	Did not evaluate. Evaluate in the next supplier audit.
54.	Are plans for validation, including methods, acceptance criteria and as appropriate, statistical techniques with rationale for sample size documented?	Did not evaluate. Evaluate in the next supplier audit.
55.	In cases where the intended use requires the medical device to be connected to, or to have an interface with, other medical device(s), does verification include confirmation that the design outputs meet design inputs when connected or interfaced?	Did not evaluate. Evaluate in the next supplier audit.
56.	Does the design validation confirm that devices conform to defined user needs and intended uses and include testing of production units under actual or simulated use conditions?	Did not evaluate. Evaluate in the next supplier audit.
57.	Is software validation considered where appropriate?	Did not evaluate. Evaluate in the next supplier audit.
58.	Are the results of design validation documented in the DHF?	Did not evaluate. Evaluate in the next supplier audit.
59.	Does the design validation include clinical evaluations, if required by national or regional regulations?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
60.	Has the design validation been completed prior to delivery to the customer or implementation of the product?	Did not evaluate. Evaluate in the next supplier audit.
61.	Review records of the validation activity results <u>and</u> any follow-up actions.	Did not evaluate. Evaluate in the next supplier audit.
	Design and Development Transfer - ISO 13485:2016 Section 7.3.8; 21 CFR 820.30; No MDD Equivalent	
62.	Has the device design been translated into production specifications as part of design transfer?	Did not evaluate. Evaluate in the next supplier audit.
63.	Are there procedures describing design transfer? Do these procedures ensure the design and development outputs are verified as suitable for manufacturing before becoming final production specifications and that production capability can meet production requirements?	Did not evaluate. Evaluate in the next supplier audit.
	Control of Design and Development changes - ISO 13485:2016 Section 7.3.9; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2, 3.4)	
64.	Does the Organization have a documented procedure describing how design and development changes are identified and documented? Do these procedures ensure the organization determines the significance of the change to function, performance, usability, safety and applicable regulatory requirements for the medical device and its intended use?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
65.	Are changes reviewed, verified, validated as appropriate, and approved before implementation?	Did not evaluate. Evaluate in the next supplier audit.
66.	Does the review of design and development changes include evaluation of the effect on component parts and products in the field, inputs or outputs of risk management and product realization processes?	Did not evaluate. Evaluate in the next supplier audit.
67.	Are records of design changes, their review and necessary action(s) maintained?	Did not evaluate. Evaluate in the next supplier audit.
	Design and Development Files, Design History File – ISO 13485:2016 Section 7.3.10; 21 CFR 820.30; MDD 93/42/EEC Annex II (3.2)	
68.	Has a design history file been established for each type of device? Does the DHF contain or reference the records necessary to demonstrate that the design was developed in accordance with the approved design plan and the requirements of design control?	Did not evaluate. Evaluate in the next supplier audit.
	Purchasing – ISO 13485:2016 Section 7.4; 21 CFR 820.50; Annex II	
	Purchasing Process – ISO 13485:2016 Section 7.4.1; 21 CFR 820.50; MDD 93/42/EEC Annex II (3.2)	
69.	Does the Organization have a documented procedure in place to ensure purchased material conforms to specified requirements?	Yes, QP-013 Rev E

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
70.	How is the type and extent of control applied to the supplier and the purchased product determined?	Purchased product must come from suppliers on the approved supplier list (ASL). Supplier are approved in accordance with QP-005.
71.	Does the Organization have a documented procedure defining the process to evaluate, select and rate suppliers, contractors, and consultants on their ability to meet specified requirements including quality requirements?	Yes, QP-005 Rev G.
72.	Are the results of supplier selections / qualification documented to include the following: a) Based on the supplier's ability to provide product that meets the organizations requirements b) Based on the performance of the supplier c) Based on the effect of the purchased product on the quality of the medical device d) Proportionate to the risk associated with the medical device	Yes, supplier selections are based on risk level of the product or service they provide. Suppliers are categorized from Category 1 (high risk) to Category 4. Performance of Category 1 and Category 2 suppliers is monitored in accordance with QP-005. Reviewed supplier qualification files for: - Steris Corp (Category 1). File included completed survey form (FM-0049 Rev B) completed on 10/28/19, ISO 13485:2016 certification from BSI (Certificate #MD 89745), Steris Assignment of Responsibility Form, Non Sterile Shipping Agreement, original approval on ECO-000534 (effective 10/29/19), and on-site audit report performed on 10/23/19. - Amcor Flexibles (Category 2). File included completed Amcor survey form completed on 2/15/21 in lieu of completing Pathway's survey form, ISO 13485:2016 certification from SGS (Certificate #US08/4926), certificate of compliance

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: 22-01

Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
		to ISO 14644-1:2015 from R Kraft (Report #AHCP201001). Amcor originally approved on 6/10/16 (documented on QP-0048 Rev B) and re-qualified on 9/9/22 per ECO-001106. - Ningbo Yingmed Medical Instruments (Category 2). File included completed survey form (FM-0049 Rev B) completed on 7/27/22, ISO 13485:2016 certification from TUV (Certificate #Q5 0992610005). Original approval on ECO-001100 effective 9/1/22. Observation #9: In review of the supplier survey form FM-0049 Rev B completed by Ningbo Yingmed Medical Instruments, they indicated that the company had been inspected by the FDA and are undergoing regulatory action, however, they did not include an explanation on the form. Recommend to follow up with the supplier. - Transcat (Category 4). File included Transcat's quality survey response form in lieu of completing Pathway's survey form, ISO/IEC 17025:2017 and ANSI/NCSL Z540-1-1994(R2002) certification from ANAB (Certificate #L2214). Transcat originally approved on 5/6/16 (documented on QP-0048 Rev B) and re-qualified on 9/1/22 per ECO-001099.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
73.	Has a list of approved suppliers, contractors, and consultants been established and is it current?	Yes, ASL is maintained in Arena.
74.	Is a system in place to ensure purchase orders are only placed with approved suppliers?	Yes, per QP-013 Rev F.
75.	How is on-going monitoring and re-evaluation of supplier performance performed? Are results of supplier monitoring data included in the re-evaluation process?	Re-evaluation is risk-based. Every 3 years a supplier survey is sent to suppliers. On-site audits are performed every 3 years on Category 1 suppliers. Regulatory certifications are reviewed annually. Performance reviews are conducted on Category 1 and selected Category 2 suppliers, as determined by management.
76.	Are suppliers notified when their product or service is found nonconforming to the specified requirements? Are issues addressed proportionate to the risk associated with the purchased product and compliance to applicable regulatory requirements?	Did not evaluate. Evaluate in the next supplier audit.
77.	Where appropriate, are suppliers requested to investigate the cause of the nonconformance and submit CAPA plans to prevent recurrences?	Did not evaluate. Evaluate in the next supplier audit.
78.	Are supplier CAPA plans reviewed for adequacy?	Did not evaluate. Evaluate in the next supplier audit.
79.	Is there a system in place to ensure suppliers provide notification of any changes that may affect the quality or performance of the product?	Did not evaluate. Evaluate in the next supplier audit.
80.	Are records of the evaluation, selection monitoring and re-evaluation of supplier capability or performance and any necessary actions arising from these activities maintained?	Yes

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
	Purchasing Information – ISO 13485:2016 Section 7.4.2; 21 CFR 820.50; MDD 93/42/EEC Annex II (3.2)	
81.	Do purchase orders/contracts include requirements for approval of product, procedures, processes and equipment?	PO's reviewed include reference to part number, description, and quantity at a minimum.
82.	Do purchase orders/contracts contain data clearly describe the product ordered, including, where applicable: a. The type, class, grade or other precise identification? b. The part number, title or other identification, and applicable revision of specifications, drawings, process requirements, inspection instructions, and other relevant technical data, including requirements for approval or qualification of product, procedures, process equipment and personnel? c. The title, number and issue of the quality system standard to be applied (if applicable)? d. Requirements for qualification of supplier personnel?	PO's contain the part number specification, description and quantity at a minimum. 1) PO-1285. Supplier: Amcor Flexibles. Description: Pouch, Chevron, 3.5" x 21" - TLP-058. Part number HDX-5001. Qty: 7,500
83.	How does the Organization ensure the adequacy of purchasing requirements before communicating them to the supplier?	Product specifications are released through document control.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
84.	Does purchasing information, include where applicable, a written agreement that the supplier notify the organization of changes in the purchased product prior to implementation of any changes that affect the ability of the purchased product to meet specified purchase requirements?	Did not evaluate.
85.	Are purchase orders/contracts documents and records maintained?	Yes
	Verification of Purchased Product – ISO 13485:2016 Section 7.4.3; 21 CFR 820.80; MDD 93/42/EEC Annex II (3.2)	
86.	Does the Organization have a documented procedure established to ensure that purchased product meets your purchasing requirements?	Yes, per QP-009 Rev E
87.	Is the extent of the organization's verification activities proportionate to the risks associated with the purchased product?	Yes
88.	If product verification is to be performed at the supplier's site, where are the verification arrangements and method of product release identified?	N/A
89.	When the organization becomes aware of any changes to the purchased product how is the affect of these changes on the medical device or product realization processes determined?	Pathway notifies customer.



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Product Realization
(ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)

#	Audit Question / Item	Evidence / Comments
90.	Are records of onsite verification maintained?	Yes. 1) HDX-5001 Rev A, Pouch (Lot # A199054). Inspection performed 3/11/22 by ST. Qty 7000. Observation #7: Silq PO not referenced. Inspected quantity does not match associated Silq PO# 000044 qty of 7500. 2) SLQ-3000 Rev A, Polymer Solution (Lot #HYX-111921A). Inspection performed 12/13/21 by ST/WH. Qty 85 liters Observation #8: In review of DHRs for Lot #'s SLQ1610-012022 and SLQ1610-012122, BOM Item No. 2 (Polymer Solution) was indicated as substitute that was used. However, based on review of inspection records for Polymer Solution (Lot #HYX-111921A), the part number and revision should be SLQ-3000 Rev A. Recommend correcting DHR's to indicate primary material designation was used. 3) SLQ-6004 Rev A Catheter (Lot #20211122). Inspection performed 1/21/22 by ST. Qty 3500 Observation #7: Silq PO not referenced. 4) SLQ-5004 Rev A, Label for pouch (Lot # not required per ID & Traceability QP-019 Rev F). No inspection required as indicated on material specification SLQ-5004 Rev A.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
	Production and Service Provision – ISO 13485:2016 Section 7.5; 21 CFR 820.70; MDD 93/42/EEC Annex II	
	Control of Production and Service Provision – ISO 13485:2016 Section 7.5.1; 21 CFR 820.70, 820.120, 820.130, 820.170, 820.184, 820.200; MDD 93/42/EEC Annex II (3.2)	
91.	Select a process for review. Review the specific procedure(s) for the process (i.e., manufacturing, inspection, etc.) and the methods for controlling and monitoring the process. Is the process controlled and monitored?	Manufacturing Procedure – SILQ ClearTract Catheter Surface Treatment using the Automation Machine MP-SLQ-002 Rev A. Process is controlled and monitored.
92.	Is the infrastructure for the manufacture of product suitable and qualified?	Yes, product is manufactured in a Class 7 certified cleanroom.
93.	Is suitable equipment used for carrying out production (or service)?	Equipment specified for production is used. Calibration and PM are scheduled as required.
94.	Are appropriate gages, etc. used in production (or service)? (See 7.6)	N/A
95.	Are packaging and labeling operations defined and implemented?	Yes, per MP-SLQ-0001 Rev A
96.	Is the process operating within specified limits? Use DHRs or other appropriate records to establish compliance.	Process is set up to run in a specified configuration and then executed. There are no limits of operation that are specified for the surface treatment other setting up the automation machine which includes selecting the catheter size and then loading the catheters into the machine for the automated process.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
		Observation #1: In review of DHRs for Lot #'s SLQ1610-012022 and SLQ1610-012122, there are empty fields for the secondary lot information. Should be marked N/A per Section 6.1.7 of the Quality Records SOP (QP-003 Rev F). Observation #4: In review of DHRs for Lot #'s SLQ1610-012022 and SLQ1610-012122, the work order number format does not conform to the format specified in Section 7.2 of Work Order procedure (WI-0006 Rev D).
97.	Is the DHR record verified and approved prior to distribution? Including the quantity manufactured and released for distribution?	DHR record is reviewed and approved prior to release to finished goods, including quantity released for distribution. Evidence: DHR's SLQ1610-012022 and SLQ1610-012122
	Cleanliness of Product and Contamination Control – ISO 13485:2016 Section 7.5.2; 21 CFR 820.70; MDD 93/42/EEC Annex II (3.2)	



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
98.	The Organization shall have documented requirements for cleanliness of product or contamination control of product: a) If product cleaned prior to sterilization or use? b) If product is supplied non-sterile is it cleaned prior to sterilization or its use? c) If product cannot be cleaned prior to sterilization or use and its cleanliness is of significance in use d) If product is supplied to be used non-sterile, and its cleanliness is of significance use e) If process agents are to be removed from product during manufacture	Requirements for cleanliness of product or contamination control of product are established by the customer. Requirements for cleanliness of product are established in the approved manufacturing procedures and travelers.
	Installation Activities – ISO 13485:2016 Section 7.5.3; 21 CFR 820.170; No MDD Equivalent	
99.	Does the Organization have documented requirements with acceptance criteria for installing and verifying the installation?	N/A. Pathway is excluded from Clause 7.5.3 Installation Activities.
100.	Does the Organization have documented requirements for installation and verification if installation is performed by a third party?	N/A. Pathway is excluded from Clause 7.5.3 Installation Activities.
101.	Are records of installation and verification maintained?	N/A. Pathway is excluded from Clause 7.5.3 Installation Activities.
	Servicing Activities – ISO 13485:2016 Section 7.5.4; 21 CFR 820.200; No MDD Equivalent	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
102.	Does the Organization have documented procedures, work instructions and reference materials and reference measurement procedures for performing servicing activities?	N/A to Silq product
103.	Does the organization analyze records of service activities, including if these are carried out by a supplier for the following: a) To determine if the information is to be handled as a complaint b) As appropriate, for input into the improvement process	N/A to Silq product
104.	Are records of servicing activities maintained?	N/A to Silq product
	Particular requirements for sterile medical devices – ISO 13485:2016 Section 7.5.5; No QSR Equivalent; No MDD Equivalent	
105.	Are records of the process parameters for the sterilization process documented and maintained for each sterilization batch?	Yes, DHR's SLQ1610-012022 and SLQ1610-012122
106.	Sterilization records traceable to each production batch of medical devices?	Yes, DHR's SLQ1610-012022 and SLQ1610-012122
	Production and Process Controls: Automated Processes – No ISO Equivalent; 21 CFR 820.70; No MDD Equivalent	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
107.	Is software that automates production processes or implements the quality system validated for its intended purpose? a. Are changes controlled and approved? b. Is the validation documented?	Did not evaluate. Evaluate in the next supplier audit.
108.	Does the Organization have a documented procedure that defines the requirements for validating software?	Did not evaluate. Evaluate in the next supplier audit.
109.	Does the software procedure address: • Management of all S/W validation activities (who, what, how, when, etc.) • 21 CFR Part 11 (electronic records & signatures) • Software requirements and traceability • Risk/hazard analysis • Off-the-shelf S/W, custom S/W, and in-house developed S/W • Design • Testing/validation • Changes and re-validation	Did not evaluate. Evaluate in the next supplier audit.
110.	Are S/W requirements evaluated for accuracy, completeness, consistency, testability, and clarity?	Did not evaluate. Evaluate in the next supplier audit.
111.	Were requirements and protocols approved prior to test commencement?	Did not evaluate. Evaluate in the next supplier audit.
112.	Is the S/W validation activity commensurate with the risk and complexity of the S/W?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
113.	Do the completed validation reports contain examples of sample outputs?	Did not evaluate. Evaluate in the next supplier audit.
114.	If any S/W issues were observed, were they handled in accordance with the protocol?	Did not evaluate. Evaluate in the next supplier audit.
115.	For programs that create, save and archive records electronically, do provisions exist in the software requirements document, protocol, test scripts, or validation results to address 21 CFR Part 11 requirements?	Did not evaluate. Evaluate in the next supplier audit.
	Validation of Processes for Production and Service Provision – ISO 13485:2016 Section 7.5.6; 21 CFR 820.75; MDD 93/42/EEC Annex II (3.2)	
116.	Does the Organization have a documented procedure describing process validation, including the requirements and responsibilities?	Did not evaluate. Evaluate in the next supplier audit.
117.	If processes can't be fully verified, are they validated in accordance with the procedure?	Did not evaluate. Evaluate in the next supplier audit.
118.	Is there a system in place to assure all processes and/or major equipment requiring validation are identified, continually monitored and controlled, and revalidated as required?	Did not evaluate. Evaluate in the next supplier audit.
119.	Are process validation protocols (IQ, OQ, PQ) written with acceptance criteria and approved prior to conducting the process validation?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
120.	Are process validation results documented providing objective evidence the process and/or equipment consistently produces the desired results? Did the same personnel who approved the protocols approve the results?	Did not evaluate. Evaluate in the next supplier audit.
121.	Have procedures been established and maintained for the monitoring and control of validated process parameters to ensure the specific requirements continue to be fulfilled?	Did not evaluate. Evaluate in the next supplier audit.
122.	Are validated processes operated by qualified personnel? Do quality records exist to identify the procedures used, as well as, the identity and date of the operator performing the task?	Did not evaluate. Evaluate in the next supplier audit.
123.	Are changes to validated processes evaluated for the necessity to re-validate or justify when no re-validation is required?	Did not evaluate. Evaluate in the next supplier audit.
124.	When applicable, have valid statistical techniques been applied? Including rationale for sample plans	Did not evaluate. Evaluate in the next supplier audit.
125.	Does the Organization have a documented procedure for validation of computer software?	Did not evaluate. Evaluate in the next supplier audit.
126.	Review records of software validation.	Did not evaluate. Evaluate in the next supplier audit.
	Particular Requirements for Sterile Medical Devices – ISO 13485:2016 Section 7.5.7; No QSR Equivalent; MDD 93/42/EEC Annex VII (3, 8.4)	



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
127.	Does the Organization have documented procedures defining the requirements for the validation of sterilization processes and sterile barrier systems?	Did not evaluate. Evaluate in the next supplier audit.
128.	Review sterilization validation and packaging sealer validation protocols / reports.	Did not evaluate. Evaluate in the next supplier audit.
129.	Are processes for sterilization validation and sterile barrier packaging validated prior to implementation and following product or process changes as appropriate? (See ISO 11607-1 & -2)	Did not evaluate. Evaluate in the next supplier audit.
130.	Records to support these processes maintained? including validation conclusions and any necessary actions?	Did not evaluate. Evaluate in the next supplier audit.
	Identification – ISO 13485:2016 Section 7.5.8; 21 CFR 820.60; MDD 93/42/EEC Annex I (13), Annex II (3.2)	
131.	Does the Organization have documented procedures defining how the product is identified throughout manufacturing processes (i.e., in production, storage, segregation areas, etc.)?	Yes, per QP-019 Rev F
132.	Does the Organization have documented procedures to ensure that medical devices returned to the organization are identified and distinguished from conforming product?	Yes, per QP-008 Rev C
133.	Review actual product in various stages of manufacturing and returns to verify proper identification.	During the facility tour, product at various stages of manufacturing was properly identified with part number, revision, lot number, quantity, and acceptance status.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
134.	How is product inspection status identified (verify in production, storage, segregation areas, etc.)?	Acceptance status label
135.	Does the organization have documented procedures for assigning a unique device identification (UDI) as defined by the applicable regulatory requirements?	Did not evaluate. Evaluate in the next supplier audit.
	Traceability – ISO 13485:2016 Section 7.5.9; 21 CFR 820.65; MDD 93/42/EEC Annex I (13), Annex II (3.2)	
	General – ISO 13485:2016 Section 7.5.9.1; 21 CFR 820.65; MDD 93/42/EEC Annex I (13), Annex II (3.2)	
136.	Does the Organization have documented procedures defining product traceability requirements? including regulatory requirements?	Yes, per QP-019 Rev F
137.	Is the level of traceability appropriate to facilitate corrective and preventive action?	Yes
138.	Do records (i.e., DHR) of traceability include components and materials used as well as specified environmental conditions where traceability is deemed necessary?	Yes, DHRs SLQ1610-012022 and SLQ1610-012122
	Particular Requirements for Active Implantable Medical Devices and Implantable Medical Devices – ISO 13485:2016 Section 7.5.9.2; 21 CFR 820.65; MDD 93/42/EEC Annex I (13), Annex II (3.2)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
139.	Where materials/products are intended for surgical implantation, are procedures established for the identification of the product with a unique batch or lot number? Is such identification documented?	N/A. Pathway does not manufacture implantable medical devices.
140.	Do records include traceability to components, materials and conditions for the work environment used, if these could cause the medical device not to satisfy its specified safety and performance requirements?	N/A. Pathway does not manufacture implantable medical devices.
141.	Are records of the name and address of shipping package consignee maintained?	N/A. Pathway does not manufacture implantable medical devices.
Customer Property – ISO 13485:2016 Section 7.5.10; No QSR Equivalent; No MDD Equivalent		
142.	Does the Organization use any customer-owned property (i.e., product, packaging, drawings, tooling, gages, etc. Note: this can include intellectual property or confidential health information)?	Did not evaluate. Evaluate in the next supplier audit.
143.	How does the Organization ensure that customer-owned property is identified, verified, protected, and safeguarded?	Did not evaluate. Evaluate in the next supplier audit.
144.	If any customer property is lost, damaged etc., how is it reported to the customer?	Did not evaluate. Evaluate in the next supplier audit.
145.	Are records pertaining to customer property maintained?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
	Preservation of Product – ISO 13485:2016 Section 7.5.11; 21 CFR 820.120, 820.130, 820.140, 820.150, 820.160; MDD 93/42/EEC Annex I (5, 13)	
146.	How does the Organization preserve the conformity of product during internal processing and delivery (distribution)?	Did not evaluate. Evaluate in the next supplier audit.
147.	How do identification, handling, packaging, storage and protection address the preservation of product?	Did not evaluate. Evaluate in the next supplier audit.
148.	Does this also apply to component parts?	Did not evaluate. Evaluate in the next supplier audit.
149.	Does the organization have documented procedures for protecting product from alteration, contamination or damage when exposed to expected conditions and hazards during processing, storage handling and distribution by:	Did not evaluate. Evaluate in the next supplier audit.
150.	a) Designing and constructing suitable packaging and shipping containers?	Did not evaluate. Evaluate in the next supplier audit.
151.	b) Documenting requirements for special conditions needed if packaging alone cannot provide preservation? If special conditions are required are these controlled and recorded?	Did not evaluate. Evaluate in the next supplier audit.
	Labeling and Packaging Control – No ISO Equivalent; 21 CFR 820.120, 820.130; No MDD Equivalent	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
152.	Does the Organization have documented procedures defining the requirements for control of labeling activities?	Yes, per MP-0016 Rev A
153.	Have measures been taken to ensure labels are able to withstand the customary conditions of processing, storage, handling, distribution, and where appropriate; use, while remaining affixed and legible?	Labels are stored at ambient warehouse conditions. During manufacturing, labels are applied in controlled environment and inspected and documented in SLQ-8004 A Inspection Record Template.
154.	Is customized labeling inspected for accuracy prior to its release for storage or use?	Yes
155.	a. Is evidence of this inspection, including the date and signature of the individual(s) performing the inspection, documented in the DHR?	Yes, DHRs SLQ1610-012022 and SLQ1610-012122
156.	Is customized labeling stored in a manner to prevent mix-ups and to ensure proper identification?	Customized labeling generated for a manufacturing lot is used or scrapped during the production run and not stored.
157.	Are labeling and packaging operations separated, physically or otherwise, to the degree necessary to assure there are no mix-ups between similar products or labels?	Yes
158.	Is the label and labeling used for each batch or lot documented in the DHR?	Yes, DHRs SLQ1610-012022 and SLQ1610-012122

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
159.	Have measures been taken to ensure device packaging is able to withstand customary conditions of processing, storage, handling, and distribution, while protecting the device from alteration or damage?	Yes, per TR-0068 Rev A and TR-0074 Rev A
160.	Are all changes to device labeling and packaging made in accordance with a formal change control system?	Yes, per QP-002.
161.	Does the DHR document the person(s) performing final labeling activities?	Yes, DHRs SLQ1610-012022 and SLQ1610-012122
	Control of Monitoring and Measuring Devices – ISO 13485:2016 Section 7.6; 21 CFR 820.70, 820.72; MDD 93/42/EEC Annex II (3.2)	
162.	Does the Organization have a documented procedure to ensure inspection, measurement and test equipment is regularly calibrated, inspected and maintained?	Yes, per QP-020 Rev D
163.	Are the following activities documented:	
164.	a. Does it contain provisions for handling, preservation, and storage of equipment in order to maintain its accuracy and fitness for use?	Yes, per QP-020 Rev D
165.	b. Does it contain a method by which equipment will be identified with a calibration label or other suitable means of identifying the calibration status of the equipment?	Asset number is assigned per WI-0008.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
166.	c. Does it contain the method for establishing calibration intervals and adjusting the intervals when necessary?	Yes, per QP-020 Rev D
167.	d. Does it contain a corrective action process when out-of-tolerance conditions are found, including a requirement to evaluate any potential adverse effect on product quality? Are there provisions for assessing the previous measurements?	Yes, per QP-020 Rev D
168.	e. Are calibration standards traceable to national, international, or other independent reproducible standards? If no standards exist, is the basis used for calibration or verification recorded?	Yes, calibration standards should be traceable to NIST or equivalent recognized agency
169.	f. Is the test hardware and software safeguarded from adjustments which could invalidate calibrations?	Yes, per QP-020 Rev D
170.	Is inspection, measuring and test equipment suitable for its intended purpose?	Yes,
171.	Is all measurement/test equipment within its current calibration interval?	Yes
172.	If any items are found beyond their calibration due date, are they conspicuously identified and segregated to prevent unintended use?	Yes, per QP-020 Rev D

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
173.	Is computer software used for monitoring and measurement? If so, is its ability to perform that function confirmed prior to initial use and reconfirmed as necessary? Are verification and validation activities proportionate to the risks?	No
174.	Are records of calibration, verification and validation activities maintained including conclusions and any necessary actions? (See ISO 10012)	Yes. 1) Impulse Sealer (ES-0099-01): Last calibrated on 8/12/22 by Transcat (Calibration report #91678957). Received and returned in tolerance. 2) Force gauge, pouch peel test (ES-0028-01): Last calibrated on 1/26/22 by Transcat (Calibration report #91662335). Received and returned in tolerance. 3) No calibration file for Automation Machine Coating (ES-0091-01). Calibration and PM are not required per TR-0117 Rev A, IQ Checklist . Observation #3: Supporting documentation to support the operation and maintenance requirements could not be found in document control. 4) No calibration file for Spectrophotometer (ES-0092-01). Calibration and PM are not required per TR-0137 Rev A, IQ Checklist.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Product Realization (ISO 13485:2016 Section 7; 21 CFR 820 Subpart B, C, E, F, G, H, L, N; MDD 93/42/EEC Annex I, II)		
#	Audit Question / Item	Evidence / Comments
Additional Observations / Comments:		
Auditor Signature / Date:  19 Oct 2022		

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
	Measurement, Analysis and Improvement: General – ISO 13485:2016 Section 8.1; 21 CFR 820.80, 820.250; MDD 93/42/EEC Annex II (3.2)	
1.	Review previous audit results and audit action effectiveness for this section.	N/A
2.	How does the organization plan and implement measurement, analysis and improvement processes needed: a. to demonstrate conformity of the product? b. to ensure conformity of the quality management system? c. to maintain the effectiveness of the quality management system?	Per management review QP-001 Rev G
3.	How does the organization determine what monitoring measurement, and analysis methods to use, including statistical techniques?	Identified in the quality manual QM-01 Rev E
4.	How does the organization determine the extent of their use?	Per applicable procedures outlined in the quality manual.
	Monitoring and measurement – ISO 13485:2016 Section 8.2; 21 CFR 820.198)	
	Monitoring and measurement: Feedback – ISO 13485:2016 Section 8.2.1; 21 CFR 820.198; MDD 93/42/EEC Annex II (3.1)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
5.	Is there a system in place to monitor whether customer requirements have been met? Are the methods for obtaining and using this information documented?	Did not evaluate. Evaluate in the next supplier audit.
6.	Is the feedback system documented?	Did not evaluate. Evaluate in the next supplier audit.
7.	Does it provide early warning of quality problems?	Did not evaluate. Evaluate in the next supplier audit.
8.	Does it provide input into the CAPA process?	Did not evaluate. Evaluate in the next supplier audit.
9.	Is review of experience gained from post-production part of the feedback system?	Did not evaluate. Evaluate in the next supplier audit.
10.	Does the information serve as input into risk management for monitoring and maintaining product requirements as well as product realization for improvement processes?	Did not evaluate. Evaluate in the next supplier audit.
11.	Where appropriate if regulatory agencies require the organization to gain specific experience from post-production activities, is the review of this experience part of the feedback process? (For example, review of post market clinical studies).	Did not evaluate. Evaluate in the next supplier audit.
	Monitoring and measurement: Complaint Handling – ISO 13485:2016 Section 8.2.2; 21 CFR 820.198; MDD 93/42/EEC Annex II (3.1)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
12.	Does the Organization have a documented procedure to ensure timely complaint handling in accordance with applicable regulatory requirements? Including minimum requirements and responsibilities for: a) Receiving and recording information b) Evaluating information to determine if feedback constitutes a complaint c) Investigating Complaints d) Determining the need to report the information to the appropriate regulatory authorities e) Handling of complaint related product f) Determining the need to initiate corrections or corrective actions	Did not evaluate. Evaluate in the next supplier audit.
13.	If a complaint is not investigated is the reason justified?	Did not evaluate. Evaluate in the next supplier audit.
14.	Does the organization have procedures for corrections or corrective action resulting from the complaint handling process, is this documented?	Did not evaluate. Evaluate in the next supplier audit.
15.	In the event the complaint indicates an outside organization contributed to the complaint, is relevant information exchanged between the organization and the third party?	Did not evaluate. Evaluate in the next supplier audit.
16.	Are complaint records maintained?	Did not evaluate. Evaluate in the next supplier audit.



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
	Monitoring and measurement: Reporting to Regulatory Authorities – ISO 13485:2016 Section 8.2.3; See 21 CFR 803 and 806; MDD 93/42/EEC Annex II (3.1)	
17.	Does the organization have procedures documenting that, when required, regulatory authorities are notified of complaints that meet specified reporting criteria of adverse events or issuance of advisory notices?	Did not evaluate. Evaluate in the next supplier audit.
18.	Are these records maintained?	Did not evaluate. Evaluate in the next supplier audit.
19.	Ask if there have been any advisory notices/recalls.	Did not evaluate. Evaluate in the next supplier audit.
	Monitoring and measurement: Internal audit – ISO 13485:2016 Section 8.2.4; 21 CFR 820.22, 820.70; MDD 93/42/EEC Annex II (3.1, 3.2)	
20.	Does the Organization have a documented procedure to ensure internal quality system audits are conducted?	Did not evaluate. Evaluate in the next supplier audit.
21.	Are internal audits being conducted at planned intervals? Including consideration of status and importance of areas to be audited?	Did not evaluate. Evaluate in the next supplier audit.
22.	Do individuals who do not have direct responsibility for the matters being audited perform the audits?	Did not evaluate. Evaluate in the next supplier audit.
23.	Do internal audits determine whether the QMS conforms to the requirements of ISO 13485 and to the other regulatory requirements established by Organization?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
24.	Can you show me an audit plan that takes into consideration the importance of the processes and areas to be audited, and the results of previous audits?	Did not evaluate. Evaluate in the next supplier audit.
25.	Where are the audit criteria, scope, frequency and methods defined?	Did not evaluate. Evaluate in the next supplier audit.
26.	Does the organization's management review audit results?	Did not evaluate. Evaluate in the next supplier audit.
27.	When audit deficiencies are found: a. Are they documented appropriately? and b. Are corrective action(s) taken when necessary?	Did not evaluate. Evaluate in the next supplier audit.
28.	Who ensures that actions are taken to eliminate detected nonconformities and their causes? Are they being taken care of in a timely manner?	Did not evaluate. Evaluate in the next supplier audit.
29.	Are dates and results of internal quality system audits, including follow-up activities, documented?	Did not evaluate. Evaluate in the next supplier audit.
	Monitoring and measurement of processes - ISO 13485:2016 Section 8.2.5; 21 CFR 820.70, 820.250; MDD 93/42/EEC Annex II (3.2)	
30.	What methods are used to monitor and measure the QMS processes?	Did not evaluate. Evaluate in the next supplier audit.
31.	Is there documented evidence that they have achieved the desired results?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
32.	When the desired results are not achieved, what actions are taken to ensure that the product meets requirements?	Did not evaluate. Evaluate in the next supplier audit.
	Monitoring and measurement of product – ISO 13485:2016 Section 8.2.6; 21 CFR 820.70, 820.80, 820.120, 820.184, 820.250; MDD 93/42/EEC Annex II (3.2)	
33.	Have procedures been established for acceptance activities?	Yes, Receiving Inspection (QP-009), In-Process Inspection (QP-031), and Final Product Inspection (QP-010)
34.	At what stages of the product realization process do monitoring and measuring activities take place?	Monitoring and measuring activities take place at all stages of the product realization process.
35.	Do receiving inspection procedures contain adequate instructions to personnel regarding: • Requirements that incoming product will be inspected, tested, or otherwise verified as conforming to specified requirements? • Any special handling or processing requirements (e.g. expiration date, certificates of conformance, etc.)? • Any inspection characteristics, critical dimensions, dimensional tolerances, etc? • The need for specialized inspection, measurement, and test equipment?	Yes, applicable requirements, inspection characteristics, critical dimensions with tolerances, sampling size, and equipment are specified on inspection procedures and forms. Example: T-0002 Rev A for Amcor Flexible pouch

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
36.	Where incoming product is released for urgent production purposes prior to inspection/verification, is it identified and recorded in order to permit immediate recall in the event of nonconformity to specified requirements?	No provision for this in Receiving Inspection, QP-009 Rev E
37.	When sampling is conducted: - Is a statistically valid sampling plan being used? - Do records indicate each supplier's lot was individually sampled and inspected? - Are samples selected randomly from each entire lot?	Yes, per AQL specified in QP-009 Rev E
38.	Is product awaiting inspection or found nonconforming adequately identified and segregated to prevent unintended use?	Yes, per QP-009 Rev E
39.	Are personnel conducting acceptance activities properly trained in such processes as: - The use of required inspection equipment? - The impact of the nonconformance on product performance if the nonconformance is not found through proper inspection or testing?	Yes, personnel are trained to procedures identified in the training matrix according to their role.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
40.	Do acceptance records contain the following information: - The acceptance activity performed and the results? - The acceptance date(s), signature of person(s) conducting the acceptance, lot size, sample size, quantity accepted/rejected, and where appropriate, the equipment used?	Yes, acceptance records contain the appropriate information of the acceptance activity performed. - HDX-5001 Rev A, Pouch (Lot #950792). Inspection performed 3/11/19 by B Tamayo. Qty 6500. Part number released as PSM-5000 which is substitute part in DMR. SLQ-3000 Rev A, Polymer Solution (Lot #HYX-111921A). Inspection performed 12/13/21 by ST/WH. Qty 85 liters - SLQ-6004 Rev A Catheter (Lot #20211122). Inspection performed 1/21/22 by ST. Qty 3500
41.	Are controls in place to ensure finished devices are not released for distribution until: - All DMR activities are complete? - All associated data and documentation is reviewed? - The release is authorized by the signature of a designated individual? - The authorization is dated?	Yes, DMR and travelers include final product inspection to ensure all DMR activities are complete and required records have been reviewed and are present. DHRs SLQ1610-012022 and SLQ1610-012122
42.	Are acceptance and rejection records maintained as part of the DHR?	Yes, DHRs SLQ1610-012022 and SLQ1610-012122
43.	Do records provide evidence product has been inspected and/or tested and clearly show the product has passed or failed?	Yes, DHRs SLQ1610-012022 and SLQ1610-012122

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
44.	Are controls in place to ensure the identification of acceptance status is maintained throughout manufacturing, packaging, and labeling and that only product which has passed required acceptance activities is distributed or used?	Yes
45.	Are final inspection and testing carried out in accordance with the quality plan and/or documented procedures to complete the evidence of product conformance to specified requirements?	Yes, final inspection and testing is carried out in accordance with established procedures and traveler/forms.
46.	Does the quality plan and/or documented procedures for final inspection and testing require all specified inspections and tests, including those specified either on product receipt or in-process, are carried out and the results meet specified requirements?	Yes. SLQ-8004 A Inspection record in DHRs SLQ1610-012022 and SLQ1610-012122
47.	If product fails to pass any inspection and/or test requirements, are the nonconforming product procedures applied?	Yes
	Control of nonconforming product and rework – ISO 13485:2016 Section 8.3.1 & 8.3.4; 21 CFR 820.90; MDD 93/42/EEC Annex II (3.2)	
48.	Does the organization have a documented procedure defining the controls for dealing with nonconforming product?	Yes, per QP-008 Rev C

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
49.	Does the nonconforming material procedure describe the system for controlling nonconforming material, including: segregation, identification, documentation, evaluation, disposition, responsibilities/authorities and notification of those responsible for the nonconformance?	Yes
50.	Does the system prevent unintended use or delivery of nonconforming product?	Yes
51.	Is the responsibility for disposition of nonconforming material clearly defined?	Yes, disposition by MRB
52.	Is the disposition of nonconforming product documented, including justification of the use of nonconforming product?	Yes, risk evaluation is performed for UAI dispositions
53.	When nonconforming product is corrected, can you demonstrate that it is re-verified to ensure it conforms to requirements including regulatory requirements (8.3.4)?	Yes, required of Rework dispositions
54.	Is a determination made and documented of any adverse effect of rework upon the product prior to its authorization and approval? (ISO 8.3.4)	Yes
55.	Are rework work instructions approved by same authority as the original work instruction? (ISO 8.3.4)	Yes
56.	Do records show evidence of determination of adverse effects of rework? (8.3.4)	The NCR records reviewed did not include rework disposition.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
57.	When the non-conformity can be attributed to a third party, do procedures require notification?	Yes
58.	Does the organization maintain records properly identifying the nature of the nonconformity, subsequent action taken, including evaluation, any investigation and rationale for decisions?	Yes 1) NCR #21-0006 2) NCR #21-0007 3) NCR #21-0011
	Actions in Response to Nonconforming Product Detected Before Delivery –ISO 13485:2016 Section 8.3.2; 21 CFR 820.90; MDD 93/42/EEC Annex II (3.2)	
59.	Does the organization deal with nonconforming product by: a) Taking action to eliminate the detected nonconformity? b) Taking action to preclude its original intended use or application c) Authorizing its use, release or acceptance under concession?	Yes, per QP-008 Rev C
60.	In cases where the product is accepted under concession does the organization provide justification? Is this approved and are applicable regulatory requirements met? Is the person authorizing the concession documented?	Yes, per QP-008 Rev C
	Actions in Response to Nonconforming Product Detected After Delivery – ISO 13485:2016 Section 8.3.3; 21 CFR 820.90; MDD 93/42/EEC Annex II (3.2)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
61.	When nonconforming product is detected after shipment, what actions are taken, such as containment? Are these actions documented and records maintained?	Did not evaluate. Evaluate in the next supplier audit.
62.	Does the organization have documented procedures for issuing advisory notices in accordance with applicable regulatory requirements? Can these be executable any time?	Did not evaluate. Evaluate in the next supplier audit.
63.	Are records of relating to advisory notices maintained?	Did not evaluate. Evaluate in the next supplier audit.
	Analysis of data - ISO 13485:2016 Section 8.4; 21 CFR 820.100, 820.250; No MDD Equivalent	
64.	What data is collected and analyzed to demonstrate the suitability and effectiveness of the QMS and to evaluate where continual improvement of its effectiveness can be made?	Data includes but not limited to: feedback/complaint handling/reporting to regulatory agencies, results from audits/inspections, monitoring and measurement of processes/products, supplier performance, CAPAs, review of changes that could affect the QMS, applicable new or revised regulations, and resource needs
65.	Does the organization have documented procedures that describe this activity? Including identification of appropriate methods, including statistical techniques and the extent of their use?	Yes, per Management Review QP-001 Rev G

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
66.	What information does this analysis provide relating to: • feedback? (section 5.6) • conformity to product requirements? (section 5.6) • characteristics and trends of processes and products including opportunities for improvement? (section 5.6) • suppliers? (section 7.4.1) • service reports, as appropriate	Information provides data in order for management to assess the suitability and effectiveness of the QMS and opportunity for necessary actions to improve.
67.	Does the organization have documented procedures to ensure that, in the event the analysis shows that the quality management system is not suitable, effective or adequate, the organization uses this as input for improvement? (see 8.5)	Yes, per QP-001 Rev G
68.	Are records of analysis maintained?	Yes. Evidence of most recent Management Review Meeting held on May 6, 2022. Agenda of topics discussed was provided.
	Improvement: General – ISO 13485:2016 Section 8.5.1; 21 CFR 820.20, 820.198; MDD 93/42/EEC Annex II (3.1, 3.2)	
69.	Is there documented evidence to demonstrate that the Organization identifies and implements changes to ensure continued QMS effectiveness? What tools are used? (See 5.6, 8.2.2, 8.4, 8.5.2, 8.5.3)	CAPA system (QP-023 Rev D) is used to demonstrate changes implemented to address systemic issues in order to ensure continued QMS effectiveness.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
70.	Does the Organization have documented procedures for the issue and implementation of advisory notices?	Did not evaluate. Evaluate in the next supplier audit.
71.	Are there procedures for receiving, reviewing, evaluation and maintaining customer complaints?	Did not evaluate. Evaluate in the next supplier audit.
72.	Does a formally designated unit manage complaints?	Did not evaluate. Evaluate in the next supplier audit.
73.	If complaints are determined as not needing investigation, is the reason documented? Is the person responsible for not investigating documented?	Did not evaluate. Evaluate in the next supplier audit.
74.	Do complaint records contain all information required by FDA and ISO 13485?	Did not evaluate. Evaluate in the next supplier audit.
75.	Does a procedure exist governing the activity for corrections and removals?	Did not evaluate. Evaluate in the next supplier audit.
	Improvement: Corrective action – ISO 13485:2016 Section 8.5.2; 21 CFR 820.100; MDD 93/42/EEC Annex II (3.1)	

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
76.	Do corrective actions records identify and address root cause(s)? (Do root causes match actions?)	Yes. 1) CAPA 21-0010: Initiated as a result of internal audit finding where ESD continuous monitors were not calibrated or on the Cal/PM list. The CAPA plan was approved and corrective actions implemented. The effectiveness checks confirmed the corrective action addressed the root cause. 2) CAPA 21-0011: Initiated as a result of viable particulates exceeding the action limit in Areas 1 and 4 and exceeding specification in Areas 2, 3, and 5 per QP-018.; also, viable particulate for cleanroom suite 105 exceeded action limit of the gowning room. CAPA plan was approved and corrective actions implemented. Effectiveness checks (3 consecutive monitoring results meeting acceptance criteria) confirmed the corrective actions addressed the root cause.
77.	Are actions taken appropriate to the severity of the problem?	Yes. CAPA 21-0010 and CAPA 21-0011



DOCUMENT TITLE:
Internal Quality Audit Checklist

Revision:
A

DOCUMENT #:
FM2-QM.SLQ017

Effective Date:
25Sep2020

Audit #: 22-01

Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
78.	Does the Organization have a documented procedure defining requirements for each of the following? a. reviewing nonconformities (including customer complaints) b. determining the causes of nonconformities c. evaluating the need for action to ensure that nonconformities do not recur d. planning and documenting, and implementing action needed, as appropriate updating documentation e. verifying that the corrective action does not adversely affect the ability to meet the applicable regulatory requirements or safety and performance of the medical device f. reviewing the corrective action taken and its effectiveness	Yes, Nonconforming Materials SOP (QP-008 Rev C) and CAPA SOP (QP-023 Rev D)
79.	Is corrective/preventive action verified/validated for effectiveness, including training? Does the system ensure finished devices are not adversely affected?	Yes. CAPA 21-0010 and CAPA 21-0011
80.	Review records of investigation and corrective actions taken.	CAPA 21-0010 and CAPA 21-0011 Observation #10 In review of CAPA No's 21-0010 and 21-0011, it is not clear if CAPAs were closed. It is recommended to update the CAPA form (FM-0020 Rev A) to include a


	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
		CAPA closure section to conform with Section 7.7 of the CAPA SOP (QP-023 Rev D).
81.	For existing problems/corrective actions, were quality data sources identified?	Yes
	Improvement: Preventive action – ISO 13485:2016 Section 8.5.3; 21 CFR 820.100; No MDD Equivalent	
82.	How does the organization determine potential nonconformities to take action on? Do preventive action records identify and address root cause(s)?	In accordance with CAPA SOP (QP-023 Rev D)
83.	Are actions taken appropriate to the severity of the problem?	Unable to determine based on CAPA's reviewed.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
84.	Does the organization have a documented procedure defining requirements for each of the following? a. determining potential nonconformities and their causes, b. evaluating the need for action to prevent occurrence of nonconformities, c. planning and documenting, and implementing action needed, as appropriate updating documentation d. recording of the results of any investigation and of action taken (see 4.2.5), e. verifying that the corrective action does not adversely affect the ability to meet the applicable regulatory requirements or safety and performance of the medical device	Yes, per QP-023 Rev D
85.	Review records of preventive actions taken.	CAPAs (CAPA 21-0010 and CAPA 21-0011) reviewed involved corrective actions taken.
	Statistical Techniques – ISO 13485:2016 Section 8.1, 8.4; 21 CFR 820.250; MDD 93/42/EEC Annex II (3.2)	
86.	Are sampling plans or other statistical methods used where appropriate?	Did not evaluate. Evaluate in the next supplier audit.
87.	Are the techniques based on valid statistical rationale?	Did not evaluate. Evaluate in the next supplier audit.

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
Measurement, Analysis and Improvement (ISO 13485:2016 Section 8; 21 CFR 820 Subpart B, G, H, I, J, M, O; MDD 93/42/EEC Annex II)		
#	Audit Question / Item	Evidence / Comments
88.	Are the techniques reviewed in consideration for nonconforming product, quality audit reports, and/or other feedback information?	Did not evaluate. Evaluate in the next supplier audit.
89.	Are the statistical methods documented?	Did not evaluate. Evaluate in the next supplier audit.
90.	When statistical methods are applied, are they evaluated to be adequate for their intended use, and are they reviewed when changes occur?	Did not evaluate. Evaluate in the next supplier audit.
Additional Observations / Comments:		
Auditor Signature / Date:  19 Oct 2022		

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
MDD 93/42/EEC Requirements		
#	Audit Question / Item	Evidence / Comments
1.	Has the device(s) been divided into Class I, IIa, IIb, or III according to the classification rules as defined in Annex IX? Is a rationale available for the classification of the device(s)?	N/A
2.	Has a technical file been generated to allow assessment of the conformity of the product with the requirements of the Directive? Does the technical file contain the particulars defined in the Annex VII part 3?	N/A
3.	Has a written declaration of conformity been completed by the manufacturer in accordance with Annex II/V/VI, covering the concerned products manufactured that meets the provisions of the Directive which apply to them?	N/A
4.	Does the EC type-examination certificate contain all relevant information for the identification of the type approved? Has appropriate documentation been annexed to the certificate and a copy kept by the notified body? Do the provisions in the quality system require that the notified body which issued the EC type-examination certificate be made aware of any significant change made to the approved product?	N/A
5.	Has the manufacturer identified the relevant Essential Requirements that apply to the device?	N/A
6.	Are the characteristics and performances, referred to in Essential Requirements Sections 1, 2 and 3, of the device(s) qualified during the lifetime of the product as stated by the manufacturer to not be adversely affected to such a degree that safety of the patient, and when applicable, of other persons are compromised when subjected to stresses during normal conditions of use?	N/A

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
MDD 93/42/EEC Requirements		
#	Audit Question / Item	Evidence / Comments
7.	Are the required quality assurance records available and retained for a period ending at least 5 years after the 1st product was manufactured as described in Annex II Section 6.1, Annex V Section 5.1, or Annex VI Section 5.1? (e.g., record retention, etc.)	N/A
8.	Are all records relating to notified bodies and competent authorities available and retained for the duration as described in Annex II Section 6.1, Annex V Section 5.1, or Annex VI Section 5.1?	N/A
9.	If applicable, does the device(s) meet all requirements to eliminate or reduce as far as possible infection and material contamination as stated in Essential Requirement Section 8?	N/A
10.	Does the manufacturer institute and keep up to date a systematic procedure to review experience gained from devices in the post production phase and to implement the necessary means to apply required corrective action?	N/A
11.	Does the manufacturer have a systematic procedure or process in place for notification of competent authorities upon the discovery of a malfunction or deterioration in the characteristics and/or performance of a device(s), inadequacy in the labeling or instructions for use which might lead to or might have led to the death or serious injury of a patient or user?	N/A

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
MDD 93/42/EEC Requirements		
#	Audit Question / Item	Evidence / Comments
12.	Does the manufacturer have a systematic procedure or process in place for notification of competent authorities upon the discovery of a technical or medical reason connected with the characteristic or performance of the device(s) for the reasons referred to in question 11 above that has led to a systematic recall of the device(s)?	N/A
13.	Has the registration of persons responsible for placing the device(s) on the market been performed in accordance with Article 14 of the directive?	N/A
14.	Does the labeling of the device(s) contain the authorized EU representative established within the community?	N/A
15.	Has the manufacturer of the device(s) informed the notified body which approved the quality system of any plan for substantial changes to the quality system or the product-range covered? Has the notified body assessed the proposed changes and verified whether the quality system still meets the requirements referred to in Annex II/V/VI Section 3.4? Has the manufacturer received a decision on the proposed changes from the notified body?	N/A
16.	Does the information supplied (e.g. labeling, IFU, etc.) by the manufacturer for the device(s) meet the requirements as specified in Essential Requirements 13 of the directive? Does the manufacturer have procedures identifying the requirements of labeling and instructions for use as defined in MDD, Annex I Section 13, and is regulations for CE marking included in these procedures?	N/A

	DOCUMENT TITLE: Internal Quality Audit Checklist	Revision: A
	DOCUMENT #: FM2-QM.SLQ017	Effective Date: 25Sep2020

Audit #: <u>22-01</u>		
MDD 93/42/EEC Requirements		
#	Audit Question / Item	Evidence / Comments
17.	Are language requirements defined in procedures for the information identified in MDD, Annex I Section 13 for the applicable markets?	N/A
18.	For measuring devices, verify that metrological requirements are complied with.	N/A