

Pathway Medtech, LLC.		Audit Response			Audit Date:	11/28/2023 - 11/29/2023	
Audit ID Reference #	AUD-23-012		Auditor Name/Company:		SILQ		

Below is a summary of the audit findings and/or recommendation for improvement for which Pathway MedTech LLC will address.

Major Findings

Finding Reference ID #	Audit Severity Level (per QP-006)	Description of Finding	Applicable QP/ Regulation/ Standard/ Product Document (s)	Revision	Owner	Target Completion Date	Status (In process/ Completed)	Corrective Action (Reference CAPA/SCAR)
N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A

Minor Findings

Finding Reference ID #	Audit Severity Level (per QP-006)	Description of Finding	Applicable QP/ Regulation/ Standard/ Product Document (s)	Revision	Owner	Target Completion Date	Status (In process/ Completed)	Corrective Action (Reference ECO/NCR)
#1	Severity Level 3 (Minor)	Per Section 8.3.4 of Training procedure (QP-017 Rev E), "Retraining shall take place yearly based on previous training date per procedure." Based on review of employee training records, this requirement has not been fulfilled There was no evidence of annual training (i.e., within the last 12 months) for Sean Tamayo for document #'s QM-01, QP-008, QP-017, QP-019, QP-026, and MLP-SLQ-0001. There was no evidence of annual training for Aaron Rogers for document #'s QM-01, QP-001, QP-008, QP-017, and QP-023. There was no evidence of annual training for Larry Dacoron for document #'s QM-01, QP-001., QP-005, QP-008, QP-009, and QP-023. There is no evidence of annual training for James Romero for QM-01.	21 CFR 820.25 Personnel (b) ISO 13485:2016 cl. 6.2	2016	Jacqueline Kim	Q12026	In-Process	CAPA-24-001

Pathway Medtech, LLC.		Audit Response				Audit Date:	11/28/2023 - 11/29/2023	
Audit ID Reference #	AUD-23-012		Auditor Name/Company:		SILQ			
Finding Reference ID #	Audit Severity Level (per QP-006)	Description of Finding	Applicable QP/ Regulation/ Standard/ Product Document (s)	Revision	Owner	Target Completion Date	Status (In process/ Completed)	Corrective Action (Reference ECO/NCR
#2	Severity Level 3 (Minor)	The calibration frequency for pin gauge set (ES-0047-01) and temperature gauge (ES-0119-03) do not comply with requirements outlined in QP-020 Table 7.1.5. There is no documentation or impact analysis to justify change in calibration frequency During review of calibration records for pin gauge set (ES-0047-01) and temperature gauge (ES-0119-03), the calibration frequency does not align with requirements in Calibration procedure (OP-020 Rev D Table 7.1.5).	21 CFR 820.70(g)(1)(g)(3) 21 CFR 820.72 (a) ISO 13485:2016 cl.7.5	2016	Jacqueline Kim	Q12026	In-Process	CAPA-23-0008
#3	Severity Level 3 (Minor)	Failed pressurization test for cleanroom to gowning room. Pressure recorded was 0.018. Acceptance criteria is 0.02.(Note: Report references acceptance criteria from QP-018 Rev H, however, the current revision is Rev P.) CERTS cleanroom and gowning room (Suite 104) monitoring report dated 7/31/23.Clean room to gowning room pressurization. No evidence of internal Pathway review of results. No evidence of corrective action with regard to failure in pressurization test.	21 CFR 820.70(c), (f) ISO 13485:2016 cl. 6.3, 6.4, 6.4.1	2016	Larry Dacoron	04/01/2024	Completed	CAPA-24-003
#4	Severity Level 3 (Minor)	During review of inspection records for SLQ-5004, the inspection requirements were not fulfilled. a) No evidence of a packing slip with acceptance label per Section 7.8.3 of QP-009 Rev E. b) No evidence of shipment inspection per Section 7.4 of QP-009 Rev E.	21 CFR 820.50 (a)(2) 21 CFR 820.80 (a)(b) ISO 13485:2016 cl.7.4.3,8.4	2016	Jacqueline Kim	Q12026	In-Process	CAPA-24-005

Pathway Medtech, LLC.		Audit Response				Audit Date:		11/28/2023 - 11/29/2023	
Audit ID Reference #	AUD-23-012			Auditor Name/Company:		SILQ			

Finding Reference ID #	Audit Severity Level (per QP-006)	Description of Finding	Applicable QP/ Regulation/ Standard/ Product Document (s)	Revision	Owner	Target Completion Date	Status (In process/ Completed)	Corrective Action (Reference ECO/NCR
		The receiving inspection record for label stockSLQ-5004(order placed on 1/24/22) did not include documented evidence of a packing slip or shipment inspection as required per Section 7.8.3 and Section 7.4 of the Receiving Inspection procedure (OP-009 Rev E), respectively						

Recommendations/Opportunities for Improvement

Recommendation Reference ID #	Description of Recommendation	Pathway Response	Owner	Target Completion Date	Status (In process/ Completed)	Corrective Action (Reference #)
#1	In review of the last two management review meeting minutes (12/21/22 and 6/9/23), several opportunities for improvement were identified: a) It is not clear from reviewing the meeting minutes what time period is being evaluated. Document the evaluation period in the Management Review Meeting Minutes form (FM-0063 Rev A). b) Record official job title and department on the attendance sheet (job titles should align with the organization chart). c) Ensure all topics reviewed in management review have their own section in the meeting minutes form (e.g., Review of Quality Objectives).	Pathway agrees with the suggestion to update the management review meeting minutes form (FM-0063) for parts a), b) and c).	Aaron Rogers	01/20/2025	Completed	ECO-001763
	d) Only the supplier performance of Class I suppliers was reviewed in the management review. It is recommended to specify which suppliers are required to be reviewed for the supplier performance metric in QP-014, not just the management review slides. ISO 13485:2016 cl.5.1(d), 5.6; 21 CFR 820.20 (c)	The requirements for supplier performance monitoring is defined in QP-005 §6.8.6. At minimum, Category 1 suppliers are monitored in Management Review. However, Pathway agrees to consolidate monitoring requirements into QP-014.	Aaron Rogers	03/31/2025	In Process	ECO-002027

Document Name: AUD-23-012_Audit Response Report (1).pdf

Transaction ID: CBJCHBCAABaPySaCcyGWE3mszeYeqxrHfKTHTVc9UHF

Pathway Medtech, LLC.		Audit Response			Audit Date:	11/28/2023 - 11/29/2023
Audit ID Reference #	AUD-23-012		Auditor Name/Company:	SILQ		

Recommendation Reference ID #	Description of Recommendation	Pathway Response	Owner	Target Completion Date	Status (In process/ Completed)	Corrective Action (Reference #)
#2	Clearly identify supplier statuses in the Supplier Qualification procedure (QP-005 Rev B). Consider adding a "conditional" status for suppliers that are undergoing supplier evaluation and may be used conditionally. Example: Erica Livingston performed internal audit from 5/11–5/12 but was not approved until 5/18 ISO 13485:2016 cl.7.4.1;21 CFR 820.50(a)(3)	Pathway agrees with the recommendation to revise the procedure.	Aaron Rogers	03/29/2024	Completed	ECO-001634
#2	Several GDP issues were noted during review of the DHR: a) Item 7 in Work Order form not checked (Lot #SLQ-81000406231) b) Not all corrections are initialed and dated (Lot #SLQ-81000406231 and Lot#SLQ-81041201231) c) "N/A" not applied per requirements outlined in Quality Record SOP QP-003 (Lot #SLQ-81000406231). d) Some BOM items do not indicate type—document or material (Lot #SLQ-81000406231 and Lot#SLQ-81041201231). ISO 13485:2016 cl. 8.3;21 CFR 820.70	Pathway agrees to update the impacted DHRs.	Jacqueline Kim	03/14/2025	In Process	N/A
#4	Re-execute the OQ and PQ (TP-HDX-0005 and TP-HDX-0006) with established CV and rinse test procedures and acceptance criteria ISO 13485:2016 cl. 7.5.6;21 CFR 820.70 (i)	TR-SLQ-0013 and TR-SLQ-0014, OQ and PQ for the 14, 16, and 18 Fr catheters, were performed.	Will Hail	05/16/2024	Completed	ECO-001685
#5	Record equipment ID and calibration due date of Magnehelic gauge when recording daily monitoring values for positive pressure ISO 13485:2016 cl. 6.3, 6.4; 21 CFR 820.70 (c),(f), (g)	Pathway to update the form (FM-0030) to identify equipment ID and calibration due date.	Larry Dacoron	04/19/2024	Completed	ECO-001660
#6	It is unclear what part revision is being inspected. Include recording of part revision to receiving inspection form ISO 13485:2016 cl. 7.4.3, 8.4; 21 CFR 820.50	Receiving Inspection Record FM-0088 includes a section for part revision. No further actions needed on this recommendation.	N/A	N/A	N/A	N/A

Pathway Medtech, LLC.		Audit Response		Audit Date:	11/28/2023 - 11/29/2023
Audit ID Reference #	AUD-23-012	Auditor Name/Company:	SILQ		

Recommendation Reference ID #	Description of Recommendation	Pathway Response	Owner	Target Completion Date	Status (In process/ Completed)	Corrective Action (Reference #)
	(a)(2), 820.80 (b)					
#7	Ensure CAPA forms are completed with up-to-date, accurate data and activities. Ensure CAPA data captured in multiple sources are accurate ISO 13485:2016 cl.8.4, 8.5;21 CFR 820.100 (a)(1)	Pathway agrees with the recommendation to revise the CAPA procedure and form accordingly. Pathway will maintain all records in Arena.	Aaron Rogers	02/13/2025	Complete	ECO002021
#8	Ensure CAPA documentation is completed in a timely, accurate manner. Ensure actions taken were documented, signed/initialed, and dated ISO 13485:2016 cl. 8.2.5, 8.5.2, 8.5.3; 21 CFR 820.100 (a)(3),(a)(4), (a)(5), (b)	Pathway agrees with the recommendation to revise the CAPA protocol and form accordingly.	Aaron Rogers	02/13/2025	Complete	ECO002021

Pathway Medtech, LLC.	Audit Response		Audit Date:	11/28/2023 - 11/29/2023
Audit ID Reference #	AUD-23-012	Auditor Name/Company:	SILQ	

Completed By: **Jacqueline Kim**


Jacqueline Kim (10-Mar-2025 10:21 PDT)

Print Name

Signature

Date

Reviewed By: **Aaron Rogers**


Aaron Rogers (10-Mar-2025 12:07 PDT)

Print Name

Signature

Date

AUD-23-012_Audit Response Report (1)

Final Audit Report

2025-03-10

Created:	2025-03-10 (Pacific Daylight Time)
By:	Jacqueline Kim (jkim@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAApySaCCyGWE3mszeyeqxrHfKTHTVc9UHF
Number of Documents:	1
Document page count:	6
Number of supporting files:	0
Supporting files page count:	0

"AUD-23-012_Audit Response Report (1)" History

-  Document created by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-03-10 - 10:20:22 AM PDT
-  Document emailed to Aaron Rogers (arogers@pathwaymedtech.com) for approval
2025-03-10 - 10:21:04 AM PDT
-  Document emailed to Jacqueline Kim (jkim@pathwaymedtech.com) for approval
2025-03-10 - 10:21:04 AM PDT
-  Agreement viewed by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-03-10 - 10:21:05 AM PDT
-  Jacqueline Kim (jkim@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC
2025-03-10 - 10:21:13 AM PDT
-  Document approved by Jacqueline Kim (jkim@pathwaymedtech.com)
Approval Date: 2025-03-10 - 10:21:13 AM PDT - Time Source: server
-  Email viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-03-10 - 12:06:09 PM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email
2025-03-10 - 12:06:39 PM PDT

 Agreement viewed by Aaron Rogers (arogers@pathwaymedtech.com)

2025-03-10 - 12:06:40 PM PDT

 Aaron Rogers (arogers@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC

2025-03-10 - 12:07:01 PM PDT

 Document approved by Aaron Rogers (arogers@pathwaymedtech.com)

Approval Date: 2025-03-10 - 12:07:01 PM PDT - Time Source: server

 Agreement completed.

2025-03-10 - 12:07:01 PM PDT