

PHASE 4 – CORRECTIVE AND PREVENTATIVE ACTION RESULT(S) & VERIFICATION OF EFFECTIVENESS (VoE) PLAN(S)	
CAPA Result(s): (Attach any supporting documents. If needed, complete the SCAR Form, FM-0087)	Preventative Action - Update IQ form: FM-0055 Updated and released via ECO-002169 - Create WI for Equipment IQ QP-012 updated and released via ECO-002171 - Update Equipment log to include production or engineering use FM-0051 updated and released via ECO-002221 Form created to populate the equipment log for check-in of new equipment Form created to request changes such as removal/edits to the equipment from the equipment log*
	Corrective Actions - Label all unlabeled equipment - CAL/PM/ES/Eng Only All unlabeled equipment items were either removed from the building or labeled accordingly. All unlabeled equipment items marked with “Engineering Only” intended to be used for production will require relabeling, Arena release, and IQ. Equipment items without PM/CAL labeling were relabeled to include NO CAL and/or NO PM as needed. PM labeling was added to all equipment with >1month PM cadence - Perform IQ on all equipment (as needed) All MP, TP, WI, etc. for active clients in production was reviewed for equipment. The equipment items were evaluated for IQ released in Arena. Approximately 125 equipment items did not have IQ. Equipment was evaluated for IQ requirements, and XXX items were determined to need IQ. IQ was performed and released in Arena under ECO-XXXXX.* QP-012 was updated to include instructions on determining when IQ is required and released via ECO-002288. ES items in Arena for all active production equipment moved to “Production/DV” in equipment log - Add all unlabeled/missed equipment to log Equipment found with numbers not in log added to log. Unlabeled equipment was labeled with “Engineering Only” and moved off-site. Equipment items found to be numbered, but not in log were added to log. - Remove lost equipment from service (Retired) Equipment unable to be located marked as “Retired” in log

CAPA Results(s) Completion Date:	tbd		
Deviations: (Add any deviations with justifications)	N/A		
Verification of Effectiveness (VoE) Plan(s): (Describe any effectiveness plans)	tbd		
Phase 5 Target Due Date:	tbd		
PHASE 4 - APPROVAL SIGNATURES			
CAPA Owner (Name/Title):	Amber O'Brien	Signature/Date:	tbd
Executive Management (Name/Title):	Aaron Rogers	Signature/Date:	tbd
Quality Assurance (Name/Title):	Larry Dacoron	Signature/Date:	tbd
CAPA EXTENSION REQUEST (☐ N/A if not applicable)			
CAPA Extension Due Date:	11/14/2025		
Justification for Extension:	*The Engineering team started IQ activities for all incoming equipment for production/DV activities since June 2025. This increase to their scope reduced the Engineering team's ability to produce IQ for past equipment. Review of documentation for active customers to create a complete list of production/DV equipment reduced the overall scope of the IQ requirements from >400 to <100. However, equipment called out in documentation is used for DV/Production activities was not assessed. There was also a change in scope to add forms to the equipment log. The form to edit/remove items cannot be generated with the current Pathway system. Additional time is requested to allow for the following: 1. The Engineering team to complete IQ on all documented equipment items; 2. The Engineering team to review the equipment specified in documents without numbers and update work instruction/procedures/plans accordingly; 3. The Engineering team to review complete equipment list (500 items) to identify any outliers from the initial gap assessment; 4. The Quality team to create a process for removing/updating equipment		
CAPA EXTENSION APPROVAL SIGNATURES (☐ N/A if not applicable)			
CAPA Owner (Name/Title):	Amber O'Brien	Signature/Date:	 Amber O'Brien (29-Sep-2025 08:53:19 PDT)
Quality Assurance (Name/Title):	Larry Dacoron	Signature/Date:	 Larry Dacoron (02-Oct-2025 13:56:55 PDT)

CAPA-23-0008 Phase 4 Extension

Final Audit Report

2025-10-02

Created:	2025-09-25 (Pacific Daylight Time)
By:	Jacqueline Kim (jkim@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAAQZWk2wNMRXJZJ0QfsohYF3WC-Xh5YdAl
Number of Documents:	1
Document page count:	2
Number of supporting files:	0
Supporting files page count:	0

"CAPA-23-0008 Phase 4 Extension" History

-  Document created by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-09-25 - 5:07:55 PM PDT
-  Document emailed to Amber O'Brien (aobrien@pathwaymedtech.com) for approval
2025-09-25 - 5:08:17 PM PDT
-  Document emailed to Larry Dacoron (ldacoron@pathwaymedtech.com) for approval
2025-09-25 - 5:08:17 PM PDT
-  Agreement viewed by Amber O'Brien (aobrien@pathwaymedtech.com)
2025-09-29 - 8:50:41 AM PDT
-  Amber O'Brien (aobrien@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email
2025-09-29 - 8:51:12 AM PDT
-  Amber O'Brien (aobrien@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC
2025-09-29 - 8:53:19 AM PDT
-  Document approved by Amber O'Brien (aobrien@pathwaymedtech.com)
Approval Date: 2025-09-29 - 8:53:19 AM PDT - Time Source: server
-  Agreement viewed by Larry Dacoron (ldacoron@pathwaymedtech.com)
2025-09-29 - 10:06:09 AM PDT

- ✓ Larry Dacoron (ldacoron@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email

2025-10-02 - 1:56:00 PM PDT

- ✓ Larry Dacoron (ldacoron@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC

2025-10-02 - 1:56:55 PM PDT

- ✓ Document approved by Larry Dacoron (ldacoron@pathwaymedtech.com)

Approval Date: 2025-10-02 - 1:56:55 PM PDT - Time Source: server

- ✓ Agreement completed.

2025-10-02 - 1:56:55 PM PDT