

PHASE 4 – CORRECTIVE AND PREVENTATIVE ACTION RESULT(S) & VERIFICATION OF EFFECTIVENESS (VoE) PLAN(S)

CAPA Result(s): (Attach any supporting documents.
If needed, complete the SCAR Form, FM-0087)

1. Train all employees to missing protocols
Training files were reviewed, and a corrective training file was created. Refer to Training files.
2. Add training file review to line clearance (in addition to current FM-0080)
 - a. Ensure QA reviewer is included in FM-0080
FM-0080 updated and released via ECO-002170
3. Update FM-0080, FM-0004, and FM-0005 to include the following note: "Trainer must countersign each DHR that the trainee signs. The trainer is responsible for ensuring compliance to the manufacturing instructions for all materials produced by the trainee."
QP-017 updated to include verbiage released via ECO-002170
4. Add FM-0080 details to QP-017 Training and WO WI
QP-017 updated to include verbiage released via ECO-002170
5. Add training levels to training QP-017 for WI/MP/TP
 - a. Specify that training levels above "read and understand" require a trainer
 - b. Update FM-0004 & FM-0005 to include training level requirements
QP-017 updated to include verbiage released via ECO-002170
6. Add instructions on forms FM-0004 & 0005 in QP-017
QP-017 updated to include verbiage released via ECO-002170
7. Update FM-0080 to specify released documents only
FM-0080 updated and released via ECO-002170
8. Add expiration for retraining at 18mo for appropriate WI/MP/TPs
QP-017 updated to include verbiage released via ECO-002170
9. Add controlled GMP/GDP training module to Arena
FD-0043 - GMP Video Training released via ECO-002170
10. Add client training requirements to QP-017
 - a. Review client training record options with Operations
 - b. Add a Cleanroom WI training requirement for all multi-use guests
 - c. Ensure users are trained on any material they sign
 - d. Make a note that limited use guests do not require training, as the host, who is trained, is responsible for oversight of their guest
QP-017 updated to include verbiage released via ECO-002170

	11. Add monthly training review to QP-017 for assemblers and new clients a. Update the Document Control QP for review requirements (if applicable) QP-017 updated to include verbiage released via ECO-002170		
CAPA Results(s) Completion Date:	07/11/2025		
Deviations: (Add any deviations with justifications)	N/A		
Verification of Effectiveness (VoE) Plan(s): (Describe any effectiveness plans)	Next 6 DHRs initiated after July 11 th , 2025, show no issues with training.		
Phase 5 Target Due Date:	12 Dec 2025		
PHASE 4 - APPROVAL SIGNATURES			
CAPA Owner (Name/Title):	Sean Tamayo / Pathway Manufacturing	Signature/Date:	 Sean Tamayo (10-Jul-2025 13:19 PDT)
Executive Management (Name/Title):	Aaron Rogers / Pathway Director of Regulatory and Quality	Signature/Date:	 Aaron Rogers (10-Jul-2025 12:53 PDT)
Quality Assurance (Name/Title):	Larry Dacoron / Pathway Senior Quality Engineer	Signature/Date:	 Larry Dacoron (15-Jul-2025 14:11 PDT)
CAPA EXTENSION REQUEST (☒ N/A if not applicable)			
CAPA Extension Due Date:	N/A		
Justification for Extension:	N/A		
CAPA EXTENSION APPROVAL SIGNATURES (☒ N/A if not applicable)			
CAPA Owner (Name/Title):	N/A	Signature/Date:	N/A
Quality Assurance (Name/Title):	N/A	Signature/Date:	N/A

CAPA-24-001 Phase 4

Final Audit Report

2025-07-15

Created:	2025-07-10 (Pacific Daylight Time)
By:	Jacqueline Kim (jkim@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAAPsCXpNBvS1mgEtWxZ7aCkzrwigUDt7pJ
Number of Documents:	1
Document page count:	2
Number of supporting files:	0
Supporting files page count:	0

"CAPA-24-001 Phase 4" History

-  Document created by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-07-10 - 12:05:16 PM PDT
-  Document emailed to Larry Dacoron (ldacoron@pathwaymedtech.com) for approval
2025-07-10 - 12:06:23 PM PDT
-  Document emailed to Aaron Rogers (arogers@pathwaymedtech.com) for approval
2025-07-10 - 12:06:23 PM PDT
-  Document emailed to Sean Tamayo (stamayo@pathwaymedtech.com) for approval
2025-07-10 - 12:06:23 PM PDT
-  Email viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-07-10 - 12:50:32 PM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email
2025-07-10 - 12:51:10 PM PDT
-  Agreement viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-07-10 - 12:51:10 PM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC
2025-07-10 - 12:53:35 PM PDT



Document approved by Aaron Rogers (arogers@pathwaymedtech.com)

Approval Date: 2025-07-10 - 12:53:35 PM PDT - Time Source: server



Email viewed by Sean Tamayo (stamayo@pathwaymedtech.com)

2025-07-10 - 1:18:19 PM PDT



Sean Tamayo (stamayo@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email

2025-07-10 - 1:18:53 PM PDT



Agreement viewed by Sean Tamayo (stamayo@pathwaymedtech.com)

2025-07-10 - 1:18:55 PM PDT



Sean Tamayo (stamayo@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC

2025-07-10 - 1:19:06 PM PDT



Document approved by Sean Tamayo (stamayo@pathwaymedtech.com)

Approval Date: 2025-07-10 - 1:19:06 PM PDT - Time Source: server



Larry Dacoron (ldacoron@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email

2025-07-15 - 2:10:03 PM PDT



Agreement viewed by Larry Dacoron (ldacoron@pathwaymedtech.com)

2025-07-15 - 2:10:04 PM PDT



Larry Dacoron (ldacoron@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC

2025-07-15 - 2:11:19 PM PDT



Document approved by Larry Dacoron (ldacoron@pathwaymedtech.com)

Approval Date: 2025-07-15 - 2:11:19 PM PDT - Time Source: server



Agreement completed.

2025-07-15 - 2:11:19 PM PDT