

Pathway Medtech, LLC.		CAPA Form		24-003	
INITIATION					
Initiated By:		Melissa Garcia		Date: 01/24/2024	
Source of Nonconformity:		SILQ Audit			
Description of Issue / Problem Statement:		Finding No.: Finding #3 Finding: Failed pressurization test for cleanroom to gowning room. Pressure recorded was 0.018. Acceptance criteria is 0.02. (Note: Report references acceptance criteria from QP-018 Rev H, however, the current revision is Rev P.			
INVESTIGATION					
Assigned To:		Larry Dacoron		Due Date: 02/16/2024	
Extension Rationale:		If applicable. Add for each extension. N/A			
Methodology:		5 Whys See Attached: ☐ Yes ☒ No			
Investigation Results:		Per QP-030 Rev D, Section 6.4 Rounding rule will apply for values extending beyond the significant digit of the specification. On a measurement scale of zero (0) to nine (9), values less than or equal to four (4) will be rounded down, while values greater than or equal to five (5) will be rounded up. Based on Pathway's procedure this does not fail the pressurization test. The pressure that was recorded was 0.018 and when applying the rounding rule the value would be 0.02 which is equivalent to the acceptance criteria.			
Root Cause(s):		There is no root cause for the failed pressurization test. After review of the Statistical Technique SOP, the value is rounded up and it meets the specification.			
Risk Analysis Review:		No risk analysis required as there is no failed pressurization test.			
CAPA Plan Actions:		Action		Responsible	
		Provide newest version of QP-018 to CERTS to address the note in their Finding #3.		Larry	
Completed By:		Larry Dacoron		Date: 02/13/2024 02/13/2024	
Quality Approval:		Aaron Rogers		Date: 02/13/2024 02/13/2024	
IMPLEMENTATION					
Assigned To:		Larry Dacoron		Due Date: 02/16/2024	
Actions Taken:		Emailed CERTS with the current revision P for QP-018			
Completed By:		Larry Dacoron		Date: 02/13/2024 02/13/2024	
Evidence of Completion:		See attached email dated 02/13/2024 to CERTS			
Verified By:		Larry Dacoron		Date: 02/13/2024 02/13/2024	
Quality Approval:		Aaron Rogers		Date: 02/13/2024 02/13/2024	
This document contains confidential, proprietary information of Pathway Medtech and shall not be copied, reproduced, or otherwise disclosed to any third party without prior written permission there from. If any portion of this document is in hardcopy form, it must be considered and treated as uncontrolled "For Reference Only".					
Page 1 of 2					

EFFECTIVENESS CHECK			
Assigned To:	Larry Dacoron	Due Date:	06/28/2024
Methodology:	Review 3 reports from CERTS to verify they have the current revision, P for QP-018 identified in their reports.		
Effectiveness Check Outcome:	Reviewed 3 test reports approved on 03/01/2024 for January testing and they all reference revision P for QP-018	Effective:	☒ Yes ☐ No
	1. Microbial Testing Suite 103 January 2024		
	2. Microbial Testing Lower-Level Cleanroom January 2024		
	3. Microbial Testing Upper-Level Cleanroom January 2024		
Completed By:	Larry Dacoron	Date:	04/01/2024 04/01/2024
Quality Approval:	Aaron Rogers	Date:	04/01/2024 04/01/2024

CAPA CLOSURE	
Quality Approval:	Aaron Rogers
	Date: 04/01/2024