

FDA QSR & ISO 13485:2016 QMS Internal Audit Checklist

Audit #: 23-002	Legend: NC = Non-Conformance OFI = Opportunity for Improvement A = Acceptable NR = Not reviewed NA = Not applicable
Dates: November 28 – 29, 2023	
Lead Auditor: Bryan Wong, QA Director (Silq)	

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
Management Controls (main subsystem)						
1	Ensure Quality Manual defines scope of QMS, procedures (or reference to) within QMS, and description of the interaction of processes within QMS	ISO 13485:2016: 4.1, 4.2.2	Review quality manual; review QMS metrics; review critical processes and procedures	Quality Manual defines scope of QMS, identifies the QMS processes, and outlines the interaction of processes.	QM-01 Rev E (Quality Manual) QP-014 Rev D (Analysis of QMS Data) contains quality objectives (i.e., QMS metrics) in Section 7.1.	A
2	Verify criteria and methods are in place to monitor and control processes for effectiveness	ISO 13485:2016: 4.1.3(a), 4.2.1(d), 8.4	Review QMS metrics; review management reviews	Methods and processes used to monitor and control processes/QMS for effectiveness outlined in Management Review SOP. Criteria defined in quality objectives.	QP-001 Rev G (Management Review) QP-014 Rev D (Analysis of QMS Data)	A
3	Verify firm has established and conducts Management Reviews, at least annually	ISO 13485:2016: 5.1(d), 5.6; 21 CFR 820.5, 820.20(c)	Request procedure in advance; review management reviews	Last management review held on 6/9/2023. Evaluation period: Dec '22 – May '23. Meeting minutes documented on FM-0063 Rev A. Previous management review held on 12/21/2022. Evaluation period: May '22 – Nov '22 Opportunity for Improvement: 1) Document the evaluation period in the Management Review Meeting Minutes form (FM-0063 Rev A). 2) Record official job title and department on the attendance sheet (should line up with the organization chart). 3) Ensure all topics reviewed in management review have their own section in the meeting minutes form (e.g., Review of Quality Objectives). 4) Only the supplier performance of Class I suppliers was reviewed in the management review. It is recommended to specify which suppliers are required to be reviewed for the supplier performance metric in QP-014.	QP-001 Rev G (Management Review) Management Review Meeting Minutes form Review date: 6/9/2023 Management Review Meeting Minutes form Review date: 12/21/2022	OFI
4	Confirm management reviews examine suitability and effectiveness of quality systems, improvements needed because of customer requirements, and resource needs	ISO 13485:2016: 4.1.3(c), 5.6.1, 5.6.3, 6.1, 8.4; 21 CFR 820.20(c)	Review procedure	Management Review SOP includes decisions and actions pertaining to: - Improvements needed to maintain the suitability, adequacy and effectiveness of the QMS and its processes	QP-001 Rev G (Management Review)	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
				- Improvement to product related to customer requirements - Changes needed to respond to new or revised regulatory requirements - Resource needs		
5	Ensure management review addresses audit results, customer feedback, process performance, CAPAs, previous management reviews, changes to QMS, recommendations for improvement, and new or revised regulatory requirements	ISO 13485:2016: 5.6.2	Review management reviews	Management reviews address the required inputs.	Management Review Meeting Minutes form Review date: 6/9/2023 Management Review Meeting Minutes form Review date: 12/21/2022	A
6	Verify firm has established a Quality Manual and Quality System Procedures and Instructions that are appropriate	ISO 13485:2016: 4.1.2(a), 4.2.1(b), (c); 21 CFR 820.5, 820.20(c), (d), (e), 820.22	Request quality manual and procedures in advance; review documents	A quality manual and adequate supporting quality system procedures have been established.	QM-01 Rev E (Quality Manual). Appendix 1 & 2 include list of QS procedures	A
7	Verify firm has established Quality Plan	ISO 13485:2016: 4.2.1(d), 5.4; 21 CFR 820.20(d)	Request quality plan in advance	Quality plans are in the form of quality inspection plans/forms for product	SLQ-81XX_A Inspection Record	A
8	Confirm that Quality Planning addresses QMS needs and Quality Objectives	ISO 13485:2016: 5.4.2; 21 CFR 820.20(a), (d)	Review quality plan	Quality plans address customer requirements related to product.	SLQ-81XX_A Inspection Record	A
9	Verify firm has implemented Quality Policy and Quality Objectives	ISO 13485:2016: 4.2.1(a), 5.1(b), (c), 5.3, 5.4.1 21 CFR 820.20(a), (d)	Interview employees about quality policy; review training records	Quality policy and quality objectives have been established in Pathway's quality system.	Quality policy in Quality Manual QM-01 Rev E. Quality objectives in QP-014 Rev D	A
10	Verify firm has established Quality Audit procedures and conducts audits	ISO 13485:2016: 4.2, 8.2.4; 21 CFR 820.20(c), 820.22	Request procedure in advance; review audit schedule and documents; review auditor training	Quality Audits procedure has been implemented. Audit schedule is in Smartsheet "Quality Calendar" and is maintained by Director of Quality. The last audit was performed on 5/11 – 5/12 by Erica Livingston, RAC. 3 minor NC's. CAPA's initiated to resolve minor NC's.	Quality Audits QP-006 Rev D Report for Audit #2023-May	A
11	Ensure quality audits examine compliance and effectiveness	ISO 13485:2016: 4.1.3(c), 4.2.1(d), 8.2.4; 21 CFR 820.22	Review procedure; review audit records	Objective of audit is to assess the adequacy, effectiveness, and conformance of the QMS.	QP-006 Rev D (Quality Audits) Report for Audit #2023-May	A
12	Verify that auditors are trained	ISO 13485:2016: 6.2, 8.2.4; 21 CFR 820.22	Review audit records; review training records	Internal auditor (Erica Livingston) has lead auditor training certificate from BSI, RAC certification, and has been working in medical device industry since 2006.	Supplier qualification file for Erica Livingston	A
13	Ensure that audits are conducted by objective parties	ISO 13485:2016: 8.2.4; 21 CFR 820.22	Review audit records; review training records	Internal auditor is contracted out to 3 rd party	Supplier qualification file for Erica Livingston	A
14	Confirm quality audits are linked to CAPA	ISO 13485:2016: 8.2.4; 21 CFR 820.22, 820.100	Review procedures	Section 7.1.8 of QP-006 links to CAPA SOP	QP-006 Rev D (Quality Audits)	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
15	Review organizational structure of firm; confirm resources are available to support processes	ISO 13485:2016: 4.1.3(b), 5.1(e), 5.5.1, 5.5.2, 6.1, 6.2; 21 CFR 820.20(b), 820.25	Request organizational chart(s) in advance	Adequate resources are available to support the QMS and its processes	FM-0053 Rev L (Org Chart)	A
16	Verify firm has defined a management representative with executive responsibility for implementing and reporting quality management system	ISO 13485:2016: 5.1, 5.5.1, 5.5.2, 6.1, 6.2; 21 CFR 820.20(b)(3), 820.25	Ask management representative to identify responsibility for: -changes to procedures, device designs, manufacturing processes -review of quality audit results -oversight and interaction with CAPA activities	Management representative is assigned in Org Chart: Director of Regulatory/Quality (Aaron Rogers). Responsibilities of management representative are outlined in Section 7.5 of the quality manual (QM-01 Rev E).	FM-0053 Rev L (Org Chart) QM-01 Rev E (Quality Manual)	A
17	Verify appropriate responsibilities, authority, and resources are in place for quality system activities	ISO 13485:2016: 5.1(e), 5.5.1, 5.5.2, 6.1, 6.2; 21 CFR 820.5(b)(1)-(2), 820.20(b), 820.25	Interview management representative about resource allocation	Confirmed in org chart	FM-0053 Rev L (Org Chart)	A
18	Verify firm has established procedures for identifying training needs; ensure personnel are trained to perform assigned responsibilities	ISO 13485:2016: 6.2; 21 CFR 820.25(b)	Review procedures; review training records	Training procedure has been implemented at Pathway. Training plans and records reviewed for Sean Tamayo, Aaron Rogers, Larry Dacoron, and James Romero. The training plans are appropriate for the responsibilities outlined in each personnel's job description. Non-conformance: 1) Per Section 8.3.4, "Retraining shall take place yearly based on previous training date per procedure." Based on review of the training records, this requirement has not been fulfilled.	QP-017 Rev E (Training) Reviewed Sean Tamayo's (Production Manager JD-007 Rev C) training plan TRP-000010 (in Arena) and training records for: - QM-01 Rev E (trained on 5/11/22) - QP-008 Rev C (trained on 5/11/22) - QP-017 Rev D (trained on 5/11/22) - QP-019 Rev E (trained on 5/11/22) - QP-026 Rev E (trained on 5/11/22) - MLP-SLQ-0001 Rev A (trained on 5/11/22) Reviewed Aaron Roger's (Reg. Affairs JD-002 Rev B and Quality Assurance JD-001 Rev C) training plan TRP-000003 (in Arena) and training records for: - QM-01 Rev E (trained on 5/9/22) - QP-001 Rev G (trained on 6/3/22) - QP-004 Rev D (trained on 2/17/23) - QP-008 Rev C (trained on 6/7/22) - QP-017 Rev E (trained on 6/7/22) - QP-023 Rev D (trained on 6/7/22) Reviewed Larry Dacoron's (Sr. Quality Engineer JD-016 Rev C) training plan TRP-000005 (in Arena) and training records for: - QM-01 Rev E (trained on 12/8/22) - QP-001 Rev G (trained on 5/13/22) - QP-005 Rev G (trained on 5/17/22)	NC

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
					- QP-008 Rev C (trained on 5/17/22) - QP-009 Rev E (trained on 6/13/22) - QP-023 Rev D (trained on 6/2/22) Reviewed James Romero's (JD-030 Rev A) training plan TRP-000021 (in Arena) and training records for: - QM-01 Rev E (trained on 8/5/22) - QP-008 Rev C (trained on 1/18/23) - QP-009 Rev E (trained on 1/18/23) - QP-021 Rev D (trained on 1/18/23) - MP-SLQ-0001 Rev B (trained on 1/18/23) - MP-SLQ-0003 Rev A (trained on 1/18/23)	
19	AT AUDIT CONCLUSION ... Determine if executive management ensures adequate and effective quality system is implemented. Ensure management is committed to and communicates importance of meeting customer requirements, regulatory requirements, and QMS.	ISO 13485:2016: 5.1(a), 5.2, 5.5.3	Interview executive management; provide confirmation or failures of quality system; review other subsystems and return to management controls			
Design & Development / Design Controls (main subsystem)						
1	Verify products are subject to design controls	ISO 13485:2016: 7.1, 7.3; 21 CFR 820.30(a)	Review procedure; review products	Product development is subject to Design Controls SOP (QP-007 Rev F)	QP-007 Rev F (Design Controls)	A
2	Verify design control and risk management procedures are established and applied	ISO 13485:2016: 7.3; 21 CFR 820.30(a) - (j)	Ensure procedures address all design control elements	Design Control and Risk Management procedures have been implemented at Pathway.	QP-007 Rev F (Design Controls) QP-004 Rev D (Risk Management)	A
3	Ensure design and development stages are identified; confirm that review, verification, validation, and design transfer activities at each stage are appropriate; verify responsibilities for design and development are defined	ISO 13485:2016: 7.3.2; 21 CFR 820.30	Review procedures	Appropriate design & development phases have been identified including responsibilities and design review requirements for each phase. Design transfer activities are outlined in Phase V.	QP-007 Rev F (Design Controls)	A
4	Select a design project		Selection criteria: -contains software -single product focus -risk based -result of complaints, problems -most recent -cover product range -recent 510(k), PMA, CE	Hydrophilix Foley Catheter	Project 00017	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
			mark			
5	Review the project design & development plan, responsibilities, and interfaces	ISO 13485:2016: 7.3.2; 21 CFR 820.30(b)	Review procedure; assess plan's -milestones -phases -responsibilities -risk management -interfaces	Design Development Plan (PL-0013 Rev A) identifies the product description and configuration, indication for use, goals and objectives, internal resources from key functional areas including project leader, external resources, roles & responsibilities, and schedule.	PL-0013 Rev A (Design Development Plan, Hydrophilix Foley Catheter)	A
6	Verify design & development plan is updated, reviewed, and approved	ISO 13485:2016: 7.3.2; 21 CFR 820.30(b)	Review plan revisions; review and approval procedures	Design Development plan latest revision B	PL-0013 Rev B (Design Development Plan, Hydrophilix Foley Catheter)	A
7	Confirm design input requirements were established, reviewed, and approved; ensure customer requirements are captured; ensure inputs include functional, performance, safety, and statutory and regulatory requirements	ISO 13485:2016: 7.2.1, 7.3.3; 21 CFR 820.30(c)	Review procedure; ensure requirements address -intended use -functional, performance, and safety requirements -applicable statutory and regulatory requirements -user and patient needs -other essential requirements	Design input requirements are outlined in Design Control SOP (QP-007 Rev F). Customer requirements are captured in the Product Requirement Specification and reviewed in design review (PS-0006 Rev A). Specific design inputs were reviewed in the applicable Phase II Design Review	QP-007 Rev F (Design Control) PS-0006 Rev A (Product Requirement Specification) Project 00017 Phase II Design Review	A
8	Incomplete, ambiguous, and/or conflicting requirements were addressed	ISO 13485:2016: 7.3.3; 21 CFR 820.30(c)	Review procedure; review resolutions	Section 7.5.4 addresses incomplete, ambiguous, or conflicting requirements	QP-007 Rev F (Design Control)	A
9	Confirm design & development outputs are established, verifiable, reviewed, and approved	ISO 13485:2016: 7.3.4(a), (c); 21 CFR 820.30(d)	Review procedure; review drawings, specifications, labeling, packaging, work instructions, IFUs	Design outputs are outlined in Design Control SOP (QP-007 Rev F). Specific design outputs for Project 0017 were reviewed in the applicable Phase III Design Review	QP-007 Rev F (Design Control) Project 00017 Phase III Design Review	A
10	Ensure design & development outputs are appropriate for purchasing, production, and servicing	ISO 13485:2016: 7.3.4(b); 21 CFR 820.30(d)	Review procedure	Raw material specifications were reviewed during Phase III Design Review	Project 00017 Phase III Design Review	A
11	Verify essential design & development outputs are identified	ISO 13485:2016: 7.3.4(d);	Review procedure; review drawings, specifications, labeling,	Design outputs identified in Phase III Design Review	Project 00017 Phase III Design Review	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
		21 CFR 820.30(d)	packaging, work instructions, IFUs			
12	Confirm acceptance criteria is referenced by design & development outputs and was defined prior to design verification and design validation activities	ISO 13485:2016: 7.3.4(c), 7.3.6; 21 CFR 820.30(d) & (f)	Review procedure; review drawings, specifications, labeling, packaging, work instructions, IFUs; review verification activities	Appropriate acceptance criteria are defined in the Design Outputs phase and established prior to design V&V activities	QP-007 Rev F (Design Control) Project 00017 Phase III Design Review Project 00017 Phase IV Design Review	A
13	Determine if design verification confirmed design outputs met design input requirements	ISO 13485:2016: 7.3.6; 21 CFR 820.30(f)	review procedure; Review verification activities	Confirmed through traceability matrix.	PS-0006 Rev B	A
14	Confirm design validation results prove device met predetermined user needs and intended uses	ISO 13485:2016: 7.3.7; 21 CFR 820.30(g)	Review procedure; review validation activities	Design validation studies were not scoped for Silq design project	Not applicable	NR
15	confirm design validation did not leave unresolved discrepancies	ISO 13485:2016: 7.3.7; 21 CFR 820.30(g)	Assess design and specification changes	Design validation studies were not scoped for Silq design project	Not applicable	NR
16	if required by national or regional regulations, confirm clinical evaluations and/or evaluation of device performance were performed	ISO 13485:2016: 7.3.7; 21 CFR 820.30(g)	Review procedure; review evaluation data	Design validation studies were not scoped for Silq design project	Not applicable	NR
17	If device contains software, confirm software was validated	ISO 13485:2016: 7.3.2, 7.3.7; 21 CFR 820.30(g), 820.75	Ensure software component have satisfied design, validation, and change control requirements	Not applicable to Silq devices	Not applicable	NR
18	Determine if initial production units (or equivalents) were used for design validation	ISO 13485:2016: 7.3.7; 21 CFR 820.30(g)	Review procedure; evaluate prototype / production records	Design validation studies were not scoped for design project	Not applicable	NR
19	Confirm risk management activities were performed	ISO 13485:2016: 7.1; ISO 14971:2000; 21 CFR 820.30(g)	Review procedure; review risk management file; ensure risk analysis, evaluation, and control steps are addressed	Risk management activities were performed in accordance with QP-004 Rev D. The risk management file contains the risk management plan, hazards analysis, DFMEA, PFMEA, and production/post-production report	QP-004 Rev D (Risk Management) RM-0018 Rev D (RM Plan) RM-0019 Rev D (HA) RM-0020 Rev D (DFMEA) RM-0021 Rev D (PFMEA) RM-0094 Rev A (Production & Post-production)	A
20	Confirm design changes were controlled and validated (or where appropriate, verified)	ISO 13485:2016: 7.3.2, 7.3.6, 7.3.9; 21 CFR 820.30(i), 820.70(b), 820.75(c)	Review procedure; review design changes and documentation decisions	Design changes are assessed using form T-007 Rev A (Design Change Assessment template). Changes are controlled, verified, and validated, as needed.	Change assessments and design reviews for Project 00047, Project 00049, Project 00053, and Project 00056	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
21	Confirm design changes have been reviewed for effect on components and product previously made	ISO 13485:2016: 7.3.2, 7.3.6, 7.3.9; 21 CFR 820.30(i), 820.70(b)	Review design changes and documentation decisions	Change assessments include review of effect on components and product.	Change assessments and design reviews for Project 00047, Project 00049, Project 00053, and Project 00056	A
22	Determine if design reviews were conducted at appropriate stages of design & development	ISO 13485:2016: 7.2.2, 7.3.2, 7.3.5; 21 CFR 820.30(e)	Review procedure; review design review documentation	Design reviews are conducted in accordance with Design Control SOP (QP-007 Rev F).	QP-007 Rev F (Design Control) Phase I – V design reviews (Project 00017)	A
23	Confirm design review attendees were appropriate for stage and included independent reviewer	ISO 13485:2016: 7.3.2, 7.3.5; 21 CFR 820.30(e)	Review design review documentation	Phase I – V design reviews were attended by appropriate personnel and an independent reviewer	Phase I – V design reviews (Project 00017)	A
24	Determine if design was correctly transferred to production	ISO 13485:2016: 7.3.2, 7.3.8; 21 CFR 820.30(h)	Review procedure; review DMR	Design was transferred into production and reviewed during Phase V design review	Phase V design review (Project 00017)	A
25	Ensure DHF contains design control documentation	ISO 13485:2016: 7.3.10 21 CFR 820.30(b) - (j)	Review DHF	DHF is the responsibility of the customer. Design control documentation is maintained in Arena PLM.	Arena PLM	A
Production and Process Controls (main subsystem)						
1	Verify product realization processes are planned; confirm that risk management occurs throughout product realization	ISO 13485:2016: 7.1; ISO 14971:2000 21 CFR 820.70	Review procedures	Procedures have been implemented for the planning and control of product realization. As outlined in the scope, risk management activities occur throughout a product's lifecycle.	QP-004 Rev D (Risk Management) QP-026 Rev F (Production & Process Controls) QP-019 Rev F (ID & Traceability) QP-009 Rev E Receiving & Receiving Inspection QP-037 Rev B Exposure Control Plan WI-0012 C Gowning & Prohibited Materials, Cleanroom WI-0014 Rev A Line Clearance QP-008 Rev C Nonconforming Materials	A
2	Verify planning of product realization is consistent with requirements of other processes of QMS	ISO 13485:2016: 7.1; 21 CFR 820.30, 820.50, 820.80, 820.181	Review procedures; review product realization documents	DMR is the output of the product realization planning. It includes objectives and acceptance criteria, work environment, process instructions, and in-process and final acceptance activities. Device History Records are maintained in accordance with requirements outlined in Production and Process Controls SOP (QP-026 Rev F) and Quality Records (QP-003 Rev F).	SLQ-211410SPT Rev A SLQ-211605SPT Rev A SLQ211610SPT Rev A SLQ-211810SPT Rev A QP-026 Rev F (Production and Process Controls) SP-003 Rev F (Quality Records)	A
3	Verify requirements have been defined for suppliers, contractors, and consultants; ensure suppliers, contractors, and consultants are selected on ability to meet requirements	ISO 13485:2016: 7.1, 7.4.2 21 CFR 820.50(a)	Review procedures; review supplier records	The Supplier Qualification SOP has been implemented outlining the requirements and process for qualifying suppliers and their continual monitoring.	QP-005 Rev G (Supplier Qualification)	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
4	Ensure firm maintains records of acceptable suppliers, contractors, and consultants	ISO 13485:2016: 7.4.1; 21 CFR 820.50(a)(3)	Review supplier records	Classification 1 suppliers Steris AST: Last on-site audit performed on 11/22/2022 by Aaron Rogers and Larry Dacoron. Summary report reviewed. Audit Plan 11/17/22 ISO Cert; Assignment of Responsibility Form (signed 10/29/19); Most recent supplier survey completed 11/29/22 (RM-0049 Rev B). ECO-001211 (includes the following activities: On-site Audit, supplier Survey, ISO certification, Agreements). Approved on and Effective on 12/9/22. DeRoyal Industries (sterilization): Requalified on ECO-001237 on 1/26/23. Activities completed: on-site audit (report on 1/18/23), supplier survey (completed 3/14/22), ISO cert (EN ISO 11135:2014, exp 2/25/25 from Presafe Denmark A/S) and (ISO 13485:2016 exp 2/25/25 from Presafe Denmark A/S), Agreement (sterilization services agreement effective 9/14/22 signed by Green OR, but no signature from Deroyal) Classification 2 supplier AMCOR: Requalified on ECO-001106 on 9/9/22. Activities completed: Supplier survey (Standard customer survey data sheet and company info), ISO cert (ISO 13485:2016 exp 3/16/24 by SGS) Classification 4 suppliers Erica Livingston (internal auditor): Approved on ECO-000976. Includes: RAC certificate obtained Dec 2017, BSI lead auditor training (Nov-Dec 2018), Gantus training certificate ISO 14971 dated 9/30/2018; BS - General Bio 12/14/2002); Resume (working in med device since 2006). Approved on 5/18/22 Transcat (Calibration services). Re-qualification on ECO-001099 approved on 9/1/22. Activities completed: assessment (Transcat quality survey response), quality statement, ISO certification ISO 17025:2017 exp 9/1/24 Opportunity For Improvement: 1) Clearly identify supplier statuses in the Supplier Qualification procedure (QP-005 Rev B). Consider adding a "conditional" status for suppliers that are undergoing supplier evaluation and may be used conditionally. Example: Erica Livingston performed internal audit from 5/11 - 5/12 but was not approved until 5/18.	Supplier files for Steris AST, DeRoyal Industries , AMCOR, Erica Livingston, and Transcat	OFI
5	Verify that data supporting supplier requirements is maintained; verify that suppliers, contractors, and consultants agree to notify firm of changes in products and/or services	ISO 13485:2016: 7.4; 21 CFR 820.40, 820.50(a)(3), (b)	Review records	No change agreement is included in the supplier survey	Steris AST Supplier Survey (completed 11/29/22)	A
6	Verify procedures for identifying product during all stages of receipt, production,	ISO 13485:2016: 7.5.8, 7.5.9;	Review procedures	ID & Traceability procedure is implemented that outlines the requirements for identifying product throughout the product	QP-019 Rev F ID & Traceability	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
	distribution, and installation are in place	21 CFR 820.60		realization stages.		
7	Ensure firm maintains procedures and records for traceability of each unit, lot, or batch of finished devices and components NOTE: may not be required for all devices	ISO 13485:2016: 7.5.9; 21 CFR 820.65	Review procedures; review DHRs	ID & Traceability procedure outlines the traceability requirements of raw material, in-process materials, and finished product.	QP-019 Rev F ID & Traceability	A
8	Select a process to review		Selection criteria: -CAPA indicators of process issues -process for higher risk device -degree of risk for process to cause device failures -lack of familiarity and experience with process -process used for multiple devices -variety in process technologies -processes not covered during previous inspections	Reviewed the manufacturing process flow (MP-SLQ-0003 Rev A) and surface treatment manufacturing procedure (MP-SLQ-0003 Rev B)	MP-SLQ-0003 Rev A (Manufacturing Process Flow, SILQ Products) MP-SLQ-0003 Rev B (Manufacturing Procedure – SILQ ClearTract Catheter Surface Treatment Using the Automation Machine)	A
9	Verify process is controlled and monitored	ISO 13485:2016: 7.5, 7.6, 8.2.5, 8.2.6, 8.4; 21 CFR 820.50, 820.70(a), 802.70(e), 820.70(f)-(h), 820.72, 820.75(b), 820.80	Review specific procedures, instructions, drawings, etc.; may include in-process and/or finished device acceptance activities	Process flow and manufacturing procedure are controlled in Pathway's document control system	MP-SLQ-0003 Rev A (Manufacturing Process Flow, SILQ Products) MP-SLQ-0003 Rev B (Manufacturing Procedure – SILQ ClearTract Catheter Surface Treatment Using the Automation Machine)	A
10	Verify the equipment used has been adjusted, calibrated, and maintained	ISO 13485:2016: 7.5; 21 CFR 820.70(g)(3), 820.72(a), 820.70(g)(1)	Review equipment records; review procedures	Calibration is performed in accordance with the Calibration procedure (QP-020 Rev D). Equipment ES-0012-10 Caliper: Calibrated on 1/16/23; cal due 1/16/24. By Fox Velley Metrology. ES-0012-14 Caliper: Calibrated on 2/28/23; cal due 2/28/24. By Transcat. ES-0047-01 Gage pins: Calibrated on 2/4/21; cal due 2/4/24. By	QP-020 Rev D Calibration Calibration files/certificates for ES-0012-10, ES-0012-14, ES-0047-01, ES-0099-01, and ES-0119-03.	NC

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
				Anmar/Transcat. ES-0099-01 (Pouch sealer): Calibrated on 8/9/23; cal due 8/9/24. By Transcat ES-0119-03 (temperature gauge): Calibrated on 12/8/21; cal due 12/8/23. By Control Co. (mfg cert) Non-conformance: 1) The calibration frequency for pin gauge set (ES-0047-01) and temperature gauge (ES-0119-03) do not comply with requirements outlined in QP-020 Table 7.1.5. There is no documentation or impact analysis to justify change in calibration frequency.		
11	Identify control and oversight activities; ensure control of inspection, measuring, test equipment, and calibration	ISO 13485:2016: 7.6, 8.4; 21 CFR 820.50(a)(2), 820.72	Review production, equipment, maintenance, & calibration records related to: -in-process acceptance criteria & acceptance -finished device acceptance criteria & acceptance -environmental control systems -contamination control systems	ES-0012-10 Caliper: Calibrated on 1/16/23; cal due 1/16/24. By Fox Valley Metrology. ES-0012-14 Caliper: Calibrated on 2/28/23; cal due 2/28/24. By Transcat. ES-0047-01 Gage pins: Calibrated on 2/4/21; cal due 2/4/24. By Anmar/Transcat. ES-0099-01 (Pouch sealer): Calibrated on 8/9/23; cal due 8/9/24. By Transcat ES-0119-03 (temperature gauge): Calibrated on 12/8/21; cal due 12/8/23. By Control Co. (mfg cert)	Calibration files/certificates for ES-0012-10, ES-0012-14, ES-0047-01, ES-0099-01, and ES-0119-03.	A
12	Verify firm has established procedures for production and process changes; ensure changes are verified or validated, as needed	ISO 13485:2016: 7.3.9, 7.5.6; 21 CFR 820.70(b), 820.75(c)	Review procedures	The Production & Process Controls procedure (QP-026 Rev F) outlines the requirements when production and process changes occur (Section 6.2).	QP-026 Rev F (Production & Process Controls)	A
13	Review device history record (DHR) to identify rejects and/or non-conformances	ISO 13485:2016: 8.3; 21 CFR 820.70	Review DHRs	Part # SLQ-211410SPT, Lot #SLQ-81000406231 Exp 2024-04-12; Qty released: 72 cases; Final inspection T-002 Rev A; Work Order form (FM-0039 Rev F) Part # SLQ-211610SPT, Lot #SLQ-81041201231 Exp 2026-01-12; Qty released: 750; Final Product Inspection T-002 Rev A; Work Order form (FM-0039 Rev F) Opportunity for Improvement: 1) Several GDP issues were noted during review of the DHR: - Item 7 in Work Order form not checked (Lot #SLQ-81000406231) - Not all corrections are initialed and dated (Lot #SLQ-81000406231 and Lot##SLQ-81041201231) "N/A" not applied per requirements outlined in Quality Record SOP QP-003 (Lot #SLQ-81000406231) - Some BOM items do not indicate type - document or material (Lot #SLQ-81000406231 and Lot##SLQ-81041201231)	DHR for Lot #SLQ-81000406231 DHR for Lot##SLQ-81041201231	OFI

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
14	Verify that defects, rejects, non-conformances, and removal of materials were handled properly	ISO 13485:2016: 8.3; 21 CFR 820.50, 820.70(h), 820.90, 820.100	Review material records; review DHRs; determine if -properly handled -result of equipment calibration failures -result of equipment maintenance failures -result of validation failures	Unable to determine. DHR's reviewed did not contain nonconformances or deviations.	Unable to determine.	NR
15	Ensure processes that cannot be fully verified are validated	ISO 13485:2016: 7.5.6; 21 CFR 820.75(a)	Review procedures; identify processes that cannot be verified; review validation records to ensure: -all operators have documented qualification -full change control of all processes -calibration and maintenance of all instruments -equipment is properly installed, adjusted, & maintained -predetermined product specifications are established -test sampling and plans are performed according to statistically valid rationale -process tolerance limits are challenged	Several processes that required validation were reviewed: sterilization and pouch sealing. Sterilization adoption was analyzed for 8Fr and 24Fr catheters to assure 10 ⁻⁶ SAL. Sterilization process spec PWI-7374 Rev 1 (full/routine cycle processing spec). See report TR-SLQ-0022 Rev A (Test Report, Sterilization Validation/Adoption). Original sterilization validation executed and reported in TR-0067 Rev A. Pouch seal validation executed on Equipment ID # ES-0099-001. TR-0129 Rev A IQ Checklist (FM-0055 Rev C); TR-0126 Rev A OQ TR-HDX-0004 Rev A PQ Validation passed established acceptance criteria	Sterilization validation TR-SLQ-0022 Rev A Test Report, Sterilization Validation/Adoption, Siq Cleartract. TR-0067 Rev A Sterilization Validation, EO, PSM Foley Catheter TR-0129 Rev A. IQ Checklist OQ (TR-0126 Rev A) PQ (TR-HDX-0004 Rev A)	A
16	Ensure automated or software driven processes are validated for intended uses	ISO 13485:2016: 7.5.6; 21 CFR 820.70(i)	Review validation records	Coating validation executed using Equipment # ES-0091-01 TR-0117 Rev A IQ Checklist TR-HDX-0005 Rev A OQ TR-HDX-0006 Rev A PQ Validation passed established criteria (OQ passed with deviation). Opportunity for Improvement: 1) Re-execute the OQ and PQ with established CV and rinse test	TR-0117 Rev A IQ TR-HDX-0005 Rev A OQ TR-HDX-0006 Rev A PQ	OFI

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
				procedures and acceptance criteria		
17	Verify that validations are documented and conducted by qualified personnel	ISO 13485:2016: 7.5.6; 21 CFR 820.75(b)(1)	Review procedures; review validation records	Will Hail trained on TP-HDX-0005 and TP-HDX-0006	TR-0117 Rev A IQ TR-HDX-0005 Rev A OQ TR-HDX-0006 Rev A PQ Individual Training record. Trainee: Will Hail. Date: 9/23/2021(TR-HDX-0005 Attachment 8.2 and TR-HDX-0006 Attachment 8.4.1)	A
18	Review personnel records to document personnel are trained per manufacturing processes and aware of potential defects	ISO 13485:2016: 6.2; 21 CFR 820.20(b)(2), 820.25, 820.70, 820.70(d), 820.75(b)(1)	Review personnel records	Employees are trained to manufacturing processes.	James Romero. Training Plan TRP-000021 in Arena. MP-SLQ-0001 Rev B and MP-SLQ-0003 trained on 1/18/23	A
19	Ensure that monitoring and control methods, data, date performed, individuals performing the process, and the major equipment used is documented	ISO 13485:2016: 7.1, 8.4; 21 CFR 820.75(b)(2)	Review validation records	Monitoring and control process parameters are documented in the DHR of validation run, including identification of equipment used, individuals performing the process and date(s) performed.	TR-0117 Rev A IQ TR-HDX-0005 Rev A OQ TR-HDX-0006 Rev A PQ	A
20	Determine linkages to other processes	ISO 13485:2016: 4.1, 4.2; 21 CFR 820.20, 820.25, 820.30, 820.40, 820.72, 820.90, 820.100, 820.180	Review procedures; review key processes	Linkages to other processes are identified in the DMR.	DHR for Lot #SLQ-81000406231 DHR for Lot##SLQ-81041201231	A
21	Ensure the infrastructure and work environment are appropriate and controlled	ISO 13485:2016: 6.3, 6.4; 21 CFR 820.70(c), (f), (g)	Review procedures; review records	Last cleanroom certification/qualification performed by Controlled Environmental Regulatory Testing Services (CERTS). Cleanroom and gowning room (Suite 104). Class 7. Performed on 7/31/23. Expires 7/31/24. No signatures on cert and particulate count results from testing service. Non-Conformance: 1) Failed pressurization test for Cleanroom to gown room at 0.018 (criteria = 0.02). Note: CERTS monitoring procedure references acceptance criteria from QP-018 Rev H. The current revision of QP-018 is Rev P. No evidence of internal Pathway review of results. No evidence of corrective action with regard to failure in pressurization test noted in 7/31/23 report. Microbial test results Tested on 7/31/23. Job 2307187. Viable surface and particulate were below Alert limits. No acceptance criteria established in SOP. Previous testing on: 4/28/23, 2/1/23 Daily monitoring log – positive pressure	Cleanroom certification/qualification report performed by Controlled Environmental Regulatory Testing Services (CERTS) of the cleanroom and gowning room (Suite 104). Performed on 7/31/23. Expires 7/31/24. Microbial test results. Tested on 7/31/23. Job 2307187. Previous testing on: 4/28/23, 2/1/23 Daily monitoring log – positive pressure: 11/3 – 11/29	NC OFI

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
				Reviewed log from 11/3 – 11/29. Both locations pass: Gowning room to ambient (0.04); Cleanroom to gowning (0.06). Opportunity for Improvement: 1) Record equipment ID and calibration due date of Magnehelic gauge		
22	Confirm that maintenance schedules, routine inspections, and adjustments to equipment occur	ISO 13485:2016: 6.3, 7.5.1, 7.5.6, 7.6; 21 CFR 820.70(g)	Review procedures; review records	Cleaning and monitoring of the cleanroom is outlined in Infrastructure and Environmental Controls procedure (QP-018 Rev P)	QP-018 Rev P (Infrastructure and Environmental Controls)	A
23	Verify procedures are in place for contamination control and cleanliness	ISO 13485:2016: 6.4.2, 7.5.2; 21 CFR 820.70(e)	Review procedures	Pathway has implemented procedures to control contamination and cleanliness	WI-0012 Rev C (Gowning and Prohibited Materials) QP-018 Rev P (Infrastructure and Environmental Controls)	A
24	Determine if verification of purchased products is adequate	ISO 13485:2016: 7.4.3, 8.4; 21 CFR 820.50(a)(2), 820.80(b)	Review procedures; review records	Verification of purchased material/product is performed in accordance with Receiving Inspection procedure (QP-009 Rev E). A sample of receiving inspection records were reviewed. Non-Conformance: 1) In review of inspection records for SLQ-5004, requirements for inspection were not fulfilled: a) No evidence of a packing slip with acceptance label (Section 7.8.3 of QP-009 Rev E). b) No evidence of shipment inspection (Section 7.4 of QP-009 Rev E) Opportunity for Improvement: 1) Include part revision to inspection form	QP-009 Rev E (Receiving Inspection) Receiving inspection records: P/N: SLQ-5004 Part Desc: Label stock No lot #. Qty 3. Order placed 1/24/2022 P/N: HDX-5001, No revision recorded Part Desc: AMCOR Tyvek Pouch, 3.5" x 21" Lot #: A199054. Qty 7000. Inspected 3/11/2022, P/N: SLQ-6004 Rev A, No revision recorded Part Desc: 16Fr x 10mL Foley Catheter Lot #:20220902, Qty 25,000. Inspected 11/30/2022	NC OFI
25	Ensure procedures define receiving, in-process, and final acceptance activities.	ISO 13485:2016: 7.5.11, 8.4; 21 CFR 820.80(a) - (d)	Review procedures	Pathway has implemented procedures defining receiving, in-process, and final acceptance activities.	QP-031 Rev C (In-process Inspection) QP-010 Rev F (Final Product Inspection)	A
26	Confirm receiving, in-process, and final acceptance activity records exist	ISO 13485:2016: 8.4; 21 CFR 820.80(e)	Review records	Reviewed sample of receiving inspection records and DHR's containing in-process and final acceptance activities.	Receiving inspection records: P/N: SLQ-5004 Part Desc: Label stock No lot #. Qty 3. Order placed 1/24/2022 P/N: HDX-5001, No revision recorded Part Desc: AMCOR Tyvek Pouch, 3.5" x 21" Lot #: A199054. Qty 7000. Inspected 3/11/2022, P/N: SLQ-6004 Rev A, No revision recorded Part Desc: 16Fr x 10mL Foley Catheter Lot #:20220902, Qty 25,000. Inspected 11/30/2022 Product DHR's reviewed (Lot #'s SLQ-81000406231 and	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
					SLQ-81041201231)	
27	Verify that procedures exist and that acceptance status of product is indicated	ISO 13485:2016: 7.1, 8.2.6; 21 CFR 820.86	Review acceptance criteria; review procedures; review product identification	Pathway has implemented the Identification & Traceability procedure that defines the acceptance status of raw materials, in-process materials, and finished product (QP-019 Rev F). Acceptance criteria for Silq product is defined in the inspection record	QP-019 Rev F (Identification and Traceability) SLQ-81XX Inspection Record	A
28	Ensure procedures define labeling activities, including integrity, inspection, storage, operations, and control numbers	ISO 13485:2016: 7.5.11; 21 CFR 820.120	Review procedures	Pathway has implemented the Labeling Control procedure (QP-038 Rev A) that outlines labeling activities within the organization.	QP-038 Rev A (Labeling Control)	A
29	Confirm that product packaging and shipping containers adequately protect device during processing, storage, handling, shipping, and distribution	ISO 13485:2016: 7.5.11; 21 CFR 820.130	Review procedures; review packaging and shipping containers	Packaging validation testing performed on Silq product configuration per TP-0016 and TP-0058. Packaged product underwent environmental conditioning and transportation simulation prior to sterile barrier testing.	TR-0068 Rev A (Report, Transportation Challenge, PSM Foley Catheter) TR-0074 Rev A (Report, Transportation Challenge Adoption, HDX Foley Catheter)	A
30	Verify procedures exist to prevent mix-ups, damage, deterioration, contamination, or other adverse effects to product during handling	ISO 13485:2016: 7.5.11; 21 CFR 820.140, 820.150	Review procedures	Pathway has implemented procedures to prevent mix-ups, damage, deterioration, contamination and other adverse effects to product and materials during handling.	QP-026 Rev F (Production and Process Control) QP-021 Rev D (Handling and Storage) WI-0014 Rev A (Line clearance WI)	A
31	Verify procedures exist for product distribution; confirm distribution records include name and address of consignee, identification and quantity shipped, date of shipment, and identification numbers	ISO 13485:2016: 4.2.3, 7.1, 7.5.8, 7.5.9.2, 7.5.11; 21 CFR 820.160	Review procedures; review distribution records	Pathway has implemented a procedure for product distribution (QP-016 Rev F).	QP-016 Rev F (Contract Review and Distribution)	A
32	Ensure installation and inspection procedures exist (if applicable); verify installation records are maintained	ISO 13485:2016: 7.5.3; 21 CFR 820.170	Review procedures; review installation records	Not applicable	Not applicable	NA
33	Ensure servicing procedures exist (if applicable); verify servicing records are maintained	ISO 13485:2016: 7.5.4; 21 CFR 820.200	Review procedures; review servicing records	Not applicable	Not applicable	NA
34	Verify firm identifies, verifies, protects, and safeguards customer property under its care	ISO 13485:2016: 7.5.10	Review procedures; identify customer property	Pathway has implemented the Customer Property procedure (QP-034 Rev A) that outlines handling of customer property.	QP-034 Rev A (Customer Property)	A
35	Review sterilization process procedures; verify sterilization process is validated	ISO 13485:2016: 7.5.7; 21 CFR 820.75(a), (c)	Review procedures; review validation records to ensure processes are effective in: -obtaining SAL -product performance not adversely affected -packaging not adversely affected	Pathway has implemented an EO Sterilization procedure (QP-0032 Rev F). Sterilization of Silq product has been performed and results are documented in reports TR-0067 and TR-0075. Sterilization validation testing demonstrates that a minimum of 10 ⁶ SAL has been achieved with the EO process. Transportation challenge testing on sterilized product demonstrates that packaging is not adversely affected	TR-0067 Rev A (Test Report, Sterilization Validation, EO, PSM Foley Catheter) TR-0075 Rev B (Test Report, Sterilization Product Adoption, HDX Foley Catheter) TR-0068 Rev A (Report, Transportation Challenge, PSM Foley Catheter) TR-0074 Rev A (Report, Transportation Challenge Adoption, HDX Foley Catheter)	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
Purchasing Controls						
1	Review supplier evaluation procedures	ISO 13485:2016: 7.4.1; 21 CFR 820.50	Review procedures	Pathway has implemented the Supplier Qualification procedure (QP-005 Rev G) to evaluate, approve, and monitor suppliers.	QP-005 Rev G (Supplier Qualification)	A
2	Ensure suppliers are evaluated for ability to meet specified requirements	ISO 13485:2016: 7.4.1; 21 CFR 820.50(a)(1)	Review procedures	Pathway has implemented the Supplier Qualification procedure (QP-005 Rev G) to evaluate, approve, and monitor suppliers.	QP-005 Rev G (Supplier Qualification)	A
3	Ensure adequacy of specifications of materials and/or services provided by supplier is confirmed	ISO 13485:2016: 7.4.2; 21 CFR 820.50(b)	Review procedures	Did not review	Did not review	NR
4	Confirm purchasing information identifies requirements for approval of product, procedures, processes, and equipment, requirements for personnel qualification, and QMS requirements	ISO 13485:2016: 7.4.2; 21 CFR 820.50	Review purchasing records; review procedures	Did not review	Did not review	NR
5	Verify supplier evaluation records are maintained	ISO 13485:2016: 7.4.1; 21 CFR 820.50(a)(3)	Review procedures; review supplier evaluation records	Supplier evaluation records are maintained. See Production and Process Controls subsystem Item #4	Supplier files for Steris AST, DeRoyal Industries , AMCOR, Erica Livingston, and Transcat	A
6	Determine that verification and acceptance of purchased materials and/or services is adequate	ISO 13485:2016: 7.4.3; 21 CFR 820.50(a)(2), 820.80(a), 820.80(b)	Review procedures; review acceptance records	Verification of purchased material/product is performed in accordance with Receiving Inspection procedure (QP-009 Rev E). A sample of receiving inspection records were reviewed. A non-conformance and opportunity for improvement was identified in Production and Process Controls subsystem Item #24 above.	QP-009 Rev E (Receiving Inspection) Receiving inspection records: P/N: SLQ-5004 Part Desc: Label stock No lot #. Qty 3. Order placed 1/24/2022 P/N: HDX-5001, No revision recorded Part Desc: AMCOR Tyvek Pouch, 3.5" x 21" Lot #: A199054. Qty 7000. Inspected 3/11/2022, P/N: SLQ-6004 Rev A, No revision recorded Part Desc: 16Fr x 10mL Foley Catheter Lot #:20220902, Qty 25,000. Inspected 11/30/2022	A (See Item #24 in Production & Process Controls subsystem)
Documentation and Records						
1	Review procedures for identification, storage, protection, retrieval, retention time, control, approval, distribution, disposition, and changes of documents and records	ISO 13485:2016: 4.2.4, 4.2.5; 21 CFR 820.40, 820.180	Review procedures	Control of documents and quality records are outlined in Document Control procedure (QP-002 Rev G) and Quality Records procedure (QP-003 Rev F)	QP-002 Rev G (Document Control) QP-003 Rev F (Quality Records)	A
2	Ensure documents and changes are approved prior to use	ISO 13485:2016: 4.2.4;	Review procedures; review documents and	Documents are released and training occurs before use.	MP-SLQ-0001, MP-SLQ0003	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
		21 CFR 820.40	records; review change management records			
3	Verify documents and records are legible and identifiable	ISO 13485:2016: 4.2.4(e), 4.2.5	Review documents and records	Records reviewed during this audit are legible and identifiable.	See records reviewed during this audit.	A
4	Ensure documents of external origin are identified with controlled distribution	ISO 13485:2016: 4.2.4(f)	Review external documents and records	External documents are controlled in accordance with Document Control procedure (QP-002 Rev G).	QP-002 Rev G (Document Control)	A
5	Confirm that documents and records are retained for required length of time (this includes retention of obsolete controlled documents and records)	ISO 13485:2016: 4.2.1, 4.2.4, 4.2.5 21 CFR 820.100(b), 820.180(b), 820.181, 820.184, 820.186, 820.198(a), 820.200(d)	Review procedures; review documents and records	Documents and records are retained in accordance with Quality Records procedure (QP-003 Rev F).	QP-003 Rev F (Quality Records)	A
6	Ensure change records are reviewed and approved by the same functions that performed original review and approval	ISO 13485:2016: 4.2.4, 7.3.9; 21 CFR 820.40(b)	Review procedures; review change records	Did not review	Did not review	NR
7	Verify change records include a description of change, identification of affected documents, approval signatures, approval date, and effective date	ISO 13485:2016: 7.3.9; 21 CFR 820.40(b)	Review procedures; review change records	Did not review	Did not review	NR
8	Ensure documents are available at point of use and obsolete document are not in use	ISO 13485:2016: 4.2.4(d), (h); 21 CFR 820.40(a)	Review procedures; review document distribution; review change management records	Current revision of documents are available in Arena. Obsolete revisions/documents are removed from point of use.	Various procedures accessed throughout the audit	A
9	Verify that firm maintains DMRs for each type of device	ISO 13485:2016: 4.2.1; 21 CFR 820.181	Review DHRs	Confirmed DMR exists for each variant of Silq's catheter product line	SLQ-211410SPT Rev A SLQ-211610 Rev A SLQ-211610SPT Rev A	A
10	Ensure that DMRs contain or make reference to device specifications, production process specifications, quality assurance procedures and specifications (including acceptance criteria), packaging and labeling specifications (including acceptance criteria), and installation, maintenance, and servicing procedures	ISO 13485:2016: 4.2.1; 21 CFR 820.181(a) - (e)	review DHRs	DMRs contains the specific specifications and procedures to manufacture product.	SLQ-211410SPT Rev A SLQ-211610 Rev A SLQ-211610SPT Rev A	A
11	Verify that DHRs are maintained and devices are manufactured according to DMR; ensure realization processes and	ISO 13485:2016: 7.1, 8.2.6; 21 CFR 820.184	review DHRs	Devices are manufactured in accordance with the DMR.	SLQ-211410SPT Rev A SLQ-211610 Rev A	A

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
	product meet requirements				SLQ-211610SPT Rev A	
12	Confirm that DHRs contain or make reference to dates of manufacture, quantity manufactured, quantity released for distribution, acceptance records demonstrating the device was manufactured per DMR, primary identification label and labeling used for each unit, and device identification and/or control numbers used.	ISO 13485:2016: 8.2.6; 21 CFR 820.184(a) - (f)	Review DHRs	DHR contains date of manufacture, quantity manufactured, quantity released for distribution, and copies of product labels used.	SLQ-211410SPT Rev A SLQ-211610 Rev A SLQ-211610SPT Rev A	A
13	Ensure firm maintains records for education, training, skills, and experience of resources	ISO 13485:2016: 6.2 (e)	Review training records	Records for training, skill, experience and education are maintained in Arena for each employee.	Training records for Sean Tamayo, Aaron Rogers, Larry Dacoron, and James Romero	A
14	Verify firm maintains purchasing and supplier records	ISO 13485:2016: 7.4.1, 7.4.3; 21 CFR 820.50	Review purchasing and supplier records	Records for suppliers were reviewed.	Supplier files for Steris AST, DeRoyal Industries, AMCOR, Erica Livingston, and Transcat	A
15	Ensure sterilization process parameters and records are maintained for each batch; ensure sterilization validation records are maintained	ISO 13485:2016: 7.5.5, 7.5.7	Review sterilization records	Sterilization process parameters and records are maintained for each batch.	SLQ-211410SPT Rev A SLQ-211610 Rev A SLQ-211610SPT Rev A	A
Corrective & Preventive Actions (CAPA) (main subsystem)						
1	Verify CAPA procedures comply with regulatory requirements	ISO 13485:2016: 4.1, 4.2, 8.5; 21 CFR 820.100(a)	Review procedures	Pathway has implemented a CAPA procedure that complies with regulatory requirements outlined in 21 CFR 820 and ISO 13485:2016.	QP-023 Rev D (Corrective Action Preventative Action)	A
2	Verify non-conforming product and CAPA procedures determine the need for investigation and notification	ISO 13485:2016: 8.3, 8.5; 21 CFR 820.90(a), 820.100(a)(2)	Review procedures	Nonconforming materials and CAPA procedures include the need for investigation and notifications.	QP-023 Rev D (Corrective Action Preventative Action) QP-008 Rev C (Nonconforming Materials)	A
3	Verify non-conforming product and CAPA procedures define responsibilities for review and disposition	ISO 13485:2016: 8.3, 8.5; 21 CFR 820.90(b)(1)	Review procedures	Nonconforming materials and CAPA procedures define the responsibilities for review and disposition.	QP-023 Rev D (Corrective Action Preventative Action) QP-008 Rev C (Nonconforming Materials)	A
4	Ensure that procedures for rework, retesting, and re-evaluation of nonconforming product exist and are followed	ISO 13485:2016: 8.3, 8.5; 21 CFR 820.90(b)(2)	Review procedures; review DHRs (of nonconforming products)	Nonconforming materials that are dispositioned for rework are to have rework instructions indicated within or attached to the NCR.	QP-008 Rev C (Nonconforming Materials)	A
5	Verify that appropriate records of quality problems have been created and used	ISO 13485:2016: 8.3, 8.5; 21 CFR 820.100(a)(1)	Review records of acceptance activities, production test failures, returned products, service records,	The DHR's reviewed during the audit did not have any NCR's associated with the batches. SLQ-211410SPT Rev A SLQ-211610 Rev A	Not reviewed	NR

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
			complaints	SLQ-211610SPT Rev A		
6	Determine if trend analysis data indicates quality problems; determine if data used for CAPA decisions	ISO 13485:2016: 8.1, 8.2.5, 8.4, 8.5; 21 CFR 820.100(a)(1), 820.250	Review procedures; review records of incoming products, components, testing, SPC data	Data analyzed in management review did not indicate any quality problems based on trend analysis	Management Review Meeting Minutes form Review date: 6/9/2023 Management Review Meeting Minutes form Review date: 12/21/2022	A
7	Verify CAPA data is complete, accurate, and timely; compare results across multiple data sources to identify quality problems	ISO 13485:2016: 8.4, 8.5; 21 CFR 820.100(a)(1)	Review data sources; use data tables to determine sampling plan; compare results	CAPA data recorded on CAPA form do not align with the CAPA data entered into Smartsheet (e.g., 23-0003, 23-0004). It is not always clear when corrective actions were completed or implemented (e.g., 22-0006, 22-0007, 23-0005). CAPA data and activities should be accurate, timely, and complete. Opportunity for Improvement: 1) Ensure CAPA forms are completed with up-to-date, accurate data and activities. 2) Ensure CAPA data captured in multiple sources are accurate.	CAPA #'s: 22-0006, 22-0007, 23-0003, 23-0004, 23-0005	OFI
8	Verify appropriate statistical techniques are implemented	ISO 13485:2016: 8.1, 8.2.5, 8.4; 21 CFR 820.100(a)(1), 820.250	Review procedures; review techniques used	Pathway implements statistical techniques in accordance with QP-030 Rev D.	QP-030 Rev D (Statistical Techniques)	A
9	Verify device failure investigations determine root cause	ISO 13485:2016: 8.3, 8.5; 21 CFR 820.100(a)(2)	Review procedures; review investigations	Failure investigations documented on CAPA and nonconforming material reports.	CAPA #'s: 22-0006, 22-0007, 23-0003, 23-0004, 23-0005 NCR #'s: 22-0012, 22-0020, 23-0022	A
10	Verify failure investigations are commensurate with risks	ISO 13485:2016: 8.3, 8.5; 21 CFR 820.100(a)(2), 820.90(b)	Review procedures; review investigations	The failure investigations are commensurate with risks.	CAPA #'s: 22-0006, 22-0007, 23-0003, 23-0004, 23-0005	A
11	Verify controls exist to prevent non-conforming product from being released	ISO 13485:2016: 8.3; 21 CFR 820.90(b)	Review investigations; review non-conformance records	Nonconforming materials procedure includes controls to ensure nonconforming product is released.	QP-008 Rev C (Nonconforming Materials)	A
12	Verify appropriate actions were taken for quality problems	ISO 13485:2016: 8.2.5, 8.5.2, 8.5.3; 21 CFR 820.100(a)(3), 820.100(a)(5); 820.100(a)(4), 820.100(b)	Review procedure; review CAPA records	Appropriate actions were taken for the quality problems. However, documentation of CAPA activities was not maintained effectively. Observations noted below. #22-0006: - Not clear if review of HR files was performed per action plan - No dates when actions were completed or implemented. Dates should be included with signatures/initials of when dates were closed, effectiveness check performed, etc. #22-0007:	CAPA #'s: 22-0006, 22-0007, 23-0003, 23-0004, 23-0005	OFI

Item	Subsystem / Assessment Detail	FDA / ISO reference	Auditor Notes	Auditor Observation	Objective Evidence	NC, OFI, or A?
				- No documentation on implementation or effectiveness on CAPA form - Add dates to signatures #23-0003: - Dispositions and quantities on NCR 23-0022, 22-0012, and 22-0020 were not clear. Disposition was selected as both UAI and scrap but quantities do not add up to total affected quantities. - Due dates not recorded in CAPA that line up with Smartsheet #23-0004: - Correction was established on CAPA but no corrective action. - CAPA implemented on 6/2/23 (i.e., Coc and checklist was updated and redlined) but the implementation was not documented on CAPA form #23-0005: - Document that a replacement solution was provided and will be stored per storage requirements. - No evidence of return of solution documented Opportunity for Improvement: 1) Ensure CAPA documentation is completed in a timely, accurate manner. Ensure actions taken were documented, signed/initialed, and dated.		
13	Determine CAPA actions were effective, verified, validated, documented, and implemented appropriately	ISO 13485:2016: 8.5; 21 CFR 820.100(a)(4), 820.100(a)(5), 820.100(b)	Review procedure; review CAPA records	Documentation of CAPA actions can be improved. See observations noted in #12 above. Opportunity for Improvement: 1) See OFI from #12 above.	CAPA #'s: 22-0006, 22-0007, 23-0003, 23-0004, 23-0005	OFI
14	Verify CAPAs and nonconformities were disseminated to personnel responsible for ensuring quality and prevention of problems	ISO 13485:2016: 8.3, 8.5; 21 CFR 820.100(a)(6)	Review CAPA and non-conformance records	CAPA's and nonconformities are disseminated to personnel responsible for ensuring quality. Personnel are retrained to procedures and quality issues. Pathway implemented additional forms and checklists as necessary.	CAPA #22-0006, 22-0007, 23-0005	A
15	Verify quality issues and CAPAs were disseminated for Management Review	ISO 13485:2016: 5.6.3, 8.3, 8.5; 21 CFR 820.100(a)(6), 820.100(a)(7)	Review procedure; review CAPA records	CAPA's and nonconformances are reviewed during management review meetings.	Management Review Meeting Minutes form Review date: 6/9/2023 Management Review Meeting Minutes form Review date: 12/21/2022	A
16	Verify firm has procedures for handling complaints and investigation of advisory notices / recalls; ensure provisions exist to feed in CAPA system	ISO 13485:2016: 7.2.3, 8.2.1, 8.2.2, 8.2.3; 21 CFR 820.100, 820.198	Review procedures	Pathway has implemented a procedure for handling customer complaints (QP-024 Rev F).	QP-024 Rev F (Complaint Handling)	A

Completed by:



28 Dec 2023