



Memo to File

To: Receiving Inspection file for M-SLQ.001 Rev B, Lot No: LBSEE

Subject: Attribute testing (Receiving Inspection)

Date: 01 November 2022

Prepared By: Ethan Rao

A batch of Polymeric Dispersion Solution, Unpurified (M.SLQ001 Rev B, Lot # LBSEE) was received from TCI on September 23, 2022. TCI did not perform testing to the following attributes specified on the source control specification: residue on evaporation, specific gravity, and pH. To confirm the batch of unpurified polymeric dispersion solution meets specifications, these tests were conducted at Silq using released procedure QC-C.SLQ001 and traceable equipment and supplies.

Results (Lot # LBSEE)

Attribute	Result	Acceptance criteria
Residue on Evaporation	18.44%	10% - 20%
Specific gravity	1.32	1.2 - 1.5
pH	6.0	5 - 8

These test results along with the test results documented in the supplier's COA demonstrate that Lot No. LBSEE meets the established material specifications and is accepted.

Procedure used: QC-C.SLQ001 Rev A

Equipment and Supplies Used:

- A. Residue on Evaporation test:
1. Centrifuge tube, 50 mL, Polypropylene, Flat Cap (S.SLQ002)
 2. Analytical Balance (Equipment ID ST-012, E.SLQ013)
 3. Syringe, 10mL (S.SLQ004)



B. Specific Gravity test:

1. Centrifuge tube, 50 mL, Polypropylene, Flat Cap (S.SLQ002)
2. H-B Instrument (Bel Arts), Durac Hydrometer, Catlog No. B61806-0600 (Specific gravity range 1.000 – 1.4000)

C. pH test

1. pH Strip (S.SLQ005), Expiration date: 02/01/2026

Certified by:

Ethan Rao

Ethan Rao

03 Nov 2022

Print Name

Signature

Date