

PHASE 1 – NCR PRE-INVESTIGATION				
NCR INITIATION				
NCR Owner:	Karim Natafji	Date Issued:	11/14/2025	
Part Number/Rev/Description:	SLQ-211610SPT / A / SILQ 2-way Cleartract SPT 16 Fr x 10 ml Catheter, Sterile, 10-Pack	Lot Qty Received:	655 912	
Supplier Name:	Pathway Medtech, LLC.	Lot/Serial No.:	SLQ-11202024 SLQ-05012025	
Client Name:	SILQ Technologies	Client Approval Required?	☒ Yes ☐ No	
Acceptance Criteria:	The check valve (Item no. 5) should be properly secured to the connector. The total length from sleeve to the end of the check valve should not exceed 26mm. Reference SLQ-6004 rev A.			
Description of Nonconformance:	Two lots of finished goods are suspected of containing devices where the check valve (Item no. 5) assembly fails to meet the specified acceptance criteria. Specifically, the total length from the sleeve to the end of the check valve may exceed 26mm, suggesting the check valve is not properly secured to the connector. This suspected failure mode is highly relevant as it corresponds to field reports previously received by SILQ. Over a period of three years, there have been five reported instances where the port connecting the saline syringe became unseated from the catheter. In all cases, this resulted in the inability to inflate the catheter balloon, necessitating a device replacement (no patient injury or adverse event was reported). Affected Scope: The nonconforming devices are present in Lots SLQ-11202024 and SLQ-05012025.			
NCR Source Type(s):	☐ Audits (internal & external) ☐ CAPAs ☐ Complaints ☐ Continuous Improvements ☐ SCARs ☐ Documentation Reviews ☐ Incoming Inspections ☐ In-process Inspections ☐ Other NCRs ☐ Final Inspections ☐ Suppliers ☐ Management Reviews ☒ Other: Client Request to Review Finished Goods (Contract Manufacturing)			
Source Reference #: (e.g., CAPA-24-001, Audit-24-001)	N/A			
RISK ANALYSIS				
Product Risk Assessment (to be completed by Customer) ☐ N/A				
Product Severity Rating:	☒ (1) No Patient Effect ☐ (2) Minor Injury ☐ (3) Moderate Injury ☐ (4) Serious Injury ☐ (5) Life Threatening Injury/Death/Disability			
Probability Rating:	☐ (1) Improbable ☒ (2) Remote ☐ (3) Occasional ☐ (4) Likely ☐ (5) Frequent			
Detection Rating:	☐ (1) Very High ☒ (2) High ☐ (3) Moderate ☐ (4) Low ☐ (5) Remote			
Process Risk Assessment (to be completed by Pathway) ☒ N/A				
QMS Risk Level / Severity Type:	☐ (1) Minor / Business Process Risk ☐ (2) Minor / Internal Compliance Risk ☐ (3) Minor / Document Compliance Risk ☐ (4) Major / Quality System Compliance ☐ (5) Major / Regulatory Approval or Warning Risk			
CORRECTION / CONTAINMENT				
Containment: <i>Interim containment actions taken to limit the spread of the problem.</i>	FG Material Isolation: Initiate a physical hold on all FG stock (summarized below).			
	Part Number	Lot	Qty	Status
	SLQ-211610SPT	SLQ-11202024	655	Released
			Location	106 WH

	SLQ-211610SPT	SLQ-05012025	912	Released	103 CEA
Interim Disposition:	☐ Inspect to New Revision ☐ Increase Sampling ☐ Hold ☐ Rework ☐ Use-as-Is ☒ Sort-to-Accept-As-Is ☒ Sort-to-Scrap ☐ Sort-to-Return to Supplier ☐ N/A: Documentation Only				
Initial Analysis and Rationale for Disposition:	This nonconformance was initiated due to a direct Client Request to Review Finished Goods (per NCR Source). The client's concern relates to inconsistencies in the check valve assembly. Internal reviews at SILQ previously noted variability in the placement and security of the valves across production samples. Based on the client's requirements, their Corrective and Preventive Action (CAPA) was initiated with the intent to review all catheters in inventory that were manufactured using untreated catheters obtained prior to November 2024. This expansive action, mandated by the client, necessitates an interim disposition to ensure no potentially affected product is released. The specified inspection scope and acceptance criteria for this sorting activity were provided by the client SILQ.				
Disposition Instructions:	The following instructions are to be executed by SILQ personnel within the Pathway clean room environment (CEA): Training and Qualification: Prior to inspection, all SILQ personnel performing the visual inspection and measurement (Step 3) must be formally trained and qualified on the specific acceptance criteria defined below. Training records must be documented and retained before work begins. Quarantine: Affected lots (SLQ-11202024 and SLQ-05012025) shall be brought into the clean room under quarantine for inspection. Inspection Method: The inspection will be performed visually by <client> personnel. The sterile barrier of the device must not be opened during this process Inspection Criteria (Acceptance): Each unit will be inspected to confirm two criteria: - The check valve (Item no. 5) is properly secured to the connector. - The total length from the sleeve (Item no. 6) to the end of the check valve does not exceed 26mm. Disposition of Acceptable Units: Units that meet the specified acceptance criteria (check valve secured AND length \$1e 26\$mm) will be removed from finished goods inventory, dispositioned, and transferred for use in research and development purposes (Sort-to-R&D/Scrap). Disposition of Nonconforming Units: Units that fail either of the specified acceptance criteria will be immediately dispositioned as Scrap and tagged appropriately.				
PHASE 1 FINAL DISPOSITION					
NCR Determination:	☒ NCR is Required – Proceed to Phase 2 ☐ NCR is not Required – Proceed to Phase 3				
PHASE 1 APPROVAL					
NCR Owner (Name/Title):	Karim Natafji / Director of Operations	Signature/Date:	 Karim Natafji (14-Nov-2025 13:33:21 PST)		
Quality Assurance (Name/Title):	Aaron Rogers / Director of Regulatory and Quality	Signature/Date:	 Aaron Rogers (14-Nov-2025 13:37:52 PST)		
Customer (Name/Title): ☐ N/A	Ethan Rao / Director of R&D	Signature/Date:			

NCR-25-079 SLQ-211610SPT Phase 1

Final Audit Report

2025-11-14

Created:	2025-11-14 (Pacific Standard Time)
By:	Aaron Rogers (arogers@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAACHlnBd1oY7INjcbSmBZ7nMpa7wUu51JY
Number of Documents:	1
Document page count:	2
Number of supporting files:	0
Supporting files page count:	0

"NCR-25-079 SLQ-211610SPT Phase 1" History

-  Document created by Aaron Rogers (arogers@pathwaymedtech.com)
2025-11-14 - 12:38:17 PM PST
-  Document emailed to knatafji@pathwaymedtech.com for approval
2025-11-14 - 12:39:03 PM PST
-  Email viewed by knatafji@pathwaymedtech.com
2025-11-14 - 1:32:46 PM PST
-  knatafji@pathwaymedtech.com authenticated with Email OTP by verifying one-time code sent via email
2025-11-14 - 1:33:01 PM PST
-  Agreement viewed by knatafji@pathwaymedtech.com
2025-11-14 - 1:33:02 PM PST
-  Signer knatafji@pathwaymedtech.com entered name at signing as Karim Natafji
2025-11-14 - 1:33:19 PM PST
-  Karim Natafji (knatafji@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC
2025-11-14 - 1:33:21 PM PST
-  Document approved by Karim Natafji (knatafji@pathwaymedtech.com)
Approval Date: 2025-11-14 - 1:33:21 PM PST - Time Source: server
-  Document emailed to Ethan Rao (ethanr@silq.tech) for approval
2025-11-14 - 1:33:24 PM PST

 Email viewed by Ethan Rao (ethanr@silq.tech)

2025-11-14 - 1:35:34 PM PST

 Ethan Rao (ethanr@silq.tech) authenticated with Email OTP by verifying one-time code sent via email

2025-11-14 - 1:36:03 PM PST

 Agreement viewed by Ethan Rao (ethanr@silq.tech)

2025-11-14 - 1:36:05 PM PST

 Ethan Rao (ethanr@silq.tech) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC

2025-11-14 - 1:36:30 PM PST

 Document approved by Ethan Rao (ethanr@silq.tech)

Approval Date: 2025-11-14 - 1:36:30 PM PST - Time Source: server

 Document emailed to Aaron Rogers (arogers@pathwaymedtech.com) for approval

2025-11-14 - 1:36:31 PM PST

 Email viewed by Aaron Rogers (arogers@pathwaymedtech.com)

2025-11-14 - 1:37:07 PM PST

 Agreement viewed by Aaron Rogers (arogers@pathwaymedtech.com)

2025-11-14 - 1:37:31 PM PST

 Aaron Rogers (arogers@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC

2025-11-14 - 1:37:52 PM PST

 Document approved by Aaron Rogers (arogers@pathwaymedtech.com)

Approval Date: 2025-11-14 - 1:37:52 PM PST - Time Source: server

 Agreement completed.

2025-11-14 - 1:37:52 PM PST