

PHASE 2 – NCR INVESTIGATION					
Investigation Plan(s)		Perform 100% visual/dimensional inspection.			
Investigation Result(s)		On 17-NOV-2025, SILQ personnel Ethan Rao performed the 100% visual/dimensional inspection of the SLQ-211610SPT finished goods with results summarized below:			
		Lot Number	Qty Inspected	Qty Accepted	Qty Rejected
		SLQ-11202024	655	585	70
		SLQ-05012025	912	870	42
Is a CAPA or SCAR required?		☐ Yes, Insert CAPA or SCAR Reference Here ☒ No, provide justification: SILQ will handle any corrective actions with their supplier through SILQ’s respective QMS.			
Defect Codes:		Choose all that Apply: Code 2 – Device Defect (Any defect with actual product including damaged packaging) _____ _____ _____			
Root Cause Codes: <i>check all that apply</i>		Choose all that Apply: 605 - Supplier Quality Issue _____ _____ _____ _____			
Corrective Action(s) identified to address the root cause(s): <i>Attach any supporting documents.</i>		Supplier Ningbo Yingmed has implemented an improved valve to address the placement and dislodgement.			
Corrective Action(s) Reference: <i>ECO #, CAPA #, SCAR #, None</i>		N/A – SILQ will handle corrective actions for their supplier through their respective QMS.			
Final Disposition:		☐ Rework (See Rework Instructions below) ☐ Return to Supplier ☒ Scrap ☒ Use-as-Is ☐ N/A – No change from initial disposition			
Rationale for Disposition:		The Sort-to-Accept and Sort-to-Scrap interim disposition was successfully executed on Lots SLQ-11202024 and SLQ-05012025 by qualified SILQ personnel, validating the integrity of the remaining units against the client's mandated visual and dimensional acceptance criteria. Units that passed the inspection are now deemed compliant with specification and fit for commercial use. Units that failed inspection have been confirmed as nonconforming material and must be returned to the client for final disposal.			
Disposition Instructions:		Scrap Unit Return (Nonconforming Material): Ship the scrap material and attach Packing List to this NCR record. Inventory Release (Conforming Material): The Released Units (those that passed inspection) shall be formally moved out of quarantine status. These units are authorized for release into the client's defined usable finished goods inventory.			
REWORK/CORRECTION INSTRUCTIONS (☒ N/A if not applicable)					
Provide Rework Instructions (include pictures as required):					
N/A					

Pathway Medtech, LLC.	Nonconformance Report	NCR No.:	25-079
-----------------------	-----------------------	----------	--------

PHASE 2 – NCR INVESTIGATION			
Rework Impact / Risk Assessment: (Determine the impact/risk of the rework process or potential problem/incident on the product, processes, systems, documentation, etc. Determine the level of severity and probability of each risk using QP-004, Risk Assessment)	(☒ N/A if not applicable) Insert risk associated with rework		
PHASE 2 APPROVAL			
NCR Owner (Name/Title):	Karim Natafji / Director of Operations	Signature/Date:	 Karim Natafji (19-Nov-2025 11:32:17 PST)
MFG/Operations (Name/Title):	Karim Natafji / Director of Operations	Signature/Date:	 Karim Natafji (19-Nov-2025 11:32:17 PST)
Engineering (Name/Title):	Bryce Sandoval / Mechanical Engineer	Signature/Date:	 Bryce Sandoval (19-Nov-2025 11:56:22 PST)
Quality Assurance (Name/Title):	Aaron Rogers / Director of Regulatory and Quality	Signature/Date:	 Aaron Rogers (19-Nov-2025 14:35:37 PST)
Regulatory Affairs (Name/Title): ☐ N/A	Aaron Rogers / Director of Regulatory and Quality	Signature/Date:	 Aaron Rogers (19-Nov-2025 14:35:37 PST)
Customer (Name/Title): ☐ N/A	Ethan Rao / Director of R&D	Signature/Date:	

<i>Pathway Medtech, LLC.</i>	Nonconformance Report	NCR No.:	25-079
------------------------------	------------------------------	----------	---------------

REWORK OUTCOME (☒ N/A if not applicable)						
Part / Equipment / MP #	Part Rev	Part / Equipment / MP Description	Lot #	NCR Qty	NCR Qty	Final Disposition
Insert Number Ref	Insert Rev	Insert item description	Insert Lot#	Insert quantity of nonconforming material	☐ Pass: ____ ☐ Fail: ____	☐ Accept ☐ Scrap ☐ Material Destruction Completed ☐ Return to Supplier
Insert Number Ref	Insert Rev	Insert item description	Insert Lot#	Insert quantity of nonconforming material	☐ Pass: ____ ☐ Fail: ____	☐ Accept ☐ Scrap ☐ Material Destruction Completed ☐ Return to Supplier
Insert Number Ref	Insert Rev	Insert item description	Insert Lot#	Insert quantity of nonconforming material	☐ Pass: ____ ☐ Fail: ____	☐ Accept ☐ Scrap ☐ Material Destruction Completed ☐ Return to Supplier
REWORK OUTCOME / FINAL APPROVAL						
Inspected By (Name/Title):		N/A	Signature/Date:		N/A	
MFG/Operations (Name/Title):		N/A	Signature/Date:		N/A	
Engineering (Name/Title):		N/A	Signature/Date:		N/A	
Quality Assurance (Name/Title):		N/A	Signature/Date:		N/A	

Pathway Medtech, LLC.	Type:	Form	Doc. No.:	FM-0019	Revision:	H
Nonconformance Report (NCR)						

PHASE 3 – NCR CLOSURE			
Verify all information is accurate, required documentation is completed and attached?	☒ Yes ☐ No or N/A, provide Justification below		
Justification:	N/A		
Corrective action(s) is properly referenced and completed?	☒ Yes ☐ No or N/A, provide Justification below		
Justification:	N/A		
Affected material disposition is completed?	☒ Yes ☐ No or N/A, provide Justification below		
Justification:	N/A		
Notify Supplier/Customer of NCR:	☐ Yes, provide a copy of completed NCR to Supplier/Customer ☒ N/A		
NCR Log Verification:	☒ Verify NCR Log is updated		
PHASE 3 APPROVAL			
Quality Assurance (Name/Title):	Jacqueline Kim / QE	Signature/Date:	 Jacqueline Kim (19-Nov-2025 15:07:52 PST)
Comments (if any):	N/A		
NCR CANCELLATION (☒ N/A if not applicable)			
Cancellation?	☐ Yes, provide justification below		
Justification:	N/A		
NCR CANCELLATION APPROVAL (☒ N/A if not applicable)			
Quality Assurance (Name/Title):	N/A	Signature/Date:	N/A

Aaron Rogers (19-Nov-2025 14:35:37 PST)

Date Inspected	Part No.	Rev.	Description
11/17/2025	SLQ-211610SPT	A	SILQ 2-way Cleartract SPT 16 Fr x 10 ml Catheter, Sterile, 10-Pack
Lot No. SLQ-11202024	Lot QTY 655	Approved Supplier(s) N/A	
Specification	The check valve is properly secured to the connector. Report the total quantity	The total length from the sleeve to the end of the check valve does not exceed 26mm.	
Revision	N/A	N/A	
Lower Spec Limit (LSL)	P/F	N/A	
Upper Spec Limit (USL)	P/F	26.000	
UOM	EA	mm	
Measuring Equipment Type	N/A	Caliper / Ruler	
Asset No.	N/A	ES-YLB-003-02	
Cal Due Date	N/A	30 Jan 2026	
AQL	100%	100%	
Sample Size	655	655	
Sample No.	Variable/ Attribute Data		
Result Summary			
Pass QTY	610	630	
Fail QTY	45	25	
Result (P/F)	P	P	
Statistical Summary			
Min			
Max			
Mean			
StDev			
Comments			
Disposition:	Accept	NCMR No:	25-079
Inspected By:	Had	Date:	17 Nov 2025
Verified By:	Amor Rodriguez	Date:	11/18/2025



INDIVIDUAL TRAINING RECORD

Doc No. FM-0004
Rev. C
Page 1 of 1

Trainee(s)				
Print Name		Signature		Date
Ethan Rao		<i>[Signature]</i>		17 Nov 2025
Training Level 1				☐ N/A
Document No.	Rev	Title or Subject		
NCR-25-079	N/A	NCR-25-079 SLR-211610SPT Phase 1		
☐ I have read and understood the above document(s)				
Training Level 2				☒ N/A
Document No.	Rev	Title or Subject	Training Method	
			☐ Presentation/ Verbal Walkthrough ☐ Hand-on Demonstration ☐ Other:	
☐ Personnel were made aware of non-conformances that may occur from improper performance				
Trainer				
Print Name		Signature		Date
Training Level 3				☒ N/A
Document No.	Rev	Title or Subject	Training Completed	
			Initial Training	
			Training Method	Date
			☐ Presentation/ Verbal Walkthrough ☐ Other:	☐ Hand-on Demonstration
			Trainer Initials	
			Date	
			Number of Supervised builds required:	
			Supervised build DHR Ref:	
			Final Approval	
			Trainer Initials	Date
			☐ Personnel were made aware of non-conformances that may occur from improper performance	
Trainer				
Print Name		Signature		Date

This document contains confidential, proprietary information of Pathway Innovation and shall not be copied or reproduced without prior written permission there from. If any portion of this document is in hardcopy form, it must be considered and treated as an uncontrolled document. Users of this document must verify that they are using the most current revision of the controlled document before performing work.

Dear Valued Customer,

Thank you for notifying us about the issues that you observed in the most recent shipment of parts. At Ningbo Yingmed, we strive to provide high quality parts and will make improvements to meet your quality needs. To address the valve placement and valve dislodgement issue for the part, we have now implemented an improved valve that ensures secure placement and prevents dislodgement. No material changes have been made. We have also included additional processes with our line-assemblers to ensure that any parts that have excess particulates, dust, or glue are rejected from being shipped. Because these improvements will incur additional manufacturing costs, please understand that increased cost will be reflected in new pricing.

Date: Nov 6th,2024

Signature:

Stamp

宁波盈迈医疗器械有限公司
NINGBO YINGMED MEDICAL INSTRUMENTS CO.,LTD

陈海祥

NCR-25-079 SLQ-211610SPT Phase 2-3

Final Audit Report

2025-11-19

Created:	2025-11-19 (Pacific Standard Time)
By:	Aaron Rogers (arogers@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAA787XA-P4WgpSwlhtN9gKnnOioo8Ki_kv
Number of Documents:	3
Document page count:	8
Number of supporting files:	0
Supporting files page count:	0

"NCR-25-079 SLQ-211610SPT Phase 2-3" History

-  Document created by Aaron Rogers (arogers@pathwaymedtech.com)
2025-11-19 - 10:25:26 AM PST
-  Document emailed to knatafji@pathwaymedtech.com for approval
2025-11-19 - 10:27:43 AM PST
-  Agreement viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-11-19 - 10:35:10 AM PST
-  Email viewed by knatafji@pathwaymedtech.com
2025-11-19 - 11:30:29 AM PST
-  knatafji@pathwaymedtech.com authenticated with Email OTP by verifying one-time code sent via email
2025-11-19 - 11:31:41 AM PST
-  Agreement viewed by knatafji@pathwaymedtech.com
2025-11-19 - 11:31:42 AM PST
-  Signer knatafji@pathwaymedtech.com entered name at signing as Karim Natafji
2025-11-19 - 11:32:15 AM PST
-  Karim Natafji (knatafji@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC
2025-11-19 - 11:32:17 AM PST
-  Document approved by Karim Natafji (knatafji@pathwaymedtech.com)
Approval Date: 2025-11-19 - 11:32:17 AM PST - Time Source: server

-  Document emailed to Bryce Sandoval (bsandoval@pathwaymedtech.com) for approval
2025-11-19 - 11:32:20 AM PST
-  Email viewed by Bryce Sandoval (bsandoval@pathwaymedtech.com)
2025-11-19 - 11:54:21 AM PST
-  Bryce Sandoval (bsandoval@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email
2025-11-19 - 11:55:10 AM PST
-  Agreement viewed by Bryce Sandoval (bsandoval@pathwaymedtech.com)
2025-11-19 - 11:55:10 AM PST
-  Bryce Sandoval (bsandoval@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC
2025-11-19 - 11:56:22 AM PST
-  Document approved by Bryce Sandoval (bsandoval@pathwaymedtech.com)
Approval Date: 2025-11-19 - 11:56:22 AM PST - Time Source: server
-  Document emailed to Ethan Rao (ethanr@silq.tech) for approval
2025-11-19 - 11:56:25 AM PST
-  Email viewed by Ethan Rao (ethanr@silq.tech)
2025-11-19 - 12:17:42 PM PST
-  Ethan Rao (ethanr@silq.tech) authenticated with Email OTP by verifying one-time code sent via email
2025-11-19 - 12:17:55 PM PST
-  Agreement viewed by Ethan Rao (ethanr@silq.tech)
2025-11-19 - 12:17:57 PM PST
-  Ethan Rao (ethanr@silq.tech) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC
2025-11-19 - 12:18:15 PM PST
-  Document approved by Ethan Rao (ethanr@silq.tech)
Approval Date: 2025-11-19 - 12:18:15 PM PST - Time Source: server
-  Document emailed to Aaron Rogers (arogers@pathwaymedtech.com) for approval
2025-11-19 - 12:18:20 PM PST
-  Email viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-11-19 - 2:34:56 PM PST

 Aaron Rogers (arogers@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC

2025-11-19 - 2:35:37 PM PST

 Document approved by Aaron Rogers (arogers@pathwaymedtech.com)

Approval Date: 2025-11-19 - 2:35:37 PM PST - Time Source: server

 Document emailed to Jacqueline Kim (jkim@pathwaymedtech.com) for approval

2025-11-19 - 2:35:39 PM PST

 Email viewed by Jacqueline Kim (jkim@pathwaymedtech.com)

2025-11-19 - 2:54:04 PM PST

 Jacqueline Kim (jkim@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email

2025-11-19 - 2:54:34 PM PST

 Agreement viewed by Jacqueline Kim (jkim@pathwaymedtech.com)

2025-11-19 - 2:54:34 PM PST

 Jacqueline Kim (jkim@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech, LLC

2025-11-19 - 3:07:52 PM PST

 Document approved by Jacqueline Kim (jkim@pathwaymedtech.com)

Approval Date: 2025-11-19 - 3:07:52 PM PST - Time Source: server

 Agreement completed.

2025-11-19 - 3:07:52 PM PST