

	DOCUMENT TITLE: Management Review Meeting Minutes	Revision: A
	DOCUMENT #: FM1-QM.SLQ018	Effective Date: 25Sep2020

Meeting Date / Time: 11 Feb 2025 1PM PST

Timeframe under review: 2024

Agenda Items: (check all that apply to this meeting)


- ☒ Review of minutes and action items from the previous meeting (required at each meeting)
- ☒ Review of Quality Policy
- ☒ Review of general quality goals and objectives
- ☒ Results, actions and status of internal audits, regulatory agency inspections and notified body audits
- ☒ Status of the Corrective and Preventive Action System
- ☒ Review of Product Complaints trends
- ☒ Review of Production Data including nonconformities
- ☒ Review of training system performance
- ☒ Review of Post Market Surveillance / Customer Feedback Data
- ☒ Review of reportings to regulatory authorities
- ☒ Review of statistical techniques (sampling plans)
- ☒ Review of supplier performance
- ☒ Changes that could affect the quality management system
- ☒ Recommendations for improvement
- ☒ New or revised regulatory requirements
- ☒ Review of Risk Management Activities
- ☒ Review of Design Control Activities
- ☒ Review of Resource Levels

Attendees:

Signature

Date

CEO:



Date: 13 Feb 2025

QA Management:



Date: 13Feb2025

R&D/Engineering Management:



Date: 13 Feb 2025

	DOCUMENT TITLE: Management Review Meeting Minutes	Revision: A
	DOCUMENT #: FM1-QM.SLQ018	Effective Date: 25Sep2020

Manufacturing Management:	_____	Date: _____
Other: Finance Management	_____	Date: _____
Other: Marketing Management	_____	Date: _____
Other: RA Management	_____	Date: _____
Other: R&D (Director)		Date: <u>18Feb2025</u>
Other: Manufacturing (Director)	_____	Date: _____
Other: QA (Director)	_____	Date: _____
Other: Independent Reviewer	_____	Date: _____

(This page to be used to demonstrate execution of MR for regulatory purposes)

Data Attachments:

- ☒ CAPA System Metric
- ☒ Audit Reports / Audit System Metrics
- ☒ Complaint Trend Data
- ☒ Post Market Surveillance / Customer Feedback Data
- ☒ Non-Conforming Material Report Data
- ☒ Training System Metrics
- ☒ Supplier Quality Data

List Other Attachments:

2024 Management Review slides 11Feb2025.pptx



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Discussion Topics and Minutes:

Topic: **Review of minutes and action items from the previous meeting**

Topic Minutes/Notes

The management review team reviewed the 2023 Management Review Meeting Minutes. The meeting minutes are maintained in FileHold. There were ten (10) action items from previous management review. 6 action items were completed. 2 action items are postponed. 2 action items were addressed through a different corrective action. A summary of each action item is provided below.

- Topic: Review of minutes and action items from the previous meeting
 - Action item 1 complete: Completed action items 1, 2, 3, 4, and 7 from 2023 Management Review meeting
- Topic: Review of the quality policy
 - Action item 2 complete: Quality policy updated (QM.SLQ027 Rev D, QM.SLQ035 Rev C)
- Topic: Review of general quality goals and objectives
 - Action item 3 complete: Quality objectives released (QM.SLQ037 Rev A)
 - Action item 4 complete: 5% complaint rate objective met; based on assessment in Feb 2024, complaint rate is 0.55%.
- Topic: Review of product complaint trends
 - Action items 5 and 6 were addressed through a design change: Ningbo implemented a new valve that ensures placement and prevents dislodging.
- Topic: Status of the production data including nonconformities
 - Action item 7 complete: Confirmed with Pathway all product lines are at a 3 year shelf life. All finished goods inventory has been relabeled accordingly.
- Topic: Changes that could affect the quality management system
 - Action item 8 postponed: Complete Gage R&R study on in-process quality control test methods for UV Light absorbance and Rinse Test. Postponed due to limited manufacturing activity during 2024.
 - Action item 9 postponed: Complete test method validation for 1H and 19F spectra readings. Postponed due to limited manufacturing activity during 2024.
- Topic: General recommendations for improvement
 - Action item 10 complete: Extend shelf life of suspension. Suspension shelf-life has been extended to two (2) years.

No action required.

Action Items

1. Action Item:

Assigned to:

N/A

DP
14Feb2025

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

Assigned to:

N/A

SP
14Feb2025

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of the Quality Policy

Topic Minutes/Notes

The team reviewed the quality policy established in QM.SLQ027 Rev D and QM.SLQ035 Rev C.

The team determined that the quality policy is suitable as is.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A *DP* 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of general quality goals and objectives

Topic Minutes/Notes

There were two quality objectives established previously to review.
The team reviewed the 2 quality objectives for 2024 involving product quality.

1. Objective: Incoming quality

- a. Goal:
The lot acceptance rate of incoming material used in finished product shall be greater than 80%.
- b. Results: Objective met
Acceptance rate = 100%. Two (2) lots of catheters received from Ningbo in 2024. Both lots accepted.

2. Objective: Finished product quality

- a. Goal:
The complaint rate of failed catheters shall be less than 5% of the overall distributed catheters.
- b. Results: Objective met
The complaint rate in 2024 was 0.3%. A total of 1,737 catheters (from all sizes) were sold in 2024. There was a total of five failed catheters.

The team agreed that the quality objectives are suitable as is.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

DP 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Results, actions and status of internal audits, regulatory agency inspections and notified body audits

Topic Minutes/Notes

The team reviewed summary of audits in 2024 which included 1 internal audit.

- 1) The internal audit was performed by Dr. Christopher Devine on January 16th through 17th, 2025. The scope of the audit covered ISO 13485:2016 and 21 CFR Parts 803, 806 and 820. The audit was completed with 0 findings, 0 recommendations and 0 concerns. Reference audit report # 24-01.

Team discussed the upcoming 2025 audit schedule which included:

- 2025 Internal Audit schedule

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

SP 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Status of the Corrective and Preventive Action System

Topic Minutes/Notes

The team reviewed the CAPA system status.
No CAPA's were initiated in 2024.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

Assigned to:

SP

14Feb2025

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of Product Complaints trends

Topic Minutes/Notes

Team reviewed the four product complaints received in 2024.

- PCF# 2024001 (Status: closed)
 - o Complainant: Oliver Estrada, Ranchos Los Amigos National Rehabilitation Center
 - o Number of units: > 2 (16Fr x 10mL)
 - o Complaint description: Two (2) ClearTract catheters had a dark spot on the distal end of the catheter itself and is visible through the clear sterile packaging.
 - o Patient injury: None
 - o Reportability determination: Not a reportable complaint
 - o Failure investigation: Insufficient cleaning of the machinery which led to a build-up of Silq's surface treatment polymer in the underlying manifold which resulted in surface agglomerates.
 - o Corrective action: 1) Added cleaning of the machine prior to each production run. 2) Pathway implemented a visual inspection of coated catheters to catch any with surface defects and agglomerations.
 - o **Action item: Update complaint to include cleaning of machine occurs before every production run.**
- PCF# 2024002 (Status: closed)
 - o Complainant: Tousif Hussain Sheikh, HPFY
 - o Number of units: 1 (18Fr x 10mL)
 - o Complaint description: Report of defective catheter where the retaining collar broke off.
 - o Patient injury: None
 - o Reportability determination: Not a reportable complaint
 - o Failure investigation: Defective catheter was not returned for investigation
- PCF# 2024003 (Status: closed)
 - o Complainant: Gina Koehler, Cleveland Clinic
 - o Number of units: 1 (16Fr x 10mL)
 - o Complaint description: When the nurse was filling the balloon with saline, the plastic end of the tube where the syringe is inserted came off. As a result, the nurse was unable to adequately fill the balloon and had to use another catheter.
 - o Patient injury: None
 - o Reportability determination: Not a reportable complaint
 - o Failure investigation: The catheter was returned to Silq for analysis, however the Luer lock connector which malfunctioned was not present as it had been discarded at the time of the malfunction. After discussion with the complainant, when the syringe containing saline was connected to the Luer lock connector, the Luer lock became unseated from the catheter neck.
 - o Corrective action: Ningbo implemented a new valve to ensure proper placement and prevents dislodging.
- PCF# 2024004 (Status: closed)
 - o Complainant: Gurav Pandey, HPFY
 - o Number of units: 1 (18Fr x 10mL)



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

- Complaint description: According to the customer, the catheter (lot #SLQ-81020515241) was defective and had a crack.
- Patient injury: None
- Reportability determination: Not a reportable complaint
- Failure investigation: Aside from normally observed encrustation on Foley catheters, no other apparent defect was observed. No crack or tear could be found on the catheter after evaluation of the catheter. There was also no apparent issue with the tip or valve of the catheter based on a visual inspection. Additional testing was performed to confirm if there was a crack or leak in the device. The catheter balloon was filled with 5mL of water for 1 minute before being increased to 10mL for an additional 1 minute. No issues or leaks were observed.

Action required.

Action Items

1. Action Item: Update complaint 2024001 to include cleaning of machine occurs before every production run.

Assigned to: B. Wong

- ☐ CAPA Required
☐ Opportunity for Improvement
☒ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

3. Action Item:

Assigned to:

N/A

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: **Status of the Production Data including nonconformities**

Topic Minutes/Notes

Team reviewed the build summary for finished devices and suspension.

- 4 lots of finished devices were manufactured in 2024. 3 of the 4 lots were accepted. 1 lot was rejected for LAL failure (Model SLQ-211810SPT, Lot # SLQ-11182024).
- SLQ-211410SPT: 22.1 cases of were released
- SLQ-211610SPT: 94.4 cases of were released
- SLQ-211810SPT: 40.5 cases of were released
- No Suspension was manufactured in 2024

Reviewed the nonconformance summary

- No NCR's were initiated by Silq or Pathway in 2024.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

SP

14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: **Review of training system performance**

Topic Minutes/Notes

Team reviewed status of the training system. All employee training is up to date. No new employees were added in 2024.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

3. Action Item:

N/A

DP 14Feb2025

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of Post Market Surveillance / Customer Feedback

Topic Minutes/Notes

Team reviewed post-market surveillance data for 2024.

- A) Post-market clinical trials
 - Analysis report on the primary endpoint of the RCT Study (analysis of biofilm) was submitted by Technomics on 19 Dec 2024. Results were favorable indicating the ClearTract had significantly less biofilm than the BARDEX and Rusch Gold Standard catheters.
- B) The evaluation of the MAUDE database: Database search from Jan 1, 2024 to Dec 31, 2024 for reports in the following Product Classes.
 - Catheter, Urological: 1,275 reports (1,135 malfunctions, 139 patient injuries, 1 death)
 - Catheter, Urological (Antimicrobial) and Accessories: 1,978 reports (1,914 malfunctions, 64 patient injuries, 0 deaths)
 - Common malfunctions included material puncture/hole/rupture, sharp edges, material twisted/bent, difficult to remove, and misassembled during manufacturing.
 - Common injuries resulted from burst container/vessel, difficult to remove, material puncture/hole/rupture, device handling, and material fragmentation.
- C) There were four (4) product complaints in 2024. Two (2) complaints confirmed to involve Silq catheter defects. Reviewed in the Review of Product Complaints section of MR.
- D) There has been no customer feedback reported.
- E) There have been no product returns, field safety corrective action/recalls, or public health advisories of safety notices from regulatory authorities.
- F) There were no NCRs initiated.
- G) There have been no device-related CAPA's.
- H) The yield for finished devices was 72%. 3 of 4 lots were accepted. 60.3 cases rejected. 94.4 cases of 2-way 16Fr x 10ml. 22.1 cases of 2-way 14Fr x 10ml released. 38 cases of 2-way 18Fr x 10ml released.
- I) Process validations: Coating 14Fr and 18Fr process validation on automated coating machine completed in 2023. Report released in May 2024 (TR-SLQ-014)
Sealing process with Impulse Sealer PQ completed with another pouch manufacturer in 2023. Report released in 2024 (TR-SLQ-0011).
- J) Supplier issues: No supplier issues or SCARs reported for 2024.
- K) No issues with training to report.

No action required.

Action Items

1. Action Item:

Assigned to:

N/A

☐ CAPA Required

☐ Opportunity for Improvement

☐ Simple Action

2. Action Item:



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

Assigned to:

N/A

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

SP 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: **Review of reportings to regulatory authorities**

Topic Minutes/Notes

There were no reportings to regulatory authorities in 2024.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A
OP 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of statistical techniques (sampling plans) used to evaluate the quality of components, materials and finished goods to assure continued applicability

Topic Minutes/Notes

Team reviewed statistical techniques (documented in QM.SLQ011 Rev. A) and agreed they are suitable and adequate for Design V&V and for process qualifications and validation.

The statistical techniques for production sampling plans are suitable and adequate.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of supplier performance

Topic Minutes/Notes

Team reviewed supplier performance. Summary of their performance and re-evaluation can be found in attached management review slides.

There were no changes to supplier approval statuses.

Team discussed the inclusion of HPFY on current ASL list and after discussion, it was agreed to remove HPFY from the ASL list as they are technically a customer and not a supplier.

Action item: Remove HPFY from Approved Supplier List.

Action Items

1. Action Item: Remove HPFY from Approved Supplier List.

Assigned to: B. Wong

- ☐ CAPA Required
☐ Opportunity for Improvement
☒ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

3. Action Item:

N/A *JP* 14Feb2025

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Changes that could affect the quality management system

Topic Minutes/Notes

The team reviewed the changes to the following regulatory standards:

- ISO 15223-1:2021 replaces ISO 15223-1:2016 (Medical devices — Symbols to be used with information to be supplied by the manufacturer, Part 1: General requirements)
 - o Addition of symbol for UDI (Unique Device Identifier). Product labels for pouch and box of Silq product (SLQ-LB-4000, SLQ-LB-4002, SLQ-LB-4008, SLQ-LB-4009) contain UDI. No action required.
- ASTM D4169-23e1 replaces ASTM D4169-22 (Standard Practice for Performance Testing of Shipping Containers and Systems)
 - o Changes to Schedule C (Vehicle Stacking) for small packages (smaller than 2.0 ft³ and weighs between 10.0 lbs and 30.0 lbs). Does not apply to the corrugated container used in the packaging validation testing (1.25 ft³, 7.88lb).
 - o No impact on testing performed in accordance with old revision of standard
- ASTM D4332-22 replaces ASTM D41332-14 (Standard Practice for Conditioning Containers, Packages, or Packaging Components for Testing)
 - o No significant changes. No impact on testing performed in accordance with old revision of standard
- ASTM F1886/F1886M-16(2024) replaces ASTM F1886/F1886M-16 (Standard Test Method for Determining Integrity of Seals for Flexible Packaging by Visual Inspection)
 - o No changes. 2024 revision is a reapproval of the 2016 revision. No impact on testing performed in accordance with old revision of standard

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

DP 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4.

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: **General recommendations for improvement**

Topic Minutes/Notes

Team discussed recommendations for improvement.

For the product:

- **Action item: Complete Gage R&R study on the following in-process quality control test methods: UV Light Absorbance (TM-HDX-001) and Rinse Test (TM-HDX-002).
Action item: Complete test method validation for ¹H and ¹⁹F NMR spectra readings.**

For the QMS:

- None.

Action required.

Action Items

1. Action Item: Complete Gage R&R study on in-process quality control test methods for UV Light Absorbance and Rinse Test

Assigned to: E. Rao

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☒ Simple Action

2. Action Item: Complete test method validation for ¹H and ¹⁹F NMR spectra readings.

Assigned to: E. Rao

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☒ Simple Action

3. Action Item: Extend shelf-life of suspension

Assigned to: E. Rao

N/A

DP 18Feb2025

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☒ Simple Action

4. Action Item:

Assigned to:

N/A

DP 14Feb2025

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☒ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: **New or revised regulatory requirements**

Topic Minutes/Notes

See "Changes that could affect the quality management system" section above for discussion of new or revised regulatory requirements.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

DP 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of Risk Management Activities

Topic Minutes/Notes

Based on review of data obtained from post-market surveillance activities, no new hazards have been identified. Potential change to an existing hazard #10.1 regarding three failure modes:

1. Balloon deflates or does not inflate – Valve malfunction or leaks (wall thickness issue)
 2. Balloon does not inflate – valve malfunctions
 3. Catheter breaks under normal tensile stresses – Material or dimensional issues
- The probability of each failure mode is currently set at '1'. **Action item to evaluate the probability rating and determine if the rating should increase based on complaints seen in the field.**

Action required.

Action Items

1. Action Item: Evaluate the probability rating and determine if the rating should increase based on complaints seen in the field.

Assigned to: B. Wong

- ☐ CAPA Required
☐ Opportunity for Improvement
☒ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action

3. Action Item:

Assigned to:

N/A

- ☐ CAPA Required
☐ Opportunity for Improvement
☒ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
☐ Opportunity for Improvement
☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of Design Control Activities

Topic Minutes/Notes

Team reviewed the summary of design control activities for Suspension (C.SLQ001).

Shelf-life extension for Suspension (C.SLQ001)

- Study performed in accordance with Protocol #VV.SLQ028. Catheter samples were coated with 2 year real-time aged Suspension C.SLQ001 (Lot #HXY-825021)
- Coating adherence verification test and crystal violet absorbance test were performed on coated samples. Results met acceptance criteria confirming a shelf life of 2 years for Suspension. Results documented in report #VV.029.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

DP
14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Topic: Review of Resources

Topic Minutes/Notes

Team reviewed current resource/staffing of the organization.

- Silq
 - No change in the organization/personnel in 2024.
 - Transition quality and regulatory responsibilities to Pathway in 2025.
- CMO's
 - Pathway: Manufacture of coated catheters, labeling, final packaging, sterilization and distribution.

No action required.

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

DP

14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action



DOCUMENT TITLE:
Management Review Meeting Minutes

Revision:
A

DOCUMENT #:
FM1-QM.SLQ018

Effective Date:
25Sep2020

Additional Topic: N/A

Topic Minutes/Notes

N/A

Action Items

1. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

2. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

3. Action Item:

N/A

DP 14Feb2025

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

4. Action Item:

Assigned to:

- ☐ CAPA Required
- ☐ Opportunity for Improvement
- ☐ Simple Action

	DOCUMENT TITLE: Management Review Meeting Minutes	Revision: A
	DOCUMENT #: FM1-QM.SLQ018	Effective Date: 25Sep2020

Meeting Date / Time: 11 Feb 2025 1PM PST

Timeframe under review: 2024

Conclusions: (check the appropriate conclusion based on group consensus)

- ☐ The quality system has been determined effective, suitable and adequate without changes required at this time
- ☒ The quality system has been determined effective, suitable and adequate. However, actions are required to maintain the system.
- ☐ The quality system has been determined effective, suitable and adequate with opportunities for improvement identified at this meeting
- ☐ The quality system has significant deficiencies and is currently either not effective, not suitable or not adequate. Immediate corrective actions are required.

Changes: (check the appropriate conclusion based on group consensus)

- ☒ Proposed changes made to the Quality System are not substantive.
- ☐ Proposed changes made to the Quality System are substantive and require notification of the Notified Body.
- ☐ There are no proposed changes to the Quality System at this time.

Approximate Next Meeting Date: January 2026