

PHASE 2 – CAPA INVESTIGATION RESULTS & ROOT CAUSE ANALYSIS																																																					
CAPA Investigation Results Assigned to:	Jacqueline Kim	CAPA Investigation Results Target Due Date:	05/23/2025	CAPA Investigation Results Completion Date:	05/22/2025																																																
CAPA Investigation Results: <i>(Describe the investigation and results that determined what cause (or will cause) the incident/problem to occur. Attach any supporting documents. If needed, complete and attach the SCAR Form, FM-0087)</i>		1. CAPA Failure Root Cause a. Omar (Warehousing) was not trained to the protocol (QP-009), so when the protocol was updated, he did not get the retraining request Note: Training plans were reviewed and updated in CAPA- 24-001 2. Review quality agreements of current clients to determine inspection/receiving responsibilities <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th colspan="2" style="text-align: left; padding: 2px;">Table 1: Quality agreements for active clients</th> </tr> <tr> <td style="width:20%; padding: 2px;">Arga</td> <td style="padding: 2px;">QA-ARG-0001 Rev. A Section 7.10 states that incoming inspection, material identification/controls, and acceptance status are Pathway's responsibilities. Signed 01/18/2024. Earliest received inventory: 03/27/2024</td> </tr> <tr> <td style="padding: 2px;">Guidestar</td> <td style="padding: 2px;">QA-GMD-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 08/10/2022. Earliest received inventory: 10/19/2023</td> </tr> <tr> <td style="padding: 2px;">Mark 2 Medical</td> <td style="padding: 2px;">QA-M2M-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 10/15/2024. Earliest received inventory: May 2024</td> </tr> <tr> <td style="padding: 2px;">SILQ</td> <td style="padding: 2px;">QA-SLQ-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 03/17/2022. Earliest received inventory: 01/23/2024</td> </tr> <tr> <td style="padding: 2px;">SYLKE</td> <td style="padding: 2px;"> QA-SYL-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 03/27/2025. QA-SDS-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 05/31/2023. Earliest received inventory: 01/24/2024 </td> </tr> <tr> <td style="padding: 2px;">Adler</td> <td style="padding: 2px;">QA-ADL-0001 Rev. A states that incoming inspection, material identification/controls, and acceptance status are Pathway's responsibilities. Signed 02/19/2020. Earliest received inventory: 12/14/2023</td> </tr> <tr> <td style="padding: 2px;">Singlepass</td> <td style="padding: 2px;">QA-SPS-0001 Rev. A Section 3.10 states that Pathway is responsible for conducting testing and/or conformance inspection. Signed 07/16/2024. Earliest received inventory: 06/07/2024</td> </tr> <tr> <td style="padding: 2px;">Neurgain</td> <td style="padding: 2px;">QA-NGN-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 06/20/2022. Earliest received inventory: 01/10/2022</td> </tr> <tr> <td style="padding: 2px;">RetractOrtho</td> <td style="padding: 2px;">No agreement required; RetractOrtho is under fulfillment only and has a service agreement with Pathway</td> </tr> <tr> <td style="padding: 2px;">Livsmed</td> <td style="padding: 2px;">No agreement required; LivsMed is under fulfillment only and has a service agreement with Pathway</td> </tr> <tr> <td style="padding: 2px;">Convergent Genomics</td> <td style="padding: 2px;">No agreement required; Convergent does not produce medical devices</td> </tr> </table> There were multiple instances of receiving performed before quality agreement was signed. Updates to the general receiving process were updated in April 2024. CHR's should be available for all inventory post April 2024. <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th colspan="6" style="text-align: left; padding: 2px;">Table 2: Inventory Sampling for Inspection Documentations</th> </tr> <tr> <th style="width:25%; padding: 2px;">PN</th> <th style="width:10%; padding: 2px;">Date Received</th> <th style="width:15%; padding: 2px;">Inspection Record Matches Specification?</th> <th style="width:10%; padding: 2px;">MTS/ COTS</th> <th style="width:10%; padding: 2px;">Accepted Label Attached?</th> <th style="width:30%; padding: 2px;">Comments</th> </tr> <tr> <td style="padding: 2px;">ARG-100004 Lot: 240726-04</td> <td style="padding: 2px;">07/26/2024</td> <td style="padding: 2px;">Yes</td> <td style="padding: 2px;">MTS</td> <td style="padding: 2px;">Yes</td> <td style="padding: 2px;">Duplicate records in Arena</td> </tr> <tr> <td style="padding: 2px;">ARG-100102</td> <td style="padding: 2px;">12/16/2024</td> <td style="padding: 2px;">Yes</td> <td style="padding: 2px;">COTS</td> <td style="padding: 2px;">Yes</td> <td style="padding: 2px;">N/A</td> </tr> </table>				Table 1: Quality agreements for active clients		Arga	QA-ARG-0001 Rev. A Section 7.10 states that incoming inspection, material identification/controls, and acceptance status are Pathway's responsibilities. Signed 01/18/2024. Earliest received inventory: 03/27/2024	Guidestar	QA-GMD-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 08/10/2022. Earliest received inventory: 10/19/2023	Mark 2 Medical	QA-M2M-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 10/15/2024. Earliest received inventory: May 2024	SILQ	QA-SLQ-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 03/17/2022. Earliest received inventory: 01/23/2024	SYLKE	QA-SYL-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 03/27/2025. QA-SDS-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 05/31/2023. Earliest received inventory: 01/24/2024	Adler	QA-ADL-0001 Rev. A states that incoming inspection, material identification/controls, and acceptance status are Pathway's responsibilities. Signed 02/19/2020. Earliest received inventory: 12/14/2023	Singlepass	QA-SPS-0001 Rev. A Section 3.10 states that Pathway is responsible for conducting testing and/or conformance inspection. Signed 07/16/2024. Earliest received inventory: 06/07/2024	Neurgain	QA-NGN-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 06/20/2022. Earliest received inventory: 01/10/2022	RetractOrtho	No agreement required; RetractOrtho is under fulfillment only and has a service agreement with Pathway	Livsmed	No agreement required; LivsMed is under fulfillment only and has a service agreement with Pathway	Convergent Genomics	No agreement required; Convergent does not produce medical devices	Table 2: Inventory Sampling for Inspection Documentations						PN	Date Received	Inspection Record Matches Specification?	MTS/ COTS	Accepted Label Attached?	Comments	ARG-100004 Lot: 240726-04	07/26/2024	Yes	MTS	Yes	Duplicate records in Arena	ARG-100102	12/16/2024	Yes	COTS	Yes	N/A
Table 1: Quality agreements for active clients																																																					
Arga	QA-ARG-0001 Rev. A Section 7.10 states that incoming inspection, material identification/controls, and acceptance status are Pathway's responsibilities. Signed 01/18/2024. Earliest received inventory: 03/27/2024																																																				
Guidestar	QA-GMD-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 08/10/2022. Earliest received inventory: 10/19/2023																																																				
Mark 2 Medical	QA-M2M-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 10/15/2024. Earliest received inventory: May 2024																																																				
SILQ	QA-SLQ-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 03/17/2022. Earliest received inventory: 01/23/2024																																																				
SYLKE	QA-SYL-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 03/27/2025. QA-SDS-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 05/31/2023. Earliest received inventory: 01/24/2024																																																				
Adler	QA-ADL-0001 Rev. A states that incoming inspection, material identification/controls, and acceptance status are Pathway's responsibilities. Signed 02/19/2020. Earliest received inventory: 12/14/2023																																																				
Singlepass	QA-SPS-0001 Rev. A Section 3.10 states that Pathway is responsible for conducting testing and/or conformance inspection. Signed 07/16/2024. Earliest received inventory: 06/07/2024																																																				
Neurgain	QA-NGN-0001 Rev. A Section 7.7 states that Pathway is responsible for all testing and/or conformance inspection. Signed 06/20/2022. Earliest received inventory: 01/10/2022																																																				
RetractOrtho	No agreement required; RetractOrtho is under fulfillment only and has a service agreement with Pathway																																																				
Livsmed	No agreement required; LivsMed is under fulfillment only and has a service agreement with Pathway																																																				
Convergent Genomics	No agreement required; Convergent does not produce medical devices																																																				
Table 2: Inventory Sampling for Inspection Documentations																																																					
PN	Date Received	Inspection Record Matches Specification?	MTS/ COTS	Accepted Label Attached?	Comments																																																
ARG-100004 Lot: 240726-04	07/26/2024	Yes	MTS	Yes	Duplicate records in Arena																																																
ARG-100102	12/16/2024	Yes	COTS	Yes	N/A																																																

PHASE 2 – CAPA INVESTIGATION RESULTS & ROOT CAUSE ANALYSIS

	Lot: 241216-01					
	ARG-100079 Lot: 230426-01	Received multiple batches of the same lot	Yes	MTS	Yes	N/A
	ARG-100086 Lot: 250206-01	Received multiple batches of the same lot	No	MTS	No	Lot CHRs not in Arena. Lot 240307-02 is in Arena, but the Fishbowl incoming inventory does not match the Arena CHRs
	GMD1-V14-06 Lot: 5670020240513	11/11/2024 (2 batches in lot)	Yes	MTS	No label in CHR	Quantities and lot numbers do not match between Arena and Fishbowl
	GMD-5001 Lot: 609416	01/12/2024	Yes	MTS	Yes	Quantities and lot numbers do not match between Arena and Fishbowl
	GMD1-V14-17 Lot: 22342	01/17/2024	No	MTS	No	No CHRs available
	GMD1-V14-10 Lot: 1013_11122024	10/19/2023	No	MTS	No	Quantities and lot numbers do not match between Arena and Fishbowl
	M2M-1100 Lot: 891509	05/12/2024	Yes	MTS	Yes	Quantities and lot numbers do not match between Arena and Fishbowl
	M2M-4000 Lot: N/A	05/20/2024	Yes	MTS	No	Lot missing from Fishbowl
	M2M-5000 Lot: N/A	10/11/2024	Yes	MTS	Yes	File is missing from Item in Arena. Lot missing from Fishbowl
	SLQ-3000 Lot: 2302-01-1	1/24/2024	Yes	MTS	Yes	Quantities and lot numbers do not match between Arena and Fishbowl
	SLQ-5013 Lot: 18768	1/23/2024	Yes	MTS	Yes	Quantities do not match between Arena and Fishbowl
	SLQ-6009 Lot: 20211122	1/24/2024	No	MTS	No	No CHRs available
	SYL-LB-4004 Lot: 12292023-01	01/24/2024	Yes	MTS	Yes	Quantities and lot numbers do not match between Arena and Fishbowl
	SYL-5150 Lot: 1526-103123	01/24/2024	Yes	MTS	Yes	None
	SYL-WC-32-010 Lot: 32-0101212231	01/24/2024	Yes	MTS	No	Quantities and lot numbers do not match between Arena and Fishbowl
	ADL-2003 Lot: AKT/PATH/ADL/2003-A/03	12/14/2023	Yes	MTS	Yes	Quantities and lot numbers do not match between Arena and Fishbowl
	ADL-3000 Lot: 2711	12/14/2023	Yes	COTS	Yes	Quantities and lot numbers do not match between Arena and Fishbowl

PHASE 2 – CAPA INVESTIGATION RESULTS & ROOT CAUSE ANALYSIS

ADL-5000 Lot: 14355	12/14/2023	Yes	COTS	No	Quantities do not match between Arena and Fishbowl
SPS-C-014-0049 Lot: 469914	07/29/2024	Yes	MTS	Yes	Quantities do not match between Arena and Fishbowl
SPS-C-001-0058 Lot: 20240404-014-01	06/07/2024	Yes	MTS	Yes	Quantities do not match between Arena and Fishbowl
SPS-C-014-0067 Lot: 933523	06/07/2024	Yes	MTS	Yes	Quantities do not match between Arena and Fishbowl. Expiration date missing.
SPS-C-014-0145 Lot: 69438	10/29/2024	Yes	MTS	Yes	Quantities do not match between Arena and Fishbowl
NGN-1100 Lot: 030725-18891	03/10/2021	Yes	MTS	Yes	None
NGN-3003 Lot: N/A	05/2023	Yes	COTS	Yes	Lot assigned in CHR, but not in Fishbowl
NGN-3032 Lot: N/A	None available	No	COTS	No	No CHRs available

Note:

- Some CHR items had duplicate records
 - Arena contains CHRs with lots that are not listed in Fishbowl and vice versa
 - Actual dates of acceptance activities were multiple years sooner than the Fishbowl and Arena record dates listed. For example, SLQ-5013 acceptance label is dated 09/02/2022, but inventory adjustment in Fishbowl is 1/24/2024
 - Names of items in Arena have incorrect revisions
 - CHR's are not released to records.
 - Some receiving activities were dated after initial inventory adjustment in Fishbowl (ie SPS-S-014-0145)
3. Protocol/Work Instructions Directions
- There are no released QPs for Fishbowl
 - There are no protocols or work instructions for:
 - How to receive and assign inventory in Fishbowl
 - Assuring Fishbowl and Arena align
 - Releasing CHR's in Arena
 - Reviewing CHR's in DHR process

Impact / Risk Assessment: *(Determine the impact/risk of the problem/incident or potential problem/incident on the product, processes, systems, documentation, etc. Determine the level of severity and probability of each risk using QP-004, Risk Assessment)*

Missing inspection records may lead to noncompliance to ISO 13485. The addition of receiving inspection templates and protocols mitigates this risk.

In the event of a component recall, loss of traceability may increase the scope of product recalls. This risk is a business risk and the probability is low.

Failure to inspect upon receiving may lead to failures in production. Increased failure in production would cause increased scrap. This risk is a business risk only.

Deviations: *(Were there any deviations from the Investigation Plan? If yes, provide justification. If no deviations, indicate "N/A")*

N/A

PHASE 2 – CAPA INVESTIGATION RESULTS & ROOT CAUSE ANALYSIS

Root Cause: (Describe the root cause(s) by performing the following and attaching any supporting documents. Attach any supporting documents. If needed, complete and attach the SCAR Form, FM-0087)

- Identify the problem
- Gather the data
- Analyze the problem
- Solve the problem

Use 5 WHY Analysis, Fishbone Diagram, Pareto Analysis, 6M (Man, Method, Material, Machine, Measurement, Mother Nature).

Fishbone

1. Problem: Component inspection records are not available in Arena
2. People
 - a. Warehousing not trained to updated protocol
3. CHRs
 - a. CHRs are stored in Sharepoint
 - b. CHRs are not reviewed in production
 - i. A CHR is required to enter materials to inventory, but the record is not required to be entered into Arena before inventory is released to production
 - c. CHRs are uploaded as Files and not Items in Arena
4. Fishbowl
 - a. Fishbowl is independent from Arena
 - b. Fishbowl lacks traceability of components received from clients, so most materials are listed as inventory adjustments
 - i. POs are not generated for component traceability
5. Process
 - a. Inspection plans vary between clients
 - b. No responsible party for uploading CHRs
 - c. No work instruction or protocol for Fishbowl inventory management

Root Cause Code(s): (check all that apply)

- ☒ Code 1 – Man – Inadequate Training
- ☐ Code 2 – Man – Procedure Not Used/Used Improperly/Step Omitted
- ☐ Code 3 – Man – Human Error
- ☐ Code 4 – Man – Lack of Qualification, Demonstrated Proficiency
- ☒ Code 5 – Method – Incomplete Procedure/Wrong/Ambiguous/Non-Existent
- ☐ Code 6 – Method – Manufacturing Process Controls
- ☐ Code 7 – Method – Validation/Verification Issue
- ☐ Code 8 – Material – Damaged/Contaminated
- ☐ Code 9 – Material – Invalid Testing
- ☐ Code 10 – Material – Testing not met
- ☐ Code 11 – Material – Expired/Damaged
- ☐ Code 12 – Material – Specification Incomplete/Wrong/Ambiguous
- ☐ Code 13 – Material – Device Biocompatibility
- ☐ Code 14 – Material – Material Compatibility
- ☐ Code 15 – Material – Visual Specification not met/Unusual characteristics
- ☐ Code 16 – Material – Incorrect Storage
- ☐ Code 17 – Machine – Defective Machine/Tool
- ☐ Code 18 – Machine – Machine/Tool Capacity
- ☐ Code 19 – Machine – Operational Failure
- ☐ Code 20 – Machine – Equipment Controls and Calibration
- ☐ Code 21 – Machine – Inadequate Maintenance
- ☐ Code 22 – Environmental Controls – Facilities/Utilities/Clean Room
- ☐ Code 23 – Environmental Controls – Temperature/Humidity
- ☐ Code 24 – Environmental Controls – Cleanliness/Particulate
- ☐ Code 25 – Environmental Controls – Microbial

PHASE 2 – CAPA INVESTIGATION RESULTS & ROOT CAUSE ANALYSIS			
		☐ Code 26 – Other – Documentation Issue ☐ Code 27 – Other – Labeling Issue ☐ Code 28 – Other – Manufacturing Issue ☐ Code 29 – Other – Supplier Issue ☐ Code 30 – Other – Design Issue: _____ ☐ Attachments	
Phase 3 Target Due Date:		05/30/2025	
PHASE 2 – APPROVAL SIGNATURES			
CAPA Owner (Name/Title):	Larry Dacoron	Signature/Date:	 Larry Dacoron (28-May-2025 09:57 PDT)
Executive Management (Name/Title):	Aaron Rogers	Signature/Date:	 Aaron Rogers (23-May-2025 12:20 PDT)
Quality Assurance (Name/Title):	Larry Dacoron	Signature/Date:	 Larry Dacoron (28-May-2025 09:57 PDT)
CAPA EXTENSION REQUEST (☑ N/A if not applicable)			
CAPA Extension Due Date:	N/A		
Justification for Extension:	N/A		
CAPA EXTENSION APPROVAL SIGNATURES (☑ N/A if not applicable)			
CAPA Owner (Name/Title):	N/A	Signature/Date:	N/A
Quality Assurance (Name/Title):	N/A	Signature/Date:	N/A

CAPA-24-005 Phase 2

Final Audit Report

2025-05-28

Created:	2025-05-23 (Pacific Daylight Time)
By:	Jacqueline Kim (jkim@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAAQRDnB8taJgEJEFBQA926c94L-I15SDfQ
Number of Documents:	1
Document page count:	5
Number of supporting files:	0
Supporting files page count:	0

"CAPA-24-005 Phase 2" History

-  Document created by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-05-23 - 12:05:40 PM PDT
-  Document emailed to Larry Dacoron (ldacoron@pathwaymedtech.com) for approval
2025-05-23 - 12:06:17 PM PDT
-  Document emailed to Aaron Rogers (arogers@pathwaymedtech.com) for approval
2025-05-23 - 12:06:17 PM PDT
-  Email viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-05-23 - 12:16:08 PM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email
2025-05-23 - 12:16:54 PM PDT
-  Agreement viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-05-23 - 12:16:54 PM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC
2025-05-23 - 12:20:45 PM PDT
-  Document approved by Aaron Rogers (arogers@pathwaymedtech.com)
Approval Date: 2025-05-23 - 12:20:45 PM PDT - Time Source: server

✓ Larry Dacoron (ldacoron@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email

2025-05-28 - 9:57:28 AM PDT

📄 Agreement viewed by Larry Dacoron (ldacoron@pathwaymedtech.com)

2025-05-28 - 9:57:28 AM PDT

📄 Larry Dacoron (ldacoron@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC

2025-05-28 - 9:57:38 AM PDT

📄 Document approved by Larry Dacoron (ldacoron@pathwaymedtech.com)

Approval Date: 2025-05-28 - 9:57:38 AM PDT - Time Source: server

✓ Agreement completed.

2025-05-28 - 9:57:38 AM PDT