

PHASE 4 – CORRECTIVE AND PREVENTATIVE ACTION RESULT(S) & VERIFICATION OF EFFECTIVENESS (VoE) PLAN(S)

CAPA Result(s): (Attach any supporting documents. If needed, complete the SCAR Form, FM-0087)

1. Train Warehousing personnel to updated receiving protocol(s)

Completed. Warehouse personnel trained on QP-009 Rev H. Refer to Training Plan in Arena.

2. Audit CHRs in Arena to ensure files are used in proper items

Pathway reviewed current inventory required for current client production based on BOMs released in Arena. The BOM inventory was gapped to inventory in Fishbowl. It was discovered that Engineering build inventory and fulfillment only inventory is not required to be tracking in Arena. QP-041 updated per ECO-002263 to reflect inventory requirements for services provided by Pathway.

There were multiple instances of files in Arena that were not released to records. Records were manually released to records by Pathway QA, and via ECO-002249.

Refer to item #4 below for records that were not available.

3. Assign and train personnel to uploading CHR forms once completed

Completed. QA personnel trained on QP-009 Rev H. Refer to Training Plan in Arena.

4. Audit all current inventory for CHR files & create a MEMO for incomplete or missing files

The following component files were not available in Arena:

Part Number	Missing Lot Record
GMD1-V14-12	Lot 5003522217
GMD1-V14-17	LN 22342
GMD1-V14-18	Lot 20250205
GMD1-V14-27	Lot 1955_112624

There were also several instances of missing records for packing materials, but it was determined that the packing material is not critical for inventory tracking. These requirements are reflected in QP-041.

Please refer to the MEMO attached for rationale on missing files.

It was noted that some parts were received with the same lot number.QP-019 Traceability was updated to add sequential numbers to differentiate parts received under the same PO on the same date. Refer to ECO-02270.

CAPA Results(s) Completion Date:

9/12/2025

Deviations: (Add any deviations with justifications)

N/A

Verification of Effectiveness (VoE) Plan(s): (Describe any effectiveness plans)

All CHRs countersigned by QA are released to records weekly for 6 weeks. The records will be gapped to the QA inbox log to ensure all records passed to QA are released in Arena. No CHRs shall be missing for the BOM used in the next 6 DHRs.

Phase 5 Target Due Date:

11/28/2025

PHASE 4 - APPROVAL SIGNATURES

CAPA Owner (Name/Title):

Larry Dacoron

Signature/Date:




Larry Dacoron (11-Sep-2025 15:50:32 PDT)

This document contains confidential, proprietary information of Pathway Medtech and shall not be copied, reproduced, or otherwise disclosed to any third party without prior written permission there from. If any portion of this document is in hardcopy form, it must be considered and treated as uncontrolled "For Reference Only".	Page 10 of 13
--	---------------

FM-0020 rev E

Document Name: CAPA-24-005 Phase 4.pdf

Transaction ID: CBJCHBCAABAAJmWO7LtPN5tL5NIKW10FJI0wFtTZQ-GM

Executive Management (Name/Title):	Aaron Rogers	Signature/Date:	 Aaron Rogers (10-Sep-2025 10:48:51 PDT)
Quality Assurance (Name/Title):	Larry Dacoron	Signature/Date:	 Larry Dacoron (11-Sep-2025 15:50:32 PDT)
CAPA EXTENSION REQUEST (☐ N/A if not applicable)			
CAPA Extension Due Date:	9/12/2025		
Justification for Extension:	Increase in scope to audit to include inventory reconciliation per CAPA-25-001. Additional time requested to complete audit.		
CAPA EXTENSION APPROVAL SIGNATURES (☐ N/A if not applicable)			
CAPA Owner (Name/Title):	Larry Dacoron	Signature/Date:	07/22/2025
Quality Assurance (Name/Title):	Aaron Rogers	Signature/Date:	07/22/2025

Memo

CAPA-24-005 Rationale for Missing CHRs

PURPOSE

The purpose of this MEMO is to provide rationale for missing CHR records in Arena for inventory listed in Fishbowl.

SCOPE

This MEMO applies to inventory active in Fishbowl for active clients in production or DV only.

BACKGROUND

CAPA-24-005 was opened due to multiple audits in 2023 that found receiving inspection documentation was unable to be provided during the audit(s). Per the CAPA investigation, Pathway confirmed that various inventory items did not have proper component history record (CHR) documentation released in Arena. After auditing physical and Sharepoint files, most of the CHRs were found and uploaded to Arena as records under ECO-002249. Several items listed as available in Fishbowl did not have corresponding CHRs. This MEMO provides rationale for those missing records.

DISCUSSION

The following records were unable to be found:

Part Number	Missing Lot Record
GMD1-V14-12	Lot 5003522217
GMD1-V14-17	LN 22342
GMD1-V14-18	Lot 20250205
GMD1-V14-27	Lot 1955_112624

The lots were unable to be located, which implies that the items were already used or removed from inventory without updating Fishbowl. The final quantity in Fishbowl shall be reduced to reflect the actual quantity available. Pathway reviewed the impacted lots to determine risk and rationale for missing records.

The following items are off-the-shelf, so have no inspection requirements: GMD1-V14-12, GMD1-V14-17, GMD1-V14-18, GMD1-V14-27. The risk associated with OTS parts is low.

CONCLUSION



Pathway Medtech, LLC.
8779 Cottonwood Ave, Suite 105
Santee, CA 92079
www.pathwaymedtech.com

The listed items with missing records are low risk (OTS). This MEMO provides appropriate rationale for the missing records. Per CAPA-24-005, all future component history records shall be uploaded and released weekly via ECO.

CAPA-24-005 Phase 4

Final Audit Report

2025-09-11

Created:	2025-09-10 (Pacific Daylight Time)
By:	Jacqueline Kim (jkim@pathwaymedtech.com)
Status:	Approved
Transaction ID:	CBJCHBCAABAAJmWO7LtPN5tL5NIkW1OFJI0wFtTZQ-GM
Number of Documents:	2
Document page count:	4
Number of supporting files:	0
Supporting files page count:	0

"CAPA-24-005 Phase 4" History

-  Document created by Jacqueline Kim (jkim@pathwaymedtech.com)
2025-09-10 - 10:33:47 AM PDT
-  Document emailed to Aaron Rogers (arogers@pathwaymedtech.com) for approval
2025-09-10 - 10:47:34 AM PDT
-  Document emailed to Larry Dacoron (ldacoron@pathwaymedtech.com) for approval
2025-09-10 - 10:47:34 AM PDT
-  Email viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-09-10 - 10:47:53 AM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email
2025-09-10 - 10:48:26 AM PDT
-  Agreement viewed by Aaron Rogers (arogers@pathwaymedtech.com)
2025-09-10 - 10:48:26 AM PDT
-  Aaron Rogers (arogers@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC
2025-09-10 - 10:48:51 AM PDT
-  Document approved by Aaron Rogers (arogers@pathwaymedtech.com)
Approval Date: 2025-09-10 - 10:48:51 AM PDT - Time Source: server

 Agreement viewed by Larry Dacoron (ldacoron@pathwaymedtech.com)

2025-09-11 - 3:37:11 PM PDT

 Larry Dacoron (ldacoron@pathwaymedtech.com) authenticated with Email OTP by verifying one-time code sent via email

2025-09-11 - 3:50:23 PM PDT

 Larry Dacoron (ldacoron@pathwaymedtech.com) has agreed to the terms of use and to do business electronically with Pathway Medtech LLC

2025-09-11 - 3:50:32 PM PDT

 Document approved by Larry Dacoron (ldacoron@pathwaymedtech.com)

Approval Date: 2025-09-11 - 3:50:32 PM PDT - Time Source: server

 Agreement completed.

2025-09-11 - 3:50:32 PM PDT