



2-Way 100% Silicone ClearTract SPT Catheter

Please read all instructions before using this device

Not made with natural rubber latex

Foley Catheter is MR Safe.

Model	Size
211410SPT	14 FR 10 mL
211605SPT	16 FR 5 mL
211610SPT	16 FR 10 mL
211810SPT	18 FR 10 mL

Contraindications

This device is contraindicated in the presence of conditions which create unacceptable risk during catheterization.

Device Description

The 2-Way 100% Silicone ClearTract Catheter is a standard single-use 2-way Foley catheter that is constructed of medical grade silicone with a surface modification. It incorporates two (2) lumens, one for inflation/deflation of the balloon and the other for drainage of the urinary tract.

The surface of the ClearTract Catheter has a surface modification that reduces surface friction.

Intended Use

The 2-Way 100% Silicone ClearTract Catheter is intended for drainage of the urinary tract. Catheterization is generally accomplished by inserting the catheter through the urethra and into the bladder. However, drainage is sometimes accomplished by suprapubic placement or other placement of the catheter, such as nephrostomy tract. Intended population is adults and pediatrics.

Directions

Installation & Inflation

1. When placement of a urethral catheter is contraindicated or unsuccessful, percutaneous suprapubic urinary bladder catheterization or nephrostomy can be used in accordance with the procedures and policies of the hospital or healthcare facility.
2. Prepare application site in a sterile fashion.
3. Visually inspect the catheter for any signs of damage or deterioration prior to use.
4. If a trial inflation is conducted, follow steps 7 & 8 on the sterile field then evacuate the sterile water with the syringe. Ensure the balloon is empty prior to insertion.
5. Apply medical grade lubricant to distal end of the catheter.
6. Gently advance the catheter through the urethra until distal tip is beyond the bladder neck.
7. Confirm the position of the distal tip. The return of urine typically indicates the distal tip is in the bladder.
8. Fill Luer-tipped syringe with sterile water and connect to the inflation valve.
9. Fill the catheter balloon to the labeled volume. Do NOT exceed labeled volume.
10. Slowly withdraw the catheter until resistance is felt. This is a sign the balloon is resting at the bladder neck.



Deflation & Removal

1. Gently connect Luer-tipped syringe to the inflation valve. Gently aspirate while allowing the pressure in the balloon to assist filling the syringe with water occupying the balloon. Remove and reconnect the syringe in the inflation valve if there is no or slow deflation. Rapid deflation could collapse the lumen and prevent adequate deflation of the balloon.
2. If during the removal process the balloon ruptures, ensure all fragments are carefully removed from patient.
3. Discard after use.

Precautions

1. Sterile technique shall be applied during installation and removal of the catheter.
2. It is recommended to be replaced after 7 days of use, however, shall not exceed 30 days.

Cautions

1. Federal law restricts the device to sale by or on the order of a physician.
2. Never clamp the catheter. Use catheter plug or valve if required.

Warnings

1. Single Use ONLY. Do NOT reuse. Reuse may cause infection, illness, injury or death.
2. Read the instructions carefully before using the device.
3. Do not use this device beyond the labeled expiration date.
4. Do not use the device if the packaging is punctured or damaged.
5. Do not force catheter as injury to the urethra can occur.
6. Do not use needle.
7. Do not exceed recommended capacities.

Potential Adverse Events

With all Foley Catheters, the following potential adverse events could occur:

1. Urinary tract infection
2. Perforation or laceration of the urethra
3. Perforation or laceration of the bladder
4. Abrasions

Symbol Description

	Catalog Number		Do not resterilize
	Lot Number		Do not use if package is damaged
	Use by		Prescription Use Only (US)
	Consult instructions for use		Foley Catheter is MR Safe
	Single Use		Manufacturer Contact Information
	Sterilized by Ethylene Oxide		Device is non-pyrogenic



SILQ Technologies Corp.
323 Sunny Isles Blvd, 7th Floor
Sunny Isles Beach, FL 33160
(424) 309-8523