CLOSED EN

UNDERLAY TT

SIZE TABLE

Smart Size Guide

Click here

This table provides an overview of all Underlay TT closed sizes and their product dimensions. Please refer to the following pages for the patient measurement protocols.

UNDERLAY TT LENGTH* *Length from bottom of the product to the top	UD23 23 cm	UD28 28 cm	UD35 35 cm
LG 32 cm			
SH 22 cm			


A **C** at the end of the SKU indicates a closed underlay.

UNDERLAY TT CIRCUMFERENCE*


*Product circumference at bottom of the product

INTENDED FOR:


Sealing ring




Pin



Lanyard



Sleeve



OPEN EN

UNDERLAY TT

SIZE TABLE

Smart Size Guide

Click here

This table provides an overview of all Underlay TT open sizes and their product dimensions. Please refer to the following pages for the patient measurement protocols.


UNDERLAY TT LENGTH* *Length from bottom of the product to the top	UD23 23 cm	UD28 28 cm	UD35 35 cm
XL 37 cm			
LG 32 cm			
MD 27 cm			
SH 22 cm			

UNDERLAY TT CIRCUMFERENCE*


*Product circumference at bottom of the product

INTENDED FOR:


Sealing ring



Pin



Lanyard



CLOSED **EN**

Compatible with pin, lanyard, sealing ring and sleeve suspension systems.

UNDERLAY TT

SIZING CHART


To determine the most suitable size for the patient, measure the circumference at 4 cm from the distal end over the skin. Then measure the length of the trimmed liner.

Smart Size Guide

Click here


STEP 1
Choose suspension type.

Suspension
Compatible with all of the below



STEP 2
Measure circumference.


Circumference
Measure at **4** cm from the distal end (over the skin)



Measurement	23-28 cm	28-35 cm	35-42 cm
SKU	23	28	35

STEP 3
Measure length.


Length
Measure the length of the trimmed liner



Measurement	27-37 cm	37+ cm
SKU	SH	LG

Suspension	Circumference	Length	SKU
Pin Lanyard Sealing ring Sleeve	23-28 cm	27-37 cm	UDTT-23-SH-C
	28-35 cm		UDTT-28-SH-C
	35-42 cm		UDTT-35-SH-C
	23-28 cm	37+ cm	UDTT-23-LG-C
	28-35 cm		UDTT-28-LG-C
	35-42 cm		UDTT-35-LG-C

Important:



Underlay Closed must sit at least **5 cm (2 in)** below the top of the liner for suspension.

ETHNOCARE™

Please see next page for **OPEN** sizing

OPEN **EN**

Compatible with pin, lanyard and sealing ring suspension systems.

UNDERLAY TT

SIZING CHART

To determine the most suitable size for the patient, measure two circumferences over the skin: at 10 cm from the distal end and just above the trimmed liner. Then measure the length of the trimmed liner.


Smart Size Guide

Click here

STEP 1

Choose suspension type.


Suspension
Compatible with all of the below



STEP 2

Measure distal circumference (C_1).

Circumference (C_1)
Measure at **10** cm from the distal end (over the skin)




Measurement	23-28 cm	28-35 cm	35-42 cm
SKU	23	28	35

STEP 3

Measure proximal circumference (C_2).

Circumference (C_2)
Measure just above the top edge of the trimmed liner (over the skin)



Starting with your **C_1** measurement, measure **C_2** to confirm whether the limb shape is compatible with **Underlay Open**.


If **C_2** is much larger than **C_1**, the limb shape is too conical. Choose the **Underlay Closed** instead.

STEP 2	If **C_1** measured **Size 23**, and:		If **C_1** measured **Size 28**, and:		If **C_1** measured **Size 35**, and:	
Measurement	**C_2** is less than 35 cm	**C_2** is greater than 35 cm	**C_2** is less than 45 cm	**C_2** is greater than 45 cm	**C_2** is less than 55 cm	**C_2** is greater than 55 cm
SKU	Keep **Size 23**	Underlay Closed	Keep **Size 28**	Underlay Closed	Keep **Size 35**	Underlay Closed

STEP 4

Measure length.

Length
Measure the length of the trimmed liner




Measurement	23-28 cm	28-33 cm	33-38 cm	38+ cm
SKU	SH	MD	LG	XL

Suspension	Circumference (C_1)	Length	SKU
Pin Lanyard Sealing ring	23-28 cm	23-28 cm	UDTT-23-SH
	28-35 cm		UDTT-28-SH
	35-42 cm		UDTT-35-SH
	23-28 cm	28-33 cm	UDTT-23-MD
	28-35 cm		UDTT-28-MD
	35-42 cm		UDTT-35-MD

Suspension	Circumference (C_1)	Length	SKU
Pin Lanyard Sealing ring	23-28 cm	33-38 cm	UDTT-23-LG
	28-35 cm		UDTT-28-LG
	35-42 cm		UDTT-35-LG
	23-28 cm	38+ cm	UDTT-23-XL
	28-35 cm		UDTT-28-XL
	35-42 cm		UDTT-35-XL

Important:



Underlay Open must sit at least **10 cm (4 in)** from the distal end for suspension.

ETHNOCARE™

Please see previous page for **CLOSED** sizing