인증번호(Certificate No.) : 제KTR-2025-05호

성능인증서
(Certificate of performance for PM-2.5 air sensor)

상호 (Company Name)	㈜에프티랩	
대표자 성명 (Representative Name)	고 재 준	
사업장 소재지 (Address/Plant)	경기도 안산시 단원구 해봉로 330번길 8, 정우벤처1차 503호 (전화번호(Tel): 070-4906-4702)	
인증내용 (Certification Detail)	**제작자(Manufacturer)** ㈜에프티랩	**제작국가(Country of manufacture)** 대한민국
	기기명칭(Device name) 미세먼지 간이측정기	**측정방식(Measurement method)** 광산란 방식
	상품명(고유명칭, Product name) DSM101	**측정범위(Measurement range)** 최소(Minimum): 0 μg/m^3 최대(Maximum): 200 μg/m^3
	최소눈금 Minimum resolution (단위, unit) 1 μg/m^3	**공인측정 오차범위(Error range)** 정확도 : 90.1 %
	성능인증 등급(Grade of performance certification) ■ 1등급(Grade of 1)* □ 등급외(Out of Grade) * A sensor, which is certified as Grade 1, shows over 80 % accuracy by matching the reference monitoring data	

「미세먼지 저감 및 관리에 관한 특별법」 제24조제1항 및 같은 법 시행규칙 제16조제3항에 따라 위와 같이 성능인증서를 발급합니다.

This is to issue the above mentioned Certificate of Performance in accordance with Article 24-1 of the「Special Act on the Reduction and Management of Fine Dust」 and Article 16-3 of the Enforcement Regulations of the same Act.

2025년 04월 24일
(Year / Month / Date)

KTR 한국화학융합시험연구원장

Head of Korea Testing & Research Institute