

Durchlasstest

Hersteller

Produktname

Probenname


Datum

VIP Mask

S013_2

2020-05-22 13:53:44.610000

Anzahl Regenerationszyklen	1.0
Differenzdruck Maske	0.39 mbar
Volumenstrom	94.2 L/min
Nozzle Pressure	200.0 mbar
NaCl Konzentration	3.0 w%
Setup	Maskenteststand v1.0
Charge	-
Durchlass	76.31
Aerosolkonzentration	22.76 mg/m ³
Messzeit	180 sec



Optische Mängel:

Holes around the mask edges are not completely closed, but it probably does not affect the measurement

Anmerkungen: