

A Prospective, Multicentre Pivotal Study of the HeartLoop Implantable Cardiac Monitor

Protocol CRD-07, Revision C

FICTIONAL EXAMPLE PROTOCOL - FOR SOFTWARE DEMONSTRATION ONLY. All names and details are invented.

Primary endpoint: detection of atrial fibrillation episodes over 12 months.

5. Investigators and Training

Only physicians who have completed HeartLoop implant training and are listed on the delegation log may perform implants.

An implant performed by a physician who is not trained and listed is an important protocol deviation.

6. Device Accountability

The serial number of every device received, implanted, explanted or returned must be recorded in the device accountability log.

8. Follow-up

Follow-up visits take place at 3, 6 and 12 months after implant, each within plus or minus 14 days.

10. Safety Reporting

Serious adverse events and device deficiencies that could have led to a serious adverse event must be reported to the sponsor within 3 calendar days of the site becoming aware of them.