

Proton Dose Calculation with LSTM Networks in a Magnetic Field

ICCR 2024 | Lyon | 10.07.2024

Fan Xiao¹, Domagoj Radonic^{1,2}, Niklas Wahl^{3,4}, Luke Voss^{3,4,5}, Ahmad Neishabouri^{4,6}, Nikolaos Delopoulos¹,
Stefanie Corradini¹, Claus Belka^{1,7,8}, George Dedes², Christopher Kurz¹, Guillaume Landry¹

¹Department of Radiation Oncology, LMU University Hospital, LMU Munich, Munich, Germany

²Department of Medical Physics, LMU Munich, Munich, Germany

³Department of Medical Physics in Radiation Oncology, German Cancer Research Center (DKFZ), Heidelberg, Germany

⁴National Center for Radiation Oncology (NCRO), Heidelberg Institute for Radiation Oncology (HIRO), Heidelberg, Germany.

⁵Ruprecht Karl University of Heidelberg, Institute of Computer Science, Heidelberg, Germany

⁶Clinical Cooperation Unit Radiation Oncology, German Cancer Research Center (DKFZ), Heidelberg, Germany

⁷German Cancer Consortium (DKTK), partner site Munich, a partnership between DKFZ and LMU University Hospital, Munich, Germany

⁸Bavarian Cancer Research Center (BZKF), Munich, Germany

Background

- Proton therapy:

- Superior dose conformity compared with photon therapy
- Limitations in achieving high-precision due to range uncertainty



Need for real-time
image-guidance
during PT

- MRI:


- High soft tissue contrast, non-ionizing radiation, real-time imaging





Combination of MRI
and PT

- Online MR-guided adaptive proton therapy:

- Treatment plan can be **adapted or re-optimized** based on the latest patient geometry
- Improve the target dose coverage and reduce the dose to surrounding normal tissues



Online plan re-optimization
based on MRI





Online plan re-optimization
needs accurate and fast
dose calculation

Images courtesy of Netherlands Cancer Institute, NKI

Background

- Proton dose calculation in a **magnetic field**:
 - Lorentz force causes charged protons to be deflected
- Monte Carlo (MC) methods:
 - Accurate but speed is usually slow
 - GPU MC methods are still in the level of several seconds
- Analytical algorithms:
 - Speed can be closer to real-time
 - Lack the accuracy of MC methods, especially in inhomogeneous tissues



1.5T B field

Ultimate need for accurate
real-time plan re-optimization
is still unmet!






Dose calculation with MC
accuracy at sub-second speeds


Background: DL dose calculation

Category1: Convert low-cost physical inputs to MC dose



PB->MC^[1]



Noisy MC dose



MC denoising^[2]



[1] Wu C 2021 Improving Proton Dose Calculation Accuracy by Using Deep Learning *Mach Learn Sci Technol*


[2] Javaid U 2021 Denoising proton therapy Monte Carlo dose distributions in multiple tumor sites: A comparative neural networks architecture study *Phys Med*

Background: DL dose calculation


Category2: Slice sequence modelling from Beam's Eye View

- Models focus on the individual beamlet from BEV
- Extract the beamlet and use LSTM or transformer models to process dose slice sequence modelling

LSTM method^[3]



Transformer method^[4]




[3] Neishabouri A 2021 Long short-term memory networks for proton dose calculation in highly heterogeneous tissues *Med Phys*

[4] Pastor-Serrano O 2022 Millisecond speed deep learning based proton dose calculation with Monte Carlo accuracy *Phys Med Biol*

Methods & Materials


Dataset

- Patients:
 - 35 prostate patients treated at MRIdian in LMU Hospital
 - Deformed CTs (from planning CT-MRI registration)
 - Resampled voxel size: $1.5 \times 1.5 \times 1.5 \text{ mm}^3$
 - Dataset split: training/validation/testing=20/5/10
- Geant4 MC dose simulation:
 - Convert CTs to **RSPs** with respect to water using Geant4
 - **1.5T** B field within a cylinder of a 30 cm radius
 - Number of histories: 1M
 - Angular region: $[40^\circ, 140^\circ] \cup [220^\circ, 320^\circ]$, $\Delta\alpha = 16.6^\circ$
 - 216 beamlets were simulated per patient
 - 3 beamlet datasets for 3 energies (**150/175/200 MeV**)
 - Cuboid extraction from a predefined distance before beamlet origin




Methods & Materials

Single energy (SE) model




Methods & Materials

Multi-energy (ME) model



Energy Embedding layer



Methods & Materials


Model training

- Models
 - **3 single energy models** were trained **separately** on 150/175/200 MeV datasets
 - LSTM + two fully connected layers
 - **One multi-energy model** were trained on a **combined** 150/175/200 MeV dataset
 - Energy embedding layer + LSTM + two fully connected layers
- Training parameters
 - Optimizer: Adam, Loss: MSE, LR: 1e-5, batchsize: 8
 - Hardware: NVIDIA RTX A6000 GPU (48 GB)
 - Training GPU memory cost: within **1 GB**
 - Training time: 5 days

Methods & Materials



Metrics

- 10 test patients in 150/175/200 MeV datasets (216 beamlets per patient)
- 3D dose distribution comparison:
 - 3D global gamma passing rate γ_{PR} (2%/2 mm, $D > 10\% D_{max}$)
- Dose range comparison:
 - Laterally integrated **dose-depth profiles** of two doses with relative error ϵ_{rel}
 - Depth difference ΔR_{D80} of the distal dose falloff to 80% of Bragg peak (mm&%)
- Magnetic dose deflection comparison:
 - Lateral profiles of two doses
 - Shift difference in the center of mass F_{COM} at each depth of two doses (mm)



Results


Typical examples of SE models




Results

Metrics comparison between SE models and ME model


150 MeV



175 MeV



200 MeV



		γ_{PR} [%]		ΔR_{D80} [%]		F_{COM} [mm]	
Model	E[MeV]	min	mean	max	mean	max	mean
SE	150	95.28	99.88	2.70	0.28	2.15	0.39
	175	94.21	99.87	3.03	0.23	1.84	0.34
	200	94.38	99.88	3.15	0.19	2.03	0.30
ME	150	94.78	99.82	4.59	0.29	2.06	0.45
	175	94.74	99.83	3.68	0.22	2.09	0.37
	200	89.69	99.78	2.54	0.17	5.36	0.35

Results

Runtime

- Total runtime:
 - Measurement device:
an Intel(R) Xeon(R) Gold 6354 3.00GHz CPU & an NVIDIA RTX A6000 GPU
 - Cuboid extraction time: 55ms
 - Model inference time:


Model	Mean (SD) [ms]
SE (150 MeV)	9 (1)
SE (175 MeV)	9 (1)
SE (200 MeV)	10 (1)
ME	10 (2)

Appendix

Preliminary experiment for an energy range (ER) model

- The previous training/validation dataset for a single energy:
 - 216 beamlets per patient, the energy of each beamlet is fixed (e.g., 150 MeV)
- **A larger training/validation dataset for an energy range** (based on the same patients) :
 - 216×4 beamlets for each patient, the energy of each beamlet is randomly selected from **125-200MeV** (76 integers)
 - Test results of the ER model on previous 150/175/200 MeV test datasets:

Model	E[MeV]	γ_{PR} [%]		ΔR_{D80} [%]		F_{COM} [mm]	
		min	mean	max	mean	max	mean
ME (ER)	150	93.32	99.76	3.83	0.35	2.73	0.53
	175	92.09	99.68	3.51	0.27	2.04	0.55
	200	89.89	99.54	3.62	0.28	3.14	0.67



Conclusion

- We developed an LSTM-based method for proton dose calculation in a magnetic field
- An extended LSTM model for multi-energy dose calculation was proposed and showed its feasibility

Outlook

- Full plan test is ongoing
- Extend more sites and spot sizes
- Introduce uncertainties quantifying method into the model, e.g., Bayesian neural networks

Thank you for your attention!
Acknowledgements



Contact: Fan.Xiao@med.uni-muenchen.de