

## ⟨1225⟩ VALIDATION OF COMPENDIAL PROCEDURES

Test procedures for assessment of the quality levels of pharmaceutical articles are subject to various requirements. According to Section 501 of the Federal Food, Drug, and Cosmetic Act, assays and specifications in monographs of the *USP–NF* constitute legal standards. The Current Good Manufacturing Practice regulations [21 CFR 211.194(a)] require that test methods, which are used for assessing compliance of pharmaceutical articles with established specifications, must meet proper standards of accuracy and reliability. Also, according to these regulations [21 CFR 211.194(a)(2)], users of analytical methods described in *USP–NF* are not required to validate the accuracy and reliability of these methods, but merely verify their suitability under actual conditions of use. Recognizing the legal status of *USP* and *NF* standards, it is essential, therefore, that proposals for adoption of new or revised compendial analytical procedures be supported by sufficient laboratory data to document their validity.

The text of this information chapter harmonizes, to the extent possible, with the International Council for Harmonisation (ICH) tripartite guideline *Validation of Analytical Procedures* and the *Methodology* extension text, which are concerned with analytical procedures included as part of registration applications submitted within the EC, Japan, and the USA.

### SUBMISSIONS TO THE COMPENDIA

Submissions to the compendia for new or revised analytical procedures should contain sufficient information to enable members of the USP Council of Experts and its Expert Committees to evaluate the relative merit of proposed procedures. In most cases, evaluations involve assessment of the clarity and completeness of the description of the analytical procedures, determination of the need for the procedures, and documentation that they have been appropriately validated. Information may vary depending upon the type of method involved. However, in most cases a submission will consist of the following sections.

#### Rationale

This section should identify the need for the procedure and describe the capability of the specific procedure proposed and why it is preferred over other types of determinations. For revised procedures, a comparison should be provided of limitations of the current compendial procedure and advantages offered by the proposed procedure.

#### Proposed Analytical Procedure

This section should contain a complete description of the analytical procedure sufficiently detailed to enable persons “skilled in the art” to replicate it. The write-up should include all important operational parameters and specific instructions such as preparation of reagents, performance of system suitability tests, description of blanks used, precautions, and explicit formulas for calculation of test results.

#### Data Elements

This section should provide thorough and complete documentation of the validation of the analytical procedure. It should include summaries of experimental data and calculations substantiating each of the applicable analytical performance characteristics. These characteristics are described in the following section.

### VALIDATION

Validation of an analytical procedure is the process by which it is established, by laboratory studies, that the performance characteristics of the procedure meet the requirements for the intended analytical applications. Typical analytical performance characteristics that should be considered in the validation of the types of procedures described in this document are listed in *Table 1*. Because opinions may differ with respect to terminology and use, each of the performance characteristics is defined in the next section of this chapter, along with a delineation of a typical method or methods by which it may be measured. The definitions refer to “test results”. The description of the analytical procedure should define what the test results for the procedure are. As noted in ISO 5725-1 and 3534-1, a test result is “the value of a characteristic obtained by carrying out a specified test method. The test method should specify that one or a number of individual measurements be made, and their average, or another appropriate function (such as the median or the standard deviation), be reported as the test result. It may also require standard corrections to be applied, such as correction of gas volumes to standard temperature and pressure. Thus, a test result can be a result calculated from several observed values. In the simple case, the test result is the observed value itself.” A test result also can be, but need not be, the final, reportable value that would be compared to the acceptance criteria of a specification. Validation of physical property methods may involve the assessment of chemometric models. However, the typical analytical characteristics used in method validation can be applied to the methods derived from the use of the chemometric models.

**Table 1. Typical Analytical Characteristics Used in Method Validation**

Accuracy
Precision

**Table 1. Typical Analytical Characteristics Used in Method Validation** (*continued*)

Specificity
Detection limit
Quantitation limit
Linearity
Range
Robustness

The effects of processing conditions and potential for segregation of materials should be considered when obtaining a representative sample to be used for validation of procedures.

In the case of compendial procedures, revalidation may be necessary in the following cases: a submission to the USP of a revised analytical procedure or the use of an established general procedure with a new product or raw material (see below in *Data Elements Required for Validation*).

The ICH documents give guidance on the necessity for revalidation in the following circumstances: changes in the synthesis of the drug substance, changes in the composition of the drug product, and changes in the analytical procedure.

This chapter is intended to provide information that is appropriate to validate a wide range of compendial analytical procedures. The validation of compendial procedures may use some or all of the suggested typical analytical characteristics used in method validation as outlined in *Table 1* and categorized by type of analytical method in *Table 2*. For some compendial procedures the fundamental principles of validation may extend beyond characteristics suggested in this chapter. For these procedures the user is referred to the individual compendial chapter for those specific analytical validation characteristics and any specific validation requirements.

## Analytical Performance Characteristics

### ACCURACY

**Definition:** The accuracy of an analytical procedure is the closeness of test results obtained by that procedure to the true value. The accuracy of an analytical procedure should be established across its range. [A note on terminology: The definition of accuracy in this chapter and ICH Q2 corresponds to unbiasedness only. In the International Vocabulary of Metrology (VIM) and documents of the International Organization for Standardization (ISO), “accuracy” has a different meaning. In ISO, accuracy combines the concepts of unbiasedness (termed “trueness”) and precision.]

**Determination:** In the case of the assay of a drug substance, accuracy may be determined by application of the analytical procedure to an analyte of known purity (e.g., a Reference Standard) or by comparison of the results of the procedure with those of a second, well-characterized procedure, the accuracy of which has been stated or defined.

In the case of the assay of a drug in a formulated product, accuracy may be determined by application of the analytical procedure to synthetic mixtures of the drug product components to which known amounts of analyte have been added within the range of the procedure. If it is not possible to obtain samples of all drug product components, it may be acceptable either to add known quantities of the analyte to the drug product (i.e., “to spike”) or to compare results with those of a second, well-characterized procedure, the accuracy of which has been stated or defined.

In the case of quantitative analysis of impurities, accuracy should be assessed on samples (of drug substance or drug product) spiked with known amounts of impurities. Where it is not possible to obtain samples of certain impurities or degradation products, results should be compared with those obtained by an independent procedure. In the absence of other information, it may be necessary to calculate the amount of an impurity based on comparison of its response to that of the drug substance; the ratio of the responses of equal amounts of the impurity and the drug substance (relative response factor) should be used, if known.

Accuracy is calculated as the percentage of recovery by the assay of the known added amount of analyte in the sample, or as the difference between the mean and the accepted true value, together with confidence intervals.

The ICH documents recommend that accuracy should be assessed using a minimum of nine determinations over a minimum of three concentration levels, covering the specified range (i.e., three concentrations and three replicates of each concentration).

Assessment of accuracy can be accomplished in a variety of ways, including evaluating the recovery of the analyte (percent recovery) across the range of the assay, or evaluating the linearity of the relationship between estimated and actual concentrations. The statistically preferred criterion is that the confidence interval for the slope be contained in an interval around 1.0, or alternatively, that the slope be close to 1.0. In either case, the interval or the definition of closeness should be specified in the validation protocol. The acceptance criterion will depend on the assay and its variability and on the product. Setting an acceptance criterion based on the lack of statistical significance of the test of the null hypothesis that the slope is 1.0 is not an acceptable approach.

Accuracy of physical property methods may be assessed through the analysis of standard reference materials, or alternatively, the suitability of the above approaches may be considered on a case-by-case basis.

### PRECISION

**Definition:** The precision of an analytical procedure is the degree of agreement among individual test results when the procedure is applied repeatedly to multiple samplings of a homogeneous sample. The precision of an analytical procedure is usually expressed as the standard deviation or relative standard deviation (coefficient of variation) of a series of measurements. Precision may be a measure of either the degree of reproducibility or of repeatability of the analytical procedure under normal operating conditions. In this context, reproducibility refers to the use of the analytical procedure in different laboratories, as

in a collaborative study. Intermediate precision (also known as ruggedness) expresses within-laboratory variation, as on different days, or with different analysts or equipment within the same laboratory. Repeatability refers to the use of the analytical procedure within a laboratory over a short period of time using the same analyst with the same equipment.

**Determination:** The precision of an analytical procedure is determined by assaying a sufficient number of aliquots of a homogeneous sample to be able to calculate statistically valid estimates of standard deviation or relative standard deviation (coefficient of variation). Assays in this context are independent analyses of samples that have been carried through the complete analytical procedure from sample preparation to final test result.

The ICH documents recommend that repeatability should be assessed using a minimum of nine determinations covering the specified range for the procedure (i.e., three concentrations and three replicates of each concentration) or using a minimum of six determinations at 100% of the test concentration.

## SPECIFICITY

**Definition:** The ICH documents define specificity as the ability to assess unequivocally the analyte in the presence of components that may be expected to be present, such as impurities, degradation products, and matrix components. Lack of specificity of an individual analytical procedure may be compensated by other supporting analytical procedures. [NOTE—Other reputable international authorities (IUPAC, AOAC-I) have preferred the term “selectivity”, reserving “specificity” for those procedures that are completely selective.] For the tests discussed below, the above definition has the following implications.

*Identification tests:* Ensure the identity of the analyte.

*Purity tests:* Ensure that all of the analytical procedures performed allow an accurate statement of the content of impurities of an analyte (e.g., related substances test, heavy metals limit, or organic volatile impurities).

*Assays:* Provide an exact result, which allows an accurate statement on the content or potency of the analyte in a sample.

**Determination:** In the case of qualitative analyses (identification tests), the ability to select between compounds of closely related structure that are likely to be present should be demonstrated. This should be confirmed by obtaining positive results (perhaps by comparison to a known reference material) from samples containing the analyte, coupled with negative results from samples that do not contain the analyte and by confirming that a positive response is not obtained from materials structurally similar to or closely related to the analyte.

In the case of analytical procedures for impurities, specificity may be established by spiking the drug substance or product with appropriate levels of impurities and demonstrating that these impurities are determined with appropriate accuracy and precision.

In the case of the assay, demonstration of specificity requires that it can be shown that the procedure is unaffected by the presence of impurities or excipients. In practice, this can be done by spiking the drug substance or product with appropriate levels of impurities or excipients and demonstrating that the assay result is unaffected by the presence of these extraneous materials.

If impurity or degradation product standards are unavailable, specificity may be demonstrated by comparing the test results of samples containing impurities or degradation products to a second well-characterized procedure (e.g., a pharmacopeial or other validated procedure). These comparisons should include samples stored under relevant stress conditions (e.g., light, heat, humidity, acid/base hydrolysis, and oxidation). In the case of the assay, the results should be compared; in the case of chromatographic impurity tests, the impurity profiles should be compared.

The ICH documents state that when chromatographic procedures are used, representative chromatograms should be presented to demonstrate the degree of selectivity, and peaks should be appropriately labeled. Peak purity tests (e.g., using diode array or mass spectrometry) may be useful to show that the analyte chromatographic peak is not attributable to more than one component.

For validation of specificity for qualitative and quantitative determinations by spectroscopic methods, chapters related to topics such as near-infrared spectrophotometry, Raman spectroscopy, and X-ray powder diffraction should be consulted.

## DETECTION LIMIT

**Definition:** The detection limit is a characteristic of limit tests. It is the lowest amount of analyte in a sample that can be detected, but not necessarily quantitated, under the stated experimental conditions. Thus, limit tests merely substantiate that the amount of analyte is above or below a certain level. The detection limit is usually expressed as the concentration of analyte (e.g., percentage or parts per billion) in the sample.

**Determination:** For noninstrumental procedures, the detection limit is generally determined by the analysis of samples with known concentrations of analyte and by establishing the minimum level at which the analyte can be reliably detected.

For instrumental procedures, the same approach may be used as for noninstrumental procedures. In the case of procedures submitted for consideration as official compendial procedures, it is almost never necessary to determine the actual detection limit. Rather, the detection limit is shown to be sufficiently low by the analysis of samples with known concentrations of analyte above and below the required detection level. For example, if it is required to detect an impurity at the level of 0.1%, it should be demonstrated that the procedure will reliably detect the impurity at that level.

In the case of instrumental analytical procedures that exhibit background noise, the ICH documents describe a common approach, which is to compare measured signals from samples with known low concentrations of analyte with those of blank samples. The minimum concentration at which the analyte can reliably be detected is established. Typically acceptable signal-to-noise ratios are 2:1 or 3:1. Other approaches depend on the determination of the slope of the calibration curve and the standard deviation of responses. Whatever method is used, the detection limit should be subsequently validated by the analysis of a suitable number of samples known to be near, or prepared at, the detection limit.

## QUANTITATION LIMIT

**Definition:** The quantitation limit is a characteristic of quantitative assays for low levels of compounds in sample matrices, such as impurities in bulk drug substances and degradation products in finished pharmaceuticals. It is the lowest amount of analyte in a sample that can be determined with acceptable *Precision* and *Accuracy* under the stated experimental conditions. The quantitation limit is expressed as the concentration of analyte (e.g., percentage or parts per billion) in the sample.

**Determination:** For noninstrumental procedures, the quantitation limit is generally determined by the analysis of samples with known concentrations of analyte and by establishing the minimum level at which the analyte can be determined with acceptable *Accuracy* and *Precision*.

For instrumental procedures, the same approach may be used as for noninstrumental procedures. In the case of procedures submitted for consideration as official compendial procedures, it is almost never necessary to determine the actual quantitation limit. Rather, the quantitation limit is shown to be sufficiently low by the analysis of samples with known concentrations of analyte above and below the quantitation level. For example, if it is required that an analyte be assayed at the level of 0.1 mg/tablet, it should be demonstrated that the procedure will reliably quantitate the analyte at that level.

In the case of instrumental analytical procedures that exhibit background noise, the ICH documents describe a common approach, which is to compare measured signals from samples with known low concentrations of analyte with those of blank samples. The minimum concentration at which the analyte can reliably be quantified is established. A typically acceptable signal-to-noise ratio is 10:1. Other approaches depend on the determination of the slope of the calibration curve and the standard deviation of responses. Whatever approach is used, the quantitation limit should be subsequently validated by the analysis of a suitable number of samples known to be near, or prepared at, the quantitation limit.

## LINEARITY AND RANGE

**Definition of linearity:** The linearity of an analytical procedure is its ability to elicit test results that are directly, or by a well-defined mathematical transformation, proportional to the concentration of analyte in samples within a given range. Thus, in this section, "linearity" refers to the linearity of the relationship of concentration and assay measurement. In some cases, to attain linearity, the concentration and/or the measurement may be transformed. [NOTE—The weighting factors used in the regression analysis may change when a transformation is applied.] Possible transformations may include log, square root, or reciprocal, although other transformations are acceptable. If linearity is not attainable, a nonlinear model may be used. The goal is to have a model, whether linear or nonlinear, that describes closely the concentration–response relationship.

**Definition of range:** The range of an analytical procedure is the interval between the upper and lower levels of analyte (including these levels) that have been demonstrated to be determined with a suitable level of precision, accuracy, and linearity using the procedure as written. The range is normally expressed in the same units as test results (e.g., percent or parts per million) obtained by the analytical procedure.

**Determination of linearity and range:** Linearity should be established across the range of the analytical procedure. It should be established initially by visual examination of a plot of signals as a function of analyte concentration or content. If there appears to be a linear relationship, test results should be established by appropriate statistical methods (e.g., by calculation of a regression line by the method of least squares). Data from the regression line itself may be helpful to provide mathematical estimates of the degree of linearity. The correlation coefficient, y-intercept, slope of the regression line, and residual sum of squares should be submitted.

The range of the procedure is validated by verifying that the analytical procedure provides acceptable precision, accuracy, and linearity when applied to samples containing analyte at the extremes of the range as well as within the range.

ICH recommends that, for the establishment of linearity, a minimum of five concentrations normally be used. It is also recommended that the following minimum specified ranges should be considered:

*Assay of a drug substance (or a finished product):* From 80% to 120% of the test concentration

*Determination of an impurity:* From 50% to 120% of the acceptance criterion

*For content uniformity:* A minimum of 70%–130% of the test concentration, unless a wider or more appropriate range based on the nature of the dosage form (e.g., metered-dose inhalers) is justified

*For dissolution testing:*  $\pm 20\%$  over the specified range (e.g., if the acceptance criteria for a controlled-release product cover a region from 30% after 1 h, and up to 90% after 24 h, the validated range would be 10%–110% of the label claim).

The traditional definition of linearity, i.e., the establishment of a linear or mathematical relationship between sample concentration and response, is not applicable to particle size analysis. For particle size analysis, a concentration range is defined (instrument- and particle size-dependent) such that the measured particle size distribution is not affected by changes in concentration within the defined concentration range. Concentrations below the defined concentration range may introduce an error due to poor signal-to-noise ratio, and concentrations exceeding the defined concentration range may introduce an error due to multiple scattering.

## ROBUSTNESS

**Definition:** The robustness of an analytical procedure is a measure of its capacity to remain unaffected by small but deliberate variations in procedural parameters listed in the procedure documentation and provides an indication of its suitability during normal usage. Robustness may be determined during development of the analytical procedure.

## SYSTEM SUITABILITY

If measurements are susceptible to variations in analytical conditions, these should be suitably controlled, or a precautionary statement should be included in the procedure. One consequence of the evaluation of *Robustness* and ruggedness should be that a series of system suitability parameters is established to ensure that the validity of the analytical procedure is maintained whenever used. Typical variations are the stability of analytical solutions, different equipment, and different analysts. In the case

of liquid chromatography, typical variations are the pH of the mobile phase, the mobile phase composition, different lots or suppliers of columns, the temperature, and the flow rate. In the case of gas chromatography, typical variations are different lots or suppliers of columns, the temperature, and the flow rate.

System suitability tests are based on the concept that the equipment, electronics, analytical operations, and samples to be analyzed constitute an integral system that can be evaluated as such. System suitability test parameters to be established for a particular procedure depend on the type of procedure being evaluated. They are especially important in the case of chromatographic procedures. Submissions to the USP should make note of the requirements in *Chromatography* (621), *System Suitability*.

## Data Elements Required for Validation

Compendial test requirements vary from highly exacting analytical determinations to subjective evaluation of attributes. Considering this broad variety, it is only logical that different test procedures require different validation schemes. This chapter covers only the most common categories of tests for which validation data should be required. These categories are as follows:

### CATEGORY I

Analytical procedures for quantitation of major components of bulk drug substances or active ingredients (including preservatives) in finished pharmaceutical products.

### CATEGORY II

Analytical procedures for determination of impurities in bulk drug substances or degradation compounds in finished pharmaceutical products. These procedures include quantitative assays and limit tests.

### CATEGORY III

Analytical procedures for determination of performance characteristics (e.g., dissolution, drug release, and others).

### CATEGORY IV

Identification tests.

For each category, different analytical information is needed. Listed in *Table 2* are data elements that are normally required for each of these categories. [NOTE—For detailed information regarding the validation of dissolution procedures, see *The Dissolution Procedure: Development and Validation* (1092).]

**Table 2. Data Elements Required for Validation**

Analytical Performance Characteristics	Category I	Category II		Category III	Category IV
		Quantitative	Limit Tests		
Accuracy	Yes	Yes	<sup>a</sup>	<sup>a</sup>	No
Precision	Yes	Yes	No	Yes	No
Specificity	Yes	Yes	Yes	<sup>a</sup>	Yes
Detection limit	No	No	Yes	<sup>a</sup>	No
Quantitation limit	No	Yes	No	<sup>a</sup>	No
Linearity	Yes	Yes	No	<sup>a</sup>	No
Range	Yes	Yes	<sup>a</sup>	<sup>a</sup>	No

<sup>a</sup> May be required, depending on the nature of the specific test.

Already established general procedures (e.g., titrimetric determination of water or bacterial endotoxins) should be verified to establish their suitability for use, such as their accuracy (and absence of possible interference) when used for a new product or raw material.

When validating physical property methods, consider the same performance characteristics required for any analytical procedure. Evaluate use of the performance characteristics on a case-by-case basis, with the goal of determining that the procedure is suitable for its intended use. The specific acceptance criteria for each validation parameter should be consistent with the intended use of the method.

Physical methods may also be classified into the four validation categories. For example, validation of a quantitative spectroscopic method may involve evaluation of *Category I* or *Category II Analytical Performance Characteristics*, depending on the method requirements. Qualitative physical property measurements, such as particle size, surface area, bulk and tapped density, which could impact performance characteristics, often best fit in *Category III*. *Category IV Analytical Performance Characteristics* usually applies to validation of qualitative identification spectroscopic methods. However, the various techniques may be used for different purposes, and the specific use of the method and characteristics of the material being analyzed should be considered when definitively applying a category to a particular type of method.

The validity of an analytical procedure can be verified only by laboratory studies. Therefore, documentation of the successful completion of such studies is a basic requirement for determining whether a procedure is suitable for its intended application(s). Current compendial procedures are also subject to regulations that require demonstration of suitability under actual conditions of use (see *Verification of Compendial Procedures* (1226) for principles relative to the verification of compendial procedures). Appropriate documentation should accompany any proposal for new or revised compendial analytical procedures.

Official