

Lecture 25

3/18/26
JRA

Housekeeping

- Take-home midterm is available
 - Pick up hard copy at end of lecture
 - Download copy from Moodle
- MT has three quantitative problems that are similar to those on practice MT
- Key to practice MT is on Moodle
- MT deadline = Friday, April 3rd, at 5pm

ACTIVITY 8

Did Problem 6 - 11 (the remaining problems)

Ultrasound (continued)

Echo-based imaging

US imaging uses a device called a transducer that both creates & detects US

- Transducer sends an US pulse into patient
- Pulse is partially reflected, partially transmitted at each interface
- Measure go-return times
- Calculate depths from kinematic equation

$$2D = V_{us} \cdot t_{echo}$$


$$D = \frac{V_{us}}{2} \cdot t_{echo}$$

$$V_{us} = 1540 \text{ m/s}$$

Measured

Depth relative to transducer

SIMULATION



- Send series of pulses
- Detect series of sets of reflections (from each interface)
- Calculate series of D_s

Build up contours in image by scanning through sample

Assign locations based on echo time

Assign brightness based on echo strength

Strong echo → Bright/White

Intermediate echo → Variable shades of gray

No echo → Dark/Black

US imaging has a variety of strengths & weaknesses:

Strengths

Very Fast - real time
Relatively inexpensive
Portable
Safe

Weaknesses

Ultrasound Frequencies

Diagnostic US has frequencies, ν , in a range from $\sim 1 \text{ MHz}$ to $\sim 20 \text{ MHz}$

This is about 1000 times the frequency range for hearing

Why does US imaging employ a range of frequencies?

Huge Punchline for US Frequencies

- Use HIGH frequencies to achieve the best resolution

High frequency $\Rightarrow$ Short wavelength $\Rightarrow$ BEST resolution

- Use LOW frequencies to get deeper penetration because high frequency US is dissipated more quickly $\therefore$ so doesn't go as deep

Bottom Line $\left\{ \begin{array}{l} \text{Surface structures / thin samples (e.g.,} \\ \text{Carotid, mammogram)} \Rightarrow \text{use high } \nu \\ \text{Deep structures (e.g., organs like liver)} \\ \Rightarrow \text{use low } \nu \end{array} \right.$

Why use frequencies in the MHz regime?

The resolution goal in medical imaging is $\sim 1\text{mm}$

0.1 mm to 20 mm (2 cm)

$\hookrightarrow$ x-rays

$\hookrightarrow$ radionuclide

Mammography is the highest resolution medical imaging technique

$$\boxed{\text{Resolution} \approx \lambda/2}$$

Rule of Thumb

Need $\lambda \approx 1 \text{ mm}$ for Resolution $\approx 0.5 \text{ mm}$

For waves, v , λ , ν are related by the equation

$$\boxed{v = \lambda \nu}$$

v = wave speed = 1540 m/s

λ = wavelength

ν = frequency

The frequency required for 0.5 mm resolution is 1.5 MHz

$$\nu = \frac{v}{\lambda} = \frac{1540 \text{ m/s}}{1 \cdot 10^{-3} \text{ m}} = 1.54 \text{ MHz} \approx 1.5 \text{ MHz}$$

Thus, need MHz US to get mm resolution

Quantifying Contrast

We know, from above, how to use echo ranging to map feature depths & learn shapes

How do you get contrast in US images?

Get contrast from strength & nature of reflections

Often get impression that US bounces off a smooth structure, but this is not the case


Most body surfaces are rough & create diffuse reflection

Medical imaging is all about interactions of waves w/ matter

Ultrasound Interactions w/ Body


There are three ways US interacts with the body that are of interest in US imaging

(1) Specular Reflection




Reflection off smooth surface.
Easiest to quantify
Example: the diaphragm

(2) Diffuse Reflection



Reflection off rough surface
Multi-directional $\rightarrow$ hard to quantify
More common than specular reflection

(3) Scattering



Reflection off small features
(e.g., red blood cells), smaller than wavelength of US
Multi-directional

Doctor is NOT always interested in quantifying interactions, but can instead interpret image based on appearance
Doctor know what's normal & what's not (& why)

Quantifying Specular Reflection

Specular reflection can be quantified using a "reflection coefficient" defined as the ratio of the reflected intensity to the incident intensity:

$$R = \frac{I_r}{I_i} = \frac{\text{Reflected Intensity}}{\text{Incident Intensity}}$$

This expression can be rewritten in terms of something called the "acoustic impedance", z

Acoustic impedance is the sound analog of the refractive index, n , used w/ light

z measures resistance to motion, like n

$$R = \frac{(z_1 - z_2)^2}{(z_1 + z_2)^2}$$

Interface Med 1 z_1
Med 2 z_2

This equation is an exact analog of the equation that describes reflection at an optical interface, w/ 'z' replacing 'n'

Look at three special cases

(1) Matched Interfaces $z_1 = z_2 \Rightarrow R = 0$

There is no reflection at matched interfaces

This situation is rare

(2) Minimally Mismatched Interfaces $\begin{cases} z_1 \approx z_2 \text{ close} \\ R = 0.1 - 1\% \end{cases}$

There is minimal but still nonzero reflection at minimally mismatched interfaces

This situation is common w/ soft tissues

Looked at z values for a variety of tissues
They were very close

(3) VERY Mismatched Interfaces $\begin{cases} z_1 \neq z_2 \text{ VERY different} \\ R \approx 16 \end{cases}$

There is significant (up to almost total) reflection at very mismatched interfaces

This situation applies to "air to skin" & to a slightly lesser extent to "soft tissue to bone"

Use a "coupling gel" positioned between the transducer & the skin to reduce mismatch and allow US to enter body

Handed out take-home midterm, covering microscopy & w/ 4 simulations on US ✓

© 2026 Bethe A. Scalettar & James R Abney