

A Statistical Perspective of Non-Inferiority Trials

Master of Science in Health Data Science

University of Galway

Áine Glynn

School of Mathematical
and Statistical Sciences

Filip Kłosowski

School of Mathematical
and Statistical Sciences

Supervised by: Prof. John Newell

June 2026

Abstract

Single-arm performance goal studies are widely used in medical device evaluation when a randomised active-control trial is impractical or ethically unjustifiable. The conclusion depends on whether sufficient evidence is provided (using a hypothesis test or confidence interval) to satisfy a specific performance goal. Despite their apparent simplicity, the conclusions of such studies are sensitive to three interconnected methodological choices: the value of the performance goal, the confidence interval method used to construct the decision rule, and the sample size calculation underlying the study design. This thesis examines each of these choices in the context of a retrospective case study of a vascular stent developed for the treatment of femoropopliteal peripheral artery disease.

Performance goals for a 30-day composite safety endpoint and a 12-month Rutherford classification efficacy endpoint were derived through a meta-analysis of three historical femoropopliteal bare nitinol stent studies comprising 116 patients in total. Confidence interval boundaries rather than point estimates were used to account for uncertainty in the historical evidence yielded a maximum acceptable 30-day composite adverse event rate of 12% and a minimum acceptable 12-month Rutherford success proportion of 88%. The limited evidence base produced wide uncertainty around both estimates, illustrating that performance goals carry inherent uncertainty that is not always visible when they are reported as fixed values.

A simulation study evaluated six confidence interval methods for binary proportions across a grid of true event proportions and sample sizes relevant to case study endpoints which are on the margins of a probabilistic range (e.g. near zero). The Wald interval was the only method to consistently fall below the nominal 95% coverage level and was judged unsuitable as a decision tool in this setting. Among the remaining methods, Wilson Score and Jeffreys achieved coverage closest to the nominal level with the narrowest intervals, while Clopper-Pearson, Agresti-Coull and the continuity-corrected Wilson Score were conservative with progressively wider intervals. These differences in coverage behaviour translated into meaningful differences in sample size requirements. Under an approved performance goal of 11%, an assumed true complication proportion of 5% and a target power of 90%, Clopper-Pearson was used as the primary method, requiring 209 evaluable patients at 89.7% achieved power. Among methods meeting the nominal coverage level, the required sample size ranged from 198 patients under Jeffreys to 247 under the continuity-corrected Wilson Score.

An interactive R Shiny application, PG-Power, was developed to implement the sample size framework and make the sensitivity of required sample size to key design assumptions accessible and reproducible. The application supports confidence interval method comparison, power curve visualisation, sensitivity analysis and report generation for single-arm performance goal studies.

The thesis concludes with an exploration of Bayesian historical borrowing as a framework for formally incorporating prior evidence into the sample size calculation, under the assumption that historical and current patient populations are sufficiently exchangeable to justify combining information across studies.

The findings demonstrate that the performance goal value, confidence interval method and sample size procedure each can alter the design requirements and the conclusion of a performance goal study independently of the underlying clinical data. Transparent derivation and pre-specification of these choices in the SAP is essential to the credibility and reproducibility of performance goal studies in medical device evaluation.

Keywords Binomial confidence intervals · Performance goals · Single-arm clinical trials · Medical device evaluation · Sample size calculation · Bayesian historical borrowing

Acknowledgments

We would like to thank Prof. John Newell for his supervision and guidance throughout this project. We also thank the lead scientist involved in the BioMimics 3D trial for providing clinical and regulatory context that informed the case study analysis.

Contents

Acknowledgments	3
1 Introduction	6
1.1 Single Arm Non-inferiority Trials	6
1.2 Performance Goals in Medical Device Studies	6
1.3 Motivation for Study	6
1.4 Case Study: BioMimics 3D Vascular Stent	7
1.5 Study Aims	7
2 Background and Statistical Framework	7
2.1 Superiority, Equivalence and Non-Inferiority	7
2.2 Non-Inferiority Margins and Performance Goals	9
2.3 Confidence Intervals for Proportions	9
2.4 Sample Size in Non-Inferiority Studies	11
2.5 Incorporating Prior Evidence: Bayesian Historical Borrowing	12
2.6 Statistical Analysis Plans in Clinical Trials	13
3 Case Study and Data Sources	13
3.1 BioMimics 3D Vascular Stent	13
3.2 Endpoints	14
3.3 Literature Search and Meta-analysis Data	14
3.4 Derivation of Performance Goals	14
4 Methods	16
4.1 Simulation Study	16
4.2 Sample Size Calculation	17
4.3 Shiny Application: PG-Power	17
4.4 Applying Bayesian Borrowing to the Safety Endpoint	17
5 Results	19
5.1 Meta-Analysis Results	19
5.2 Simulation Study: CI Method Comparison	19
5.3 Sample Size and Confidence Interval Method Comparison	20
5.4 Bayesian Historical Borrowing Results	22
6 Discussion	23
6.1 Interpretation of Main Findings	23
6.1.1 Aim 1: Deriving Evidence-Based Performance Goals Through Meta-Analysis	23
6.1.2 Aim 2: Comparing Confidence Interval Methods for Binary Endpoints	24
6.1.3 Aim 3: Evaluating the Effect of Performance Goal and Interval Choice on Sample Size	24
6.1.4 Aim 4: Developing PG-Power	26
6.1.5 Aim 5: The Value and Limits of Bayesian Borrowing	26
6.1.6 Combined Implications	27
6.2 SAP in Performance Goal Studies	28
6.3 Study Limitations	28
6.4 Directions for Future Research	30
Conclusion	31
Appendix	32
Project GitHub Repository	32
References	33

List of Figures

1	Confidence interval decision rules for superiority, non-inferiority and equivalence trials. Each row shows a representative 95% confidence interval for the treatment difference. The shaded region indicates the null hypothesis zone for each design.	8
2	Example saw-tooth power curve illustrating the discrete stepwise behaviour of binomial power calculations for a single-arm binary endpoint. Power is plotted against sample size n . Within each range of n , power decreases gradually before jumping sharply upward when an additional allowable event becomes permissible.	12
3	Proposed performance goals for the BioMimics 3D vascular stent pivotal study derived from meta-analysis of historical femoropopliteal bare nitinol stent studies.	19
4	Pooled event proportions with 95% confidence intervals for the safety and efficacy endpoints. The diamond denotes the pooled estimate and the dashed red line indicates the derived performance goal.	20
5	Heterogeneity statistics from the random-effects meta-analysis for the 30-day composite safety endpoint (Table A) and the 12-month Rutherford classification efficacy endpoint (Table B).	21
6	Coverage probability curves for six binomial confidence interval methods at $n = 50$, evaluated across true proportions p in 0.001, ..., 0.200. The dashed red line indicates the nominal 95% level.	22
7	Coverage probability curves for six binomial confidence interval methods at $n = 100$, evaluated across true proportions p in 0.001, ..., 0.200. The dashed red line indicates the nominal 95% level.	23
8	Coverage probability curves for all six methods overlaid at $n = 100$. The dashed red line indicates the nominal 95% level.	24
9	Average coverage probability against average expected interval width at $n = 20$. The dashed red line indicates the 95% target coverage.	25
10	Average coverage probability against average expected interval width at $n = 50$. The dashed red line indicates the 95% target coverage.	25
11	Average coverage probability against average expected interval width at $n = 100$. The dashed red line indicates the 95% target coverage.	26
12	Power curve for the BioMimics 3D 30-day composite safety endpoint produced by PG-Power using the Clopper-Pearson method. The green point marks the required evaluable sample size of $n = 209$, at which achieved power was 89.7%. The dashed line indicates the 90% target power. The decision rule is to reject H_0 if no more than 14 of 209 patients experience the composite endpoint.	27
13	Required evaluable sample size and achieved power for six confidence interval methods under the BioMimics 3D safety endpoint design assumptions (performance goal $p_0 = 0.11$, assumed true proportion $p_1 = 0.05$, one-sided $\alpha = 0.025$, target power 90%), produced by PG-Power. Clopper-Pearson was used as the primary method.	27
14	Confidence intervals for the observed event proportion across all six methods at $n = 209$ with 14 observed events, produced by PG-Power under the BioMimics 3D safety endpoint design assumptions. The dashed vertical line indicates the performance goal of 11%. All six methods produce an upper bound below the performance goal at the decision boundary.	28
15	Bayesian power curve for the BioMimics 3D 30-day composite safety endpoint under a Beta(2, 48) prior (performance goal $p_0 = 0.11$, assumed true proportion $p_1 = 0.05$, one-sided $\alpha = 0.025$, decision threshold 0.95, target power 90%). The red point marks the required evaluable sample size of $n = 79$	29
16	Effect of prior weight on the posterior distribution of the true event proportion θ under three Beta priors with the same event ratio (1:24) but increasing effective sample sizes of 5, 50 and 500. The dashed vertical line indicates the performance goal of 0.11. Greater prior weight concentrates the posterior further from the performance goal, reducing the required sample size but increasing sensitivity to the exchangeability assumption.	30

1 Introduction

1.1 Single Arm Non-inferiority Trials

Single Arm non-inferiority trials arise when an effective treatment or device already exists and a placebo-controlled comparison would be unethical, impractical or clinically uninformative (1–3). Rather than demonstrating superiority, the goal is to show that any loss of efficacy associated with the new intervention remains within a clinically acceptable limit (1–3). This limit, known as the non-inferiority margin, represents the largest reduction in efficacy considered tolerable in exchange for other benefits such as improved safety, simpler delivery, lower cost or better patient adherence (1,2,4).

The trial conclusion depends on both the observed treatment effect and the uncertainty surrounding it (1,2,5). In practice, this is assessed by comparing a confidence interval for the treatment effect against the pre-specified margin (1,2,6). For favourable endpoints, where higher values indicate better performance, the lower confidence bound is the relevant quantity, while for unfavourable endpoints the upper bound is used (1,2). This one-sided structure distinguishes two-arm non-inferiority trials from superiority and equivalence designs, and places the choice of margin and confidence interval method at the centre of the trial’s interpretation (1–3,6).

1.2 Performance Goals in Medical Device Studies

Performance goal studies are essentially another word for a single-arm non-inferiority designs and are particularly common in medical device evaluation (1,7,8). Rather than comparing the new intervention against an active control, the observed outcome is compared with a pre-specified benchmark derived from historical evidence and clinical judgement (7,9). This benchmark defines the minimum level of performance considered acceptable for the device under evaluation [(7); (9)]

The performance goal approach is most useful when a randomised active-control trial is difficult to justify or conduct (7,8). However, the absence of a concurrent comparator makes the choice of benchmark especially consequential (7,10). Historical studies may differ from the current trial in ways that affect the validity of the comparison, including differences in the target population, endpoint definitions, follow-up duration and broader clinical practice (7,10). These differences must be carefully considered when constructing and justifying a performance goal.

The direction of the performance goal depends on the nature of the endpoint (7,9). For efficacy endpoints, where higher event proportions reflect better performance, the estimated proportion and its confidence interval must exceed a minimum acceptable level (7,9). For safety endpoints, where lower proportions are desirable, the estimated proportion and its confidence interval must remain below a maximum acceptable level of harm (7,9). In either case, a performance goal should not be treated as a simple numerical cut-off (5,7,9). It represents a clinical and statistical threshold whose validity depends on the endpoint being measured, the quality and consistency of the historical evidence, and the level of uncertainty considered acceptable for regulatory decision-making (7,8,10).

1.3 Motivation for Study

Although non-inferiority and performance goal designs appear straightforward in principle, their conclusions can be sensitive to several methodological choices (1–3). Small variations in any of these decisions can alter the interpretation of the same clinical data and, consequently, the study conclusion (1,2).

This sensitivity is particularly consequential in medical device studies, where single-arm performance goal designs are common and the absence of a concurrent comparator means the historical benchmark carries much of the evidential weight (7,8,11). If the benchmark is too lenient, a device may appear acceptable despite a clinically meaningful loss of performance. If it is too strict, a device that performs well may require an impractically large sample size or fail to meet the criterion regardless (4,7,10). In either case, the benchmark itself shapes the conclusion before a single patient is enrolled.

Previous reviews have raised concerns about how consistently these choices are reported. Piaggio et al. found that only around one-fifth of 332 non-inferiority and equivalence trials published between 1990 and 2000 provided a suitable rationale for the chosen margin (6). A later review of 232 drug trials found that while nearly all reports specified a margin, only 24% explained how it had been determined (6). Together these findings point to a recurring problem: margins and performance goals are often presented as settled values without making the reasoning behind them transparent (2,6).

This study is motivated by that problem. Using the BioMimics 3D vascular stent trial as a case study, it examines how performance goal selection, confidence interval choice and sample size calculation interact to influence conclusions in a single-arm medical device setting (1,5,7).

1.4 Case Study: BioMimics 3D Vascular Stent

The BioMimics 3D vascular stent trial serves as the primary case study throughout this analysis. Rather than attempting to reproduce the original trial, the case study is used as a real medical device setting in which to examine how statistical choices surrounding performance goals influence the interpretation of device performance.

Clinical and regulatory context for the analysis was obtained through meetings with a lead scientist involved in the original trial. These discussions clarified how the device was intended to be evaluated in practice and highlighted that the statistical design could not be reduced to a purely numerical exercise — it had to reflect clinical judgement, regulatory expectations and practical feasibility alongside the available evidence.

1.5 Study Aims

This study has five objectives:

1. To derive evidence-based performance goals for a single-arm medical device study through meta-analysis of comparable historical device studies;
2. To compare confidence interval methods for binary endpoints in terms of coverage probability and interval width;
3. To evaluate how performance goal selection and confidence interval choice affect sample size requirements and non-inferiority conclusions;
4. To develop an interactive Shiny application, PG-Power, to support sample size planning and confidence interval method selection in performance goal studies;
5. To explore the potential role of Bayesian historical borrowing as a means of reducing the required sample size.

2 Background and Statistical Framework

2.1 Superiority, Equivalence and Non-Inferiority

Superiority, equivalence and non-inferiority trials address distinct clinical questions and should not be interpreted interchangeably (12,13). Let p_T denote the success probability in the investigational treatment group and p_C denote the success probability in the comparator group. For a binary success endpoint, the treatment effect is defined as (13,14):

$$\theta = p_T - p_C$$

A superiority trial evaluates whether the investigational treatment performs better than the comparator. The null and alternative hypotheses are (1,12,13):

$$H_0 : \theta \leq 0 \quad H_1 : \theta > 0$$

Superiority is concluded when the lower confidence bound for θ lies above zero (1).

An equivalence trial asks whether the difference between treatments is small enough to be considered clinically unimportant in either direction. Denoting the equivalence margin as Δ_E , the hypotheses are (6,12,13):

$$H_0 : \theta \leq -\Delta_E \text{ or } \theta \geq \Delta_E \quad H_1 : -\Delta_E < \theta < \Delta_E$$

Equivalence is concluded when the confidence interval for θ lies entirely within the equivalence margins (6,12,13):

$$-\Delta_E < \text{CI}_{1-\alpha}(\theta) < \Delta_E$$




Figure 1: Confidence interval decision rules for superiority, non-inferiority and equivalence trials. Each row shows a representative 95% confidence interval for the treatment difference. The shaded region indicates the null hypothesis zone for each design.

A non-inferiority trial is less restrictive, focusing only on differences in the unfavourable direction. Denoting the non-inferiority margin as Δ_{NI} , the largest clinically acceptable loss of efficacy, the hypotheses are (1,2,12,13):

$$H_0 : \theta \leq -\Delta_E, \quad \theta \geq \Delta_E \quad H_1 : -\Delta_E < \theta < \Delta_E$$

Non-inferiority is concluded when the lower confidence bound lies above the margin (1,6,12,13):

$$L_{1-\alpha}(\theta) > -\Delta_{NI}$$

These three designs are often confused in practice, but the distinction matters. Failure to demonstrate superiority does not imply equivalence or non-inferiority — a non-significant superiority test may simply reflect insufficient power or imprecision rather than genuine similarity between treatments (1,2,12). Each design therefore requires its own pre-specified margin, endpoints, analysis population and decision rule, and conclusions should not be switched between frameworks after the data have been collected (5,6,13).

2.2 Non-Inferiority Margins and Performance Goals

The non-inferiority margin is one of the central design choices in a non-inferiority trial (1,2,15). It defines the largest loss of efficacy considered clinically acceptable in exchange for other benefits offered by the investigational treatment (1,4,15). As established in Section 1.1, the selected margin affects the interpretation of the trial, the required sample size and the probability of concluding non-inferiority (1,13,15).

In an active-control non-inferiority trial, the margin is usually defined relative to the established effect of the comparator. In single-arm medical device studies, the same logic is applied through a performance goal, a pre-specified benchmark against which the device is evaluated directly, without a concurrent control group (8,16). For a binary efficacy endpoint, where higher proportions reflect better performance, the hypotheses take the form:

$$H_0 : p \leq p_0 \quad H_1 : p > p_0$$

where p is the true success probability for the investigational device and p_0 is the performance goal. Acceptable performance is concluded when the lower confidence bound exceeds the goal (7,8):

$$L_{1-\alpha}(p) > p_0$$

For an unfavourable safety endpoint, where lower proportions reflect better performance, the direction is reversed:

$$H_0 : p \geq p_0 \quad H_1 : p < p_0$$

and acceptable performance is concluded when the upper confidence bound falls below the goal (1,3,6):

$$U_{1-\alpha}(p) < p_0$$

These rules highlight why the performance goal and confidence interval method are closely linked (7,15). The same observed event proportion can produce different confidence bounds depending on the method used, and when the bound lies close to the performance goal, that choice alone may determine whether the device is judged to meet the criterion (7,17). This dependence motivates the comparison of confidence interval methods in Section 2.3 and the simulation study in Section 4.1 (1,7,17).

2.3 Confidence Intervals for Proportions

In single-arm performance goal studies with binary endpoints, the primary analysis is based on an observed event proportion (7,8). If X patients experience the outcome of interest out of n patients, the observed sample proportion is:

$$\hat{p} = \frac{X}{n}, \quad X \sim \text{Binomial}(n, p)$$

where p is the true underlying event probability. In this setting the confidence interval is not merely a descriptive summary — it forms part of the formal decision rule, and different methods applied to the same data can produce different conclusions, particularly at small sample sizes or extreme proportions (17,18).

Six methods are compared in this study. The **Wald interval** is the simplest, based on a normal approximation to the sampling distribution of $\hat{p}$ (17,18):

$$\hat{p} \pm z_{1-\alpha/2} \sqrt{\frac{\hat{p}(1-\hat{p})}{n}}$$

It is included primarily as a reference, as it is known to perform poorly near the boundaries of 0 and 1 and can produce limits outside $[0, 1]$ (17).

The **Wilson Score interval** improves on the Wald interval by inverting the score test rather than approximating the sampling distribution directly (17):

$$\frac{\hat{p} + \frac{z^2}{2n} \pm z \sqrt{\frac{\hat{p}(1-\hat{p})}{n} + \frac{z^2}{4n^2}}}{1 + \frac{z^2}{n}}$$

where $z = z_{1-\alpha/2}$. It generally achieves better coverage behaviour than the Wald interval, especially near the boundaries (17).

The **Agresti-Coull interval** is closely related to the Wilson interval but constructs the interval around an adjusted proportion (17,19):

$$\tilde{p} = \frac{X + \frac{z^2}{2}}{n + z^2}$$

$$\tilde{p} \pm z_{1-\alpha/2} \sqrt{\frac{\tilde{p}(1-\tilde{p})}{n + z^2}}$$

For a 95% interval this is approximately equivalent to adding two successes and two failures before applying the Wald formula, pulling extreme proportions away from the boundaries and improving stability (17,19).

The **Clopper-Pearson interval** is based directly on the binomial distribution rather than a normal approximation, with bounds obtained from beta distribution quantiles (20):

$$L = B^{-1}\left(\frac{\alpha}{2}; X, n - X + 1\right), \quad U = B^{-1}\left(1 - \frac{\alpha}{2}; X + 1, n - X\right)$$

It is conservative by construction, meaning true coverage typically exceeds the nominal level (11,17). This conservatism is consistent with FDA guidance on the use of exact methods in single-arm medical device studies, where the absence of a concurrent control makes protection of the Type I error rate especially important and conservative interval methods provide a natural safeguard against inflated false positive rates (7,15,21). However, this conservatism comes at a practical cost: wider intervals and larger sample size requirements relative to methods that sit closer to the nominal coverage level (17,19,22).

The **Wilson Score interval with continuity correction**, implemented in R via `prop.test()`, adjusts for the discreteness of the binomial distribution and is the default output of that function (18). It is included because its widespread availability in standard R workflows means it is commonly used in practice, often without deliberate consideration of its properties (17).

The **Jeffreys interval** is Bayesian in construction, derived from the Jeffreys prior for a binomial proportion (17,23):

$$p \sim \text{Beta}\left(\frac{1}{2}, \frac{1}{2}\right)$$

After observing X events in n patients, the posterior is:

$$p \mid X \sim \text{Beta}\left(X + \frac{1}{2}, n - X + \frac{1}{2}\right)$$

The interval is obtained from the posterior quantiles. Despite its Bayesian construction, the Jeffreys interval is frequently evaluated using frequentist coverage properties and performs competitively in simulation studies of binomial intervals (17).

The key performance measure used to compare these methods is coverage probability — the probability that the constructed interval contains the true value of p (17):

$$\text{CP}(p) = P_p(L(X) \leq p \leq U(X))$$

Because binomial outcomes are discrete, coverage fluctuates across values of p and n rather than sitting exactly at the nominal level (17). This is particularly relevant in a performance goal setting, where a method that performs acceptably on average may still undercover at the specific proportion values most relevant to the clinical decision. The simulation study in Section 4.1 evaluates these methods on this basis before they are applied to the BioMimics 3D sample size calculations in Section 5.3.

2.4 Sample Size in Non-Inferiority Studies

For a two-arm non-inferiority study with equal allocation and binary endpoints, an approximate sample size per group is given by (13,24):

$$n = \frac{\left[z_{1-\alpha} \sqrt{2\bar{p}(1-\bar{p})} + z_{1-\beta} \sqrt{p_T(1-p_T) + p_C(1-p_C)} \right]^2}{(p_T - p_C + \Delta_{NI})^2}$$

where n is the required number of patients per group, $\bar{p}$ is the average of the assumed success probabilities, $z_{1-\alpha}$ is the critical value for the one-sided type I error rate, and $z_{1-\beta}$ is the critical value corresponding to the desired power (13,24).

This formula highlights the importance of the distance between the expected treatment effect and the non-inferiority threshold (1,13):

$$p_T - p_C + \Delta_{NI}$$

When this distance is large, the investigational treatment is expected to perform comfortably above the margin and a smaller sample size may be sufficient (13). When it is small, more patients are needed to demonstrate non-inferiority with adequate power, and modest changes in the assumed proportions or margin can substantially affect the required n (4,13).

In a single-arm performance goal study, the same principle applies but performance is compared against a fixed benchmark rather than an active control. The required sample size must provide a sufficiently high probability that the relevant confidence bound satisfies the performance goal, assuming the device performs at the expected true proportion p_1 (7,8):

$$P_{p_1}(L_{1-\alpha}(X, n) > p_0) \geq 1 - \beta$$

where $L_{1-\alpha}(X, n)$ is the lower confidence bound calculated from X successes in n patients and p_0 is the performance goal. Since $X \sim \text{Binomial}(n, p)$, the power for a candidate sample size can be evaluated exactly by summing over all values of X that satisfy the decision rule (13,17):

$$\text{Power}(n) = \sum_{x=0}^n \mathbf{1}\{L_{1-\alpha}(x, n) > p_0\} \binom{n}{x} p_1^x (1-p_1)^{n-x}$$

where $\mathbf{1}\{\cdot\}$ is an indicator function equal to 1 when the confidence interval rule is satisfied and 0 otherwise. The required sample size is the smallest n achieving the target power:

$$n^* = \min \{n : \text{Power}(n) \geq 1 - \beta\}$$

This formulation directly aligns the sample size calculation with the planned primary analysis (7,15). If the final decision will be based on a Wilson, Clopper-Pearson or Jeffreys interval, the sample size calculation should use the same method — otherwise the design and analysis are operating under different statistical rules (7,17).

Power calculations for binomial endpoints follow a discrete, stepwise pattern because the decision criterion depends on an integer number of allowable events. As n increases, the maximum allowable number of events remains fixed across certain ranges and increases only at specific thresholds (17). Within each range, power decreases gradually as n grows; it then jumps upward when an additional event becomes permissible. The result is a saw-tooth power curve that reflects the discrete nature of the binomial endpoint (13,17).

The required sample size can also be estimated through simulation. For each candidate n , a large number of trial datasets are generated under the assumed true proportion p_1 , the selected confidence interval method is applied to each, and power is estimated as the proportion of simulated trials in which the decision rule is satisfied (5,13). The required sample size is the smallest n at which this simulated power meets the target (7,13,17). This approach is implemented in specialist software including PASS (25), nQuery (26) and Cytel East (27), as well as several R packages (28), and is particularly useful because it mirrors the planned analysis directly (7,13).




Figure 2: Example saw-tooth power curve illustrating the discrete stepwise behaviour of binomial power calculations for a single-arm binary endpoint. Power is plotted against sample size n . Within each range of n , power decreases gradually before jumping sharply upward when an additional allowable event becomes permissible.

2.5 Incorporating Prior Evidence: Bayesian Historical Borrowing

The frequentist sample size framework described in Section 4.2 treats each trial as an isolated exercise, making no formal use of evidence accumulated from comparable prior studies. Bayesian historical borrowing offers an alternative by incorporating this prior evidence directly into the analysis through a prior distribution, which is then updated with current trial data to produce a posterior distribution (23,29,30). In medical device studies, where single-arm designs are common and historical data from comparable devices are often available, this approach can meaningfully reduce the number of patients required in the current trial (8,10,29).

For a binary endpoint, a natural and tractable model uses the Beta-Binomial conjugate structure (18,23,31). If historical data contribute α_0 observed events and β_0 observed non-events, the prior distribution is (30):

$$p \sim \text{Beta}(\alpha_0, \beta_0)$$

After observing s events and f non-events in the current trial, the posterior distribution is obtained by conjugate updating (23,30,31):

$$p \mid s, f \sim \text{Beta}(\alpha_0 + s, \beta_0 + f)$$

This posterior combines historical and current evidence into a single updated estimate of the true event probability. The influence of the prior relative to the current data is often summarised by its effective sample size, which approximates the number of current observations the prior is equivalent to (23,30,31). A prior with a large effective sample size concentrates the posterior more tightly around the historical estimate, reducing the number of current patients needed to reach a given level of precision or power (31).

The central assumption underlying any borrowing strategy is exchangeability — the requirement that the historical and current populations are sufficiently similar that combining information across them is scientifically justified (31,32). Where full exchangeability cannot be assumed, partial borrowing approaches allow the historical contribution to be downweighted (30,32). The power prior of Ibrahim and Chen (2000) achieves this by raising the historical likelihood to a discounting factor $a_0 \in [0, 1]$ before forming the prior, with $a_0 = 1$ representing full borrowing and $a_0 = 0$ representing no borrowing (30). Adaptive borrowing

methods take a data-driven approach, adjusting the degree of borrowing based on the observed compatibility between historical and current data — increasing borrowing when the two are consistent and reducing it when they diverge (31,32).

Because borrowing affects operating characteristics including Type I error and power, regulatory guidance recommends that its properties be evaluated through simulation across a range of plausible scenarios before the design is finalised (31–33). The FDA guidance on Bayesian statistics in medical device trials acknowledges the potential for historical borrowing to reduce sample size requirements while emphasising that the assumptions underlying the prior must be clinically and statistically justified (29,31). Section 4.4 applies this framework to the BioMimics 3D safety endpoint using the historical evidence derived in Section 5.1.

2.6 Statistical Analysis Plans in Clinical Trials

A Statistical Analysis Plan (SAP) is a formal document that provides a detailed and technical description of the statistical methods to be used in the analysis of a clinical trial, elaborating on the approaches outlined in the study protocol (33,34). According to ICH E9, the SAP should be finalised before the unblinded review of data, ensuring that all analytical decisions are made independently of the observed results (5). This requirement exists because statistical analysis involves choices about endpoints, methods, populations and decision rules that can influence the conclusion of a study, and pre-specifying those choices removes the opportunity for them to be made, consciously or otherwise, in a direction favourable to a particular outcome.

Pre-specification is considered a cornerstone of reproducible clinical research and is expected by regulatory bodies including the FDA and EMA as part of any submission supporting a marketing authorisation (5,11,35). For randomised controlled trials, the discipline of the randomisation process provides some inherent protection against bias, but the SAP remains essential for ensuring that the primary analysis is defined prospectively and that any deviations from it are transparent and justified (14,33).

In single-arm performance goal studies, the SAP carries particular weight. Without a concurrent control group, the trial conclusion depends entirely on how the observed data are compared against a pre-specified benchmark, and the choices embedded in that comparison — namely the performance goal value, the confidence interval method and the decision rule — can each independently alter the conclusion (7,10,35). A SAP that specifies these choices before data collection provides the primary safeguard against the analytical flexibility that the single-arm design would otherwise permit (5,8). The practical implications of these choices for sample size and study feasibility are examined in the context of the BioMimics 3D case study throughout this thesis.

3 Case Study and Data Sources

3.1 BioMimics 3D Vascular Stent

The BioMimics 3D vascular stent was selected as the primary case study throughout this analysis. The device is a self-expanding, permanent bare-metal nitinol stent developed for the treatment of femoropopliteal peripheral artery disease, a condition characterised by atherosclerotic narrowing of the arteries supplying the lower limb (36–38). In this setting, successful treatment is judged on achieving initial vessel patency and its ability to maintain clinical benefit over time without restenosis, repeat intervention, amputation or major adverse events (9,36,39).

Unlike conventional straight stents, the BioMimics 3D stent incorporates a helical geometry intended to mimic the natural curvature and movement of the femoropopliteal artery (36,38). This design is hypothesised to promote blood flow patterns within the treated vessel and to reduce the risk of restenosis after implantation (36,38). Consequently, clinically meaningful outcomes such as target vessel revascularisation, amputation and changes in Rutherford classification are important measures of the device’s ability to provide sustained therapeutic benefit in patients with symptomatic peripheral artery disease (9,36,37).

Initial meetings with the lead scientist involved in the BioMimics development helped clarify the selection of appropriate performance goals, the purpose of the device and the statistical issues most relevant to the design of a pivotal study. These discussions informed the choice of safety and efficacy endpoints, the interpretation of clinically meaningful performance, and the practical constraints surrounding sample size. This was important because the statistical design requires clinical knowledge of the device’s intended purpose (1,5,7). Performance goals, endpoints and sample size that are mathematically convenient but clinically weak would not provide a convincing basis for evaluating the device (5,7,9).

3.2 Endpoints

Endpoint selection was guided by two criteria: clinical relevance to femoropopliteal disease and sufficient historical literature to support defensible performance goals (7,9). Candidate endpoints were identified through a targeted review of comparable peripheral vascular device trials and the 2007 guidance document by Rocha-Singh et al. on behalf of VIVA Physicians, which synthesised evidence from three studies of femoropopliteal bare nitinol stents and served as the primary regulatory reference point for this analysis (9,37,38).

For safety, candidate endpoints were assessed at 30 days and included all-cause mortality, major amputation and target vessel revascularisation. These were combined into a single composite endpoint because each component represents a clinically significant harm but individually occurs too infrequently to yield stable estimates (9,36,37).

For efficacy, candidate endpoints were assessed at 12 months. Rutherford classification, a clinical scale describing the severity of peripheral artery disease symptoms, was selected because it captures patient-level disease status rather than purely anatomical outcomes (9,37). Specifically, improvement or no deterioration in Rutherford score from post-intervention baseline was used, as this reflects whether patients experienced meaningful clinical benefit over the follow-up period (9,36).

The selected endpoints were therefore the 30-day composite of all-cause mortality, major amputation and target vessel revascularisation for safety, and improvement or no deterioration in Rutherford classification score at 12 months for efficacy.

3.3 Literature Search and Meta-analysis Data

The same guidance document was used to determine justifiable performance goal values for both the safety and efficacy endpoints.

For each relevant study, aggregate-level outcome data were extracted. This included the number of patients experiencing the endpoint and the number of patients assessed at the relevant follow-up time. Outcomes were grouped according to endpoint definition and time point before pooling. The time-point distinction was important because early safety outcomes at 30 days and longer-term efficacy outcomes at 12 months measure different aspects of device performance and cannot be combined into a single estimate (9,36).

A logit transformation was used to stabilise the variance before pooling. This was required as the outcomes were binary proportions and because the extreme event proportions expected create boundary settings where raw proportions can behave poorly (7,17).

A generalised linear mixed model with a binomial likelihood was applied to pool study-level proportions. This approach was selected because the extracted data consisted of event counts versus study sample sizes and additionally handled extreme proportions more appropriately than simple normal approximations. A random-effects structure was employed to allow for between-study heterogeneity rather than assuming that all studies estimated the same underlying event proportion. This was important because the historical studies differed in design, population, endpoint reporting and follow-up (7,10).

The resulting estimates used to construct candidate performance goals were as follows. For the 30-day safety composite endpoint, the pooled estimate was approximately 6%, with the upper confidence limit extending to approximately 12%. Values between the pooled point estimate and the upper confidence limit were considered as the safety performance goal, giving a benchmark of approximately 6–12% for the maximum acceptable proportion of patients with an early composite event. For the 12-month stable or improved Rutherford classification endpoint, the pooled estimate was approximately 95.8%, with the lower confidence limit extending to approximately 88%. Values between the pooled point estimate and the lower confidence limit were considered as the efficacy performance goal, giving a benchmark of approximately 88–95.8% for the minimum acceptable proportion of patients with stable or improved Rutherford classification.

3.4 Derivation of Performance Goals

In principle, the pooled point estimate from the historical studies represents the weighted average or best estimate of expected performance in the available evidence. However, using the point estimate directly as the performance goal would ignore uncertainty in the historical data (7). This is problematic when the available evidence may be based on a small number of studies with limited sample sizes and differing endpoint definitions (5,7,10).

For this reason, the proposed performance goals in this study were based on confidence interval boundaries rather than the pooled point estimates. For the safety endpoint, the upper confidence limit was used because it allows for uncertainty around the estimated historical event proportion (7,9). For the efficacy endpoint, the lower confidence limit was used because it allows for greater uncertainty around the estimated success proportion (7,9).

4 Methods

4.1 Simulation Study

To evaluate the confidence interval methods introduced in Section 2.3, a simulation study was designed following the approach of (22). The aim was to compare coverage probability and expected interval width across a range of true event proportions and sample sizes, before applying these methods within the performance goal framework (22).

For each scenario, the number of observed events was assumed to follow a binomial distribution,

$$X \sim \text{Binomial}(n, p)$$

where X is the number of observed events, n is the sample size and p is the true underlying event probability (17). Rather than evaluating performance at a single fixed proportion, a grid of true proportions was used so that interval behaviour could be examined across a range of clinically plausible event rates. Given that the BioMimics 3D endpoints involve either low complication rates or high success rates, the simulation focused on the range (17,22):

$$p \in \{0.001, 0.002, \dots, 0.200\}$$

Methods were evaluated at three sample sizes, $n = 20, 50$ and 100 , reflecting the small to moderate study sizes typical of single-arm performance goal studies (7,8).

For each combination of n , p and confidence interval method, the confidence limits were calculated across all possible values of X . Two operating characteristics were then used to assess performance: coverage probability and expected interval width (17,22).

Coverage probability is the probability that the constructed interval contains the true value of p (17):

$$CP(p) = P_p(L(X, n) \leq p \leq U(X, n))$$

where $L(X, n)$ and $U(X, n)$ are the lower and upper confidence limits. Since the target nominal level was 95%, a well-performing method should produce coverage close to 0.95 across the evaluated range. To summarise behaviour across the full grid, both the average coverage probability and the minimum coverage probability were recorded:

$$\overline{CP} = \frac{1}{m} \sum_{j=1}^m CP(p_j)$$

$$CP_{\min} = \min_{1 \leq j \leq m} CP(p_j)$$

The average captures typical performance, while the minimum identifies whether any method fell substantially below the nominal level at specific values of p .

Expected interval width was used as a measure of precision. For a given true proportion, sample size and method, the expected width was calculated as:

$$E_p[W(X, n)] = \sum_{x=0}^n \{U(x, n) - L(x, n)\} \binom{n}{x} p^x (1-p)^{n-x}$$

and averaged across the simulation grid to give an overall summary of interval length for each method (17,22):

$$\bar{W} = \frac{1}{m} \sum_{j=1}^m E_{p_j}[W(X, n)]$$

Results were summarised graphically. Coverage probability was plotted against the true proportion p for each method and sample size, with a dashed reference line at the nominal 95% level. Combined plots were also produced to allow direct comparison across methods on the same axes. A second set of plots compared average coverage probability against average expected interval width, providing a summary of the trade-off between reliability and precision for each method. The findings from this simulation informed the selection of confidence interval methods used in the sample size calculations in Section 4.2.

4.2 Sample Size Calculation

Sample size calculations were carried out for the BioMimics 3D case study using the single-arm performance goal framework described in Section 2.4 (7,9). The performance goal values, assumed true event proportion and endpoint were determined from the meta-analysis and clinical discussions described in Sections 3.1 and 3.3, with results reported in Section 5.3.

A target power of 90% was used throughout, which is conventional in medical device studies where 80% power, common in pharmaceutical trials, is considered insufficient to support a regulatory submission (13,16,24,31). A one-sided significance level of $\alpha = 0.025$ corresponding to a one sided 97.5% confidence interval and consistent with the decision rule described in Section 2.2 was used(3,31).

Sample size was determined through simulation rather than closed-form calculation, as exact analytical solutions are not always available when the confidence interval method forms part of the decision rule (13,17,19). Simulation provides a flexible alternative that mirrors the planned primary analysis directly and is the approach implemented in PG-Power, described in Section 4.3.

4.3 Shiny Application: PG-Power

PG-Power is an interactive R Shiny application developed as part of this thesis to implement the simulation-based sample size framework described in Section 4.2 for single-arm performance goal studies. For each candidate sample size, the application simulates a large number of trials under the assumed true event proportion, applies the selected confidence interval method, and estimates power as the proportion of simulated trials satisfying the performance goal rule. The smallest n achieving the target power is returned as the required sample size.

The application was used to carry out the BioMimics 3D sample size calculations and to compare the six confidence interval methods introduced in Section 2.3 under the same design assumptions, allowing the effect of interval choice on required sample size and achieved power to be examined consistently. Results are reported in Section 5.3. The full source code and documentation are available in the GitHub repository listed in the Appendix.

4.4 Applying Bayesian Borrowing to the Safety Endpoint

The frequentist sample size calculations in Section 4.2 made no use of the historical studies identified in Section 3.3 beyond their role in deriving the performance goal. This section examines whether formally incorporating that evidence as a prior distribution reduces the required sample size relative to the frequentist result obtained in Section 5.3 (29,31).

To construct a concrete and interpretable prior, a hypothetical Phase II dataset of 50 participants with 2 observed events was used, producing a Beta(2, 48) prior with an estimated event proportion of approximately 4% and an effective sample size of 50 (31). The choice of a hypothetical dataset rather than the historical studies directly was deliberate — the meta-analysis in Section 5.1 identified wide uncertainty around the heterogeneity estimates and only three contributing studies, making a fully justified prior difficult to construct (30,32,33). The hypothetical prior therefore serves as an illustrative example of how borrowing would operate under plausible assumptions, rather than as a formal design recommendation (29,32).

Under the Beta-Binomial model introduced in Section 2.5, incorporating s events and f non-events from the current trial updates the prior to (23,31):

$$p \mid s, f \sim \text{Beta}(2 + s, 48 + f)$$

For inference, the posterior probability that the true event proportion falls below the performance goal was evaluated against a decision threshold of 0.95 (29,31):

$$P(p < p_0 \mid s, f) \geq 0.95$$

The required sample size was determined as the smallest n for which this criterion was satisfied with at least 90% probability under the assumed true event proportion of 5%, consistent with the frequentist design assumptions in Section 4.2 (13,31).

To examine sensitivity to prior weight, three prior distributions were compared: Beta(0.2, 4.8), Beta(2, 48) and Beta(20, 480), encoding the same event ratio of approximately 1:24 but corresponding to effective sample sizes of approximately 5, 50 and 500, representing weak, moderate and strong borrowing respectively (30,32,33).

Safety & Efficacy Performance Goals				
Endpoint	Timepoint	Pooled Rate	95% CI	Performance Goal
Safety Endpoints				
Composite (Death, Amputation, TVR)	30 Days	6.0%	[2.9, 12.1%]	$\leq 12\%$
Efficacy Endpoints				
RCC: Improved or No Change	12 Months	95.8%	[87.9, 98.6%]	$\geq 88\%$

Figure 3: Proposed performance goals for the BioMimics 3D vascular stent pivotal study derived from meta-analysis of historical femoropopliteal bare nitinol stent studies.

5 Results

5.1 Meta-Analysis Results

The meta-analysis of the three femoropopliteal bare nitinol stent studies described in Section 3.3 produced pooled estimates for both endpoints. The results are summarised in Figures 3 and 4.

The heterogeneity statistics for both endpoints are reported in Figure 5.

For both endpoints, the estimated between-study variance τ^2 was 0 and the I^2 statistic was 0.0% (95% CI: 0.0–0.9%), with $H = 1.00$. Cochran’s Q test produced p -values of 0.998 and 0.520 for the safety and efficacy endpoints respectively. The implications of these heterogeneity estimates for the interpretation of the derived performance goals are discussed in Section 6.1.

5.2 Simulation Study: CI Method Comparison

Coverage probability was evaluated across a grid of true proportions $p \in \{0.001, \dots, 0.200\}$ for each of the six confidence interval methods at sample sizes $n = 50$ and $n = 100$. Results are summarised through coverage probability curves and coverage-width scatter plots.

At $n = 50$, the Wald interval produced coverage below 0.80 for true proportions near zero, only approaching the nominal level near $p = 0.20$. The Wilson Score interval oscillated around 0.95, with dips below 0.90 at very low proportions near $p = 0.001$ to 0.01. The Jeffreys interval followed a similar pattern but with more pronounced dips below nominal at very low proportions. The Wilson Score with continuity correction and Agresti-Coull interval maintained coverage above the nominal level across virtually the entire proportion range. The Clopper-Pearson interval was the most conservative, with coverage remaining above 0.97 across nearly the entire evaluated range.

At $n = 100$ the pattern was broadly similar. The Wald interval remained well below 0.95 across most of the proportion range. The remaining five methods continued to oscillate above the nominal level, with Wilson Score and Jeffreys tracking closest to 0.95 and Clopper-Pearson and Wilson Score with continuity correction remaining the most conservative.

The combined overlay for $n = 100$ shows all six methods on the same axes, illustrating the clear separation between the Wald interval and the remaining five methods across the full proportion range.

The coverage-width scatter plots summarise the trade-off between average coverage probability and average expected interval width at each sample size.

At $n = 20$, the Wald interval produced average coverage of approximately 0.73 with an average width of approximately 0.21. Wilson Score and Jeffreys sat at or just above the nominal line with average widths of approximately 0.24 and 0.25 respectively. Clopper-Pearson and Wilson Score with continuity correction sat above the nominal line with average widths of approximately 0.29 and 0.30.




Figure 4: Pooled event proportions with 95% confidence intervals for the safety and efficacy endpoints. The diamond denotes the pooled estimate and the dashed red line indicates the derived performance goal.

At $n = 50$ and $n = 100$ the five non-Wald methods converged, with all five clustered near and above the nominal line in the coverage-width plots. The ordering remained consistent across both sample sizes: Wilson Score and Jeffreys sat closest to 0.95 with average widths of approximately 0.16 and 0.11 at $n = 50$ and $n = 100$ respectively, while Clopper-Pearson produced the widest average intervals at both sample sizes.

5.3 Sample Size and Confidence Interval Method Comparison

Under the design assumptions established in Section 4.2, the required sample size was determined using the single-arm performance goal framework with an approved performance goal of $p_0 = 0.11$, an assumed true complication proportion of $p_1 = 0.05$, a one-sided significance level of $\alpha = 0.025$ and a target power of 90%. The formal hypotheses were:

$$H_0 : p \geq 0.11 \quad H_1 : p < 0.11$$

The Clopper-Pearson interval was used as the primary method for the main sample size calculation. This choice was made partly on the basis of the simulation results in Section 5.2, which showed Clopper-Pearson to be the most conservative method with guaranteed nominal coverage, but also because the exact binomial method underlying the Clopper-Pearson interval is the default approach used in widely adopted sample size software including PASS and nQuery (17,20,25,26).

The required sample size varied across the six confidence interval methods evaluated under the same design assumptions, as summarised in the table below.

The Wald interval produced the smallest required sample size of 169 patients but achieved power of only 77.3% under the assumed true complication proportion, falling well short of the 90% target. This is consistent with

Table A. Heterogeneity - 30-Day Composite (Death, Amputation, TVR)		
Measure	Value	CI or p-value
τ^2	0.00	NA
I^2	0.0%	0.0% to 0.9%
H	1.00	1.00 to 3.10
Q (Wald)	0.00	p = 0.998
Q (LRT)	NA	NA
Table B. Heterogeneity - RCC: Improved / No Change (12 Months)		
Measure	Value	CI or p-value
τ^2	0.00	NA
I^2	0.0%	0.0% to 0.9%
H	1.00	1.00 to 3.10
Q (Wald)	1.31	p = 0.520
Q (LRT)	NA	NA

Figure 5: Heterogeneity statistics from the random-effects meta-analysis for the 30-day composite safety endpoint (Table A) and the 12-month Rutherford classification efficacy endpoint (Table B).

the simulation results in Section 5.2, which showed the Wald interval systematically undercovering across the proportion range relevant to this endpoint (17,19). A sample size derived from the Wald interval would therefore be unreliable as a basis for study design.

Among the remaining methods, Jeffreys required 198 patients with achieved power of 87.7%, and Wilson Score required 208 patients at 83.9% power, both falling below the 90% target despite requiring more patients than Wald. Clopper-Pearson and Agresti-Coull were the only methods to approach the target, requiring 209 and 210 patients at achieved powers of 89.7% and 89.4% respectively. The Wilson Score with continuity correction required the largest sample size of 247 patients, achieving 92.8% power, consistent with its conservative coverage behaviour observed in Section 5.2.

Notably, Wilson Score required only one fewer patient than Clopper-Pearson but achieved considerably lower power at 83.9% versus 89.7%. This reflects the stepwise nature of the power calculation for binary endpoints discussed in Section 2.4, where a single additional patient can shift the maximum allowable number of events and produce a discrete jump in achieved power (13,17). The difference between the smallest reliable sample size and the largest, Clopper-Pearson at 209 and Wilson Score with continuity correction at 247, represents 38 additional evaluable patients or approximately 42 additional enrolments after dropout adjustment, illustrating how method choice directly affects study feasibility (14,16).

A full report generated by the PG-Power application for these design assumptions, including power curves and sensitivity analyses, is available in PDF form on the project GitHub repository (see Appendix).




Figure 6: Coverage probability curves for six binomial confidence interval methods at $n = 50$, evaluated across true proportions p in 0.001, ..., 0.200. The dashed red line indicates the nominal 95% level.

5.4 Bayesian Historical Borrowing Results

Under the $\text{Beta}(2, 48)$ prior and the design assumptions described in Section 4.4, the required evaluable sample size was $n = 79$, permitting no more than 4 observed events. This represents a reduction of 130 patients, approximately 62%, relative to the Clopper-Pearson frequentist result of 209 from Section 5.3.

The power curve illustrates how achieved power evolves across candidate sample sizes under the Bayesian model. The saw-tooth pattern reflects the same discrete stepwise behaviour described in Section 2.4, where power decreases gradually within ranges of n and increases sharply when an additional event becomes permissible. The required sample size of 79 corresponds to the first point at which the power curve crosses the 90% threshold.

The sensitivity of this result to prior weight is shown in the figure below. All three distributions, $\text{Beta}(0.2, 4.8)$, $\text{Beta}(2, 48)$ and $\text{Beta}(20, 480)$, encode the same underlying event ratio but differ substantially in how tightly they concentrate the posterior.

With $\text{ESS} = 5$ the posterior is broad and driven largely by the current data, offering modest sample size reduction relative to the frequentist result. With $\text{ESS} = 50$ the posterior concentrates more clearly around the historical event proportion of 4%, well below the performance goal of 11%, producing the reduction to $n = 79$ reported above. With $\text{ESS} = 500$ the posterior is narrow and the prior dominates, largely overriding the current trial data. While this produces the greatest reduction in required sample size, it also renders the conclusion highly sensitive to the validity of the exchangeability assumption. If the true event proportion in the current study differs from the historical estimate, a strongly informative prior will pull the posterior toward the historical value regardless of what the current data show.




Figure 7: Coverage probability curves for six binomial confidence interval methods at $n = 100$, evaluated across true proportions p in 0.001, ..., 0.200. The dashed red line indicates the nominal 95% level.

6 Discussion

6.1 Interpretation of Main Findings

This study set out to examine how performance goal selection, confidence interval choice and sample size calculation interact to influence conclusions in a single-arm medical device setting. The five aims established in Section 1.5 are addressed in turn before their combined implications are considered.

6.1.1 Aim 1: Deriving Evidence-Based Performance Goals Through Meta-Analysis

The meta-analysis of three historical femoropopliteal nitinol stent studies yielded a pooled 30-day composite complication proportion of approximately 6%, with the upper confidence limit extending to 12%, and a pooled 12-month Rutherford classification success proportion of approximately 95.8%, with the lower confidence limit at 88%. Confidence interval boundaries rather than point estimates were used to derive the proposed performance goals, acknowledging the uncertainty inherent in a small historical evidence base of only three studies and 116 patients combined. The resulting benchmarks of 12% for safety and 88% for efficacy are therefore ranges rather than fixed values, and the width of those ranges reflects the limited precision of the available evidence.

This has two practical consequences. First, where within the range the performance goal is set matters: the sensitivity of sample size requirements to the difference between the proposed 12% and approved 11% thresholds is examined in Aim 3. Second, deriving performance goals from conservative confidence interval boundaries rather than point estimates produces a less stringent benchmark, which reduces the risk of an unrealistically strict threshold but introduces the possibility of biocreep (40). If successive device generations are each evaluated against benchmarks derived from increasingly conservative historical estimates, acceptable performance standards may gradually drift from the original level of clinical effectiveness. In a single-arm study without a concurrent control, this risk is amplified because there is no active comparator to anchor the comparison (7,10). The validity of the derived benchmarks ultimately depends on clinical and regulatory judgement alongside the statistical evidence, and no amount of methodological care can substitute for that judgement.




Figure 8: Coverage probability curves for all six methods overlaid at $n = 100$. The dashed red line indicates the nominal 95% level.

6.1.2 Aim 2: Comparing Confidence Interval Methods for Binary Endpoints

The simulation study confirmed that confidence interval method choice has a meaningful and consistent effect on coverage behaviour at the low event proportions relevant to the BioMimics 3D safety endpoint. The Wald interval was the only method to consistently fall below the nominal 95% level, with average coverage of approximately 0.73 at $n = 20$, 0.84 at $n = 50$ and 0.89 at $n = 100$. Importantly, this undercoverage was not accompanied by any compensating gain in precision: the Wald interval produced average widths comparable to better-performing methods while failing to meet the nominal level throughout. Although its poor behaviour at extreme proportions is well documented in the literature, its inclusion here serves a practical purpose by demonstrating that a method commonly applied by default in standard software would produce unreliable results in exactly this setting (17,19).

Among the remaining five methods, Wilson Score and Jeffreys sat closest to the nominal level with the narrowest intervals, while Clopper-Pearson, Agresti-Coull and the continuity-corrected Wilson Score were consistently conservative with progressively wider intervals. Both Wilson Score and Jeffreys showed dips below the nominal level at very low proportions near $p = 0.001$ to 0.01 (17,22), which is directly relevant given the expected complication rate for the BioMimics 3D safety endpoint. The ordering of methods was stable across all sample sizes, providing a consistent basis for method selection in the calculations that followed.

6.1.3 Aim 3: Evaluating the Effect of Performance Goal and Interval Choice on Sample Size

The sample size results demonstrated that neither the performance goal nor the confidence interval method can be treated as a neutral design choice. Under the approved performance goal of 11% and an assumed true complication proportion of 5%, the required evaluable sample size ranged from 169 patients under the Wald interval, which failed to meet the 90% power target, to 247 under the continuity-corrected Wilson Score. Among the methods approaching the power target, Clopper-Pearson required 209 patients at 89.7% power and Agresti-Coull required 210 at 89.4%, while Jeffreys and Wilson Score fell short of 90% at their minimum required sample sizes despite requiring fewer patients.




Figure 9: Average coverage probability against average expected interval width at $n = 20$. The dashed red line indicates the 95% target coverage.




Figure 10: Average coverage probability against average expected interval width at $n = 50$. The dashed red line indicates the 95% target coverage.

This pattern reflects the saw-tooth structure of power calculations for binary endpoints discussed in Section 2.4 (17). A single additional patient can shift the maximum allowable event count and produce a discrete jump in achieved power, meaning a method requiring slightly fewer patients may achieve substantially lower power if that threshold has not been crossed. The difference between the most and least demanding reliable methods represents 38 evaluable patients, or approximately 42 additional enrolments after dropout adjustment, illustrating how a statistical design decision made at the planning stage propagates directly into recruitment targets and study cost.




Figure 11: Average coverage probability against average expected interval width at $n = 100$. The dashed red line indicates the 95% target coverage.

6.1.4 Aim 4: Developing PG-Power

The PG-Power application provided a transparent and reproducible implementation of the sample size framework, allowing the BioMimics 3D design assumptions to be evaluated consistently across all six confidence interval methods. By making the sensitivity of required sample size and achieved power to method choice immediately visible, the application supports the kind of deliberate, pre-specified decision-making that the results of Aims 2 and 3 show to be necessary. The full report generated for the BioMimics 3D analysis is available in the GitHub repository (see Appendix).

6.1.5 Aim 5: The Value and Limits of Bayesian Borrowing

Incorporating a $\text{Beta}(2, 48)$ prior based on a hypothetical Phase II dataset of 50 patients with 2 observed events reduced the required sample size from 209 to 79 patients, a reduction of approximately 62% under the same design assumptions. This illustrates the potential efficiency gains that historical borrowing can offer in a medical device setting where recruitment is costly and historical evidence from comparable devices is available.

However, the magnitude of this reduction is sensitive to the weight assigned to the prior. Moving from an effective sample size of 5 to 500 while holding the event ratio constant progressively concentrates the posterior around the historical estimate, reducing the required sample size but increasing dependence on the exchangeability assumption. When the prior is strongly informative and the historical and current populations differ, the posterior can be drawn toward the historical estimate regardless of what the current data show, inflating the Type I error rate in ways that undermine the trial conclusion (30,32,33). The prior weight is therefore a substantive design choice rather than a technical detail.

In the BioMimics 3D context this concern is not hypothetical. The meta-analysis drew on only three studies with wide uncertainty around the heterogeneity estimates, and the exchangeability of that evidence with a future pivotal study cannot be assumed. The reduction from 209 to 79 patients should be understood as an illustration of the method's potential rather than a design recommendation. Realising that potential in practice would require formal exchangeability assessment, sensitivity analyses across a range of prior specifications, and simulation studies evaluating operating characteristics under scenarios where historical and current data diverge (30,32,33).




Figure 12: Power curve for the BioMimics 3D 30-day composite safety endpoint produced by PG-Power using the Clopper-Pearson method. The green point marks the required evaluable sample size of $n = 209$, at which achieved power was 89.7%. The dashed line indicates the 90% target power. The decision rule is to reject H_0 if no more than 14 of 209 patients experience the composite endpoint.

Method	Required n	Achieved power (%)
Wald (Z-score)	169	77.3%
Wilson Score	208	83.9%
Agresti-Coull	210	89.4%
Clopper-Pearson	209	89.7%
Prop.test	247	92.8%
Jeffreys	198	87.7%

■ Green = meets or beats required n · Amber = larger n · Bold = currently selected method

Figure 13: Required evaluable sample size and achieved power for six confidence interval methods under the BioMimics 3D safety endpoint design assumptions (performance goal $p_0 = 0.11$, assumed true proportion $p_1 = 0.05$, one-sided $\alpha = 0.025$, target power 90%), produced by PG-Power. Clopper-Pearson was used as the primary method.

6.1.6 Combined Implications

Taken together, the five aims tell a coherent story about the interdependence of statistical design choices in single-arm performance goal studies. The benchmark, the confidence interval method, the sample size calculation and the incorporation of prior evidence are not independent decisions. Each constrains and is constrained by the others, and the trial conclusion reflects all of them simultaneously.

The meta-analysis established that the performance goal is a range, not a fixed value. The simulation study showed that the confidence interval method determines both the reliability of the decision rule and the sample size required to achieve a given power. The sample size analysis made this concrete, with an interval-method-driven difference of 38 evaluable patients under identical clinical assumptions. The PG-Power application demonstrated that these calculations can be made transparent and reproducible, and the Bayesian analysis showed that efficiency gains are available from historical borrowing but only when the underlying assumptions are honestly examined.




Figure 14: Confidence intervals for the observed event proportion across all six methods at $n = 209$ with 14 observed events, produced by PG-Power under the BioMimics 3D safety endpoint design assumptions. The dashed vertical line indicates the performance goal of 11%. All six methods produce an upper bound below the performance goal at the decision boundary.

The thread running through all five aims is the one identified in Section 1.3. These choices are routinely made in performance goal studies, but their rationale is not always made visible. The results of this study suggest that the concern (6) raised about non-inferiority margin reporting extends equally to the confidence interval method, the sample size procedure and the prior specification. Making these choices explicit, justified and pre-specified is not a procedural formality. It is the mechanism by which a performance goal study earns the credibility needed to support a regulatory decision.

6.2 SAP in Performance Goal Studies

As established in Section 2.6, pre-specifying the SAP before data collection is a regulatory expectation and a prerequisite for reproducible clinical research. The findings of this study make clear why this matters in the performance goal setting specifically. The performance goal value, confidence interval method and assumed true event proportion each independently altered the required sample size and the trial conclusion across Sections 5.2 through 5.3, and none of these choices is neutral. A confidence interval method selected post-hoc in light of favourable results, or a performance goal derived after reviewing the data, would undermine the validity of the study conclusion regardless of the quality of the underlying clinical evidence (6,35).

The 38-patient difference in required sample size between the most and least demanding reliable confidence interval methods identified in Section 5.3 illustrates concretely what is at stake at the design stage. PG-Power directly supports this process by allowing the sensitivity of sample size requirements to method choice, performance goal value and dropout assumptions to be examined transparently before the SAP is finalised. Documenting these outputs alongside the chosen method and its justification provides a clear and reproducible record of the design rationale, which is precisely what regulators and future reviewers need to evaluate the study's conclusions with confidence (5,33).

6.3 Study Limitations

The most important limitation of this study is the restricted historical evidence base. The meta-analysis drew on only three studies comprising 116 patients, identified through a single guidance document rather than a systematic literature search. This limits both the precision of the pooled estimates and the ability to reliably




Figure 15: Bayesian power curve for the BioMimics 3D 30-day composite safety endpoint under a Beta(2, 48) prior (performance goal $p_0 = 0.11$, assumed true proportion $p_1 = 0.05$, one-sided $\alpha = 0.025$, decision threshold 0.95, target power 90%). The red point marks the required evaluable sample size of $n = 79$.

characterise between-study heterogeneity. The wide confidence intervals around the I^2 statistics for both endpoints mean that the derived performance goals carry substantial uncertainty, and a more comprehensive search of the literature might have identified additional comparable studies that could have narrowed that uncertainty or shifted the pooled estimates (7).

Relatedly, the performance goal derivation focused on a single guidance document developed in 2007. Clinical practice in femoropopliteal intervention has evolved considerably since then, and it is possible that more recent device studies would yield different pooled estimates (36,38). The external validity of the derived benchmarks to a contemporary BioMimics 3D pivotal study therefore cannot be assumed without further review of the current literature (40).

The simulation study was conducted over a proportion range of $p \in \{0.001, \dots, 0.200\}$ at sample sizes of $n = 20, 50$ and 100 . While this range and these sample sizes are appropriate for the BioMimics 3D safety endpoint, the conclusions may not generalise to settings with higher event rates, larger studies or endpoints near $p = 0.5$. The simulation also evaluated methods solely on coverage probability and expected interval width. Other properties, such as the location of the interval relative to the true proportion or tail behaviour at extreme values, were not examined (17,22).

The Bayesian analysis used a hypothetical Phase II dataset rather than the actual historical studies as the basis for the prior. While this was a deliberate choice to illustrate the method under transparent and controlled assumptions, it means the sample size reduction reported in Section 5.4 cannot be directly interpreted as a realistic design recommendation for a BioMimics 3D pivotal study. A fully justified prior would require a more thorough assessment of exchangeability, potentially involving clinical experts and formal heterogeneity testing across a broader evidence base than was available here (30,32,33).

Finally, this study focused exclusively on single-arm performance goal designs with binary endpoints. Many medical device studies involve time-to-event endpoints, continuous outcomes or repeated measures (13,16), and the methods developed here do not extend directly to those settings. The conclusions are therefore specific to the single-arm binary endpoint framework and should not be generalised beyond it without further methodological work.




Figure 16: Effect of prior weight on the posterior distribution of the true event proportion θ under three Beta priors with the same event ratio (1:24) but increasing effective sample sizes of 5, 50 and 500. The dashed vertical line indicates the performance goal of 0.11. Greater prior weight concentrates the posterior further from the performance goal, reducing the required sample size but increasing sensitivity to the exchangeability assumption.

6.4 Directions for Future Research

The most immediate extension of this work would be a systematic literature search to replace the single 2007 guidance document used as the historical evidence base. With only three contributing studies available, the derived performance goals carry considerable uncertainty, and more recent femoropopliteal device studies may yield meaningfully different pooled estimates. A broader evidence base would also permit formal assessment of publication bias, which was not possible here (7).

The simulation study could be extended to higher event rates, larger sample sizes and a wider set of operating characteristics beyond coverage probability and expected interval width (17,22). Applying the same simulation and sample size framework to the 12-month Rutherford classification efficacy endpoint would also complete the BioMimics 3D design picture, and a joint power calculation considering both endpoints simultaneously would be a natural further step.

The Bayesian borrowing analysis would benefit from formalisation. This would require a formal exchangeability assessment drawing on clinical expert judgement, sensitivity analyses across a range of prior specifications including the power prior and robust mixture prior approaches introduced in Section 2.5, and simulation studies evaluating operating characteristics under scenarios where historical and current data diverge (30,32,33). Incorporating this into PG-Power as a Bayesian module, alongside extensions to time-to-event endpoints and two-arm designs, would substantially broaden the application's utility for medical device trial planning.

Finally, empirical validation against actual BioMimics 3D pivotal study data, if it were to become available, would allow the performance goal framework and confidence interval method rankings developed here to be tested against real clinical outcomes rather than simulated ones.

Conclusion

This thesis examined how three interconnected statistical choices, the performance goal value, the confidence interval method and the sample size calculation, interact to shape the conclusion of a single-arm performance goal study. Using the BioMimics 3D vascular stent as a case study, it demonstrated that none of these choices is neutral and that each can alter the trial conclusion independently of the underlying clinical data.

The meta-analysis of three historical femoropopliteal nitinol stent studies yielded pooled estimates of approximately 6% for the 30-day composite safety endpoint and 95.8% for the 12-month Rutherford classification efficacy endpoint. Deriving performance goals from confidence interval boundaries rather than point estimates produced benchmarks of 12% for safety and 88% for efficacy, acknowledging the uncertainty inherent in a small and imprecise historical evidence base. The width of the uncertainty around these estimates illustrated that performance goals are rarely as settled as they appear when reported as single values, and that the reasoning behind them carries as much weight as the numbers themselves.

The simulation study showed that confidence interval method choice has a consistent and meaningful effect on coverage behaviour at the low event proportions characteristic of medical device safety endpoints. The Wald interval was the only method to consistently fall below the nominal 95% coverage level and is unsuitable as a decision tool in this setting (17). Wilson Score and Jeffreys achieved coverage closest to the nominal level with the narrowest intervals, while Clopper-Pearson, Agresti-Coull and the continuity-corrected Wilson Score provided conservative coverage at the cost of wider intervals. These differences translated directly into sample size requirements, with the required evaluable sample size ranging from 169 patients under the Wald interval to 247 under the continuity-corrected Wilson Score under identical design assumptions, and Clopper-Pearson requiring 209 patients as the primary result.

The Bayesian historical borrowing analysis demonstrated that incorporating prior evidence through a $\text{Beta}(2, 48)$ prior reduced the required sample size from 209 to 79 patients, a reduction of approximately 62%. This reduction was sensitive to the weight assigned to the prior, and stronger borrowing increased dependence on the exchangeability assumption in ways that risk inflating the Type I error rate. The analysis established the potential of Bayesian borrowing as a complementary approach to the frequentist framework, contingent on careful prior specification and formal exchangeability assessment.

The PG-Power application provided a transparent and reproducible implementation of the sample size framework, making the sensitivity of required sample size to design assumptions immediately visible and supporting the pre-specified decision-making that the results consistently point toward.

Taken together, these findings reinforce a conclusion that extends beyond the BioMimics 3D case study. The concern raised in Section 1.3, that performance goals and statistical methods are often presented without explanation, applies not only to the benchmark itself but to the confidence interval method, the sample size procedure and the prior specification (6). Each of these choices shapes the trial conclusion before a patient is enrolled, and making them explicit, justified and pre-specified in the SAP is not a procedural formality. It is the foundation on which a credible and reproducible performance goal study rests.

Appendix

Project GitHub Repository

All materials associated with this thesis are publicly available at:

<https://github.com/FilipMKgit/MDX2526>

The repository contains the PG-Power R Shiny application with full source code and documentation, a PDF report generated by PG-Power for the BioMimics 3D safety endpoint analysis under the design assumptions described in Section 2.4, the R simulation code used to produce the results in Section 5.2, the thesis document in both Word and Rmd formats, and supporting project materials including the research poster and presentation slides.

References

1. Jennifer Schumi, Janet T. Wittes. Through the looking glass: Understanding non-inferiority. *Trials*. 2011 May;12(1):106. doi:[10.1186/1745-6215-12-106](https://doi.org/10.1186/1745-6215-12-106)
2. Jack Cuzick, Peter Sasieni. Interpreting the results of noninferiority trials-a review. *British journal of cancer*. 2022 Nov;127(10):1755–9. doi:[10.1038/s41416-022-01937-w](https://doi.org/10.1038/s41416-022-01937-w)
3. Simon J. W. Oczkowski. A clinician’s guide to the assessment and interpretation of noninferiority trials for novel therapies. *Open medicine : a peer-reviewed, independent, open-access journal*. 2014;8(2):e67–72.
4. Conor D Tweed, Matteo Quartagno, Michelle N Clements, Rebecca M Turner, Andrew J Nunn, David T Dunn, Ian R White, Andrew J Copas. Exploring different objectives in non-inferiority trials. *BMJ*. 2024;385. doi:[10.1136/bmj-2023-078000](https://doi.org/10.1136/bmj-2023-078000)
5. International Council for Harmonisation. Statistical Principles for Clinical Trials E9 [Guideline] [Internet]. International Council for Harmonisation (ICH); 1998 [cited 2026 May 12]. Available from: <https://www.ich.org/page/efficacy-guidelines>
6. Gilda Piaggio, Diana R. Elbourne, Stuart J. Pocock, Stephen J. W. Evans, Douglas G. Altman, for the CONSORT Group. Reporting of Noninferiority and Equivalence Randomized Trials: Extension of the CONSORT 2010 Statement. *JAMA*. 2012 Dec;308(24):2594–604. doi:[10.1001/jama.2012.87802](https://doi.org/10.1001/jama.2012.87802)
7. Chenguang Wang, Gary L. Rosner, Tingting Bao, Nelson Lu, Wei-Chen Chen, Heng Li, Ram Tiwari, Yunling Xu, Lilly Q. Yue. Leveraging real-world evidence for determining performance goals for medical device studies. *Statistics in medicine*. 2021 Dec;40(29):6577–89. doi:[10.1002/sim.9199](https://doi.org/10.1002/sim.9199)
8. Minyan Wang, Huan Ma, Yun Shi, Haojie Ni, Chu Qin, Conghua Ji. Single-arm clinical trials: Design, ethics, principles. *BMJ supportive & palliative care*. 2024 Dec;15(1):46–54. doi:[10.1136/spcare-2024-004984](https://doi.org/10.1136/spcare-2024-004984)
9. K. J. Rocha-Singh, others. Performance goals and endpoint assessments for clinical trials of femoropopliteal bare nitinol stents in patients with symptomatic peripheral arterial disease. *Catheterization and Cardiovascular Interventions*. 2007;69:910–9.
10. Shang Cao, Hongwei Kan, Mingliang Lu, Jianhong Pan. Estimating comparative effectiveness using Single-Arm trials: A challenge in the field of agnosticism. *BMC Medical Research Methodology*. 2025 Oct;25(1):233. doi:[10.1186/s12874-025-02660-9](https://doi.org/10.1186/s12874-025-02660-9)
11. Medical Devices - Clinical investigations and performance studies - European Commission [Internet]. 2026 [cited 2026 Jun 4]. Available from: https://health.ec.europa.eu/medical-devices-clinical-investigations-and-performance-studies_en
12. Emmanuel Lesaffre. Superiority, Equivalence, and Non-Inferiority Trials. *Bulletin of the Hospital for Joint Diseases* [Internet]. 2008;66(2). Available from: https://journals.lww.com/bhjd/fulltext/2008/66020/superiority,_equivalence,_and_non_inferiority.16.aspx
13. Shein-Chung Chow, Jun Shao, Hansheng Wang, Yuliya Lokhnygina. *Sample size calculations in clinical research*. chapman; hall/CRC; 2017.
14. Kenneth F Schulz, David A Grimes. Sample size calculations in randomised trials: Mandatory and mystical. *The Lancet*. 2005 Apr;365(9467):1348–53. doi:[10.1016/S0140-6736\(05\)61034-3](https://doi.org/10.1016/S0140-6736(05)61034-3)

15. U.S. Food and Drug Administration. Non-Inferiority Clinical Trials to Establish Effectiveness [Guidance for {Industry}] [Internet]. U.S. Food; Drug Administration; 2016 [cited 2026 Jan 5]. Available from: <https://www.fda.gov/media/78504/download>
16. Xiaofeng Wang, Xinge Ji. Sample Size Estimation in Clinical Research: From Randomized Controlled Trials to Observational Studies. An Overview of Study Design and Statistical Considerations. 2020 Jul;158(1, Supplement):S12–20. doi:[10.1016/j.chest.2020.03.010](https://doi.org/10.1016/j.chest.2020.03.010)
17. Lawrence D. Brown, T. Tony Cai, Anirban DasGupta. Interval Estimation for a Binomial Proportion. Statistical Science [Internet]. 2001 [cited 2026 Jun 4];16(2):101–17. Available from: <http://www.jstor.org/stable/2676784>
18. Alan Agresti. Categorical data analysis. John Wiley & Sons; 2013.
19. Alan Agresti, Brent A Coull. Approximate is better than “exact” for interval estimation of binomial proportions. The American Statistician. 1998;52(2):119–26.
20. Charles J Clopper, Egon S Pearson. The use of confidence or fiducial limits illustrated in the case of the binomial. Biometrika. 1934;26(4):404–13.
21. Center for Devices and Radiological Health. CDRH Learn. FDA [Internet]. FDA; 2026 [cited 2026 Jun 10]. Available from: <https://www.fda.gov/training-and-continuing-education/cdrh-learn>
22. Ana M Pires, Conceição Amado. Interval estimators for a binomial proportion: Comparison of twenty methods. REVSTAT-Statistical Journal. 2008;6(2):165–97.
23. Christian P Robert, Nicolas Chopin, Judith Rousseau. Harold Jeffreys’s theory of probability revisited. Statistical Science. 2009;24(2):141–72. doi:[10.1214/09-STS284](https://doi.org/10.1214/09-STS284)
24. Steven A Julious. Sample sizes for clinical trials with normal data. Statistics in medicine. 2004;23(12):1921–86.
25. PASS Sample Size Software Documentation | PASS Software Help [Internet]. [cited 2026 Jun 11]. Available from: <https://www.ncss.com/software/pass/pass-documentation/>
26. Janet D Elashoff. nQuery Advisor Version 5.0 User’s Guide.
27. East Archives [Internet]. [cited 2026 Jun 11]. Available from: <https://dev-www.cytel.com/by-topic/east/>
28. Binom.confint function - RDocumentation [Internet]. [cited 2026 Jun 11]. Available from: <https://www.rdocumentation.org/packages/binom/versions/1.1-1.1/topics/binom.confint>
29. Gregory Campbell. Bayesian statistics in medical devices: Innovation sparked by the FDA. Journal of biopharmaceutical statistics. 2011;21(5):871–87.
30. Joseph G Ibrahim, Ming-Hui Chen. Power prior distributions for regression models. Statistical Science. 2000;46–60.
31. US Food and Drug Administration. Guidance for the use of Bayesian statistics in medical device clinical trials. Maryland: US Food and Drug Administration. 2010.

32. Maria Vittoria Chiaruttini, Giulia Lorenzoni, Dario Gregori. Bayesian dynamic borrowing in group-sequential design for medical device studies. *BMC Medical Research Methodology*. 2025;25(1):78.
33. Carrol Gamble, Ashma Krishan, Deborah Stocken, Steff Lewis, Edmund Juszczak, Caroline Doré, Paula R Williamson, Douglas G Altman, Alan Montgomery, Pilar Lim. Guidelines for the content of statistical analysis plans in clinical trials. *Jama*. 2017;318(23):2337–43.
34. Marc Fradera, Xavier Serra-Aracil. Statistical Analysis Plan (SAP): What is it and how to develop it? *Cirugia Espanola*. 2024;103(1):45–8.
35. Brennan C Kahan, Gordon Forbes, Suzie Cro. How to design a pre-specified statistical analysis approach to limit p-hacking in clinical trials: The Pre-SPEC framework. *BMC medicine*. 2020;18(1):253.
36. Christos Rammos, Thomas Zeller, Michael Piorkowski, Koen Deloose, Klaus Hertting, Volker Sesselmann, Gunnar Tepe, Peter Gaines, Michael Lichtenberg. The BioMimics 3D Helical Centreline Nitinol Stent in Chronic Limb Threatening Ischaemia and Complex Lesions: Three Year Outcomes of the MIMICS-3D Registry. *European Journal of Vascular and Endovascular Surgery*. 2024;67(6):923–32. doi:<https://doi.org/10.1016/j.ejvs.2024.02.043>
37. John R. Laird, Barry T. Katzen, Dierk Scheinert, Johannes Lammer, Jeffrey Carpenter, Maurice Buchbinder, Rajesh Dave, Gary Ansel, Alexandra Lansky, Ecaterina Cristea, Tyrone J. Collins, Jeffrey Goldstein, Michael R. Jaff, for the RESILIENT Investigators. Nitinol Stent Implantation Versus Balloon Angioplasty for Lesions in the Superficial Femoral Artery and Proximal Popliteal Artery. *Circulation: Cardiovascular Interventions*. 2010 Jun;3(3):267–76. doi:[10.1161/CIRCINTERVENTIONS.109.903468](https://doi.org/10.1161/CIRCINTERVENTIONS.109.903468)
38. Thomas Zeller, Peter A. Gaines, Gary M. Ansel, Colin G. Caro. Helical Centerline Stent Improves Patency. *Circulation: Cardiovascular Interventions*. 2016;9(6):e002930. doi:[10.1161/CIRCINTERVENTIONS.115.002930](https://doi.org/10.1161/CIRCINTERVENTIONS.115.002930)
39. G. Tepe, others. Local delivery of paclitaxel to inhibit restenosis during angioplasty of the leg. *The New England Journal of Medicine*. 2008;358:689–99.
40. Siobhan Everson-Stewart, Scott S Emerson. Bio-creep in non-inferiority clinical trials. *Statistics in medicine*. 2010;29(27):2769–80.