


REVISIONS			
REV	DESCRIPTION	DATE	APPROVAL
B	RELEASE/CHANGE PER ECO-R240705	SEE_AGILE	SEE_AGILE

AUTOCAD DRAWING—ALL CHANGES MUST BE DONE IN AUTOCAD DATA BASE



ITEM	PART NO.	DESCRIPTION	SUPPLIER	QTY.
1	84729-0001	RECEPTACLE, USB PANEL MT .152M	MOLEX	1
2	JR-201S	INLET, IEC CLASS 2 FIG 8 SOLDER TAG	MULTICOMP	1
3	1085460	SOCKET, 4MM SHROUD BLK.	FARNELL	4
4	1085459	SOCKET, 4MM SHROUD RED	FARNELL	3
5	735WHI	MAINS SOCKET CIRCULAR :BS1363: 1984	MK (ELECTRIC)	1
6	BIM2009/19-GY/GY	BOX, ABS GREY 250 X 200 X 65MM	CAMDENBOSS	1
7	891430	RUBBER FEET	PB	4
8	----	22 AWG WIRE, MULTICORE, PVC INSULATED	PB	A/R
9	L712AS	DC JACK, PANEL MOUNT, 2.5MM	SWITCHCRAFT	1
10	PX0575/10/63	POWER INLET, FLANGE MOUNT, IEC SOLDER CONNECTIONS, 250VAC, 10A	BULGIN	1

NOTES: UNLESS OTHERWISE SPECIFIED.

- APPLICABLE STANDARDS/SPECIFICATIONS:  
ASME Y14.3- 2008, MULTIVIEW AND SECTIONAL VIEW DRAWINGS.  
ASME Y14.4M-2009, PICTORIAL DRAWINGS.  
ASME Y14.5-2009, DIMENSIONS AND TOLERANCES.  
ASME Y14.38-2007, ABBREVIATIONS AND ACRONYMS.
- ALL DIMENSIONS ARE IN MILLIMETERS.
- DEBURR AND REMOVE ALL SHARP EDGES.
- ALL ENCLOSED IN AN ABS BOX (250MM X 200MM X 65MM).
- UNUSED USB LINES TIED AWAY.
- EQUIVALENT PARTS MAY BE USED WITH TEST ENGINEERING APPROVAL.

GENERAL DIMENSIONAL TOLERANCES			PROPRIETARY & CONFIDENTIAL. THIS DOCUMENT MAY NOT BE COPIED, DISCLOSED, OR USED IN WHOLE OR IN PART WITHOUT THE CONSENT OF COVIDIEN.	
DEC.	DEC.	DEC.	TITLE	
.x +/-	.xx +/-	.xxx +/-	PB500 ELECTRICAL SAFETY CABLE CONTINUITY TEST BOX	
1	0.1	0.01	COVIDIEN Galway, Ireland	
HOLES	FRAC.	ANG.	© 2011 COVIDIEN ALL RIGHTS RESERVED.	
+/-	+/-	+/-	SIZE A4 FILENAME: 10022190B00.DWG	
0.05	1/16	0.5°	DRAWN MARK SWEETMAN DATE 20/02/2014	
SCALE NTS	DO NOT SCALE DRAWING		DWG NO.: 10022190	
THIRD ANGLE PROJECTION	CHECKED See Agile DATE See Agile		REV B	
	APPROVED See Agile DATE See Agile		SHEET 1 of 1	