

Table of contents

\\NEUROSPIN

<table border="1"><tr><td>NEUROSPIN C2P</td></tr></table>	NEUROSPIN C2P
NEUROSPIN C2P	

<table border="1"><tr><td>QA</td></tr></table>	QA
QA	

<table border="1"><tr><td>HeadNeck20</td></tr></table>	HeadNeck20
HeadNeck20	

<table border="1"><tr><td><table border="1"><tr><td>localizer_1sag_1cor_1axi</td></tr><tr><td>EP2D_QA_Noise_and_SNR</td></tr><tr><td>EP2D_QA_warm-up</td></tr><tr><td>EP2D_QA_Stability</td></tr></table></td></tr></table>	<table border="1"><tr><td>localizer_1sag_1cor_1axi</td></tr><tr><td>EP2D_QA_Noise_and_SNR</td></tr><tr><td>EP2D_QA_warm-up</td></tr><tr><td>EP2D_QA_Stability</td></tr></table>	localizer_1sag_1cor_1axi	EP2D_QA_Noise_and_SNR	EP2D_QA_warm-up	EP2D_QA_Stability
<table border="1"><tr><td>localizer_1sag_1cor_1axi</td></tr><tr><td>EP2D_QA_Noise_and_SNR</td></tr><tr><td>EP2D_QA_warm-up</td></tr><tr><td>EP2D_QA_Stability</td></tr></table>	localizer_1sag_1cor_1axi	EP2D_QA_Noise_and_SNR	EP2D_QA_warm-up	EP2D_QA_Stability	
localizer_1sag_1cor_1axi					
EP2D_QA_Noise_and_SNR					
EP2D_QA_warm-up					
EP2D_QA_Stability					

\\NEUROSPIN\NEUROSPIN C2P\QA\HeadNeck20\localizer_1sag_1cor_1axi

TA: 6 sec Coil Selection: Auto Voxel Size: 1.1×1.1×7.0 mm³ Acc.: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
TR	10.0 ms
TE	4.50 ms
Averages	1
Concatenations	3
AutoAlign	---

Contrast - Common

TR	10.0 ms
TE	4.50 ms
TD	0.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	40 deg

Contrast - Common

Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	1
SWI	Off
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
Base Resolution	256
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	24
Deep Resolve	Off
Phase Partial Fourier	Off
Asymmetric Echo	Off

Resolution - Filter

Raw Filter	On
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	1

Geometry - Common

Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
TR	10.0 ms
Multi-Slice Mode	Sequential
Series	Ascending
Concatenations	3

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slice-sel.
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.260524 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	10.0 ms
Segments	1
Concatenations	3

Physio - Cardiac

Tagging	None
Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	280 mm
FOV Phase	100.0 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	3

Inline - Liver

Liver Registration	Off
Save Original Images	On

Inline - Subtraction

Subtract	Off
----------	-----

Inline - Subtraction

Measurements	1
StdDev	Off
Save Original Images	On

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Soft Tissue

Wash-in	Off
Wash-out	Off
TTP	Off
PEI	Off
MIP Time	Off
Measurements	1

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - MapIt

MapIt	None
Flip Angle	40 deg
Measurements	1
Contrasts	1
TE	4.50 ms
TR	10.0 ms
Save Original Images	On

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	fl
Dimension	2D
Excitation	Slice-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Flow Compensation	None
Bandwidth	180 Hz/Px
Asymmetric Echo	Off
Segments	1

Sequence - Part 2

Introduction	On
RF Spoiling	On
Acoustic noise reduction	Off

\NEUROSPIN\NEUROSPIN C2P\QA\HeadNeck20\EP2D_QA_Noise_and_SNR

TA: 55 sec Coil Selection: Manual Voxel Size: 3.4×3.4×4.0 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	2000.0 ms
TE	26.00 ms
Averages	1
Concatenations	1
AutoAlign	---
Coil Elements	HE1-4;NE1,2

Contrast - Common

TR	2000.0 ms
TE	26.00 ms
MTC	Off
Flip Angle	90 deg
Fat-Water Contrast	Standard
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	2
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FOV Read	220 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	4.0 mm
Base Resolution	64
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	2000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	4 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260524 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2000.0 ms
Concatenations	1

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	1698 Hz/Px
Echo Spacing	0.65 ms
Free Echo Spacing	Off
EPI Factor	64

Sequence - Part 2

Introduction	Off
Ghost Reduction	Off

Sequence - Special

QA Postprocessing	Noise & SNR
Mosaic	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\NEUROSPIN\\NEUROSPIN C2P\\QA\\HeadNeck20\\EP2D_QA_warm-up

TA: 6:44 min Coil Selection: Manual Voxel Size: 3.4×3.4×4.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	27
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	2000.0 ms
TE	26.00 ms
Averages	1
Concatenations	1
AutoAlign	---
Coil Elements	HE1-4;NE1,2

Contrast - Common

TR	2000.0 ms
TE	26.00 ms
MTC	Off
Flip Angle	90 deg
Fat-Water Contrast	Standard
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	200
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FOV Read	220 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	4.0 mm
Base Resolution	64
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	27
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	2000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	134 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260524 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2000.0 ms
Concatenations	1

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	1698 Hz/Px
Echo Spacing	0.65 ms
Free Echo Spacing	Off
EPI Factor	64

Sequence - Part 2

Introduction	Off
Ghost Reduction	Off

Sequence - Special

QA Postprocessing	None
Mosaic	On

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\NEUROSPIN\NEUROSPIN C2P\QA\HeadNeck20\EP2D_QA_Stability

TA: 6:44 min Coil Selection: Manual Voxel Size: 3.4×3.4×4.0 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	27
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	2000.0 ms
TE	26.00 ms
Averages	1
Concatenations	1
AutoAlign	---
Coil Elements	HE1-4;NE1,2

Contrast - Common

TR	2000.0 ms
TE	26.00 ms
MTC	Off
Flip Angle	90 deg
Fat-Water Contrast	Standard
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	200
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FOV Read	220 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	4.0 mm
Base Resolution	64
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	27
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	2000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
CoilShim	Off
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	134 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260524 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2000.0 ms
Concatenations	1

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	1698 Hz/Px
Echo Spacing	0.65 ms
Free Echo Spacing	Off
EPI Factor	64

Sequence - Part 2

Introduction	Off
Ghost Reduction	Off

Sequence - Special

QA Postprocessing	EPI Stability
Mosaic	On

Sequence - Assistant

SAR Assistant	Off
---------------	-----