

\NEUROSPIN\C2P NeuroSpin\QA\C2P_Protocol_QA_B1map\1-ScatteringMatrix

TA: 1 sec Coil Selection: Auto Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine**Contrast - Common**

Fat-Water Contrast	Standard
--------------------	----------

Contrast - Dynamic

Dynamic Mode	Standard
--------------	----------

Resolution - Common**Resolution - Acceleration****Resolution - Filter**

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common**Geometry - AutoAlign**

Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
---------	----------

System - Tx/Rx

Frequency 1H	297.158235 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	ns_smatrix
Gradient Mode	Normal

Sequence - Part 2**Sequence - Nuclei**

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

No of dummy scans	10
No of scan per TX	1
RF Pulse type	Rect
RF Voltage	50.0
RF Duration	1000
RF TBW	2.0

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\NEUROSPIN\C2P NeuroSpin\QA\C2P_Protocol_QA_B1map\2-AutoFOV_position_for_QA_B1MAP

TA: 1 sec Coil Selection: Auto Voxel Size: 0.4x0.4x5.0 mm³ Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
FOV Read	240 mm
FOV Phase	100.0 %
TR	200.0 ms
Averages	1

Contrast - Common

TR	200.0 ms
Flip Angle	1 deg
Fat-Water Contrast	Standard
Contrasts	1

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1

Resolution - Common

FOV Read	240 mm
FOV Phase	100.0 %
Base Resolution	600

Resolution - Acceleration**Resolution - Filter**

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
FOV Read	240 mm
FOV Phase	100.0 %
TR	200.0 ms

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Initial Position	Isocenter
Phase	0.0 mm
Read	0.0 mm
Shift	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg

System - Adjust Volume

A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
---------	----------

System - Tx/Rx

Frequency 1H	297.158235 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	profile
Dimension	2D
Gradient Mode	Fast
Bandwidth	400 Hz/Px

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\NEUROSPIN\C2P NeuroSpin\QA\C2P_Protocol_QA_B1map\3-QA_B1MAP

TA: 30 sec Coil Selection: Auto Voxel Size: 5.0×5.0×2.5 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	5
Distance Factor	100 %
Position	L999.0 A999.0 F999.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	320 mm
FOV Phase	62.5 %
Slice Thickness	2.5 mm
TR	3000.0 ms
TE	2.04 ms
Averages	1
Concatenations	1
AutoAlign	---

Contrast - Common

TR	3000.0 ms
TE	2.04 ms
Magn. Preparation	None
Flip Angle	5 deg
Fat-Water Contrast	Standard
Dark Blood	Off
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement
Reordering	Centric

Resolution - Common

FOV Read	320 mm
FOV Phase	62.5 %

Resolution - Common

Slice Thickness	2.5 mm
Base Resolution	64
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off
Asymmetric Echo	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Off
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	5
Distance Factor	100 %
Position	L999.0 A999.0 F999.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	320 mm
FOV Phase	62.5 %
Slice Thickness	2.5 mm
TR	3000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice Group	1
Position	L999.0 A999.0 F999.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
AutoAlign	---
Initial Position	L999.0 A999.0 F999.0
L	999.0 mm
A	999.0 mm
F	999.0 mm
Initial Orientation	Transversal
Initial Rotation	90.00 deg

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L999.0 A999.0 F999.0 mm
Orientation	Transversal
Rotation	90.00 deg
R >> L	200 mm
A >> P	320 mm
F >> H	23 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	297.158235 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	3000.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	320 mm

Physio - Cardiac

FOV Phase	62.5 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	tfl
Dimension	2D
Excitation	Slice-sel.
RF Pulse Type	Low SAR
Gradient Mode	Whisper
Flow Compensation	None
Reordering	Centric
Bandwidth	1000 Hz/Px
Echo Spacing	3.60 ms
Asymmetric Echo	Off
Turbo Factor	40

Sequence - Part 2

Introduction	Off
RF Spoiling	On
Incr. Gradient Spoiling	Off

Sequence - Special

B1 mapping type	Interferometric fast relative
Export B1 maps	On
Mosaic recon.	On
Sat. flip angle	90 °
Time BW product	9.0
Distribute slices	On
Perform QA pTx	On
FOV shift for QA pTx	3D shift

Sequence - Assistant

SAR Assistant	Off
---------------	-----