

Tektronix (China) Co. Ltd.  
1227 Chuan Qiao Road  
Pudong New District, Shanghai, China 201206



## CERTIFICATE OF TRACEABLE CALIBRATION

Certificate No: PCXPRL68RJ

Revision: 00

Manufacturer: Tektronix

Model: TDS2014C

Serial No: C012371

Description: Oscilloscope; 100 MHz, 4 Channel

Temperature: 24 °C

Humidity: 68 %

Cal Date: 21-MAY-2011

Date Placed In Service: \_\_\_\_\_\*

Due Date: \_\_\_\_\_\*

\*Optional customer entry fields: The "Due Date" may be established by adding the calibration interval to the "Date Placed In Service." Recommended calibration intervals, for Tektronix products, are available on the Tektronix website at:  
<http://www.tek.com/service/product-support/>

### INSTRUMENT CONDITION:

Received: Not applicable, initial calibration

Returned: In tolerance, within published measurement specification

Tektronix certifies the above instrument has been calibrated using standards traceable to the PRC National Institute of Metrology (NIM) and/or other National Metrology Institutes (NIST, NPL, PTB) that are linked to the international system of units (SI). The policies and procedures used for the calibration of this product are based on ISO/IEC 17025:2005.

This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

### CALIBRATION PROCEDURE:

MANIFEST:Product\_Oscilloscopes\_TDS1K2K-SC\_Full VERSION:75

### CALIBRATION EQUIPMENT USED:

<u>MANUFACTURER/MODEL</u>	<u>MODEL DESCRIPTION</u>	<u>ID NUMBER</u>	<u>DUE DATE</u>
Keithley 2000	Digital Multimeter	CT0059	29-Feb-2012
Wavetek 9100	Calibrator	CT0046	30-Apr-2012

Issued By: *Walter Qian*

Certified By: Zhengxing Wang

Quality Manager: Qian, Walter

Date Issued: 21-MAY-2011

ISO 9001 Registered Quality System

TDS2014C C012371

Page 1 of 1

001-1387-01

